



# THE NEW ZEALAND JOINT REGISTRY

TWENTY YEAR REPORT  
JANUARY 1999 TO DECEMBER 2018

A large, stylized blue hand graphic, palm facing up, serves as a background for the lower half of the page. A series of white circles connected by a white line runs diagonally across the hand, starting from the fingers and ending at the wrist. The final circle at the wrist contains the text "20 YEARS".

**20  
YEARS**





## Registry Management Committee

|                 |   |
|-----------------|---|
| John McKie      | Chairman and Registry Supervisor, Orthopaedic Surgeon                 |
| Peter Devane    | Orthopaedic Surgeon   |
| Simon Young     | Orthopaedic Surgeon   |
| Brendon Coleman | Orthopaedic Surgeon   |
| Dawson Muir     | Orthopaedic Surgeon   |
| Perry Turner    | Secretary New Zealand Orthopaedic Association,<br>Orthopaedic Surgeon |
| Hugh Griffin    | Orthopaedic Industry Liaison Association                              |
| Peter Larmer    | Arthritis New Zealand   |
| Andrea Pettett  | CEO New Zealand Orthopaedic Association                               |
| Toni Hobbs      | Registry Coordinator  |
| Chris Frampton  | Registry Statistician   |

## Annual Report Editorial Committee

|                 |                       |
|-----------------|-----------------------|
| John McKie      | Registry Supervisor   |
| Toni Hobbs      | Registry Coordinator  |
| Chris Frampton  | Registry Statistician |
| Peter Devane    | Hips                  |
| Brendan Coleman | Shoulders             |
| Simon Young     | Knees                 |
| Dawson Muir     | Ankles                |

---

**Email:** [toni.hobbs@cdhb.health.nz](mailto:toni.hobbs@cdhb.health.nz)  
**Website:** [www.nzoa.org.nz/nz-joint-registry](http://www.nzoa.org.nz/nz-joint-registry)  
**Date of Publication:** December 2019

---

# CONTENTS

---

|  |     |
|--|-----|
| Editorial Comment  | 4   |
| Acknowledgments  | 8   |
| Participating Hospitals and Coordinators                                   | 9   |
| Development and Implementation of the New Zealand Registry                 | 11  |
| Additional analyses  | 13  |
| Hip Arthroplasty   | 14  |
| Knee Arthroplasty  | 82  |
| Unicompartmental Knee Arthroplasty   | 115 |
| Ankle Arthroplasty   | 125 |
| Shoulder Arthroplasty  | 130 |
| Elbow Arthroplasty   | 146 |
| Lumbar Disc Replacement  | 151 |
| Cervical Disc Replacement  | 153 |
| Reoperation without Replacement or Removal of any<br>Prosthetic Components | 154 |
| Appendices:  |     |
| - Appendix 1 - Oxford 12 Questionnaire References                          | 155 |
| - Appendix 2 - Publications  | 156 |
| - Appendix 3 - Data forms  | 157 |
| - Appendix 4 - Oxford 12 Questionnaire forms                               | 172 |



## EDITORIAL COMMENT

The Registry Management Committee is pleased to present the twenty year report of the New Zealand Orthopaedic Association's Joint Registry.

In this year's report the format of previous years has been followed such that each arthroplasty section is self-contained. This does, however, result in a certain amount of intersection repetition.

The total number of registered joint arthroplasties at 31st of December 2018 was 303,829, which had been performed on 204,507 individual patients, of which 46,199 (23%) have now died during the 20 year period.

The number of observed component years (ocys) contained within the Registry is now 1.9 million. The increase of 22,271 registered joints for 2018 compared to the 20,699 in 2017 represents an overall annual gain of 7.6%.

The mean BMIs are 31.26 (knees) and 29.0 (hips) but there are significant numbers of morbidly obese (BMI>40) people receiving arthroplasties.

**As for previous years, analyses of revision data has been confined to primary registered arthroplasties.**

### Hip arthroplasty

There are 135,461 conventional total hip arthroplasties with an overall revision rate of 0.72 per 100 ocys (95% confidence interval; 0.70 -0.73) with a 19 year prosthesis survival of 84.70% (cemented 86.1%; uncemented 84.5% and hybrid 84.10%).

More females than males received a hip replacement (52.93% vs 47.07%), with a slightly higher mean age (68.46 vs 65.45 years), but a very wide range for both (13 to 101 yrs.)

Most had no previous surgery (97.4%) and a diagnosis of osteoarthritis (89.1%). Most operations were performed through the posterior (67.1%) or lateral approach (24.4%).

Approximately 200 hips per year in New Zealand are performed through the anterior approach, and this number has remained steady since 2014, despite its popularity in the literature. Fully cemented hip replacement has fallen from 14% in 2012, but has stabilised at approximately 7% in the last 2 years.

The ceramic on polyethylene bearing surface continues to increase in popularity rising from 42% of the total in 2017 to 48% in 2018.

The proportion of the metal on metal articulation continues to decline and in 2018 was less than 1% of the total, all with head sizes < 32mm.

The most popular head size overall remains the 32mm and in 2017, this was used in 60% of primary arthroplasties. In 2018, this was used in 62% of primary arthroplasties.

However, the percentage use of 36mm head sizes remained steady in 2018, similar to its use in 2017.

On the other hand, metal on metal articulations fare poorly when revision rates are analysed against head size, bearing surface materials, age bands and cemented/uncemented/hybrid variants of the same prosthesis. Further reinforcement is from the survival curves for bearing surfaces.

The use of cross linked polyethylene remains the dominant choice again accounting for in excess of 90% of all polyethylene used (92% compared with 94% in 2017).

As in previous years, the three types of hip fixation have been analysed against the four age bands: less than 55 years; 55-64 years; 65-74 years, and greater than 74 years. The data shows that overall the hybrid hip has the lowest revision rate.

There are 1,092 hip prosthesis combinations in the Registry but only 227 (21%) with 50 or more registrations.

As with the nineteen year report, this years' report does not include a Table of Revisions vs Hip Prostheses Combinations Sorted on Number of Implantations, since it does not reflect what is currently being used.

Instead we have replaced it with a new Table labelled Revisions versus Hip Prostheses Combinations used in 2018, Sorted on Revision Rate.

This Table reflects prostheses combinations currently being used.

There were 100 different combinations of acetabular and femoral components used in 2018 that had more than 50 operations in the Registry.

Note that the total of prostheses used in 2017 in this table is 8,618, not 9,186, meaning that only 568 (6.2%) "new" prostheses combinations (less than 50 operations in total) were used by surgeons in 2018, a small percentage similar to the 5.4% seen in the 2018 report.


We hope surgeons use this Table to monitor performance of their chosen implant combinations.

The next table, Revision versus Hip Prostheses Combinations Sorted on Revision Rate, has an extra column which includes the number of prostheses used in 2018.

This will allow surgeons to see which prostheses combinations currently being used, may have a higher than average revision rate.

Newly introduced prostheses generally require 3 years for their revision rates to reach surgical significance.

Comparison of the 2018 Column and the No. of Ops. Column can identify these.



“ In this year's report the format of previous years has been followed such that each arthroplasty section is self-contained. This does, however, result in a certain amount of intersection repetition.”

The Corail/Pinnacle combination was again the most popular in 2018, with 1,471 primary arthroplasties.

Second most popular in 2018 was the Exeter V40/ Trident combination, with 1,049 primary arthroplasties. Both have revision rates well below the NZ mean, 0.68 and 0.43 ocys respectively.

The Exeter V40/Continuum combination identified in last years' report has improved its revision rate in 2018 and is no longer a cause for concern.

In 2018, only 23 patients had operations using prosthesis combinations which had a significantly higher revision rate, compared to 31 patients in 2017.

Otherwise New Zealand surgeons are using (and being supplied with) prostheses combinations that have good track records for revision rate.

Despite last years' report highlighting the Twinsys cemented/ Pinnacle combination having an unacceptable revision rate (10 patients in 2017), a further 8 patients had this combination in 2018.

Although there were no further revisions of this combination in 2018, (total 7), its use should still be questioned.

Similarly, surgeons chose to use the ABGII/RM Pressfit Cup in 12 patients, despite it having a significantly higher revision rate in 2017.

The 2018 revision rate of this combination show the same poor results and its continued use should be questioned.

## Resurfacing hip arthroplasty

All BHR's have increased in number in 2018. 118 compared to 94 in 2017, and the third year showing an increase from the low point of their use in 2016. The revision rate has again fallen from a rate of 1.15 ocys in 2017 to 1.06 ocys (0.90 – 1.24).

## Knee arthroplasty

110,076 conventional total knee arthroplasties have been registered totalling 753,723 ocys with the overall revision rate 0.48/100 ocys, (95% confidence interval; 0.47-0.50) and the excellent 20 year survival of 92.1%.

The number of TKA's implanted continues to increase, with 8,392 implanted in 2018, a 16.29% increase over 2017.

As was done for recent annual reports, several variants of basically the same knee prosthesis type for example, Nexgen and LCS, which are registered separately, have been merged into the one group to enable comparable statistical analyses with other prostheses which may have also had variants, but are registered as one or two prostheses.

There are 25 different knee prostheses in the Registry that have a minimum of 50 registrations.

The Triathlon remains the most popular prosthesis in 2018, with the Attune holding second place.

Calculation of revision rates for individual prostheses with a minimum of 50 arthroplasties shows that among the bigger registered numbers the Duracon, although no longer implanted, has the lowest revision rate of 0.321/100ocys.

The Nexgen has the biggest number of registrations at 19,728 with 155,715 ocys and a revision rate of 0.52/100ocys.

Four of the currently used cemented prostheses, Balansys, Persona, Trekking and the Journey and the one fully uncemented prosthesis, LCS has a higher revision rate than the overall rate of 0.48/100 ocys @ the 95% confidence interval.

It is important to note that the use of revisions per 100 component years as an outcome measure will tend to disadvantage newer prostheses such as the Persona and the Attune, as revision for infection occurs more commonly in the first year post implantation.


Although uncemented knee arthroplasty represents just 4-5% of all primary knee arthroplasties, it has a significantly higher revision rate than either fully cemented or hybrid in which the tibial component is cemented and the femoral component uncemented.

In the last two years there has been a small increase in the percentage use of fully uncemented TKA prostheses, reversing the previous trend.

The KM curves for the three types of fixation show that the uncemented curve continues to steeply diverge from the other two.

Similar to other registry findings, analysis suggests that the tibial component remains the limiting factor in uncemented TKA replacement.

The analyses comparing revision rates and survival of fixed versus mobile bearing knees continue to show that there is similar longer term survival for both versions.



Again this year, separate analyses for cruciate retaining versus posterior stabilised knee prostheses demonstrate that overall there are significantly higher revision rates for posterior stabilised prostheses. This is also graphically illustrated with KM survival graphs, and seems to hold true across almost all brands with both PS and CR versions.

There are 602 registered patello-femoral prostheses, with 71 added in 2018, compared to 65 in 2017.

58 have been revised and the revision rate at 1.91/100 ocys is nearly four times that for total knee arthroplasty. All except six were revised to a total knee arthroplasty.

Again this year revision rate tables and survival curves are included for the five different BMI groupings and, like hip arthroplasty, the morbidly obese (BMI>40) group have statistically significant poorer prosthesis survival.

## Unicompartmental Knee Arthroplasty

Unicompartmental prostheses with a total of 87,000 ocys, a mean revision rate of 1.20/100 ocys and a 17 year survival of 79.5%. Pain remains the main listed reason for revision in 52 % of cases where a reason is given.

There were 1,096 registrations in 2018, very similar to the 2017 numbers.

Once again the Oxford uncemented prosthesis was very dominant, accounting for 69% of the unicompartmental prostheses implanted in 2018.

The revision rate is 0.77/100 ocys and this drops to 0.70/100 ocys (0.58 – 0.84) when the medial Oxford UKR's are analysed separately- lateral Oxford UKR's have a revision rate of 1.67/100 ocys (1.06 – 2.51).

The lowest revision rate is currently the Zimmer unicompartmental prosthesis at 0.47/100 ocys.

The overall revision rate is 1.2/100 ocys, however surgeons who perform less than 10 UKR's per year have a significantly higher revision rate – 1.39/100 ocys (1.28- 1.51) compared to surgeons doing 10 or more procedures 1.01/100 ocys (0.92 – 1.11).

## Ankle Arthroplasty

There are 1,619 primary registered ankle prostheses with a total of 10,038 ocys.

There were 117 primary ankle arthroplasties registered in 2018.

## Shoulder Arthroplasty

There are 10,324 registered primary shoulder prostheses, with a total of 54,864 ocys. An additional 1,074 primary shoulder replacements have been performed in 2018. Over recent time, there is a 6-7% annual growth in the utilisation of shoulder arthroplasty in New Zealand.

Reverse arthroplasty remains the predominant implant in 2018, with 70% of all shoulder arthroplasties being reverse arthroplasty. There is a slow, but steady increase in the utilisation of reverse arthroplasty, predominantly at the expense of hemiarthroplasty. The decline in anatomic shoulder

replacement continues and currently represents 23% of all shoulder arthroplasties performed, a decrease of 3% from 2017.

The 10 year survival of all shoulder prostheses is 91.9%, whilst the 15 year revision free survival is 89.4%.

The revision rate of 0.95 per 100 component years for primary shoulder arthroplasty remains steady, as do the rates of total (0.94) and reverse arthroplasty (0.76). The burden of revision surgery in shoulder arthroplasty is increasing, with a 12% increase in revision workload for the 2018 year. 817 revision cases were performed, with pain remaining the primary indication for revision.

Although reverse shoulder arthroplasty has increased revision rates compared to total shoulder replacement during the first two years, reverse arthroplasty outperforms total shoulder replacement with a ten year survival of 96% compared to a rate of 92% for total shoulder replacement.

Partial resurfacing continues to have a significantly higher revision rate than all other groups, with a trend to increasing revision rate from previous years. However, only 2 cases of partial resurfacing were performed in 2018 and 6 cases of total resurfacing, so there is little utilisation of this type of procedure in the current period.

Arthroplasties utilising uncemented glenoids continue to show a 4 times revision rate compared to those having cemented glenoids.

Average Oxford scores remain unchanged from 2018. There is an improvement in scores from 6 months to 5 years, but then the scores stabilise at 10 years. The initial four-point difference in scores for total shoulder and reverse shoulder decreases at 5 years, but the total shoulder scores remain 2.5 points higher at 5 years.

An Oxford score of less than 27 results in a fivefold increase in risk of revision compared to those with a score of 34 or greater.

## Elbow Arthroplasty

There are 587 registered primary elbow prostheses with a total of 3,819 ocys.


There were 27 primary elbow prostheses registered in 2018.

Worldwide, Rheumatoid arthritis has decreased and Trauma has increased as an indication for elbow replacement.

## Deep Infection

Once again we have compared the deep infection revision rates within six months of the arthroplasty for primary hip and knee arthroplasty against the theatre environment. Six months has been chosen, as infection within this time period is highly likely to have been introduced at the time of surgery.

The registry data continues to show an increased rate of infection when exhaust suits and laminar flow ventilation is used. This data needs to continue to be interpreted with caution. The data regarding suit use is likely to be accurate and experimental evidence has supported the observation that exhaust suits are counterproductive.



Data on use of laminar flow is likely to be inaccurate with many surgeons unsure of the status of ventilation in the theatres used.

The Registry intends to record the status of all theatres used and have the theatre listed on the data capture form to improve the accuracy of this over time.

## Oxford 12 Questionnaire

Six month, 5, 10 and 15 year scores analyses of the individual score categories for primary hip and knee arthroplasties continue to demonstrate that the six-month score is indicative of the longer term outcome.

It is noteworthy that the 15 year scores still have a similar high percentage of excellent/good outcomes as the 6 month, 5 and 10 year outcomes.

As noted in previous years, the statistically significant relationship between the six month, five and ten year scores and revision within two years of the scoring date for primary hips, knees (including unicompartmental) and shoulders (six months and five years only) has again been demonstrated.

With the very large number of recorded six month Oxford hip and knee scores, the score groupings can be further broken down to demonstrate an even more convincing relationship between score and risk of revision within two years.

Once again analyses of hip and knee six month post first revision arthroplasty questionnaire data has been undertaken and it demonstrates a similar relationship between the Oxford score at six months and the second revision within two years.

This year Oxford score analyses for some of the larger number hip and knee prostheses have been undertaken and show that there is little score difference among these prostheses at six months and without exception they have higher (better) scores at five years. For all the knee scores the higher 5 year scores are not only statistically significant but also clinically significant when compared to the 6 month scores.

With regard to shoulder arthroplasty, Conventional Total and Resurfacing Head types have significantly higher six month and five year scores.

## Deceased Person's Data

A deceased person's data is valid in perpetuity for all analyses involving the time interval prior to the person's death e.g. if a person dies eight years post primary hip replacement their data is always valid for all analyses for that eight year period. Hence the rider "deceased patients censored at time of death".

---

**John McKie** – Supervisor  
**Toni Hobbs** – Coordinator  
**Chris Frampton** – Statistician

---





## ACKNOWLEDGEMENTS

The Registry is very appreciative of the support from the following:

**Canterbury District Health Board:**

For accommodation and other facilities

**Chris Lewis, Information Analyst, Ministry of Health:**

For audit compliance information

**Mike Wall, Alumni Software:**

For comprehensive Information Technology services

**DHD Creative:**

Final design

## FUNDING

The Registry wishes to acknowledge development and ongoing funding support from:

- ACCIDENT COMPENSATION CORPORATION
- CANTERBURY DISTRICT HEALTH BOARD
- MINISTRY OF HEALTH
- ORTHOPAEDIC SURGEONS

## PARTICIPATING HOSPITALS

We wish to gratefully acknowledge the support of all participating hospitals and especially the coordinators who have taken responsibility for the data forms.





# HOSPITALS AND CONTACTS

## Public Hospitals

### Auckland Hospital

Auckland 1142  
Contact: [Shelley Thomas](#)

### Burwood Hospital

Christchurch 8083  
Contact: [Diane Darley](#)

### Christchurch Hospital

Christchurch 8140  
Contact: [Ruth Hanham](#)

### Dunedin Hospital

Dunedin 9016  
Contact: [Jennifer Larsen](#)

### Elective Surgery Centre

Takapuna 0740  
Contact: [Wings Chang](#)

### Gisborne Hospital

Gisborne 4010  
Contact: [Gretel McKenzie](#)

### Grey Base Hospital

Greymouth 7840  
Contact: [Lynette Sweetman](#)

### Hawkes Bay Hospital

Hastings 4120  
Contact: [Rochelle Holder](#)

### Hutt Hospital

Lower Hutt 5040  
Contact: [Margot Clapham](#)

### Kenepuru Hospital

Porirua 5240  
Contact: [Courtney Dougherty](#)

### Manukau Surgery Centre

Auckland 2104  
Contact: [Amanda Ellis](#)

### Masterton Hospital

Masterton 5840  
Contact: [Lisa Manihera](#)

### Middlemore Hospital

Auckland 1640  
Contact: [Lalesh Deo](#)

### Nelson Hospital

Nelson 7040  
Contact: [Sadie Sheridan](#)

### North Shore Hospital,

Takapuna 0740  
Contact: [Petra Mons](#)

### Palmerston North Hospital

Palmerston North 4442  
Contact: [Maria Shaw/Karen McKie](#)

### Rotorua Hospital

Rotorua 3046  
Contact: [Janice Reynolds](#)

### Southland Hospital

Invercargill 9812  
Contact: [Helen Powley](#)

### Taranaki Base Hospital

New Plymouth 4342  
Contact: [Allison Tijssen](#)

### Tauranga Hospital

Tauranga 3143  
Contact: [David Nyhoff](#)

### Timaru Hospital

Timaru 7940  
Contact: [Carla Spence](#)

### Waikato Hospital

Hamilton 3204  
Contact: [Lorraine Grainger](#)

### Wairau Hospital

Blenheim 7240  
Contact: [Monette Johnston](#)

### Wellington Hospital

Newtown 6242  
Contact: [Brigitte Stravens](#)

### Whakatane Hospital

Whakatane 3158  
Contact: [Karen Burke](#)

### Whanganui Hospital

Whanganui 4540  
Contact: [Susan Slight](#)

### Whangarei Area Hospital

Whangarei 0140  
Contact: [Leanne Thorn](#)

## Private Hospitals

### Ascot Integrated Hospital

Remuera 1050  
Contact: [Alicia Zanders](#)

### Belverdale Hospital

Wanganui 4500  
Contact: [Donna Plumridge](#)

### Bidwill Trust Hospital

Timaru 7910  
Contact: [Kay Taylor](#)

### Boulcott Hospital

Lower Hutt 5040  
Contact: [Karen Hall](#)

### Bowen Hospital

Wellington 6035  
Contact: [Pam Kohnke](#)

**Braemar Private Hospital**

Hamilton 3204

Contact: [Phyllis Lee](#)

**Chelsea Hospital**

Gisborne 4010

Contact: [Vicki Briant](#)

**Crest Hospital**

Palmerston North 4440

Contact: [Terri Sellwood](#)

**Grace Hospital**

Tauranga 3112

Contact: [Anne Heke](#)

**Kensington Hospital**

Whangarei 0112

Contact: [Sandy Brace](#)

**Manuka Street Hospital**

Nelson 7010

Contact: [Karen Tijssen](#)

**Mercy Hospital**

Dunedin 9054

Contact: [Liz Cadman](#)

**Mercy Integrated Hospital**

Auckland 1023

Contact: [Marie Buitenhek/Janine Wells](#)

**Ormiston Hospital**

Auckland 2016

Contact: [Julie Hodgson](#)

**Royston Hospital**

Hastings 4122

Contact: [Mark Casey](#)

**Southern Cross Hospital, Brightside**

Epsom 1023

Contact: [Rachel White](#)

**Southern Cross Hospital**

Christchurch Central 8013

Contact: [Diane Kennedy](#)

**Southern Cross Hospital**

Hamilton 3216

Contact: [Daphne Van Dam](#)

**Southern Cross Hospital**

Invercargill Central 9810

Contact: [Maree Henderson](#)

**Southern Cross Hospital**

New Plymouth 4310

Contact: [Leane Belgrave/Brendan Toole](#)

**Southern Cross North Harbour**

Glenfield 0627

Contact: [Sarah Young](#)

**Southern Cross Hospital**

Rotorua 3015

Contact: [Penny Garwood](#)

**Southern Cross Hospital**

Newtown, Wellington 6021

Contact: [Marian Lee](#)

**St Georges Hospital**


Christchurch 8014

Contact: [Ali Perry](#)

**Wakefield Hospital**

Newtown, Wellington 6021

Contact: [Jo Barlow](#)



# DEVELOPMENT AND IMPLEMENTATION OF THE NEW ZEALAND JOINT REGISTRY

The year 1997 marked 30 years since the first total hip replacement had been performed in New Zealand and as a way of recognizing this milestone it was unanimously agreed by the membership of the New Zealand Orthopaedic Association (NZOA) to adopt a proposal by the then President, Alastair Rothwell, to set up a National Joint Registry.

New Zealand surgeons had always been heavily dependent upon northern hemisphere teaching, training and outcome studies for developing their joint arthroplasty practice and it was felt that it was more than timely to determine the characteristics of joint arthroplasty practice in New Zealand and compare the outcomes with northern hemisphere counterparts. It was further considered that New Zealand would be ideally suited for a National Registry with its strong and co-operative NZOA membership, close relationship with the implant supply industry and its relatively small population. Advantages of a Registry were seen to be: survivorship of different types of implants and techniques; revision rates and reasons for these; infection and dislocation rates; patient satisfaction outcomes; audit for individual surgeons, hospitals, and regions; opportunities for in-depth studies of certain cohorts and as a database for fundraising for research.

## Administrative Network

It was decided that the Registry should be based in the Department of Orthopaedic Surgery, Christchurch Hospital, and initially run by three part-time staff: a Registry Supervisor (Alastair Rothwell), the Registry Coordinator (Toni Hobbs) and the Registry Secretary (Pat Manning). As all three already worked in the Orthopaedic Department, it was a cost-effective and efficient arrangement to get the Registry underway.

New Zealand was divided into 19 geographic regions and an orthopaedic surgeon in each region was designated as the Regional Coordinator whose task was to set up and maintain the data collection network within the hospitals for that region. This network included a Theatre Nurse Coordinator in every hospital in New Zealand who voluntarily took responsibility for supervising the completion, collection and dispatch of the data forms to the Registry.

## Data Collection Forms

The clear message from the NZOA membership was to keep the forms for data collection simple and user friendly. The Norwegian Joint Register's form was used as a starting point but a number of changes were made following early trials. The forms are largely if not completely filled out by the operating theatre circulating nurse ready to be checked and signed by the surgeon at the end of the operation.

## Database

The Microsoft Access 97 database programme was chosen because it is easy to use, has powerful query functions, can cope with one patient having several procedures on one or more joints over a lifetime and has "add on" provisions. The database is expected to meet the projected requirements of the Registry for at least 20 years. It can accommodate software upgrades as required.

## Patient Generated Outcomes

The New Zealand Registry was one of the first to collect data from patient generated outcomes. The validated Oxford Hip and Knee outcomes questionnaires were chosen and questions were added to these, relating to dislocation, infection and any other complication that did not require further joint surgery. These additions have now been discontinued. It was agreed that these questionnaires should be sent to all registered patients six months following surgery and then at five yearly intervals. The initial response rate was between 70% and 75% and this has remained steady.

However, because of the large number of registered primary hip and knee arthroplasties and, on the advice of our statistician, questionnaires have been sent out on a random selection basis since July 2002 to achieve an annual response of 20% for each group. All patients in the other arthroplasty groups, including revision arthroplasty, are sent the questionnaires.

## Funding

Several sources of funding were investigated including contributions from the Ministry of Health, various funding agencies, medical insurance societies and an implant levy payable by surgeons and public hospitals to supplement a grant from the NZOA. In the early years the Registry had a "hand to mouth" existence relying on grants from the NZOA and Wishbone Trust until it received significant annual grants from the Accident Compensation Corporation.

From 2002, funding became more reliable with the surgeons paying a \$10 levy, and they now pay \$25 for each joint registered from a private hospital.

The latest MOH contract has been extended for a further 3 years with 4 six monthly payments of \$37,500 (excluding GST).

Since 2005 the Southern Cross Hospitals have contributed a grant of \$10,000 annually.



## Ethical Approval

Application was made to the Canterbury Ethical Committee early in 1998; first for approval for hospital data collection without the need for patient consent and second for the patient generated outcomes using the Oxford 12 questionnaire plus the additional questions. The first part of the application was initially readily approved but the second part required several amendments to patient information and consent forms before approval was obtained.

A reapplication had to be made when the Ethics Committee of a private hospital chain refused to allow their nurses to participate in the project unless there was prior written patient consent. This view was supported by the Privacy Commissioner on the grounds that the Registry data includes patient identification details. The approval process was eventually successful but did delay the New Zealand-wide launch.

## Surgeon and Hospital Reports

Since 2008 each surgeon receives an annual report giving their revision rate for primary registered primary arthroplasties, and this includes their questionnaire responses.

## Introduction of the Registry

The National Joint Registry was introduced as a planned staged procedure.

### Stage I: November 1997 to March 1998

The base administrative structure was established. The data forms and the database were developed and a trial was performed at Burwood Hospital.

### Stage II: April 1998 to June 1998

Further trialling was performed throughout the Christchurch Hospitals and the data forms and information packages were further refined.

### Stage III: July 1998 to March 1999

The data collection was expanded into five selected New Zealand regions for trial and assessment.

Also during this time communication networks and the distribution of information packages into the remaining regions of New Zealand were carried out.

### Stage IV: April 1st 1999

The National Joint Registry became fully operational throughout New Zealand.

## Inclusion of Other Joint Replacement Arthroplasties

At the request of the NZOA membership, the database for the Registry was expanded to include total hip replacements for fractured neck of femur, unicompartmental replacements for knees, and total joint replacements for ankles, elbows and shoulders (including hemiarthroplasty for the latter). Commencement of this data collection was in January 2000

and this information is included in the annual surgeon and hospital reports.

The validated Oxford questionnaire was available for the shoulder and derived, but not validated, questionnaires developed for the elbow and ankle joints.

In 2016 the Oxford Elbow Score (OES) and the Manchester-Oxford Foot Questionnaire were introduced replacing the former questionnaires that were not validated.

All patients receiving total arthroplasty of the above joints, as well as unicompartmental knee arthroplasties, are sent questionnaires with a response rate of 70%. As for hips and knees, the questionnaires are sent out 6M post-surgery then at 5Y, 10Y and 15Y.

## Monitoring of Data Collection

The aim of the Registry is to achieve a minimum of 90% compliance for all hospitals undertaking joint replacement surgery in New Zealand.

It is quite easy to check the compliance for public hospitals as they are required to make regular returns with details of all joint replacement surgery to the NZ Health Information Service. The registered joints from the Registry can be compared against the hospital returns for the same period and the compliance calculated. Any obvious discrepancies are checked out with the hospitals concerned and the situation remedied. It is more difficult with private hospital surgery as they are not required to file electronic returns. However, by enlisting the aid of prosthesis supply companies, it is possible to check the use of prostheses region by region and any significant discrepancy is further investigated. In addition, any change in the pattern of returns from private hospitals is checked.

Another method is to check data entry for each hospital against the previous corresponding months and if there is an obvious trend change then again this is investigated.

The most recent compliance audit in February 2018 again demonstrated a New Zealand-wide public hospital compliance of > 95% when compared to NZHIS data.

Following the introduction of the South Island PICS system at the beginning of October 2018, the Registry lost the ability to search for nationwide NHI entries and was not able to access nationwide date of death registrations.

This has now been overcome, and the data entry staff now use the MOH NHI lookup system to check NHI entries and addresses.

Also, the Registry can now access the nationwide death files through the MOH FTP server with twice monthly updates. Accurate date of death registrations are essential for our statistical analyses.

## NZJR Staff

The current staff are data entry (1.75 FTE), Registry coordinator (0.8 FTE), Registry supervisor (0.2 FTE) and statistician (0.04 FTE).



## ADDITIONAL ANALYSES

The number of registered joint replacements for the 20 year period to December 2018 was 303,829.

During this period 204,507 individual patients were registered, of which 46,199 (23%) have died.

Bilateral joint replacements carried out under the same anaesthetic;

### Bilateral hips

2,545

patients (5,090 hips)  
4% of primary hips

### Bilateral knees

4,311

patients (8,622 knees) 8% of  
primary knees

### Bilateral Unicompartmental knees

969

patients (1,938 knees) 15%  
of unicompartmental knees

### Bilateral ankles

2

patients (4 ankles)

### Bilateral shoulders

5

patients (8 shoulders)

*Trainee Surgeons: In the following analyses consultants took responsibility for their registrar surgeon procedures.*



# HIP ARTHROPLASTY

## PRIMARY HIP ARTHROPLASTY

The **twenty year** report analyses data for the period January 1999 – December 2018. There were 137,338 primary hip procedures registered including 1,877 resurfacing arthroplasties. This is an additional 9,186 compared to last year's report.

Number of operations by year



## Data Analysis

### Age and sex distribution

The average age for all patients with primary hip arthroplasty was 67 years, with a range of 13 – 101 years.

#### All hip arthroplasty

|               | Female | Male   |
|---------------|--------|--------|
| Number        | 72,690 | 64,651 |
| Percentage    | 52.93  | 47.07  |
| Mean age      | 68.46  | 65.45  |
| Maximum age   | 100.95 | 99.62  |
| Minimum age   | 13.43  | 14.64  |
| Standard dev. | 11.45  | 11.49  |

#### Conventional hip arthroplasty

|               | Female | Male   |
|---------------|--------|--------|
| Number        | 67,526 | 58,870 |
| Percentage    | 53.42  | 46.58  |
| Mean age      | 68.50  | 65.77  |
| Maximum age   | 100.95 | 99.62  |
| Minimum age   | 13.43  | 14.64  |
| Standard dev. | 11.44  | 11.37  |

### Resurfacing hip arthroplasty

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 259    | 1,618 |
| Percentage    | 13.80  | 86.20 |
| Mean age      | 50.00  | 52.18 |
| Maximum age   | 65.88  | 81.44 |
| Minimum age   | 25.72  | 17.74 |
| Standard dev. | 7.22   | 8.54  |

### Body Mass Index


For the nine year period 2010 – 2018 there were 49,340 BMI registrations for primary hip replacements. The average was 29.00 with a range of 14 – 65 and a standard deviation of 5.68.

### Previous operation

|                   |         |
|-------------------|---------|
| None              | 131,943 |
| Internal fixation | 2,471   |
| Osteotomy         | 684     |
| Arthrodesis       | 92      |

### Diagnosis

|                    |         |
|--------------------|---------|
| Osteoarthritis     | 120,688 |
| Acute fracture NOF | 5,248   |
| Avascular necrosis | 4,105   |




|                         |       |
|-------------------------|-------|
| Developmental dysplasia | 2,910 |
| Rheumatoid arthritis    | 1,660 |
| Old fracture NOF        | 1,609 |
| Other inflammatory      | 919   |
| Tumour                  | 632   |
| Post-acute dislocation  | 353   |

Approach


|                        |        |
|------------------------|--------|
| Posterior              | 90,913 |
| Lateral                | 33,007 |
| Anterior               | 4,623  |
| Minimally invasive     | 1,905  |
| Image guided surgery   | 598    |
| Trochanteric osteotomy | 225    |

Comparison of proportions of cemented vs uncemented vs hybrid by year







Comparison of different bearing surface usage over time




Comparison of head size usage over time







## Comparison usage of standard vs cross linked polyethylene over time



### Bone graft

|                      |       |
|----------------------|-------|
| Femoral autograft    | 238   |
| Femoral allograft    | 48    |
| Femoral synthetic    | 9     |
| Acetabular autograft | 1,035 |
| Acetabular allograft | 129   |
| Acetabular synthetic | 6     |

### Cement

|                      |              |
|----------------------|--------------|
| Femur cemented       | 81,923 (60%) |
| Antibiotic in cement | 56,051 (68%) |
| Acetabulum cemented  | 27,976 (20%) |
| Antibiotic in cement | 17,738 (63%) |

### Systemic antibiotic prophylaxis

|  |               |
|--|---------------|
| Patient number receiving at least one systemic antibiotic: | 131,825 (96%) |
|--|---------------|

### Operating theatre

|              |        |
|--------------|--------|
| Conventional | 83,480 |
| Laminar flow | 51,864 |
| Space suits  | 40,604 |

### ASA Class

This was introduced with the updated forms at the beginning of 2005.

### Definitions

**ASA class 1:** A healthy patient

**ASA class 2:** A patient with mild systemic disease

**ASA class 3:** A patient with severe systemic disease that limits activity but is not incapacitating

**ASA class 4:** A patient with an incapacitating systemic disease that is a constant threat to life

| ASA | Number | Percentage |
|-----|--------|------------|
| 1   | 16,778 | 16         |
| 2   | 61,489 | 59         |
| 3   | 24,314 | 24         |
| 4   | 796    | 1          |

For the fourteen year period 2005 – 2018, there were 103,460 (96%) primary hip procedures with the ASA class recorded.

### Operative time (skin to skin)

|         |            |
|---------|------------|
| Average | 78 minutes |
|---------|------------|

### Surgeon grade

The updated forms introduced in 2005 have separated the updated forms introduced in 2005 have separated advanced trainee into supervised and unsupervised. The following figures are for the thirteen year period 2005 – 2018.

|                               |        |
|-------------------------------|--------|
| Consultant                    | 93,927 |
| Advanced trainee supervised   | 8,876  |
| Advanced trainee unsupervised | 2,847  |
| Basic trainee                 | 1,990  |

### Prosthesis usage

#### Conventional primary hips

##### Top 10 femoral components used in 2018


|                      |       |
|----------------------|-------|
| Exeter V40           | 3,322 |
| Corail               | 1,582 |
| Accolade II          | 519   |
| C-Stem AMT           | 398   |
| Stemsys              | 302   |
| Polarstem uncemented | 297   |
| MS 30                | 269   |
| CPT                  | 250   |
| Twinsys cemented     | 247   |
| Twinsys uncemented   | 245   |

##### Top 10 acetabular components used in 2018


|                     |       |
|---------------------|-------|
| Pinnacle            | 2,350 |
| Trident             | 1,363 |
| RM Pressfit cup     | 1,024 |
| Continuum TM        | 820   |
| Tritanium           | 644   |
| R3 porous           | 542   |
| Fitmore             | 402   |
| Exeter X3           | 338   |
| G7 acetabular shell | 277   |
| Trilogy             | 234   |

##### Top ten combinations used in 2018


| Femur                | Acetabulum      | All Years | 2018  |
|----------------------|-----------------|-----------|-------|
| Corail               | Pinnacle        | 10,351    | 1,471 |
| Exeter V40           | Trident         | 10,390    | 1,049 |
| Exeter V40           | Tritanium       | 3,120     | 418   |
| C-Stem AMT           | Pinnacle        | 2,339     | 356   |
| Exeter V40           | Exeter X3       | 2,103     | 336   |
| Exeter V40           | RM Pressfit cup | 2,385     | 301   |
| Exeter V40           | Pinnacle        | 2,448     | 290   |
| Polarstem uncemented | R3 porous       | 1,517     | 271   |
| Twinsys uncemented   | RM Pressfit cup | 4,855     | 241   |
| Exeter V40           | Continuum TM    | 2,539     | 236   |



### Most used femoral components per year for five years 2014 – 2018



### Most used acetabular components per year for five years 2014 – 2018






**Resurfacing hips components used in 2018**

|     |     |
|-----|-----|
| BHR | 118 |
|-----|-----|

**Resurfacing Components for five years 2014 – 2018**




**Surgeon and Hospital Workload**

**Surgeons**

In 2018, 236 surgeons performed 9,169 total hip replacements, an average of 39 procedures per surgeon.

**Hospitals**

In 2018, primary hip replacement was performed in 51 hospitals, 27 public and 24 private.



## REVISION HIP ARTHROPLASTY

Revision is defined by the Registry as a new operation in a previously replaced hip joint during which one of the components is exchanged, removed, manipulated or added. It includes excision arthroplasty and amputation, but not soft tissue procedures. A two-stage procedure is registered as one revision.

### Data analysis

For the twenty year period January 1999 – December 2018, there were 19,582 revision hip procedures registered.

The average age for a revision hip replacement was 70 years, with a range of 18–100 years.

#### Revision hips

|               | Female | Male   |
|---------------|--------|--------|
| Number        | 9,465  | 10,117 |
| Percentage    | 48.34  | 51.66  |
| Mean age      | 70.45  | 70.01  |
| Maximum age   | 100.28 | 99.83  |
| Minimum age   | 17.52  | 20.57  |
| Standard dev. | 12.02  | 10.95  |

The percentage of revision hips to primary hips is 14%.

## Body Mass Index

For the 9 year period 2010 – 2018, there were 3,329 BMI registrations for revision hip replacements. The average BMI was 28.98 with a range of 15-55 with a standard deviation of 5.69.

## REVISION OF REGISTERED PRIMARY HIP ARTHROPLASTIES

This section analyses data for revisions of **registered primary hip arthroplasties** for the twenty year period.

There were 6,965 revisions of the 135,461 primary conventional hip replacements (5%) and 150 revisions of the 1,877 resurfacing hip replacements (8%) a total of 7,115 revisions.

### Conventional hip arthroplasty analyses

#### Time to revision for conventional hips

|                    |            |
|--------------------|------------|
| Average            | 2,137 days |
| Maximum            | 7,091 days |
| Minimum            | 0 days     |
| Standard deviation | 1,828 days |

#### Reason for revision

|                                |       |
|--------------------------------|-------|
| Reason for revision            |       |
| Dislocation                    | 1,489 |
| Loosening acetabular component | 1,476 |
| Loosening femoral component    | 1,169 |
| Pain                           | 1,012 |
| Deep infection                 | 865   |
| Fracture femur                 | 799   |

### Analysis of the six main reasons for revision by year after primary procedure

| Years | Dislocation |      | Loosening Acetabular |      | Loosening Femoral |      | Deep infection |      | Pain  |      | Fracture Femur |      |
|-------|-------------|------|----------------------|------|-------------------|------|----------------|------|-------|------|----------------|------|
|       | Count       | %    | Count                | %    | Count             | %    | Count          | %    | Count | %    | Count          | %    |
| 0     | 629         | 42.2 | 156                  | 10.6 | 104               | 8.9  | 375            | 43.4 | 79    | 7.8  | 270            | 33.8 |
| 1     | 171         | 11.5 | 79                   | 5.4  | 85                | 7.3  | 102            | 11.8 | 102   | 10.1 | 48             | 6.0  |
| 2     | 123         | 8.3  | 76                   | 5.1  | 78                | 6.7  | 78             | 9.0  | 90    | 8.9  | 45             | 5.6  |
| 3     | 99          | 6.6  | 86                   | 5.8  | 72                | 6.2  | 50             | 5.8  | 72    | 7.1  | 38             | 4.8  |
| 4     | 63          | 4.2  | 69                   | 4.7  | 68                | 5.8  | 39             | 4.5  | 68    | 6.7  | 54             | 6.8  |
| 5     | 67          | 4.5  | 75                   | 5.1  | 66                | 5.6  | 34             | 3.9  | 74    | 7.3  | 37             | 4.6  |
| 6     | 59          | 4.0  | 94                   | 6.4  | 90                | 7.7  | 28             | 3.2  | 69    | 6.8  | 31             | 3.9  |
| 7     | 44          | 3.0  | 85                   | 5.8  | 86                | 7.4  | 26             | 3.0  | 56    | 5.5  | 38             | 4.8  |
| 8     | 53          | 3.6  | 99                   | 6.7  | 72                | 6.2  | 28             | 3.2  | 63    | 6.2  | 42             | 5.3  |
| 9     | 32          | 2.1  | 113                  | 7.7  | 69                | 5.9  | 30             | 3.5  | 56    | 5.5  | 39             | 4.9  |
| 10    | 29          | 1.9  | 82                   | 5.6  | 88                | 7.5  | 19             | 2.2  | 51    | 5.0  | 36             | 4.5  |
| >10   | 120         | 8.1  | 462                  | 31.3 | 291               | 24.9 | 56             | 6.5  | 232   | 22.9 | 121            | 15.1 |
| Total | 1,489       | 100  | 1,476                | 100  | 1,169             | 100  | 865            | 100  | 1,012 | 100  | 799            | 100  |




### Analyses of numbers of the six main reasons for revision by year


|                  | Dislocation | Loosening Acetabular | Loosening Femoral | Deep infection | Pain | Fracture Femur |
|------------------|-------------|----------------------|-------------------|----------------|------|----------------|
|                  | No.         | No.                  | No.               | No.            | No.  | No.            |
| <b>1999-2007</b> | 450         | 239                  | 182               | 177            | 106  | 91             |
| <b>2008</b>      | 82          | 88                   | 64                | 37             | 33   | 40             |
| <b>2009</b>      | 81          | 108                  | 75                | 37             | 38   | 43             |
| <b>2010</b>      | 87          | 104                  | 79                | 49             | 67   | 45             |
| <b>2011</b>      | 106         | 116                  | 88                | 45             | 106  | 53             |
| <b>2012</b>      | 91          | 126                  | 88                | 46             | 97   | 52             |
| <b>2013</b>      | 94          | 130                  | 102               | 61             | 110  | 54             |
| <b>2014</b>      | 87          | 104                  | 96                | 62             | 74   | 72             |
| <b>2015</b>      | 102         | 125                  | 102               | 89             | 101  | 79             |
| <b>2016</b>      | 105         | 110                  | 95                | 81             | 83   | 89             |
| <b>2017</b>      | 102         | 110                  | 99                | 84             | 106  | 95             |
| <b>2018</b>      | 101         | 114                  | 99                | 97             | 91   | 86             |

### Analyses of the percentages of the six main reasons for revision by year

|                  | Dislocation | Loosening Acetabular | Loosening Femoral | Deep infection | Pain | Fracture Femur |
|------------------|-------------|----------------------|-------------------|----------------|------|----------------|
|                  | %           | %                    | %                 | %              | %    | %              |
| <b>1999-2007</b> | 37.9        | 20.1                 | 15.3              | 14.9           | 8.9  | 7.7            |
| <b>2008</b>      | 24.9        | 26.7                 | 19.5              | 11.2           | 10.0 | 12.2           |
| <b>2009</b>      | 22.2        | 29.6                 | 20.5              | 10.1           | 10.4 | 11.8           |
| <b>2010</b>      | 21.5        | 25.7                 | 19.6              | 12.1           | 16.6 | 11.1           |
| <b>2011</b>      | 20.7        | 22.6                 | 17.2              | 8.8            | 20.7 | 10.3           |
| <b>2012</b>      | 17.3        | 23.9                 | 16.7              | 8.7            | 18.4 | 9.9            |
| <b>2013</b>      | 15.9        | 21.9                 | 17.2              | 10.3           | 18.5 | 9.1            |
| <b>2014</b>      | 15.6        | 18.6                 | 17.2              | 11.1           | 13.2 | 12.9           |
| <b>2015</b>      | 16.4        | 20.1                 | 16.4              | 14.3           | 16.3 | 12.7           |
| <b>2016</b>      | 17.0        | 17.8                 | 15.4              | 13.1           | 13.4 | 14.4           |
| <b>2017</b>      | 16.6        | 17.9                 | 16.2              | 13.2           | 17.3 | 15.5           |
| <b>2018</b>      | 16.1        | 18.1                 | 15.7              | 15.4           | 14.5 | 13.7           |



### Comparison of the 6 main reasons for revision over time





## RESURFACED HIP ANALYSES

There were 1,877 resurfacing hips registered for the period 2000 – 2018, and 150 (8%) have been revised.

### Time to revision for resurfaced hips

|                    |            |
|--------------------|------------|
| Average            | 2,016 days |
| Maximum            | 4,430 days |
| Minimum            | 10 days    |
| Standard deviation | 1,157 days |

### Reason for revision

|                             |    |
|-----------------------------|----|
| Pain                        | 48 |
| Loosening acetabulum        | 17 |
| Deep infection              | 17 |
| Loosening femoral component | 16 |
| Fracture femur              | 15 |
| Dislocation                 | 2  |

### Statistical note

In the tables below there are two statistical terms readers may not be familiar with:

#### i) Observed component years

This is the number of registered primary procedures multiplied by the number of years each component has been in place.


#### ii) Rate/100 component years

This is equivalent to the yearly revision rate expressed as a percentage and is derived by dividing the number of prostheses revised by the observed component years multiplied by 100. It therefore allows for the number of years of post-operative follow up in calculating the revision rate. These rates are usually very low, hence it is expressed per 100 component years rather than per component year. Statisticians consider that this is a more accurate way of deriving a revision rate for comparison when analysing data with widely varying follow up times. It is also important to note the confidence intervals. The closer they are to the estimated revision rate/100 component years, the more precise the estimate is.

### Statistical Significance

Where it is stated that a difference among results is significant the p value is 0.05 or less. In most of these situations this is because there is no overlap of the confidence intervals (CI's) but sometimes significance can apply in the presence of CI overlap.






## Conventional Primary Hip Arthroplasties

### All Primary Total Hip Arthroplasties


| No. Ops. | Observed comp. Yrs | Number Revised | Rate/100-component-years | Exact 95% Confidence Interval |      |
|----------|--------------------|----------------|--------------------------|-------------------------------|------|
| 135,461  | 972,138            | 6,965          | 0.72                     | 0.70                          | 0.73 |


The figure below summarises the 30 Hip prostheses combinations with >1000 procedures. Showing the number of procedures for the history of the Registry and for the previous 2 years.






The figure below summarises the 30 Hip prostheses combinations with >1000 procedures. Showing the number of procedures for the previous 2 years and the historical revision rate.







The figure below summarises the 16 Hip femur prostheses with >2000 procedures. Showing the number of procedures for the history of the Registry and for the previous 2 years.






The figure below summarises the 16 Hip femur prostheses with >2000 procedures. Showing the number of procedures for the previous 2 years and the historical revision rate.







The figure below summarises the 24 Hip acetabular prostheses with >1000 procedures. Showing the number of procedures for the history of the Registry and for the previous 2 years.





The figure below summarises the 24 Hip acetabular prostheses with >1000 procedures. Showing the number of procedures for the previous 2 years and the historical revision rate.





| Femur Prosthesis         | Acetabular Prosthesis  | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |       | Procedures 2018 |
|--------------------------|------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|-----------------|
| Anthology Porous         | R3 porous              | 68      | 486.2              | 32             | 6.58                     | 4.42                          | 9.17  | 1               |
| MasterSL                 | Delta-TT Cup           | 59      | 52.6               | 2              | 3.81                     | 0.46                          | 13.75 | 36              |
| CPT                      | G7 acetabular shell    | 58      | 95.0               | 3              | 3.16                     | 0.65                          | 9.23  | 26              |
| Taperloc Complete        | Continuum TM           | 108     | 101.1              | 3              | 2.97                     | 0.61                          | 8.67  | 63              |
| ABGII                    | RM Pressfit cup        | 81      | 218.5              | 6              | 2.75                     | 1.01                          | 5.98  | 12              |
| Twinsys cemented         | Pinnacle               | 76      | 286.1              | 7              | 2.45                     | 0.98                          | 5.04  | 8               |
| Accolade II              | RM Pressfit cup        | 79      | 126.2              | 3              | 2.38                     | 0.49                          | 6.95  | 27              |
| Taperloc Complete        | RM Pressfit cup        | 168     | 266.9              | 5              | 1.87                     | 0.61                          | 4.37  | 44              |
| CPT                      | Delta-TT Cup           | 79      | 182.5              | 3              | 1.64                     | 0.34                          | 4.80  | 18              |
| Echo Bi-Metric           | Continuum TM           | 106     | 183.8              | 3              | 1.63                     | 0.34                          | 4.77  | 26              |
| C-Stem                   | Pinnacle               | 77      | 260.2              | 4              | 1.54                     | 0.42                          | 3.94  | 12              |
| Taperloc Complete        | G7 acetabular shell    | 259     | 390.5              | 6              | 1.54                     | 0.56                          | 3.34  | 96              |
| H-Max C                  | Delta-TT Cup           | 61      | 134.1              | 2              | 1.49                     | 0.18                          | 5.39  | 21              |
| Exeter V40               | G7 acetabular shell    | 112     | 212.7              | 3              | 1.41                     | 0.29                          | 4.12  | 39              |
| Friendly                 | Delta-TT Cup           | 67      | 410.6              | 5              | 1.22                     | 0.40                          | 2.84  | 1               |
| Exeter V40               | Trabecular Metal Shell | 211     | 1079.1             | 13             | 1.20                     | 0.64                          | 2.06  | 14              |
| Spectron                 | Reflection cemented    | 2956    | 29,689.6           | 353            | 1.19                     | 1.07                          | 1.32  | 2               |
| H-Max S                  | Delta-PF Cup           | 194     | 591.6              | 7              | 1.18                     | 0.42                          | 2.32  | 10              |
| Stemsys                  | Polymax                | 119     | 254.4              | 3              | 1.18                     | 0.24                          | 3.45  | 20              |
| Avenir Muller uncemented | Continuum TM           | 179     | 1,074.6            | 11             | 1.02                     | 0.51                          | 1.83  | 1               |
| Corail                   | RM Pressfit cup        | 144     | 504.8              | 5              | 0.99                     | 0.32                          | 2.31  | 7               |
| S-Rom                    | Pinnacle               | 375     | 3,633.1            | 35             | 0.96                     | 0.66                          | 1.32  | 8               |
| CPT                      | Fitmore                | 191     | 1,053.9            | 10             | 0.95                     | 0.46                          | 1.75  | 10              |
| CPT                      | Continuum TM           | 1387    | 4,915.7            | 46             | 0.94                     | 0.69                          | 1.25  | 191             |
| Stemsys                  | Agilis Ti-por          | 439     | 1,501.4            | 14             | 0.93                     | 0.49                          | 1.52  | 46              |
| H-Max S                  | Delta-TT Cup           | 737     | 2,937.5            | 27             | 0.92                     | 0.59                          | 1.32  | 48              |
| CPT                      | Trilogy                | 843     | 6,459.6            | 56             | 0.87                     | 0.65                          | 1.13  | 1               |
| C-Stem AMT               | RM Pressfit cup        | 129     | 467.7              | 4              | 0.86                     | 0.23                          | 2.19  | 23              |
| Exeter V40               | Continuum TM           | 2539    | 10,125.6           | 86             | 0.85                     | 0.68                          | 1.05  | 236             |
| Trabecular Metal Stem    | Continuum TM           | 447     | 2,151.9            | 18             | 0.84                     | 0.50                          | 1.32  | 10              |
| CLS                      | Titanium               | 80      | 362.3              | 3              | 0.83                     | 0.17                          | 2.42  | 4               |
| M/L Taper                | Continuum TM           | 1004    | 4,120.2            | 34             | 0.83                     | 0.57                          | 1.15  | 123             |
| C-Stem AMT               | Pinnacle               | 2339    | 8,132.8            | 66             | 0.81                     | 0.62                          | 1.03  | 356             |
| Accolade II              | Delta-TT Cup           | 72      | 125.6              | 1              | 0.80                     | 0.02                          | 4.43  | 18              |
| Corail                   | Fitmore                | 285     | 893.5              | 7              | 0.78                     | 0.31                          | 1.61  | 17              |
| CBC                      | RM Pressfit cup        | 444     | 2,688.8            | 21             | 0.78                     | 0.48                          | 1.19  | 12              |
| CLS                      | Continuum TM           | 719     | 2,920.8            | 22             | 0.75                     | 0.47                          | 1.14  | 75              |
| Summit                   | Pinnacle               | 2248    | 1,3418.1           | 101            | 0.75                     | 0.61                          | 0.91  | 185             |



| Femur Prosthesis         | Acetabular Prosthesis  | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      | Procedures 2018 |
|--------------------------|------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|-----------------|
| Accolade II              | Trident                | 858     | 2,158.4            | 16             | 0.74                     | 0.42                          | 1.20 | 219             |
| Synergy Porous           | R3 porous              | 1779    | 7,919.5            | 58             | 0.73                     | 0.56                          | 0.95 | 126             |
| Corail                   | Continuum TM           | 304     | 1,099.0            | 8              | 0.73                     | 0.31                          | 1.43 | 35              |
| Optimys                  | RM Pressfit cup        | 106     | 140.0              | 1              | 0.71                     | 0.02                          | 3.98 | 46              |
| CLS                      | Trabecular Metal Shell | 53      | 420.1              | 3              | 0.71                     | 0.15                          | 2.09 | 2               |
| Echo Bi-Metric           | G7 acetabular shell    | 320     | 722.1              | 5              | 0.69                     | 0.22                          | 1.62 | 95              |
| Exeter V40               | R3 porous              | 605     | 2,186.6            | 15             | 0.69                     | 0.38                          | 1.13 | 70              |
| Corail                   | Pinnacle               | 10351   | 48,382.9           | 331            | 0.68                     | 0.61                          | 0.76 | 1,471           |
| CPCS                     | R3 porous              | 326     | 889.9              | 6              | 0.67                     | 0.21                          | 1.39 | 63              |
| CLS                      | Reflection porous      | 373     | 2,990.9            | 20             | 0.67                     | 0.41                          | 1.03 | 15              |
| M/L Taper                | Trident                | 249     | 748.4              | 5              | 0.67                     | 0.22                          | 1.56 | 52              |
| Exeter V40               | Tritanium              | 3120    | 11,169.7           | 74             | 0.66                     | 0.52                          | 0.83 | 418             |
| CLS                      | RM Pressfit cup        | 567     | 3,853.8            | 25             | 0.65                     | 0.42                          | 0.96 | 27              |
| Twinsys uncemented       | RM Pressfit cup        | 4855    | 29,448.8           | 191            | 0.65                     | 0.56                          | 0.75 | 241             |
| Exeter V40               | Delta-TT Cup           | 223     | 771.1              | 5              | 0.65                     | 0.21                          | 1.51 | 34              |
| Twinsys cemented         | CCB                    | 441     | 2,331.4            | 15             | 0.64                     | 0.36                          | 1.06 | 14              |
| CPT                      | ZCA                    | 545     | 5,487.9            | 35             | 0.64                     | 0.44                          | 0.89 | 2               |
| Avenir Muller uncemented | RM Pressfit cup        | 51      | 158.0              | 1              | 0.63                     | 0.02                          | 3.53 | 2               |
| Corail                   | Trident                | 88      | 480.8              | 3              | 0.62                     | 0.13                          | 1.82 | 8               |
| CLS                      | Trilogy                | 596     | 4,236.2            | 25             | 0.59                     | 0.38                          | 0.87 | 24              |
| Accolade II              | Tritanium              | 935     | 2,381.4            | 14             | 0.59                     | 0.32                          | 0.99 | 200             |
| C-Stem AMT               | Marathon cemented      | 330     | 1,763.1            | 10             | 0.57                     | 0.27                          | 1.04 | 14              |
| Polarstem uncemented     | R3 porous              | 1517    | 4,554.6            | 25             | 0.55                     | 0.36                          | 0.81 | 271             |
| Exeter V40               | Exeter                 | 1639    | 15,405.7           | 84             | 0.55                     | 0.43                          | 0.68 | 1               |
| Lateral straight stem    | Muller PE cup          | 750     | 7,185.8            | 39             | 0.54                     | 0.39                          | 0.74 | 1               |
| Stemsys                  | Fixa Ti Por            | 707     | 2,700.6            | 14             | 0.52                     | 0.27                          | 0.85 | 77              |
| CCA                      | CCB                    | 769     | 6,078.6            | 31             | 0.51                     | 0.34                          | 0.71 | 4               |
| Avenir Muller uncemented | Fitmore                | 65      | 196.3              | 1              | 0.51                     | 0.01                          | 2.84 | 27              |
| Wagner cone stem         | Fitmore                | 73      | 803.3              | 4              | 0.50                     | 0.14                          | 1.27 | 1               |
| CLS                      | Fitmore                | 2300    | 23,622.5           | 117            | 0.50                     | 0.41                          | 0.59 | 34              |
| Twinsys cemented         | RM Pressfit cup        | 1864    | 8,557.7            | 42             | 0.49                     | 0.35                          | 0.66 | 197             |
| Corail                   | Tritanium              | 168     | 822.6              | 4              | 0.49                     | 0.13                          | 1.24 | 15              |
| M/L Taper                | Trilogy                | 215     | 1,853.2            | 9              | 0.49                     | 0.20                          | 0.89 | 3               |
| Exeter V40               | Contemporary           | 6504    | 49,605.6           | 235            | 0.47                     | 0.42                          | 0.54 | 114             |
| Summit                   | Trilogy                | 168     | 1,286.5            | 6              | 0.47                     | 0.17                          | 1.02 | 11              |
| Exeter V40               | Pinnacle               | 2448    | 11,707.0           | 53             | 0.45                     | 0.34                          | 0.59 | 290             |






| Femur Prosthesis       | Acetabular Prosthesis | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      | Procedures 2018 |
|------------------------|-----------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|-----------------|
| MS 30                  | Continuum TM          | 404     | 1,595.9            | 7              | 0.44                     | 0.16                          | 0.86 | 33              |
| Exeter V40             | CCB                   | 567     | 3,018.0            | 13             | 0.43                     | 0.23                          | 0.74 | 33              |
| Exeter V40             | Trident               | 10390   | 63,631.7           | 274            | 0.43                     | 0.38                          | 0.48 | 1049            |
| Stemsys                | RM Pressfit cup       | 325     | 1,189.5            | 5              | 0.42                     | 0.14                          | 0.98 | 38              |
| Exeter V40             | Reflection cemented   | 926     | 5,623.2            | 23             | 0.41                     | 0.26                          | 0.61 | 35              |
| Exeter V40             | Trilogy               | 2805    | 19,700.1           | 78             | 0.40                     | 0.31                          | 0.49 | 149             |
| Exeter V40             | Exeter X3             | 2103    | 7,355.7            | 28             | 0.38                     | 0.25                          | 0.54 | 336             |
| Standard straight stem | Muller PE cup stem    | 629     | 5,764.2            | 21             | 0.36                     | 0.22                          | 0.55 | 1               |
| Synergy Porous         | Reflection porous     | 1238    | 11,875.0           | 43             | 0.36                     | 0.26                          | 0.48 | 14              |
| Exeter V40             | RM Pressfit cup       | 2385    | 10,968.5           | 37             | 0.34                     | 0.23                          | 0.46 | 301             |
| CLS                    | Pinnacle              | 90      | 614.1              | 2              | 0.33                     | 0.04                          | 1.18 | 10              |
| Corail                 | Trilogy               | 207     | 953.4              | 3              | 0.31                     | 0.06                          | 0.92 | 17              |
| Spectron               | R3 porous             | 432     | 2,439.2            | 7              | 0.29                     | 0.10                          | 0.56 | 11              |
| MS 30                  | Fitmore               | 2174    | 14,197.5           | 39             | 0.27                     | 0.20                          | 0.38 | 192             |
| Stemsys                | DeltaMotion Cup       | 485     | 2,413.5            | 6              | 0.25                     | 0.08                          | 0.51 | 44              |
| Twinsys cemented       | Continuum TM          | 108     | 408.7              | 1              | 0.24                     | 0.01                          | 1.36 | 6               |
| MS 30                  | Trilogy               | 331     | 1,958.0            | 4              | 0.20                     | 0.06                          | 0.52 | 19              |
| Exeter V40             | ZCA                   | 93      | 574.3              | 1              | 0.17                     | 0.00                          | 0.97 | 8               |
| Exeter V40             | Fitmore               | 963     | 4,733.5            | 8              | 0.17                     | 0.07                          | 0.33 | 91              |
| Stemsys                | Delta-PF Cup          | 396     | 1,077.1            | 1              | 0.09                     | 0.00                          | 0.43 | 74              |
| Exeter V40             | Polymax               | 63      | 61.3               | 0              | 0.00                     | 0.00                          | 6.02 | 35              |
| Stemsys cemented       | RM Pressfit cup       | 64      | 142.1              | 0              | 0.00                     | 0.00                          | 2.60 | 18              |
| C-Stem                 | Marathon cemented     | 89      | 310.1              | 0              | 0.00                     | 0.00                          | 1.19 | 14              |
| Twinsys cemented       | Reflection porous     | 59      | 174.7              | 0              | 0.00                     | 0.00                          | 2.11 | 9               |
| Synergy Porous         | Continuum TM          | 55      | 120.5              | 0              | 0.00                     | 0.00                          | 3.06 | 6               |
| Exeter V40             | ZCA all-poly cup      | 104     | 432.0              | 0              | 0.00                     | 0.00                          | 0.85 | 4               |

**Revision versus hip prostheses combinations sorted on revision rate**  
(minimum of 50 primary registered arthroplasties)


| Femur Prosthesis      | Acetabular Prosthesis | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |       | Procedures 2018 |
|-----------------------|-----------------------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|-----------------|
| S-Rom                 | ASR                   | 130     | 801.1              | 94             | 11.73                    | 9.43                          | 14.29 | 0               |
| Coral                 | ASR                   | 156     | 1,199.7            | 83             | 6.92                     | 5.51                          | 8.58  | 0               |
| Anthology Porous      | BHR Acetabular Cup    | 93      | 726.3              | 50             | 6.88                     | 5.11                          | 9.08  | 0               |
| Anthology Porous      | R3 porous             | 68      | 486.2              | 32             | 6.58                     | 4.42                          | 9.17  | 1               |
| Summit                | ASR                   | 88      | 741.0              | 36             | 4.86                     | 3.40                          | 6.73  | 0               |
| Synergy Porous        | BHR Acetabular Cup    | 114     | 1,034.2            | 40             | 3.87                     | 2.76                          | 5.27  | 0               |
| MasterSL              | Delta-TT Cup          | 59      | 52.6               | 2              | 3.81                     | 0.46                          | 13.75 | 36              |
| CLS                   | Artek                 | 59      | 712.1              | 25             | 3.51                     | 2.22                          | 5.10  | 0               |
| CLS                   | Durom                 | 198     | 1,950.2            | 63             | 3.23                     | 2.48                          | 4.13  | 0               |
| CPT                   | G7 acetabular shell   | 58      | 95.0               | 3              | 3.16                     | 0.65                          | 9.23  | 26              |
| Taperloc Complete     | Continuum TM          | 108     | 101.1              | 3              | 2.97                     | 0.61                          | 8.67  | 63              |
| ABGII                 | RM Pressfit cup       | 81      | 218.5              | 6              | 2.75                     | 1.01                          | 5.98  | 12              |
| Twinsys cemented      | Pinnacle              | 76      | 286.1              | 7              | 2.45                     | 0.98                          | 5.04  | 8               |
| Accolade II           | RM Pressfit cup       | 79      | 126.2              | 3              | 2.38                     | 0.49                          | 6.95  | 27              |
| Lateral straight stem | Trilogy               | 69      | 555.6              | 12             | 2.16                     | 1.12                          | 3.77  | 0               |
| ABGII                 | Duraloc               | 139     | 1,929.9            | 40             | 2.07                     | 1.46                          | 2.79  | 0               |
| ABG                   | Duraloc               | 116     | 1,860.4            | 38             | 2.04                     | 1.42                          | 2.77  | 0               |
| Taperloc Complete     | RM Pressfit cup       | 168     | 266.9              | 5              | 1.87                     | 0.61                          | 4.37  | 44              |
| CPT                   | Delta-TT Cup          | 79      | 182.5              | 3              | 1.64                     | 0.34                          | 4.80  | 18              |
| Echo Bi-Metric        | Continuum TM          | 106     | 183.8              | 3              | 1.63                     | 0.34                          | 4.77  | 26              |
| Elite plus            | Duraloc               | 608     | 7,022.0            | 112            | 1.59                     | 1.31                          | 1.92  | 0               |
| Prodigy               | Duraloc               | 113     | 1,479.9            | 23             | 1.55                     | 0.99                          | 2.33  | 0               |
| M/L Taper             | Delta-TT Cup          | 64      | 323.4              | 5              | 1.55                     | 0.50                          | 3.61  | 0               |
| C-Stem                | Pinnacle              | 77      | 260.2              | 4              | 1.54                     | 0.42                          | 3.94  | 12              |
| Taperloc Complete     | G7 acetabular shell   | 259     | 390.5              | 6              | 1.54                     | 0.56                          | 3.34  | 96              |
| H-Max C               | Delta-TT Cup          | 61      | 134.1              | 2              | 1.49                     | 0.18                          | 5.39  | 21              |
| ABG                   | ABGII                 | 72      | 1,098.6            | 16             | 1.46                     | 0.83                          | 2.37  | 0               |
| CLS                   | RM cup                | 113     | 1,173.0            | 17             | 1.45                     | 0.81                          | 2.27  | 0               |
| Exeter                | Duraloc               | 553     | 7,808.0            | 113            | 1.45                     | 1.19                          | 1.74  | 0               |
| Exeter V40            | G7 acetabular shell   | 112     | 212.7              | 3              | 1.41                     | 0.29                          | 4.12  | 39              |
| CBC                   | Expansys shell        | 183     | 1,791.9            | 25             | 1.40                     | 0.90                          | 2.06  | 0               |
| Contemporary          | Contemporary          | 71      | 912.4              | 12             | 1.32                     | 0.68                          | 2.30  | 0               |
| CCA                   | Contemporary          | 74      | 762.2              | 10             | 1.31                     | 0.63                          | 2.41  | 0               |
| H-Max M               | Delta-PF Cup          | 71      | 535.0              | 7              | 1.31                     | 0.47                          | 2.57  | 0               |
| Twinsys uncemented    | Selexys TPS           | 1231    | 10,413.9           | 136            | 1.31                     | 1.10                          | 1.54  | 0               |
| Spectron              | Duraloc               | 1151    | 13,752.9           | 177            | 1.29                     | 1.10                          | 1.49  | 0               |
| Friendly              | Delta-TT Cup          | 67      | 410.6              | 5              | 1.22                     | 0.40                          | 2.84  | 1               |



| Femur Prosthesis         | Acetabular Prosthesis    | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      | Procedures 2018 |
|--------------------------|--------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|-----------------|
| Spectron                 | Muller PE cup            | 66      | 658.5              | 8              | 1.21                     | 0.52                          | 2.39 | 0               |
| Mallory-Head             | M2A                      | 105     | 1,237.7            | 15             | 1.21                     | 0.68                          | 2.00 | 0               |
| Exeter V40               | Trabecular Metal Shell   | 211     | 1,079.1            | 13             | 1.20                     | 0.64                          | 2.06 | 14              |
| Spectron                 | Reflection cemented      | 2956    | 29,689.6           | 353            | 1.19                     | 1.07                          | 1.32 | 2               |
| H-Max S                  | Delta-PF Cup             | 194     | 591.6              | 7              | 1.18                     | 0.42                          | 2.32 | 10              |
| Stemsys                  | Polymax                  | 119     | 254.4              | 3              | 1.18                     | 0.24                          | 3.45 | 20              |
| S-Rom                    | Ultima                   | 78      | 1,250.4            | 14             | 1.12                     | 0.61                          | 1.88 | 0               |
| Spectron                 | Morscher                 | 210     | 2,769.6            | 31             | 1.12                     | 0.76                          | 1.59 | 0               |
| Twinsys cemented         | Selexys TPS              | 65      | 458.4              | 5              | 1.09                     | 0.35                          | 2.55 | 0               |
| AML MMA                  | Duraloc                  | 74      | 1,034.3            | 11             | 1.06                     | 0.50                          | 1.84 | 0               |
| CLS                      | Allofit                  | 192     | 1,924.0            | 20             | 1.04                     | 0.61                          | 1.57 | 0               |
| Avenir Muller uncemented | Continuum TM             | 179     | 1,074.6            | 11             | 1.02                     | 0.51                          | 1.83 | 1               |
| CLS                      | Duraloc                  | 699     | 9,128.1            | 91             | 1.00                     | 0.80                          | 1.22 | 0               |
| Exeter                   | Contemporary             | 1551    | 18,301.8           | 182            | 0.99                     | 0.85                          | 1.15 | 0               |
| Corail                   | RM Pressfit cup          | 144     | 504.8              | 5              | 0.99                     | 0.32                          | 2.31 | 7               |
| CPT                      | Tritanium                | 85      | 611.2              | 6              | 0.98                     | 0.36                          | 2.14 | 0               |
| S-Rom                    | Pinnacle                 | 375     | 3,633.1            | 35             | 0.96                     | 0.66                          | 1.32 | 8               |
| Corail                   | Duraloc                  | 464     | 5,139.0            | 49             | 0.95                     | 0.71                          | 1.26 | 0               |
| CPT                      | Fitmore                  | 191     | 1,053.9            | 10             | 0.95                     | 0.46                          | 1.75 | 10              |
| C-Stem                   | Duraloc                  | 53      | 634.1              | 6              | 0.95                     | 0.35                          | 2.06 | 0               |
| CPT                      | Continuum TM             | 1387    | 4,915.7            | 46             | 0.94                     | 0.69                          | 1.25 | 191             |
| Stemsys                  | Agilis Ti-por            | 439     | 1,501.4            | 14             | 0.93                     | 0.49                          | 1.52 | 46              |
| Exeter V40               | Duraloc                  | 987     | 10,786.5           | 100            | 0.93                     | 0.75                          | 1.12 | 0               |
| H-Max S                  | Delta-TT Cup             | 737     | 2,937.5            | 27             | 0.92                     | 0.59                          | 1.32 | 48              |
| AML                      | Duraloc                  | 53      | 780.7              | 7              | 0.90                     | 0.36                          | 1.85 | 0               |
| SL modular stem          | RM cup                   | 322     | 4,578.4            | 40             | 0.87                     | 0.62                          | 1.19 | 0               |
| Twinsys uncemented       | RM cup                   | 122     | 1,036.2            | 9              | 0.87                     | 0.40                          | 1.65 | 0               |
| CPT                      | Trilogy                  | 843     | 6,459.6            | 56             | 0.87                     | 0.65                          | 1.13 | 1               |
| ABGII                    | Delta-PF Cup             | 107     | 1,273.7            | 11             | 0.86                     | 0.40                          | 1.49 | 0               |
| C-Stem AMT               | RM Pressfit cup          | 129     | 467.7              | 4              | 0.86                     | 0.23                          | 2.19 | 23              |
| CPT                      | Monoblock Acetabular Cup | 84      | 937.6              | 8              | 0.85                     | 0.33                          | 1.61 | 0               |
| Exeter V40               | Continuum TM             | 2539    | 10,125.6           | 86             | 0.85                     | 0.68                          | 1.05 | 236             |
| ABGII                    | Trident                  | 342     | 4,038.1            | 34             | 0.84                     | 0.58                          | 1.18 | 0               |
| CBC                      | Fitmore                  | 59      | 597.2              | 5              | 0.84                     | 0.27                          | 1.95 | 0               |
| Trabecular Metal Stem    | Continuum TM             | 447     | 2,151.9            | 18             | 0.84                     | 0.50                          | 1.32 | 10              |
| CLS                      | Tritanium                | 80      | 362.3              | 3              | 0.83                     | 0.17                          | 2.42 | 4               |
| M/L Taper                | Continuum TM             | 1004    | 4,120.2            | 34             | 0.83                     | 0.57                          | 1.15 | 123             |




| Femur Prosthesis         | Acetabular Prosthesis    | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      | Procedures 2018 |
|--------------------------|--------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|-----------------|
| Corail                   | Pinnacle                 | 10351   | 48,382.9           | 331            | 0.68                     | 0.61                          | 0.76 | 1471            |
| Polarstem uncemented     | Reflection porous        | 335     | 2,072.2            | 14             | 0.68                     | 0.37                          | 1.13 | 0               |
| Exeter                   | Osteolock                | 836     | 11,108.7           | 75             | 0.68                     | 0.53                          | 0.85 | 0               |
| CPCS                     | R3 porous                | 326     | 889.9              | 6              | 0.67                     | 0.21                          | 1.39 | 63              |
| CLS                      | Reflection porous        | 373     | 2,990.9            | 20             | 0.67                     | 0.41                          | 1.03 | 15              |
| M/L Taper                | Trident                  | 249     | 748.4              | 5              | 0.67                     | 0.22                          | 1.56 | 52              |
| Exeter V40               | Tritanium                | 3120    | 11,169.7           | 74             | 0.66                     | 0.52                          | 0.83 | 418             |
| CLS                      | RM Pressfit cup          | 567     | 3,853.8            | 25             | 0.65                     | 0.42                          | 0.96 | 27              |
| Twinsys uncemented       | RM Pressfit cup          | 4855    | 29,448.8           | 191            | 0.65                     | 0.56                          | 0.75 | 241             |
| Exeter V40               | Delta-TT Cup             | 223     | 771.1              | 5              | 0.65                     | 0.21                          | 1.51 | 34              |
| CPT                      | Duraloc                  | 212     | 2,467.9            | 16             | 0.65                     | 0.37                          | 1.05 | 0               |
| Twinsys cemented         | CCB                      | 441     | 2,331.4            | 15             | 0.64                     | 0.36                          | 1.06 | 14              |
| CPT                      | ZCA                      | 545     | 5,487.9            | 35             | 0.64                     | 0.44                          | 0.89 | 2               |
| Exeter                   | CLS Expansion            | 129     | 1,578.9            | 10             | 0.63                     | 0.30                          | 1.16 | 0               |
| Avenir Muller uncemented | RM Pressfit cup          | 51      | 158.0              | 1              | 0.63                     | 0.02                          | 3.53 | 2               |
| Elite plus               | Charnley                 | 298     | 3,661.5            | 23             | 0.63                     | 0.40                          | 0.94 | 0               |
| MS 30                    | Morscher                 | 787     | 9,737.6            | 61             | 0.63                     | 0.48                          | 0.80 | 0               |
| CLS                      | Weill ring               | 106     | 1,601.3            | 10             | 0.62                     | 0.28                          | 1.11 | 0               |
| Corail                   | Trident                  | 88      | 480.8              | 3              | 0.62                     | 0.13                          | 1.82 | 8               |
| Versys cemented          | ZCA                      | 391     | 4200.1             | 26             | 0.62                     | 0.40                          | 0.91 | 0               |
| Twinsys uncemented       | Continuum TM             | 133     | 821.0              | 5              | 0.61                     | 0.20                          | 1.42 | 0               |
| Twinsys uncemented       | Trilogy                  | 209     | 1,811.8            | 11             | 0.61                     | 0.30                          | 1.09 | 0               |
| CLS                      | Trilogy                  | 596     | 4,236.2            | 25             | 0.59                     | 0.38                          | 0.87 | 24              |
| Accolade II              | Tritanium                | 935     | 2,381.4            | 14             | 0.59                     | 0.32                          | 0.99 | 200             |
| Accolade                 | Muller PE cup            | 114     | 1,206.7            | 7              | 0.58                     | 0.21                          | 1.14 | 0               |
| C-Stem AMT               | Marathon cemented        | 330     | 1,763.1            | 10             | 0.57                     | 0.27                          | 1.04 | 14              |
| Spectron                 | Trident                  | 78      | 886.1              | 5              | 0.56                     | 0.15                          | 1.24 | 0               |
| Exeter                   | Bio-clad poly            | 113     | 1,254.7            | 7              | 0.56                     | 0.22                          | 1.15 | 0               |
| Elite plus               | Elite Plus Ogee          | 110     | 1,084.5            | 6              | 0.55                     | 0.20                          | 1.20 | 0               |
| Polarstem uncemented     | R3 porous                | 1517    | 4,554.6            | 25             | 0.55                     | 0.36                          | 0.81 | 271             |
| Exeter V40               | Exeter                   | 1639    | 15,405.7           | 84             | 0.55                     | 0.43                          | 0.68 | 1               |
| Lateral straight stem    | Muller PE cup            | 750     | 7,185.8            | 39             | 0.54                     | 0.39                          | 0.74 | 1               |
| Corail                   | Monoblock Acetabular Cup | 95      | 941.1              | 5              | 0.53                     | 0.17                          | 1.24 | 0               |
| Stemsys                  | Fixa Ti Por              | 707     | 2,700.6            | 14             | 0.52                     | 0.27                          | 0.85 | 77              |
| MS 30                    | RM Pressfit cup          | 90      | 774.5              | 4              | 0.52                     | 0.14                          | 1.32 | 0               |



| Femur Prosthesis         | Acetabular Prosthesis | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      | Procedures 2018 |
|--------------------------|-----------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|-----------------|
| Spectron                 | Fitmore               | 78      | 978.2              | 5              | 0.51                     | 0.14                          | 1.12 | 0               |
| CCA                      | CCB                   | 769     | 6,078.6            | 31             | 0.51                     | 0.34                          | 0.71 | 4               |
| Avenir Muller uncemented | Fitmore               | 65      | 196.3              | 1              | 0.51                     | 0.01                          | 2.84 | 27              |
| Wagner cone stem         | Fitmore               | 73      | 803.3              | 4              | 0.50                     | 0.14                          | 1.27 | 1               |
| CLS                      | Fitmore               | 2300    | 23,622.5           | 117            | 0.50                     | 0.41                          | 0.59 | 134             |
| Corail                   | Pinnacle              | 10351   | 48,382.9           | 331            | 0.68                     | 0.61                          | 0.76 | 1471            |
| Polarstem uncemented     | Reflection porous     | 335     | 2,072.2            | 14             | 0.68                     | 0.37                          | 1.13 | 0               |
| Exeter                   | Osteolock             | 836     | 11,108.7           | 75             | 0.68                     | 0.53                          | 0.85 | 0               |
| CPCS                     | R3 porous             | 326     | 889.9              | 6              | 0.67                     | 0.21                          | 1.39 | 63              |
| CLS                      | Reflection porous     | 373     | 2,990.9            | 20             | 0.67                     | 0.41                          | 1.03 | 15              |
| M/L Taper                | Trident               | 249     | 748.4              | 5              | 0.67                     | 0.22                          | 1.56 | 52              |
| Exeter V40               | Tritanium             | 3120    | 11,169.7           | 74             | 0.66                     | 0.52                          | 0.83 | 418             |
| CLS                      | RM Pressfit cup       | 567     | 3,853.8            | 25             | 0.65                     | 0.42                          | 0.96 | 27              |
| Twinsys uncemented       | RM Pressfit cup       | 4855    | 29,448.8           | 191            | 0.65                     | 0.56                          | 0.75 | 241             |
| Exeter V40               | Delta-TT Cup          | 223     | 771.1              | 5              | 0.65                     | 0.21                          | 1.51 | 34              |
| CPT                      | Duraloc               | 212     | 2,467.9            | 16             | 0.65                     | 0.37                          | 1.05 | 0               |
| Twinsys cemented         | CCB                   | 441     | 2,331.4            | 15             | 0.64                     | 0.36                          | 1.06 | 14              |
| CPT                      | ZCA                   | 545     | 5,487.9            | 35             | 0.64                     | 0.44                          | 0.89 | 2               |
| Exeter                   | CLS Expansion         | 129     | 1,578.9            | 10             | 0.63                     | 0.30                          | 1.16 | 0               |
| Avenir Muller uncemented | RM Pressfit cup       | 51      | 158.0              | 1              | 0.63                     | 0.02                          | 3.53 | 2               |
| Elite plus               | Charnley              | 298     | 3,661.5            | 23             | 0.63                     | 0.40                          | 0.94 | 0               |
| MS 30                    | Morscher              | 787     | 9,737.6            | 61             | 0.63                     | 0.48                          | 0.80 | 0               |
| CLS                      | Weill ring            | 106     | 1,601.3            | 10             | 0.62                     | 0.28                          | 1.11 | 0               |
| Corail                   | Trident               | 88      | 480.8              | 3              | 0.62                     | 0.13                          | 1.82 | 8               |
| Versys cemented          | ZCA                   | 391     | 4200.1             | 26             | 0.62                     | 0.40                          | 0.91 | 0               |
| Twinsys uncemented       | Continuum TM          | 133     | 821.0              | 5              | 0.61                     | 0.20                          | 1.42 | 0               |
| Twinsys uncemented       | Trilogy               | 209     | 1,811.8            | 11             | 0.61                     | 0.30                          | 1.09 | 0               |
| CLS                      | Trilogy               | 596     | 4,236.2            | 25             | 0.59                     | 0.38                          | 0.87 | 24              |
| Accolade II              | Tritanium             | 935     | 2,381.4            | 14             | 0.59                     | 0.32                          | 0.99 | 200             |
| Accolade                 | Muller PE cup         | 114     | 1,206.7            | 7              | 0.58                     | 0.21                          | 1.14 | 0               |
| C-Stem AMT               | Marathon cemented     | 330     | 1,763.1            | 10             | 0.57                     | 0.27                          | 1.04 | 14              |
| Spectron                 | Trident               | 78      | 886.1              | 5              | 0.56                     | 0.15                          | 1.24 | 0               |
| Exeter                   | Bio-clad poly         | 113     | 1,254.7            | 7              | 0.56                     | 0.22                          | 1.15 | 0               |
| Elite plus               | Elite Plus Ogee       | 110     | 1,084.5            | 6              | 0.55                     | 0.20                          | 1.20 | 0               |
| Polarstem uncemented     | R3 porous             | 1517    | 4,554.6            | 25             | 0.55                     | 0.36                          | 0.81 | 271             |
| Exeter V40               | Exeter                | 1639    | 15,405.7           | 84             | 0.55                     | 0.43                          | 0.68 | 1               |



| Femur Prosthesis         | Acetabular Prosthesis    | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      | Procedures 2018 |
|--------------------------|--------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|-----------------|
| Lateral straight stem    | Muller PE cup            | 750     | 7,185.8            | 39             | 0.54                     | 0.39                          | 0.74 | 1               |
| Corail                   | Monoblock Acetabular Cup | 95      | 941.1              | 5              | 0.53                     | 0.17                          | 1.24 | 0               |
| Stemsys                  | Fixa Ti Por              | 707     | 2,700.6            | 14             | 0.52                     | 0.27                          | 0.85 | 77              |
| MS 30                    | RM Pressfit cup          | 90      | 774.5              | 4              | 0.52                     | 0.14                          | 1.32 | 0               |
| Spectron                 | Fitmore                  | 78      | 978.2              | 5              | 0.51                     | 0.14                          | 1.12 | 0               |
| CCA                      | CCB                      | 769     | 6,078.6            | 31             | 0.51                     | 0.34                          | 0.71 | 4               |
| Avenir Muller uncemented | Fitmore                  | 65      | 196.3              | 1              | 0.51                     | 0.01                          | 2.84 | 27              |
| Wagner cone stem         | Fitmore                  | 73      | 803.3              | 4              | 0.50                     | 0.14                          | 1.27 | 1               |
| CLS                      | Fitmore                  | 2300    | 23,622.5           | 117            | 0.50                     | 0.41                          | 0.59 | 34              |
| CLS                      | Morscher                 | 1682    | 22,801.5           | 112            | 0.49                     | 0.40                          | 0.59 | 0               |
| Twinsys cemented         | RM Pressfit cup          | 1864    | 8,557.7            | 42             | 0.49                     | 0.35                          | 0.66 | 197             |
| Spectron                 | Biomex acet shell porous | 68      | 1,023.2            | 5              | 0.49                     | 0.16                          | 1.14 | 0               |
| Elite plus               | Elite Plus LPW           | 282     | 3,075.1            | 15             | 0.49                     | 0.27                          | 0.80 | 0               |
| Corail                   | Tritanium                | 168     | 822.6              | 4              | 0.49                     | 0.13                          | 1.24 | 15              |
| M/L Taper                | Trilogy                  | 215     | 1,853.2            | 9              | 0.49                     | 0.20                          | 0.89 | 3               |
| Exeter                   | Trilogy                  | 213     | 2,940.2            | 14             | 0.48                     | 0.26                          | 0.80 | 0               |
| Exeter V40               | Contemporary             | 6504    | 4,9605.6           | 235            | 0.47                     | 0.42                          | 0.54 | 114             |
| CLS                      | Monoblock Acetabular Cup | 80      | 848.4              | 4              | 0.47                     | 0.13                          | 1.21 | 0               |
| Exeter V40               | Morscher                 | 630     | 7,226.3            | 34             | 0.47                     | 0.33                          | 0.66 | 0               |
| Exeter V40               | Osteolock                | 270     | 3,198.6            | 15             | 0.47                     | 0.25                          | 0.75 | 0               |
| Summit                   | Trilogy                  | 168     | 1,286.5            | 6              | 0.47                     | 0.17                          | 1.02 | 11              |
| Accolade                 | Trident                  | 1867    | 20,510.0           | 94             | 0.46                     | 0.37                          | 0.56 | 0               |
| H-Max M                  | Delta-TT Cup             | 86      | 656.1              | 3              | 0.46                     | 0.09                          | 1.34 | 0               |
| ABGII                    | Pinnacle                 | 67      | 656.3              | 3              | 0.46                     | 0.09                          | 1.34 | 0               |
| Exeter V40               | Pinnacle                 | 2448    | 11,707.0           | 53             | 0.45                     | 0.34                          | 0.59 | 290             |
| Spectron                 | Mallory-Head             | 152     | 1,793.2            | 8              | 0.45                     | 0.19                          | 0.88 | 0               |
| SL monoblock             | Muller PE cup            | 488     | 5,462.9            | 24             | 0.44                     | 0.27                          | 0.64 | 0               |
| MS 30                    | Continuum TM             | 404     | 1,595.9            | 7              | 0.44                     | 0.16                          | 0.86 | 33              |
| Versys                   | Trilogy                  | 272     | 3,889.9            | 17             | 0.44                     | 0.25                          | 0.70 | 0               |
| Exeter                   | Morscher                 | 551     | 8,255.2            | 36             | 0.44                     | 0.31                          | 0.60 | 0               |
| Charnley                 | Charnley                 | 456     | 5,338.8            | 23             | 0.43                     | 0.27                          | 0.64 | 0               |
| Exeter V40               | CCB                      | 567     | 3,018.0            | 13             | 0.43                     | 0.23                          | 0.74 | 33              |
| Exeter V40               | Trident                  | 10390   | 63,631.7           | 274            | 0.43                     | 0.38                          | 0.48 | 1049            |
| Stemsys                  | RM Pressfit cup          | 325     | 1,189.5            | 5              | 0.42                     | 0.14                          | 0.98 | 38              |
| Lateral straight stem    | Continuum TM             | 78      | 476.3              | 2              | 0.42                     | 0.05                          | 1.52 | 0               |
| Summit                   | Duraloc                  | 101     | 1,212.1            | 5              | 0.41                     | 0.13                          | 0.96 | 0               |
| Exeter                   | Muller PE cup            | 119     | 1,465.0            | 6              | 0.41                     | 0.15                          | 0.89 | 0               |




| Femur Prosthesis         | Acetabular Prosthesis    | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      | Procedures 2018 |
|--------------------------|--------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|-----------------|
| Exeter V40               | Reflection cemented      | 926     | 5,623.2            | 23             | 0.41                     | 0.26                          | 0.61 | 35              |
| CCA                      | RM Pressfit cup          | 135     | 1,251.2            | 5              | 0.40                     | 0.13                          | 0.93 | 0               |
| Exeter V40               | Trilogy                  | 2805    | 19,700.1           | 78             | 0.40                     | 0.31                          | 0.49 | 149             |
| CPT                      | Pinnacle                 | 64      | 522.6              | 2              | 0.38                     | 0.05                          | 1.38 | 0               |
| Exeter V40               | Exeter X3                | 2103    | 7,355.7            | 28             | 0.38                     | 0.25                          | 0.54 | 336             |
| Avenir Muller uncemented | Pinnacle                 | 99      | 799.3              | 3              | 0.38                     | 0.05                          | 1.00 | 0               |
| Trabecular Metal Stem    | Monoblock Acetabular Cup | 74      | 822.4              | 3              | 0.36                     | 0.05                          | 0.97 | 0               |
| C-Stem                   | Elite Plus Ogee          | 55      | 548.6              | 2              | 0.36                     | 0.04                          | 1.32 | 0               |
| Standard straight stem   | Muller PE cup            | 629     | 5,764.2            | 21             | 0.36                     | 0.22                          | 0.55 | 1               |
| Synergy Porous           | Reflection porous        | 1238    | 11,875.0           | 43             | 0.36                     | 0.26                          | 0.48 | 14              |
| Lateral straight stem    | Weber                    | 287     | 2,865.7            | 10             | 0.35                     | 0.17                          | 0.64 | 0               |
| Standard straight stem   | ZCA all-poly cup         | 50      | 290.6              | 1              | 0.34                     | 0.00                          | 1.92 | 0               |
| Twinsys cemented         | RM cup                   | 148     | 1,464.9            | 5              | 0.34                     | 0.09                          | 0.75 | 0               |
| MS 30                    | Muller PE cup            | 462     | 4,414.8            | 15             | 0.34                     | 0.19                          | 0.56 | 0               |
| Exeter V40               | RM Pressfit cup          | 2385    | 10,968.5           | 37             | 0.34                     | 0.23                          | 0.46 | 301             |
| Corail                   | Delta-PF Cup             | 80      | 893.5              | 3              | 0.34                     | 0.07                          | 0.98 | 0               |
| Corail                   | Ultima                   | 135     | 1,205.2            | 4              | 0.33                     | 0.09                          | 0.85 | 0               |
| Exeter V40               | Muller PE cup            | 94      | 906.0              | 3              | 0.33                     | 0.07                          | 0.97 | 0               |
| CLS                      | Pinnacle                 | 90      | 614.1              | 2              | 0.33                     | 0.04                          | 1.18 | 10              |
| Exeter V40               | Monoblock Acetabular Cup | 123     | 1,575.0            | 5              | 0.32                     | 0.10                          | 0.74 | 0               |
| Echo Bi-Metric           | Exceed ABT Ringloc-X     | 57      | 316.6              | 1              | 0.32                     | 0.01                          | 1.76 | 0               |
| Corail                   | Trilogy                  | 207     | 953.4              | 3              | 0.31                     | 0.06                          | 0.92 | 17              |
| Avenir Muller uncemented | Tritanium                | 91      | 647.3              | 2              | 0.31                     | 0.04                          | 1.12 | 0               |
| Standard straight stem   | Weber                    | 134     | 1,294.7            | 4              | 0.31                     | 0.08                          | 0.79 | 0               |
| Exeter V40               | Reflection porous        | 475     | 4,076.5            | 12             | 0.29                     | 0.15                          | 0.51 | 0               |
| Friendly                 | Delta-PF Cup             | 168     | 1,700.2            | 5              | 0.29                     | 0.10                          | 0.69 | 0               |
| Spectron                 | R3 porous                | 432     | 2,439.2            | 7              | 0.29                     | 0.10                          | 0.56 | 11              |
| Accolade                 | Tritanium                | 152     | 1,063.1            | 3              | 0.28                     | 0.06                          | 0.82 | 0               |
| MS 30                    | Fitmore                  | 2174    | 14,197.5           | 39             | 0.27                     | 0.20                          | 0.38 | 192             |
| Versys cemented          | Trilogy                  | 237     | 2,636.1            | 7              | 0.27                     | 0.11                          | 0.55 | 0               |
| Stemsys                  | DeltaMotion Cup          | 485     | 2,413.5            | 6              | 0.25                     | 0.08                          | 0.51 | 44              |
| Twinsys cemented         | Continuum TM             | 108     | 408.7              | 1              | 0.24                     | 0.01                          | 1.36 | 6               |
| Lateral straight stem    | RM Pressfit cup          | 173     | 1,253.4            | 3              | 0.24                     | 0.05                          | 0.70 | 0               |
| Basis                    | Reflection porous        | 108     | 844.6              | 2              | 0.24                     | 0.03                          | 0.86 | 0               |



| Femur Prosthesis         | Acetabular Prosthesis | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      | Procedures 2018 |
|--------------------------|-----------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|-----------------|
| CPT                      | ZCA all-poly cup      | 96      | 468.8              | 1              | 0.21                     | 0.01                          | 1.19 | 0               |
| MS 30                    | Trilogy               | 331     | 1,958.0            | 4              | 0.20                     | 0.06                          | 0.52 | 19              |
| Exeter V40               | CLS Expansion         | 88      | 1,008.5            | 2              | 0.20                     | 0.02                          | 0.72 | 0               |
| MS 30                    | ZCA all-poly cup      | 94      | 506.3              | 1              | 0.20                     | 0.01                          | 1.10 | 0               |
| SL modular stem          | Muller PE cup         | 83      | 1,104.8            | 2              | 0.18                     | 0.02                          | 0.65 | 0               |
| Exeter V40               | Weber                 | 53      | 561.9              | 1              | 0.18                     | 0.00                          | 0.99 | 0               |
| Exeter V40               | ZCA                   | 93      | 574.3              | 1              | 0.17                     | 0.00                          | 0.97 | 8               |
| Exeter V40               | Fitmore               | 963     | 4,733.5            | 8              | 0.17                     | 0.07                          | 0.33 | 91              |
| Corail                   | Reflection porous     | 140     | 1,360.4            | 2              | 0.15                     | 0.02                          | 0.53 | 0               |
| Lateral straight stem    | ZCA                   | 98      | 750.9              | 1              | 0.13                     | 0.00                          | 0.74 | 0               |
| Synergy Porous           | Delta-PF Cup          | 88      | 779.7              | 1              | 0.13                     | 0.00                          | 0.71 | 0               |
| Accolade                 | Pinnacle              | 180     | 1,609.3            | 2              | 0.12                     | 0.02                          | 0.45 | 0               |
| Avenir Muller uncemented | RM cup                | 105     | 825.7              | 1              | 0.12                     | 0.00                          | 0.67 | 0               |
| Stemsys                  | Delta-PF Cup          | 396     | 1,077.1            | 1              | 0.09                     | 0.00                          | 0.43 | 74              |
| Standard straight stem   | RM Pressfit cup       | 137     | 1,090.0            | 1              | 0.09                     | 0.00                          | 0.51 | 0               |
| Twinsys uncemented       | Delta-PF Cup          | 370     | 3,000.8            | 1              | 0.03                     | 0.00                          | 0.19 | 0               |
| Exeter V40               | Polymax               | 63      | 61.3               | 0              | 0.00                     | 0.00                          | 6.02 | 35              |
| Stemsys cemented         | RM Pressfit cup       | 64      | 142.1              | 0              | 0.00                     | 0.00                          | 2.60 | 18              |
| C-Stem                   | Marathon cemented     | 89      | 310.1              | 0              | 0.00                     | 0.00                          | 1.19 | 14              |
| Twinsys cemented         | Reflection porous     | 59      | 174.7              | 0              | 0.00                     | 0.00                          | 2.11 | 9               |
| Synergy Porous           | Continuum TM          | 55      | 120.5              | 0              | 0.00                     | 0.00                          | 3.06 | 6               |
| Exeter V40               | ZCA all-poly cup      | 104     | 432.0              | 0              | 0.00                     | 0.00                          | 0.85 | 4               |
| Corail                   | DeltaMotion Cup       | 78      | 518.8              | 0              | 0.00                     | 0.00                          | 0.71 | 0               |
| Exeter                   | Trident               | 84      | 1,270.3            | 0              | 0.00                     | 0.00                          | 0.29 | 0               |
| Lateral straight stem    | ZCA all-poly cup      | 70      | 422.4              | 0              | 0.00                     | 0.00                          | 0.87 | 0               |





## Revisions versus Hip Prostheses Combinations and Fixation Method Sorted on Number of Implantations

(Minimum of 50 primary registered arthroplasties)

### Fully Cemented


| Femur Prosthesis       | Acetabular Prosthesis | No. Ops | Observed comp. yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      |
|------------------------|-----------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Exeter V40             | Contemporary          | 6,504   | 49,605.6           | 235            | 0.47                     | 0.42                          | 0.54 |
| Spectron               | Reflection cemented   | 2,956   | 29,689.6           | 353            | 1.19                     | 1.07                          | 1.32 |
| Exeter V40             | Exeter X3             | 2,103   | 7,355.7            | 28             | 0.38                     | 0.25                          | 0.54 |
| Exeter V40             | Exeter                | 1,639   | 15,405.7           | 84             | 0.55                     | 0.43                          | 0.67 |
| Exeter                 | Contemporary          | 1,551   | 18,301.8           | 182            | 0.99                     | 0.85                          | 1.15 |
| Exeter                 | Exeter                | 1,326   | 15,253.9           | 114            | 0.75                     | 0.62                          | 0.90 |
| Exeter V40             | Reflection cemented   | 926     | 5,623.2            | 23             | 0.41                     | 0.26                          | 0.61 |
| CCA                    | CCB                   | 769     | 6,078.6            | 31             | 0.51                     | 0.34                          | 0.71 |
| Lateral straight stem  | Muller PE cup         | 750     | 7,185.8            | 39             | 0.54                     | 0.39                          | 0.74 |
| Standard straight stem | Muller PE cup         | 629     | 5,764.2            | 21             | 0.36                     | 0.22                          | 0.55 |
| Exeter V40             | CCB                   | 567     | 3,018.0            | 13             | 0.43                     | 0.23                          | 0.74 |
| CPT                    | ZCA                   | 545     | 5,487.9            | 35             | 0.64                     | 0.44                          | 0.89 |
| SL monoblock           | Muller PE cup         | 488     | 5,462.9            | 24             | 0.44                     | 0.27                          | 0.64 |
| MS 30                  | Muller PE cup         | 462     | 4,414.8            | 15             | 0.34                     | 0.19                          | 0.56 |
| Charnley               | Charnley              | 456     | 5,338.8            | 23             | 0.43                     | 0.27                          | 0.64 |
| Twinsys cemented       | CCB                   | 441     | 2,331.4            | 15             | 0.64                     | 0.36                          | 1.06 |
| Versys cemented        | ZCA                   | 391     | 4,200.1            | 26             | 0.62                     | 0.39                          | 0.89 |
| C-Stem AMT             | Marathon cemented     | 330     | 1,763.1            | 10             | 0.57                     | 0.27                          | 1.04 |
| Charnley               | Charnley Cup Ogee     | 303     | 3,856.2            | 28             | 0.73                     | 0.48                          | 1.05 |
| Elite plus             | Charnley              | 298     | 3,661.5            | 23             | 0.63                     | 0.40                          | 0.94 |
| Lateral straight stem  | Weber                 | 287     | 2,865.7            | 10             | 0.35                     | 0.17                          | 0.64 |
| Elite plus             | Elite Plus LPW        | 282     | 3,075.1            | 15             | 0.49                     | 0.27                          | 0.80 |
| Exeter V40             | Bio-clad poly         | 140     | 980.3              | 7              | 0.71                     | 0.29                          | 1.47 |
| Standard straight stem | Weber                 | 134     | 1,294.7            | 4              | 0.31                     | 0.08                          | 0.79 |
| MS 30                  | Contemporary          | 128     | 1,229.7            | 10             | 0.81                     | 0.39                          | 1.50 |
| Exeter                 | Muller PE cup         | 119     | 1,465.0            | 6              | 0.41                     | 0.15                          | 0.89 |
| Exeter                 | Bio-clad poly         | 113     | 1,254.7            | 7              | 0.56                     | 0.22                          | 1.15 |
| Elite plus             | Elite Plus Ogee       | 110     | 1,084.5            | 6              | 0.55                     | 0.20                          | 1.20 |
| Exeter V40             | ZCA all-poly cup      | 104     | 432.0              | 0              | 0.00                     | 0.00                          | 0.85 |
| Lateral straight stem  | ZCA                   | 98      | 750.9              | 1              | 0.13                     | 0.00                          | 0.74 |
| CPT                    | ZCA all-poly cup      | 96      | 468.8              | 1              | 0.21                     | 0.01                          | 1.19 |
| Exeter V40             | Muller PE cup         | 94      | 906.0              | 3              | 0.33                     | 0.07                          | 0.97 |
| MS 30                  | ZCA all-poly cup      | 94      | 506.3              | 1              | 0.20                     | 0.01                          | 1.10 |
| Exeter V40             | ZCA                   | 93      | 574.3              | 1              | 0.17                     | 0.00                          | 0.97 |



| Femur Prosthesis       | Acetabular Prosthesis | No. Ops | Observed comp. yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      |
|------------------------|-----------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| C-Stem                 | Marathon cemented     | 89      | 310.1              | 0              | 0.00                     | 0.00                          | 1.19 |
| SL modular stem        | Muller PE cup         | 83      | 1,104.8            | 2              | 0.18                     | 0.02                          | 0.65 |
| CCA                    | Contemporary          | 74      | 762.2              | 10             | 1.31                     | 0.63                          | 2.41 |
| Contemporary           | Contemporary          | 71      | 912.4              | 12             | 1.32                     | 0.68                          | 2.30 |
| Lateral straight stem  | ZCA all-poly cup      | 70      | 422.4              | 0              | 0.00                     | 0.00                          | 0.87 |
| Spectron               | Muller PE cup         | 66      | 658.5              | 8              | 1.21                     | 0.52                          | 2.39 |
| C-Stem                 | Elite Plus Ogee       | 55      | 548.6              | 2              | 0.36                     | 0.04                          | 1.32 |
| Exeter V40             | Weber                 | 53      | 561.9              | 1              | 0.18                     | 0.00                          | 0.99 |
| Standard straight stem | ZCA all-poly cup stem | 50      | 290.6              | 1              | 0.34                     | 0.00                          | 1.92 |

### Uncemented


| Femur Prosthesis      | Acetabular Prosthesis | No. Ops | Observed comp. yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------------------|-----------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Corail                | Pinnacle              | 10,351  | 48,382.9           | 331            | 0.68                     | 0.61                          | 0.76 |
| Twinsys uncemented    | RM Pressfit cup       | 4,855   | 29,448.8           | 191            | 0.65                     | 0.56                          | 0.75 |
| CLS                   | Fitmore               | 2,300   | 23,622.5           | 117            | 0.50                     | 0.41                          | 0.59 |
| Summit                | Pinnacle              | 2,248   | 13,418.1           | 101            | 0.75                     | 0.61                          | 0.91 |
| Accolade              | Trident               | 1,867   | 20,510.0           | 94             | 0.46                     | 0.37                          | 0.56 |
| Synergy Porous        | R3 porous             | 1,779   | 7,919.5            | 58             | 0.73                     | 0.56                          | 0.95 |
| CLS                   | Morscher              | 1,682   | 22,801.5           | 112            | 0.49                     | 0.40                          | 0.59 |
| Polarstem uncemented  | R3 porous             | 1,517   | 4,554.6            | 25             | 0.55                     | 0.36                          | 0.81 |
| CLS                   | CLS Expansion         | 1,263   | 15,710.0           | 118            | 0.75                     | 0.62                          | 0.90 |
| Synergy Porous        | Reflection porous     | 1,238   | 11,875.0           | 43             | 0.36                     | 0.26                          | 0.48 |
| Twinsys uncemented    | Selexys TPS           | 1,231   | 10,413.9           | 136            | 1.31                     | 1.09                          | 1.54 |
| M/L Taper             | Continuum TM          | 1,001   | 4,111.7            | 34             | 0.83                     | 0.57                          | 1.16 |
| Accolade II           | Tritanium             | 935     | 2,381.4            | 14             | 0.59                     | 0.32                          | 0.99 |
| Accolade II           | Trident               | 858     | 2,158.4            | 16             | 0.74                     | 0.42                          | 1.20 |
| H-Max S               | Delta-TT Cup          | 736     | 2,936.0            | 27             | 0.92                     | 0.59                          | 1.32 |
| CLS                   | Continuum TM          | 719     | 2,920.8            | 22             | 0.75                     | 0.47                          | 1.14 |
| Stemsys               | Fixa Ti Por           | 707     | 2,700.6            | 14             | 0.52                     | 0.27                          | 0.85 |
| CLS                   | Duraloc               | 699     | 9,128.1            | 91             | 1.00                     | 0.80                          | 1.22 |
| CLS                   | Trilogy               | 596     | 4,236.2            | 25             | 0.59                     | 0.38                          | 0.87 |
| CLS                   | RM Pressfit cup       | 567     | 3,853.8            | 25             | 0.65                     | 0.42                          | 0.96 |
| Stemsys               | DeltaMotion Cup       | 485     | 2,413.5            | 6              | 0.25                     | 0.08                          | 0.51 |
| Corail                | Duraloc               | 464     | 5,139.0            | 49             | 0.95                     | 0.71                          | 1.26 |
| Trabecular Metal Stem | Continuum TM          | 447     | 2,151.9            | 18             | 0.84                     | 0.50                          | 1.32 |
| CBC                   | RM Pressfit cup       | 444     | 2,688.8            | 21             | 0.78                     | 0.48                          | 1.19 |



| Femur Prosthesis         | Acetabular Prosthesis | No. Ops | Observed comp. yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |       |
|--------------------------|-----------------------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| Stemsys                  | Agilis Ti-por         | 438     | 1,500.1            | 14             | 0.93                     | 0.49                          | 1.52  |
| Stemsys                  | Delta-PF Cup          | 396     | 1,077.1            | 1              | 0.09                     | 0.00                          | 0.43  |
| S-Rom                    | Pinnacle              | 375     | 3,633.1            | 35             | 0.96                     | 0.66                          | 1.32  |
| CLS                      | Reflection porous     | 373     | 2,990.9            | 20             | 0.67                     | 0.41                          | 1.03  |
| Twinsys uncemented       | Delta-PF Cup          | 370     | 3,000.8            | 1              | 0.03                     | 0.00                          | 0.16  |
| ABGII                    | Trident               | 342     | 4,038.1            | 34             | 0.84                     | 0.58                          | 1.18  |
| Polarstem uncemented     | Reflection porous     | 335     | 2,072.2            | 14             | 0.68                     | 0.37                          | 1.13  |
| Stemsys                  | RM Pressfit cup       | 325     | 1,189.5            | 5              | 0.42                     | 0.14                          | 0.98  |
| Echo Bi-Metric           | G7 acetabular shell   | 320     | 722.1              | 5              | 0.69                     | 0.22                          | 1.62  |
| Corail                   | Continuum TM          | 304     | 1,099.0            | 8              | 0.73                     | 0.31                          | 1.43  |
| Corail                   | Fitmore               | 285     | 893.5              | 7              | 0.78                     | 0.31                          | 1.61  |
| Versys                   | Trilogy               | 272     | 3,889.9            | 17             | 0.44                     | 0.25                          | 0.70  |
| Taperloc Complete        | G7 acetabular shell   | 259     | 390.5              | 6              | 1.54                     | 0.56                          | 3.34  |
| M/L Taper                | Trident               | 248     | 746.2              | 5              | 0.67                     | 0.22                          | 1.56  |
| M/L Taper                | Trilogy               | 215     | 1,853.2            | 9              | 0.49                     | 0.20                          | 0.89  |
| Twinsys uncemented       | Trilogy               | 209     | 1,811.7            | 11             | 0.61                     | 0.30                          | 1.09  |
| Corail                   | Trilogy               | 207     | 953.4              | 3              | 0.31                     | 0.06                          | 0.92  |
| CLS                      | Durom                 | 198     | 1,950.2            | 63             | 3.23                     | 2.48                          | 4.13  |
| H-Max S                  | Delta-PF Cup          | 193     | 589.8              | 7              | 1.19                     | 0.42                          | 2.33  |
| CLS                      | Allofit               | 192     | 1,924.0            | 20             | 1.04                     | 0.61                          | 1.57  |
| CBC                      | Expansys shell        | 183     | 1,791.9            | 25             | 1.40                     | 0.90                          | 2.06  |
| Accolade                 | Pinnacle              | 180     | 1,609.3            | 2              | 0.12                     | 0.02                          | 0.45  |
| Avenir Muller uncemented | Continuum TM          | 179     | 1,074.6            | 11             | 1.02                     | 0.51                          | 1.83  |
| Corail                   | Tritanium             | 168     | 822.6              | 4              | 0.49                     | 0.13                          | 1.24  |
| Summit                   | Trilogy               | 168     | 1,286.5            | 6              | 0.47                     | 0.17                          | 1.02  |
| Taperloc Complete        | RM Pressfit cup       | 168     | 266.9              | 5              | 1.87                     | 0.61                          | 4.37  |
| CLS                      | Trident               | 165     | 1,862.0            | 14             | 0.75                     | 0.39                          | 1.23  |
| Corail                   | ASR                   | 156     | 1,199.7            | 83             | 6.92                     | 5.51                          | 8.58  |
| Accolade                 | Tritanium             | 152     | 1,063.1            | 3              | 0.28                     | 0.06                          | 0.82  |
| Corail                   | RM Pressfit cup       | 144     | 504.8              | 5              | 0.99                     | 0.32                          | 2.31  |
| Corail                   | Reflection porous     | 140     | 1,360.4            | 2              | 0.15                     | 0.02                          | 0.53  |
| ABGII                    | Duraloc               | 139     | 1,929.9            | 40             | 2.07                     | 1.46                          | 2.79  |
| Twinsys uncemented       | Continuum TM          | 133     | 821.0              | 5              | 0.61                     | 0.20                          | 1.42  |
| S-Rom                    | ASR                   | 130     | 801.1              | 94             | 11.73                    | 9.43                          | 14.29 |
| Omnifit                  | Trident               | 126     | 1,561.3            | 12             | 0.77                     | 0.37                          | 1.30  |
| Twinsys uncemented       | RM cup                | 122     | 1,036.2            | 9              | 0.87                     | 0.40                          | 1.65  |
| Stemsys                  | Polymax               | 119     | 254.4              | 3              | 1.18                     | 0.24                          | 3.45  |
| ABG                      | Duraloc               | 116     | 1,860.4            | 38             | 2.04                     | 1.42                          | 2.77  |
| Synergy Porous           | BHR Acetabular Cup    | 114     | 1,034.2            | 40             | 3.87                     | 2.76                          | 5.27  |



| Femur Prosthesis         | Acetabular Prosthesis    | No. Ops | Observed comp. yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      |
|--------------------------|--------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| CLS                      | RM cup                   | 113     | 1,173.0            | 17             | 1.45                     | 0.81                          | 2.27 |
| Prodigy                  | Duraloc                  | 113     | 1,479.9            | 23             | 1.55                     | 0.99                          | 2.33 |
| Taperloc Complete        | Continuum TM             | 108     | 101.1              | 3              | 2.97                     | 0.61                          | 8.67 |
| ABGII                    | Delta-PF Cup             | 107     | 1,273.7            | 11             | 0.86                     | 0.43                          | 1.55 |
| CLS                      | Weill ring               | 106     | 1,601.3            | 10             | 0.62                     | 0.28                          | 1.11 |
| Echo Bi-Metric           | Continuum TM             | 106     | 183.8              | 3              | 1.63                     | 0.34                          | 4.77 |
| Optimys                  | RM Pressfit cup          | 106     | 140.0              | 1              | 0.71                     | 0.02                          | 3.98 |
| Avenir Muller uncemented | RM cup                   | 105     | 825.7              | 1              | 0.12                     | 0.00                          | 0.67 |
| Mallory-Head             | M2A                      | 105     | 1,237.7            | 15             | 1.21                     | 0.68                          | 2.00 |
| Summit                   | Duraloc                  | 101     | 1,212.1            | 5              | 0.41                     | 0.13                          | 0.96 |
| Avenir Muller uncemented | Pinnacle                 | 99      | 799.3              | 3              | 0.38                     | 0.05                          | 1.00 |
| Corail                   | Monoblock Acetabular Cup | 95      | 941.1              | 5              | 0.53                     | 0.17                          | 1.24 |
| Anthology Porous         | BHR Acetabular Cup       | 91      | 713.9              | 49             | 6.86                     | 5.08                          | 9.07 |
| Avenir Muller uncemented | Tritanium                | 91      | 647.3              | 2              | 0.31                     | 0.04                          | 1.12 |
| CLS                      | Pinnacle                 | 90      | 614.1              | 2              | 0.33                     | 0.04                          | 1.18 |
| Corail                   | Trident                  | 88      | 480.8              | 3              | 0.62                     | 0.13                          | 1.82 |
| Summit                   | ASR                      | 88      | 741.0              | 36             | 4.86                     | 3.40                          | 6.73 |
| Synergy Porous           | Delta-PF Cup             | 88      | 779.7              | 1              | 0.13                     | 0.00                          | 0.71 |
| H-Max M                  | Delta-TT Cup             | 86      | 656.1              | 3              | 0.46                     | 0.09                          | 1.34 |
| ABGII                    | RM Pressfit cup          | 81      | 218.5              | 6              | 2.75                     | 1.01                          | 5.98 |
| CLS                      | Monoblock Acetabular Cup | 80      | 848.4              | 4              | 0.47                     | 0.13                          | 1.21 |
| CLS                      | Tritanium                | 80      | 362.3              | 3              | 0.83                     | 0.17                          | 2.42 |
| Corail                   | Delta-PF Cup             | 80      | 893.5              | 3              | 0.34                     | 0.07                          | 0.98 |
| Accolade II              | RM Pressfit cup          | 79      | 126.2              | 3              | 2.38                     | 0.49                          | 6.95 |
| Corail                   | DeltaMotion Cup          | 78      | 518.8              | 0              | 0.00                     | 0.00                          | 0.71 |
| S-Rom                    | Ultima                   | 78      | 1,250.4            | 14             | 1.12                     | 0.61                          | 1.88 |
| AML MMA                  | Duraloc                  | 74      | 1,034.3            | 11             | 1.06                     | 0.50                          | 1.84 |
| Trabecular Metal Stem    | Monoblock Acetabular Cup | 74      | 822.4              | 3              | 0.36                     | 0.05                          | 0.97 |
| Wagner cone stem         | Fitmore                  | 73      | 803.3              | 4              | 0.50                     | 0.14                          | 1.27 |
| ABG                      | ABGII                    | 72      | 1,098.6            | 16             | 1.46                     | 0.83                          | 2.37 |
| Accolade II              | Delta-TT Cup             | 72      | 125.6              | 1              | 0.80                     | 0.02                          | 4.43 |
| H-Max M                  | Delta-PF Cup             | 71      | 535.0              | 7              | 1.31                     | 0.53                          | 2.70 |
| Anthology Porous         | R3 porous                | 68      | 486.2              | 32             | 6.58                     | 4.42                          | 9.17 |
| ABGII                    | Pinnacle                 | 67      | 656.3              | 3              | 0.46                     | 0.09                          | 1.34 |
| Furlong                  | Furlong                  | 66      | 789.3              | 6              | 0.76                     | 0.28                          | 1.65 |
| Avenir Muller uncemented | Fitmore                  | 65      | 196.3              | 1              | 0.51                     | 0.01                          | 2.84 |




| Femur Prosthesis         | Acetabular Prosthesis  | No. Ops | Observed comp. yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |       |
|--------------------------|------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| M/L Taper                | Delta-TT Cup           | 64      | 323.4              | 5              | 1.55                     | 0.50                          | 3.61  |
| Tri-Lock BPS             | Pinnacle               | 63      | 425.3              | 3              | 0.71                     | 0.10                          | 1.88  |
| CBC                      | Fitmore                | 59      | 597.2              | 5              | 0.84                     | 0.27                          | 1.95  |
| CLS                      | Artek                  | 59      | 712.1              | 25             | 3.51                     | 2.27                          | 5.18  |
| MasterSL                 | Delta-TT Cup           | 59      | 52.6               | 2              | 3.81                     | 0.46                          | 13.75 |
| Echo Bi-Metric           | Exceed ABT Ringloc-X   | 57      | 316.6              | 1              | 0.32                     | 0.01                          | 1.76  |
| Synergy Porous           | Continuum TM           | 55      | 120.5              | 0              | 0.00                     | 0.00                          | 3.06  |
| AML                      | Duraloc                | 53      | 780.7              | 7              | 0.90                     | 0.36                          | 1.85  |
| CLS                      | Trabecular Metal Shell | 53      | 420.1              | 3              | 0.71                     | 0.15                          | 2.09  |
| Avenir Muller uncemented | RM Pressfit cup        | 51      | 158.0              | 1              | 0.63                     | 0.02                          | 3.53  |

### Hybrid

| Femur Prosthesis      | Acetabular Prosthesis | No. Ops | Observed comp. yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------------------|-----------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Exeter V40            | Trident               | 10,390  | 63,631.7           | 274            | 0.43                     | 0.38                          | 0.48 |
| Exeter V40            | Tritanium             | 3,120   | 11,169.7           | 74             | 0.66                     | 0.52                          | 0.83 |
| Exeter V40            | Trilogy               | 2,805   | 19,700.1           | 78             | 0.40                     | 0.31                          | 0.49 |
| Spectron              | Reflection porous     | 2,755   | 27,633.0           | 228            | 0.83                     | 0.72                          | 0.94 |
| Exeter V40            | Continuum TM          | 2,538   | 10,125.6           | 86             | 0.85                     | 0.68                          | 1.05 |
| Exeter V40            | Pinnacle              | 2,448   | 11,707.0           | 53             | 0.45                     | 0.34                          | 0.59 |
| Exeter V40            | RM Pressfit cup       | 2,385   | 10,968.5           | 37             | 0.34                     | 0.23                          | 0.46 |
| C-Stem AMT            | Pinnacle              | 2,339   | 8,132.8            | 66             | 0.81                     | 0.62                          | 1.03 |
| MS 30                 | Fitmore               | 2,174   | 14,197.5           | 39             | 0.27                     | 0.20                          | 0.38 |
| Twinsys cemented      | RM Pressfit cup       | 1,864   | 8,557.7            | 42             | 0.49                     | 0.35                          | 0.66 |
| CPT                   | Continuum TM          | 1,387   | 4,915.7            | 46             | 0.94                     | 0.69                          | 1.25 |
| Spectron              | Duraloc               | 1,151   | 13,752.9           | 177            | 1.29                     | 1.10                          | 1.49 |
| Exeter V40            | Duraloc               | 987     | 10,786.5           | 100            | 0.93                     | 0.75                          | 1.12 |
| Exeter V40            | Fitmore               | 963     | 4,733.5            | 8              | 0.17                     | 0.07                          | 0.33 |
| CPT                   | Trilogy               | 843     | 6,459.6            | 56             | 0.87                     | 0.65                          | 1.13 |
| Exeter                | Osteolock             | 836     | 11,108.7           | 75             | 0.68                     | 0.53                          | 0.85 |
| MS 30                 | Morscher              | 787     | 9,737.6            | 61             | 0.63                     | 0.48                          | 0.80 |
| Exeter V40            | Morscher              | 630     | 7,226.3            | 34             | 0.47                     | 0.33                          | 0.66 |
| Elite plus            | Duraloc               | 608     | 7,022.0            | 112            | 1.59                     | 1.31                          | 1.92 |
| Exeter V40            | R3 porous             | 605     | 2,186.6            | 15             | 0.69                     | 0.38                          | 1.13 |
| Exeter                | Duraloc               | 553     | 7,808.0            | 113            | 1.45                     | 1.19                          | 1.74 |
| Exeter                | Morscher              | 551     | 8,255.2            | 36             | 0.44                     | 0.31                          | 0.60 |
| Lateral straight stem | RM cup                | 533     | 5,245.0            | 42             | 0.80                     | 0.58                          | 1.08 |



| Femur Prosthesis       | Acetabular Prosthesis    | No. Ops | Observed comp. yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      |
|------------------------|--------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Exeter V40             | Reflection porous        | 475     | 4,076.5            | 12             | 0.29                     | 0.15                          | 0.51 |
| Spectron               | R3 porous                | 432     | 2,439.2            | 7              | 0.29                     | 0.10                          | 0.56 |
| MS 30                  | Continuum TM             | 404     | 1,595.9            | 7              | 0.44                     | 0.18                          | 0.90 |
| MS 30                  | Trilogy                  | 331     | 1,958.0            | 4              | 0.20                     | 0.06                          | 0.52 |
| CPCS                   | R3 porous                | 326     | 889.9              | 6              | 0.67                     | 0.21                          | 1.39 |
| SL modular stem        | RM cup                   | 322     | 4,578.4            | 40             | 0.87                     | 0.62                          | 1.19 |
| Exeter V40             | Osteolock                | 270     | 3,198.6            | 15             | 0.47                     | 0.26                          | 0.77 |
| Versys cemented        | Trilogy                  | 237     | 2,636.1            | 7              | 0.27                     | 0.11                          | 0.55 |
| Exeter V40             | Delta-TT Cup             | 223     | 771.1              | 5              | 0.65                     | 0.21                          | 1.51 |
| Exeter                 | Trilogy                  | 213     | 2,940.2            | 14             | 0.48                     | 0.26                          | 0.80 |
| CPT                    | Duraloc                  | 212     | 2,467.9            | 16             | 0.65                     | 0.37                          | 1.05 |
| Exeter V40             | Trabecular Metal Shell   | 211     | 1,079.1            | 13             | 1.20                     | 0.64                          | 2.06 |
| Spectron               | Morscher                 | 210     | 2,769.6            | 31             | 1.12                     | 0.76                          | 1.59 |
| CPT                    | Fitmore                  | 191     | 1,053.9            | 10             | 0.95                     | 0.42                          | 1.68 |
| Lateral straight stem  | RM Pressfit cup          | 173     | 1,253.4            | 3              | 0.24                     | 0.05                          | 0.70 |
| Friendly               | Delta-PF Cup             | 168     | 1,700.2            | 5              | 0.29                     | 0.10                          | 0.69 |
| Spectron               | Mallory-Head             | 152     | 1,793.2            | 8              | 0.45                     | 0.19                          | 0.88 |
| Twinsys cemented       | RM cup                   | 148     | 1,464.9            | 5              | 0.34                     | 0.09                          | 0.75 |
| CPT                    | Trident                  | 145     | 1,620.2            | 12             | 0.74                     | 0.38                          | 1.29 |
| Standard straight stem | RM cup                   | 138     | 1,540.2            | 11             | 0.71                     | 0.33                          | 1.24 |
| Standard straight stem | RM Pressfit cup          | 137     | 1,090.0            | 1              | 0.09                     | 0.00                          | 0.51 |
| CCA                    | RM Pressfit cup          | 135     | 1,251.2            | 5              | 0.40                     | 0.13                          | 0.93 |
| Corail                 | Ultima                   | 134     | 1,196.0            | 4              | 0.33                     | 0.09                          | 0.86 |
| C-Stem AMT             | RM Pressfit cup          | 129     | 467.7              | 4              | 0.86                     | 0.23                          | 2.19 |
| Exeter                 | CLS Expansion            | 129     | 1,578.9            | 10             | 0.63                     | 0.30                          | 1.16 |
| Exeter V40             | Monoblock Acetabular Cup | 123     | 1,575.0            | 5              | 0.32                     | 0.10                          | 0.74 |
| Accolade               | Muller PE cup            | 114     | 1,206.7            | 7              | 0.58                     | 0.21                          | 1.14 |
| Exeter V40             | G7 acetabular shell      | 112     | 212.7              | 3              | 1.41                     | 0.29                          | 4.12 |
| Basis                  | Reflection porous        | 108     | 844.6              | 2              | 0.24                     | 0.03                          | 0.86 |
| Twinsys cemented       | Continuum TM             | 108     | 408.7              | 1              | 0.24                     | 0.01                          | 1.36 |
| MS 30                  | RM Pressfit cup          | 90      | 774.5              | 4              | 0.52                     | 0.14                          | 1.32 |
| Exeter V40             | CLS Expansion            | 88      | 1,008.5            | 2              | 0.20                     | 0.02                          | 0.72 |
| CPT                    | Tritanium                | 85      | 611.2              | 6              | 0.98                     | 0.36                          | 2.14 |
| CPT                    | Monoblock Acetabular Cup | 84      | 937.6              | 8              | 0.85                     | 0.33                          | 1.61 |
| Exeter                 | Trident                  | 84      | 1,270.3            | 0              | 0.00                     | 0.00                          | 0.29 |
| CPT                    | Delta-TT Cup             | 79      | 182.5              | 3              | 1.64                     | 0.23                          | 4.39 |
| Lateral straight stem  | Continuum TM             | 78      | 476.3              | 2              | 0.42                     | 0.05                          | 1.52 |



| Femur Prosthesis      | Acetabular Prosthesis    | No. Ops | Observed comp. yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------------------|--------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Spectron              | Fitmore                  | 78      | 978.2              | 5              | 0.51                     | 0.14                          | 1.12 |
| Spectron              | Trident                  | 78      | 886.1              | 5              | 0.56                     | 0.15                          | 1.24 |
| C-Stem                | Pinnacle                 | 77      | 260.2              | 4              | 1.54                     | 0.42                          | 3.94 |
| Twinsys cemented      | Pinnacle                 | 76      | 286.1              | 7              | 2.45                     | 0.98                          | 5.04 |
| Lateral straight stem | Trilogy                  | 69      | 555.6              | 12             | 2.16                     | 1.12                          | 3.77 |
| Spectron              | Biomex acet shell porous | 68      | 1,023.2            | 5              | 0.49                     | 0.16                          | 1.14 |
| Friendly              | Delta-TT Cup             | 67      | 410.6              | 5              | 1.22                     | 0.40                          | 2.84 |
| Twinsys cemented      | Selexys TPS              | 65      | 458.4              | 5              | 1.09                     | 0.35                          | 2.55 |
| CPT                   | Pinnacle                 | 64      | 522.6              | 2              | 0.38                     | 0.05                          | 1.38 |
| Stemsys cemented      | RM Pressfit cup          | 64      | 142.1              | 0              | 0.00                     | 0.00                          | 2.60 |
| Exeter V40            | Polymax                  | 63      | 61.3               | 0              | 0.00                     | 0.00                          | 6.02 |
| H-Max C               | Delta-TT Cup             | 61      | 134.1              | 2              | 1.49                     | 0.18                          | 5.39 |
| Twinsys cemented      | Reflection porous        | 59      | 174.7              | 0              | 0.00                     | 0.00                          | 2.11 |
| CPT                   | G7 acetabular shell      | 58      | 95.0               | 3              | 3.16                     | 0.65                          | 9.23 |
| MS 30                 | Duraloc                  | 55      | 776.5              | 6              | 0.77                     | 0.28                          | 1.68 |
| C-Stem                | Duraloc                  | 53      | 634.1              | 6              | 0.95                     | 0.30                          | 1.95 |


#### Prosthesis combinations based on femur in alphabetical order

| Femur Prosthesis | Acetabular Prosthesis | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      | Procedures 2018 |
|------------------|-----------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|-----------------|
| ABG              | Duraloc               | 116     | 1,860.4            | 38             | 2.04                     | 1.42                          | 2.77 | 0               |
| ABG              | ABGII                 | 72      | 1,098.6            | 16             | 1.46                     | 0.83                          | 2.37 | 0               |
| ABGII            | RM Pressfit cup       | 81      | 218.5              | 6              | 2.75                     | 1.01                          | 5.98 | 12              |
| ABGII            | Duraloc               | 139     | 1,929.9            | 40             | 2.07                     | 1.46                          | 2.79 | 0               |
| ABGII            | Delta-PF Cup          | 107     | 1,273.7            | 11             | 0.86                     | 0.40                          | 1.49 | 0               |
| ABGII            | Trident               | 342     | 4,038.1            | 34             | 0.84                     | 0.58                          | 1.18 | 0               |
| ABGII            | Pinnacle              | 67      | 656.3              | 3              | 0.46                     | 0.09                          | 1.34 | 0               |
| Accolade         | Muller PE cup         | 114     | 1,206.7            | 7              | 0.58                     | 0.21                          | 1.14 | 0               |
| Accolade         | Trident               | 1867    | 20,510.0           | 94             | 0.46                     | 0.37                          | 0.56 | 0               |
| Accolade         | Tritanium             | 152     | 1,063.1            | 3              | 0.28                     | 0.06                          | 0.82 | 0               |
| Accolade         | Pinnacle              | 180     | 1,609.3            | 2              | 0.12                     | 0.02                          | 0.45 | 0               |
| Accolade II      | RM Pressfit cup       | 79      | 126.2              | 3              | 2.38                     | 0.49                          | 6.95 | 27              |
| Accolade II      | Delta-TT Cup          | 72      | 125.6              | 1              | 0.80                     | 0.02                          | 4.43 | 18              |
| Accolade II      | Trident               | 858     | 2,158.4            | 16             | 0.74                     | 0.42                          | 1.20 | 219             |
| Accolade II      | Tritanium             | 935     | 2,381.4            | 14             | 0.59                     | 0.32                          | 0.99 | 200             |
| AML              | Duraloc               | 53      | 780.7              | 7              | 0.90                     | 0.36                          | 1.85 | 0               |
| AML MMA          | Duraloc               | 74      | 1,034.3            | 11             | 1.06                     | 0.50                          | 1.84 | 0               |
| Anthology Porous | BHR Acetabular Cup    | 93      | 726.3              | 50             | 6.88                     | 5.11                          | 9.08 | 0               |
| Anthology Porous | R3 porous             | 68      | 486.2              | 32             | 6.58                     | 4.42                          | 9.17 | 1               |



| Femur Prosthesis         | Acetabular Prosthesis    | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      | Procedures 2018 |
|--------------------------|--------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|-----------------|
| Avenir Muller uncemented | Continuum TM             | 179     | 1,074.6            | 11             | 1.02                     | 0.51                          | 1.83 | 1               |
| Avenir Muller uncemented | RM Pressfit cup          | 51      | 158.0              | 1              | 0.63                     | 0.02                          | 3.53 | 2               |
| Avenir Muller uncemented | Fitmore                  | 65      | 196.3              | 1              | 0.51                     | 0.01                          | 2.84 | 27              |
| Avenir Muller uncemented | Pinnacle                 | 99      | 799.3              | 3              | 0.38                     | 0.05                          | 1.00 | 0               |
| Avenir Muller uncemented | Titanium                 | 91      | 647.3              | 2              | 0.31                     | 0.04                          | 1.12 | 0               |
| Avenir Muller uncemented | RM cup                   | 105     | 825.7              | 1              | 0.12                     | 0.00                          | 0.67 | 0               |
| Basis                    | Reflection porous        | 108     | 844.6              | 2              | 0.24                     | 0.03                          | 0.86 | 0               |
| CBC                      | Expansys shell           | 183     | 1,791.9            | 25             | 1.40                     | 0.90                          | 2.06 | 0               |
| CBC                      | Fitmore                  | 59      | 597.2              | 5              | 0.84                     | 0.27                          | 1.95 | 0               |
| CBC                      | RM Pressfit cup          | 444     | 2,688.8            | 21             | 0.78                     | 0.48                          | 1.19 | 12              |
| CCA                      | Contemporary             | 74      | 762.2              | 10             | 1.31                     | 0.63                          | 2.41 | 0               |
| CCA                      | CCB                      | 769     | 6,078.6            | 31             | 0.51                     | 0.34                          | 0.71 | 4               |
| CCA                      | RM Pressfit cup          | 135     | 1,251.2            | 5              | 0.40                     | 0.13                          | 0.93 | 0               |
| Charnley                 | Charnley Cup Ogee        | 303     | 3,856.2            | 28             | 0.73                     | 0.48                          | 1.05 | 0               |
| Charnley                 | Charnley                 | 456     | 5,338.8            | 23             | 0.43                     | 0.27                          | 0.64 | 0               |
| CLS                      | Artek                    | 59      | 712.1              | 25             | 3.51                     | 2.22                          | 5.10 | 0               |
| CLS                      | Durom                    | 198     | 1,950.2            | 63             | 3.23                     | 2.48                          | 4.13 | 0               |
| CLS                      | RM cup                   | 113     | 1,173.0            | 17             | 1.45                     | 0.81                          | 2.27 | 0               |
| CLS                      | Allofit                  | 192     | 1,924.0            | 20             | 1.04                     | 0.61                          | 1.57 | 0               |
| CLS                      | Duraloc                  | 699     | 9,128.1            | 91             | 1.00                     | 0.80                          | 1.22 | 0               |
| CLS                      | Titanium                 | 80      | 362.3              | 3              | 0.83                     | 0.17                          | 2.42 | 4               |
| CLS                      | Continuum TM             | 719     | 2,920.8            | 22             | 0.75                     | 0.47                          | 1.14 | 75              |
| CLS                      | Trident                  | 165     | 1,862.0            | 14             | 0.75                     | 0.39                          | 1.23 | 0               |
| CLS                      | CLS Expansion            | 1263    | 15,710.0           | 118            | 0.75                     | 0.62                          | 0.90 | 0               |
| CLS                      | Trabecular Metal Shell   | 53      | 420.1              | 3              | 0.71                     | 0.15                          | 2.09 | 2               |
| CLS                      | Reflection porous        | 373     | 2,990.9            | 20             | 0.67                     | 0.41                          | 1.03 | 15              |
| CLS                      | RM Pressfit cup          | 567     | 3,853.8            | 25             | 0.65                     | 0.42                          | 0.96 | 27              |
| CLS                      | Weill ring               | 106     | 1,601.3            | 10             | 0.62                     | 0.28                          | 1.11 | 0               |
| CLS                      | Trilogy                  | 596     | 4,236.2            | 25             | 0.59                     | 0.38                          | 0.87 | 24              |
| CLS                      | Fitmore                  | 2300    | 23,622.5           | 117            | 0.50                     | 0.41                          | 0.59 | 34              |
| CLS                      | Morscher                 | 1682    | 22,801.5           | 112            | 0.49                     | 0.40                          | 0.59 | 0               |
| CLS                      | Monoblock Acetabular Cup | 80      | 848.4              | 4              | 0.47                     | 0.13                          | 1.21 | 0               |
| CLS                      | Pinnacle                 | 90      | 614.1              | 2              | 0.33                     | 0.04                          | 1.18 | 10              |
| Contemporary             | Contemporary             | 71      | 912.4              | 12             | 1.32                     | 0.68                          | 2.30 | 0               |
| Corail                   | ASR                      | 156     | 1,199.7            | 83             | 6.92                     | 5.51                          | 8.58 | 0               |
| Corail                   | RM Pressfit cup          | 144     | 504.8              | 5              | 0.99                     | 0.32                          | 2.31 | 7               |






| Femur Prosthesis | Acetabular Prosthesis    | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      | Procedures 2018 |
|------------------|--------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|-----------------|
| Corail           | Duraloc                  | 464     | 5,139.0            | 49             | 0.95                     | 0.71                          | 1.26 | 0               |
| Corail           | Fitmore                  | 285     | 893.5              | 7              | 0.78                     | 0.31                          | 1.61 | 17              |
| Corail           | Continuum TM             | 304     | 1,099.0            | 8              | 0.73                     | 0.31                          | 1.43 | 35              |
| Corail           | Pinnacle                 | 10351   | 48,382.9           | 331            | 0.68                     | 0.61                          | 0.76 | 1471            |
| Corail           | Trident                  | 88      | 480.8              | 3              | 0.62                     | 0.13                          | 1.82 | 8               |
| Corail           | Monoblock Acetabular Cup | 95      | 941.1              | 5              | 0.53                     | 0.17                          | 1.24 | 0               |
| Corail           | Tritanium                | 168     | 822.6              | 4              | 0.49                     | 0.13                          | 1.24 | 15              |
| Corail           | Delta-PF Cup             | 80      | 893.5              | 3              | 0.34                     | 0.07                          | 0.98 | 0               |
| Corail           | Ultima                   | 135     | 1,205.2            | 4              | 0.33                     | 0.09                          | 0.85 | 0               |
| Corail           | Trilogy                  | 207     | 953.4              | 3              | 0.31                     | 0.06                          | 0.92 | 17              |
| Corail           | Reflection porous        | 140     | 1,360.4            | 2              | 0.15                     | 0.02                          | 0.53 | 0               |
| Corail           | DeltaMotion Cup          | 78      | 518.8              | 0              | 0.00                     | 0.00                          | 0.71 | 0               |
| CPCS             | R3 porous                | 326     | 889.9              | 6              | 0.67                     | 0.21                          | 1.39 | 63              |
| CPT              | G7 acetabular shell      | 58      | 95.0               | 3              | 3.16                     | 0.65                          | 9.23 | 26              |
| CPT              | Delta-TT Cup             | 79      | 182.5              | 3              | 1.64                     | 0.34                          | 4.80 | 18              |
| CPT              | Tritanium                | 85      | 611.2              | 6              | 0.98                     | 0.36                          | 2.14 | 0               |
| CPT              | Fitmore                  | 191     | 1,053.9            | 10             | 0.95                     | 0.46                          | 1.75 | 10              |
| CPT              | Continuum TM             | 1387    | 4,915.7            | 46             | 0.94                     | 0.69                          | 1.25 | 191             |
| CPT              | Trilogy                  | 843     | 6,459.6            | 56             | 0.87                     | 0.65                          | 1.13 | 1               |
| CPT              | Monoblock Acetabular Cup | 84      | 937.6              | 8              | 0.85                     | 0.33                          | 1.61 | 0               |
| CPT              | Trident                  | 145     | 1,620.2            | 12             | 0.74                     | 0.38                          | 1.29 | 0               |
| CPT              | Duraloc                  | 212     | 2,467.9            | 16             | 0.65                     | 0.37                          | 1.05 | 0               |
| CPT              | ZCA                      | 545     | 5,487.9            | 35             | 0.64                     | 0.44                          | 0.89 | 2               |
| CPT              | Pinnacle                 | 64      | 522.6              | 2              | 0.38                     | 0.05                          | 1.38 | 0               |
| CPT              | ZCA all-poly cup         | 96      | 468.8              | 1              | 0.21                     | 0.01                          | 1.19 | 0               |
| C-Stem           | Pinnacle                 | 77      | 260.2              | 4              | 1.54                     | 0.42                          | 3.94 | 12              |
| C-Stem           | Duraloc                  | 53      | 634.1              | 6              | 0.95                     | 0.35                          | 2.06 | 0               |
| C-Stem           | Elite Plus Ogee          | 55      | 548.6              | 2              | 0.36                     | 0.04                          | 1.32 | 0               |
| C-Stem           | Marathon cemented        | 89      | 310.1              | 0              | 0.00                     | 0.00                          | 1.19 | 14              |
| C-Stem AMT       | RM Pressfit cup          | 129     | 467.7              | 4              | 0.86                     | 0.23                          | 2.19 | 23              |
| C-Stem AMT       | Pinnacle                 | 2339    | 8,132.8            | 66             | 0.81                     | 0.62                          | 1.03 | 356             |
| C-Stem AMT       | Marathon cemented        | 330     | 1,763.1            | 10             | 0.57                     | 0.27                          | 1.04 | 14              |
| Echo Bi-Metric   | Continuum TM             | 106     | 183.8              | 3              | 1.63                     | 0.34                          | 4.77 | 26              |
| Echo Bi-Metric   | G7 acetabular shell      | 320     | 722.1              | 5              | 0.69                     | 0.22                          | 1.62 | 95              |
| Echo Bi-Metric   | Exceed ABT Ringloc-X     | 57      | 316.6              | 1              | 0.32                     | 0.01                          | 1.76 | 0               |
| Elite plus       | Duraloc                  | 608     | 7,022.0            | 112            | 1.59                     | 1.31                          | 1.92 | 0               |
| Elite plus       | Charnley                 | 298     | 3,661.5            | 23             | 0.63                     | 0.40                          | 0.94 | 0               |




| Femur Prosthesis | Acetabular Prosthesis    | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      | Procedures 2018 |
|------------------|--------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|-----------------|
| Elite plus       | Elite Plus Ogee          | 110     | 1,084.5            | 6              | 0.55                     | 0.20                          | 1.20 | 0               |
| Elite plus       | Elite Plus LPW           | 282     | 3,075.1            | 15             | 0.49                     | 0.27                          | 0.80 | 0               |
| Exeter           | Duraloc                  | 553     | 7,808.0            | 113            | 1.45                     | 1.19                          | 1.74 | 0               |
| Exeter           | Contemporary             | 1551    | 18,301.8           | 182            | 0.99                     | 0.85                          | 1.15 | 0               |
| Exeter           | Exeter                   | 1326    | 15,253.9           | 114            | 0.75                     | 0.62                          | 0.90 | 0               |
| Exeter           | Osteolock                | 836     | 11,108.7           | 75             | 0.68                     | 0.53                          | 0.85 | 0               |
| Exeter           | CLS Expansion            | 129     | 1,578.9            | 10             | 0.63                     | 0.30                          | 1.16 | 0               |
| Exeter           | Bio-clad poly            | 113     | 1,254.7            | 7              | 0.56                     | 0.22                          | 1.15 | 0               |
| Exeter           | Trilogy                  | 213     | 2,940.2            | 14             | 0.48                     | 0.26                          | 0.80 | 0               |
| Exeter           | Morscher                 | 551     | 8,255.2            | 36             | 0.44                     | 0.31                          | 0.60 | 0               |
| Exeter           | Muller PE cup            | 119     | 1,465.0            | 6              | 0.41                     | 0.15                          | 0.89 | 0               |
| Exeter           | Trident                  | 84      | 1,270.3            | 0              | 0.00                     | 0.00                          | 0.29 | 0               |
| Exeter V40       | G7 acetabular shell      | 112     | 212.7              | 3              | 1.41                     | 0.29                          | 4.12 | 39              |
| Exeter V40       | Trabecular Metal Shell   | 211     | 1,079.1            | 13             | 1.20                     | 0.64                          | 2.06 | 14              |
| Exeter V40       | Duraloc                  | 987     | 10,786.5           | 100            | 0.93                     | 0.75                          | 1.12 | 0               |
| Exeter V40       | Continuum TM             | 2539    | 10,125.6           | 86             | 0.85                     | 0.68                          | 1.05 | 236             |
| Exeter V40       | Bio-clad poly            | 140     | 980.3              | 7              | 0.71                     | 0.29                          | 1.47 | 0               |
| Exeter V40       | R3 porous                | 605     | 2,186.6            | 15             | 0.69                     | 0.38                          | 1.13 | 70              |
| Exeter V40       | Tritanium                | 3120    | 11,169.7           | 74             | 0.66                     | 0.52                          | 0.83 | 418             |
| Exeter V40       | Delta-TT Cup             | 223     | 771.1              | 5              | 0.65                     | 0.21                          | 1.51 | 34              |
| Exeter V40       | Exeter                   | 1639    | 15,405.7           | 84             | 0.55                     | 0.43                          | 0.68 | 1               |
| Exeter V40       | Contemporary             | 6504    | 49,605.6           | 235            | 0.47                     | 0.42                          | 0.54 | 114             |
| Exeter V40       | Morscher                 | 630     | 7226.3             | 34             | 0.47                     | 0.33                          | 0.66 | 0               |
| Exeter V40       | Osteolock                | 270     | 3,198.6            | 15             | 0.47                     | 0.25                          | 0.75 | 0               |
| Exeter V40       | Pinnacle                 | 2448    | 11,707.0           | 53             | 0.45                     | 0.34                          | 0.59 | 290             |
| Exeter V40       | CCB                      | 567     | 3,018.0            | 13             | 0.43                     | 0.23                          | 0.74 | 33              |
| Exeter V40       | Trident                  | 10390   | 63,631.7           | 274            | 0.43                     | 0.38                          | 0.48 | 1049            |
| Exeter V40       | Reflection cemented      | 926     | 5,623.2            | 23             | 0.41                     | 0.26                          | 0.61 | 35              |
| Exeter V40       | Trilogy                  | 2805    | 19,700.1           | 78             | 0.40                     | 0.31                          | 0.49 | 149             |
| Exeter V40       | Exeter X3                | 2103    | 7,355.7            | 28             | 0.38                     | 0.25                          | 0.54 | 336             |
| Exeter V40       | RM Pressfit cup          | 2385    | 10,968.5           | 37             | 0.34                     | 0.23                          | 0.46 | 301             |
| Exeter V40       | Muller PE cup            | 94      | 906.0              | 3              | 0.33                     | 0.07                          | 0.97 | 0               |
| Exeter V40       | Monoblock Acetabular Cup | 123     | 1,575.0            | 5              | 0.32                     | 0.10                          | 0.74 | 0               |
| Exeter V40       | Reflection porous        | 475     | 4,076.5            | 12             | 0.29                     | 0.15                          | 0.51 | 0               |
| Exeter V40       | CLS Expansion            | 88      | 1,008.5            | 2              | 0.20                     | 0.02                          | 0.72 | 0               |
| Exeter V40       | Weber                    | 53      | 561.9              | 1              | 0.18                     | 0.00                          | 0.99 | 0               |
| Exeter V40       | ZCA                      | 93      | 574.3              | 1              | 0.17                     | 0.00                          | 0.97 | 8               |
| Exeter V40       | Fitmore                  | 963     | 4,733.5            | 8              | 0.17                     | 0.07                          | 0.33 | 91              |
| Exeter V40       | Polymax                  | 63      | 61.3               | 0              | 0.00                     | 0.00                          | 6.02 | 35              |



| Femur Prosthesis      | Acetabular Prosthesis | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |       | Procedures 2018 |
|-----------------------|-----------------------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|-----------------|
| Exeter V40            | ZCA all-poly cup      | 104     | 432.0              | 0              | 0.00                     | 0.00                          | 0.85  | 4               |
| Friendly              | Delta-TT Cup          | 67      | 410.6              | 5              | 1.22                     | 0.40                          | 2.84  | 1               |
| Friendly              | Delta-PF Cup          | 168     | 1,700.2            | 5              | 0.29                     | 0.10                          | 0.69  | 0               |
| Furlong               | Furlong               | 66      | 789.3              | 6              | 0.76                     | 0.28                          | 1.65  | 0               |
| H-Max C               | Delta-TT Cup          | 61      | 134.1              | 2              | 1.49                     | 0.18                          | 5.39  | 21              |
| H-Max M               | Delta-PF Cup          | 71      | 535.0              | 7              | 1.31                     | 0.47                          | 2.57  | 0               |
| H-Max M               | Delta-TT Cup          | 86      | 656.1              | 3              | 0.46                     | 0.09                          | 1.34  | 0               |
| H-Max S               | Delta-PF Cup          | 194     | 591.6              | 7              | 1.18                     | 0.42                          | 2.32  | 10              |
| H-Max S               | Delta-TT Cup          | 737     | 2,937.5            | 27             | 0.92                     | 0.59                          | 1.32  | 48              |
| Lateral straight stem | Trilogy               | 69      | 555.6              | 12             | 2.16                     | 1.12                          | 3.77  | 0               |
| Lateral straight stem | RM cup                | 533     | 5245.0             | 42             | 0.80                     | 0.58                          | 1.08  | 0               |
| Lateral straight stem | Muller PE cup         | 750     | 7,185.8            | 39             | 0.54                     | 0.39                          | 0.74  | 1               |
| Lateral straight stem | Continuum TM          | 78      | 476.3              | 2              | 0.42                     | 0.05                          | 1.52  | 0               |
| Lateral straight stem | Weber                 | 287     | 2,865.7            | 10             | 0.35                     | 0.17                          | 0.64  | 0               |
| Lateral straight stem | RM Pressfit cup       | 173     | 1,253.4            | 3              | 0.24                     | 0.05                          | 0.70  | 0               |
| Lateral straight stem | ZCA                   | 98      | 750.9              | 1              | 0.13                     | 0.00                          | 0.74  | 0               |
| Lateral straight stem | ZCA all-poly cup      | 70      | 422.4              | 0              | 0.00                     | 0.00                          | 0.87  | 0               |
| M/L Taper             | Delta-TT Cup          | 64      | 323.4              | 5              | 1.55                     | 0.50                          | 3.61  | 0               |
| M/L Taper             | Continuum TM          | 1004    | 4,120.2            | 34             | 0.83                     | 0.57                          | 1.15  | 123             |
| M/L Taper             | Trident               | 249     | 748.4              | 5              | 0.67                     | 0.22                          | 1.56  | 52              |
| M/L Taper             | Trilogy               | 215     | 1,853.2            | 9              | 0.49                     | 0.20                          | 0.89  | 3               |
| Mallory-Head          | M2A                   | 105     | 1,237.7            | 15             | 1.21                     | 0.68                          | 2.00  | 0               |
| MasterSL              | Delta-TT Cup          | 59      | 52.6               | 2              | 3.81                     | 0.46                          | 13.75 | 36              |
| MS 30                 | Contemporary          | 128     | 1,229.7            | 10             | 0.81                     | 0.39                          | 1.50  | 0               |
| MS 30                 | Duraloc               | 55      | 776.5              | 6              | 0.77                     | 0.28                          | 1.68  | 0               |
| MS 30                 | Morscher              | 787     | 9,737.6            | 61             | 0.63                     | 0.48                          | 0.80  | 0               |
| MS 30                 | RM Pressfit cup       | 90      | 774.5              | 4              | 0.52                     | 0.14                          | 1.32  | 0               |
| MS 30                 | Continuum TM          | 404     | 1,595.9            | 7              | 0.44                     | 0.16                          | 0.86  | 33              |
| MS 30                 | Muller PE cup         | 462     | 4,414.8            | 15             | 0.34                     | 0.19                          | 0.56  | 0               |
| MS 30                 | Fitmore               | 2174    | 14,197.5           | 39             | 0.27                     | 0.20                          | 0.38  | 192             |
| MS 30                 | Trilogy               | 331     | 1,958.0            | 4              | 0.20                     | 0.06                          | 0.52  | 19              |
| MS 30                 | ZCA all-poly cup      | 94      | 506.3              | 1              | 0.20                     | 0.01                          | 1.10  | 0               |
| Omnifit               | Trident               | 149     | 1,814.5            | 13             | 0.72                     | 0.36                          | 1.19  | 0               |
| Optimys               | RM Pressfit cup       | 106     | 140.0              | 1              | 0.71                     | 0.02                          | 3.98  | 46              |
| Polarstem uncemented  | Reflection porous     | 335     | 2,072.2            | 14             | 0.68                     | 0.37                          | 1.13  | 0               |



| Femur Prosthesis       | Acetabular Prosthesis    | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |       | Procedures 2018 |
|------------------------|--------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|-----------------|
| Polarstem uncemented   | R3 porous                | 1517    | 4,554.6            | 25             | 0.55                     | 0.36                          | 0.81  | 271             |
| Prodigy                | Duraloc                  | 113     | 1,479.9            | 23             | 1.55                     | 0.99                          | 2.33  | 0               |
| SL modular stem        | RM cup                   | 322     | 4,578.4            | 40             | 0.87                     | 0.62                          | 1.19  | 0               |
| SL modular stem        | Muller PE cup            | 83      | 1,104.8            | 2              | 0.18                     | 0.02                          | 0.65  | 0               |
| SL monoblock           | Muller PE cup            | 488     | 5,462.9            | 24             | 0.44                     | 0.27                          | 0.64  | 0               |
| Spectron               | Duraloc                  | 1151    | 13,752.9           | 177            | 1.29                     | 1.10                          | 1.49  | 0               |
| Spectron               | Muller PE cup            | 66      | 658.5              | 8              | 1.21                     | 0.52                          | 2.39  | 0               |
| Spectron               | Reflection cemented      | 2956    | 29,689.6           | 353            | 1.19                     | 1.07                          | 1.32  | 2               |
| Spectron               | Morscher                 | 210     | 2,769.6            | 31             | 1.12                     | 0.76                          | 1.59  | 0               |
| Spectron               | Reflection porous        | 2755    | 27,633.0           | 228            | 0.83                     | 0.72                          | 0.94  | 0               |
| Spectron               | Trident                  | 78      | 886.1              | 5              | 0.56                     | 0.15                          | 1.24  | 0               |
| Spectron               | Fitmore                  | 78      | 978.2              | 5              | 0.51                     | 0.14                          | 1.12  | 0               |
| Spectron               | Biomex acet shell porous | 68      | 1023.2             | 5              | 0.49                     | 0.16                          | 1.14  | 0               |
| Spectron               | Mallory-Head             | 152     | 1,793.2            | 8              | 0.45                     | 0.19                          | 0.88  | 0               |
| Spectron               | R3 porous                | 432     | 2,439.2            | 7              | 0.29                     | 0.10                          | 0.56  | 11              |
| S-Rom                  | ASR                      | 130     | 801.1              | 94             | 11.73                    | 9.43                          | 14.29 | 0               |
| S-Rom                  | Ultima                   | 78      | 1,250.4            | 14             | 1.12                     | 0.61                          | 1.88  | 0               |
| S-Rom                  | Pinnacle                 | 375     | 3,633.1            | 35             | 0.96                     | 0.66                          | 1.32  | 8               |
| Standard straight stem | RM cup                   | 138     | 1,540.2            | 11             | 0.71                     | 0.33                          | 1.24  | 0               |
| Standard straight stem | Muller PE cup            | 629     | 5,764.2            | 21             | 0.36                     | 0.22                          | 0.55  | 1               |
| Standard straight stem | ZCA all-poly cup         | 50      | 290.6              | 1              | 0.34                     | 0.00                          | 1.92  | 0               |
| Standard straight stem | Weber                    | 134     | 1,294.7            | 4              | 0.31                     | 0.08                          | 0.79  | 0               |
| Standard straight stem | RM Pressfit cup          | 137     | 1,090.0            | 1              | 0.09                     | 0.00                          | 0.51  | 0               |
| Stemsys                | Polymax                  | 119     | 254.4              | 3              | 1.18                     | 0.24                          | 3.45  | 20              |
| Stemsys                | Agilis Ti-por            | 439     | 1,501.4            | 14             | 0.93                     | 0.49                          | 1.52  | 46              |
| Stemsys                | Fixa Ti Por              | 707     | 2,700.6            | 14             | 0.52                     | 0.27                          | 0.85  | 77              |
| Stemsys                | RM Pressfit cup          | 325     | 1,189.5            | 5              | 0.42                     | 0.14                          | 0.98  | 38              |
| Stemsys                | DeltaMotion Cup          | 485     | 2,413.5            | 6              | 0.25                     | 0.08                          | 0.51  | 44              |
| Stemsys                | Delta-PF Cup             | 396     | 1,077.1            | 1              | 0.09                     | 0.00                          | 0.43  | 74              |
| Stemsys cemented       | RM Pressfit cup          | 64      | 142.1              | 0              | 0.00                     | 0.00                          | 2.60  | 18              |
| Summit                 | ASR                      | 88      | 741.0              | 36             | 4.86                     | 3.40                          | 6.73  | 0               |
| Summit                 | Pinnacle                 | 2248    | 13,418.1           | 101            | 0.75                     | 0.61                          | 0.91  | 185             |
| Summit                 | Trilogy                  | 168     | 1,286.5            | 6              | 0.47                     | 0.17                          | 1.02  | 11              |
| Summit                 | Duraloc                  | 101     | 1,212.1            | 5              | 0.41                     | 0.13                          | 0.96  | 0               |
| Synergy Porous         | BHR Acetabular Cup       | 114     | 1,034.2            | 40             | 3.87                     | 2.76                          | 5.27  | 0               |
| Synergy Porous         | R3 porous                | 1779    | 7,919.5            | 58             | 0.73                     | 0.56                          | 0.95  | 126             |




| Femur Prosthesis      | Acetabular Prosthesis    | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      | Procedures 2018 |
|-----------------------|--------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|-----------------|
| Synergy Porous        | Reflection porous        | 1238    | 11,875.0           | 43             | 0.36                     | 0.26                          | 0.48 | 14              |
| Synergy Porous        | Delta-PF Cup             | 88      | 7,79.7             | 1              | 0.13                     | 0.00                          | 0.71 | 0               |
| Synergy Porous        | Continuum TM             | 55      | 120.5              | 0              | 0.00                     | 0.00                          | 3.06 | 6               |
| Taperloc Complete     | Continuum TM             | 108     | 101.1              | 3              | 2.97                     | 0.61                          | 8.67 | 63              |
| Taperloc Complete     | RM Pressfit cup          | 168     | 266.9              | 5              | 1.87                     | 0.61                          | 4.37 | 44              |
| Taperloc Complete     | G7 acetabular shell      | 259     | 390.5              | 6              | 1.54                     | 0.56                          | 3.34 | 96              |
| Trabecular Metal Stem | Continuum TM             | 447     | 2,151.9            | 18             | 0.84                     | 0.50                          | 1.32 | 10              |
| Trabecular Metal Stem | Monoblock Acetabular Cup | 74      | 822.4              | 3              | 0.36                     | 0.05                          | 0.97 | 0               |
| Tri-Lock BPS          | Pinnacle                 | 63      | 425.3              | 3              | 0.71                     | 0.10                          | 1.88 | 0               |
| Twinsys cemented      | Pinnacle                 | 76      | 286.1              | 7              | 2.45                     | 0.98                          | 5.04 | 8               |
| Twinsys cemented      | Selexys TPS              | 65      | 458.4              | 5              | 1.09                     | 0.35                          | 2.55 | 0               |
| Twinsys cemented      | CCB                      | 441     | 2,331.4            | 15             | 0.64                     | 0.36                          | 1.06 | 14              |
| Twinsys cemented      | RM Pressfit cup          | 1864    | 8557.7             | 42             | 0.49                     | 0.35                          | 0.66 | 197             |
| Twinsys cemented      | RM cup                   | 148     | 1,464.9            | 5              | 0.34                     | 0.09                          | 0.75 | 0               |
| Twinsys cemented      | Continuum TM             | 108     | 408.7              | 1              | 0.24                     | 0.01                          | 1.36 | 6               |
| Twinsys cemented      | Reflection porous        | 59      | 174.7              | 0              | 0.00                     | 0.00                          | 2.11 | 9               |
| Twinsys uncemented    | Selexys TPS              | 1231    | 1,0413.9           | 136            | 1.31                     | 1.10                          | 1.54 | 0               |
| Twinsys uncemented    | RM cup                   | 122     | 1,036.2            | 9              | 0.87                     | 0.40                          | 1.65 | 0               |
| Twinsys uncemented    | RM Pressfit cup          | 4855    | 29,448.8           | 191            | 0.65                     | 0.56                          | 0.75 | 241             |
| Twinsys uncemented    | Continuum TM             | 133     | 821.0              | 5              | 0.61                     | 0.20                          | 1.42 | 0               |
| Twinsys uncemented    | Trilogy                  | 209     | 1,811.8            | 11             | 0.61                     | 0.30                          | 1.09 | 0               |
| Twinsys uncemented    | Delta-PF Cup             | 370     | 3,000.8            | 1              | 0.03                     | 0.00                          | 0.19 | 0               |
| Versys                | Trilogy                  | 272     | 3,889.9            | 17             | 0.44                     | 0.25                          | 0.70 | 0               |
| Versys cemented       | ZCA                      | 391     | 4,200.1            | 26             | 0.62                     | 0.40                          | 0.91 | 0               |
| Versys cemented       | Trilogy                  | 237     | 2,636.1            | 7              | 0.27                     | 0.11                          | 0.55 | 0               |
| Wagner cone stem      | Fitmore                  | 73      | 803.3              | 4              | 0.50                     | 0.14                          | 1.27 | 1               |

## Revision vs Bearing Surface Articulations vs Head sizes 28mm, 32mm, 36mm & >36mm

| Size | Surfaces | No. Ops. | Observed comp. yrs. | Number Revised | Rate/100 component-years | Exact 95% confidence interval |       |
|------|----------|----------|---------------------|----------------|--------------------------|-------------------------------|-------|
| <=28 | CC       | 777      | 8,017.0             | 59             | 0.74                     | 0.56                          | 0.95  |
| <=28 | CM       | 48       | 196.0               | 3              | 1.53                     | 0.21                          | 4.09  |
| <=28 | CP       | 11,780   | 11,081.8            | 750            | 0.68                     | 0.63                          | 0.73  |
| <=28 | MM       | 2,999    | 39,640.3            | 306            | 0.77                     | 0.69                          | 0.86  |
| <=28 | MP       | 46,332   | 437,788.4           | 3,109          | 0.71                     | 0.69                          | 0.74  |
| 32   | CC       | 3,800    | 30,548.7            | 167            | 0.55                     | 0.47                          | 0.63  |
| 32   | CP       | 14,155   | 53,919.0            | 297            | 0.55                     | 0.49                          | 0.62  |
| 32   | MM       | 480      | 4,816.0             | 42             | 0.87                     | 0.63                          | 1.18  |
| 32   | MP       | 29,256   | 136,243.4           | 798            | 0.59                     | 0.55                          | 0.63  |
| 36   | CC       | 7,352    | 46,365.5            | 266            | 0.57                     | 0.51                          | 0.65  |
| 36   | CM       | 443      | 3,657.0             | 27             | 0.74                     | 0.49                          | 1.07  |
| 36   | CP       | 6,553    | 22,930.6            | 145            | 0.63                     | 0.53                          | 0.74  |
| 36   | MM       | 1,002    | 10,398.2            | 129            | 1.24                     | 1.04                          | 1.47  |
| 36   | MP       | 3,940    | 15,308.0            | 111            | 0.73                     | 0.60                          | 0.87  |
| >36  | CC       | 1,835    | 8,735.1             | 44             | 0.50                     | 0.36                          | 0.67  |
| >36  | CM       | 7        | 62.5                | 0              | 0.00                     | 0.00                          | 5.90  |
| >36  | CP       | 7        | 19.1                | 0              | 0.00                     | 0.00                          | 19.34 |
| >36  | MM       | 1,648    | 14,787.0            | 550            | 3.72                     | 3.42                          | 4.04  |
| >36  | MP       | 34       | 185.3               | 1              | 0.54                     | 0.00                          | 3.01  |

## Summary Revision Rates vs Head Size

| Size | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <=28 | 61,936  | 596,423.4          | 4,227          | 0.71                     | 0.69                          | 0.73 |
| 32   | 47,691  | 225,527.1          | 1,304          | 0.58                     | 0.55                          | 0.61 |
| 36   | 19,290  | 98,659.4           | 678            | 0.69                     | 0.64                          | 0.74 |
| >36  | 3,531   | 23,789.1           | 595            | 2.50                     | 2.30                          | 2.71 |



### Revision Comparison Standard vs Cross linked Polyethylene

| Surfaces | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| CC       | 13,783  | 93,689.5           | 536            | 0.57                     | 0.52                          | 0.62 |
| CM       | 500     | 3,919.1            | 30             | 0.77                     | 0.52                          | 1.09 |
| CP       | 32,551  | 187,712.9          | 1,193          | 0.64                     | 0.60                          | 0.67 |
| PS       | 7,025   | 80,720.8           | 618            | 0.77                     | 0.71                          | 0.83 |
| PX       | 25,526  | 106,992.1          | 575            | 0.54                     | 0.49                          | 0.58 |
| MM       | 6,135   | 69,665.7           | 1,029          | 1.48                     | 1.39                          | 1.57 |
| MP       | 79,578  | 589,627.2          | 4,019          | 0.68                     | 0.66                          | 0.70 |
| PS       | 36,988  | 353,017.5          | 2,699          | 0.76                     | 0.74                          | 0.79 |
| PX       | 42,590  | 236,609.8          | 1,320          | 0.56                     | 0.53                          | 0.59 |

### Revision vs Bearing Surfaces of Uncemented Prostheses

| Surfaces | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| CC       | 10,789  | 74,417.0           | 445            | 0.60                     | 0.54                          | 0.66 |
| CM       | 480     | 3,877.6            | 29             | 0.75                     | 0.49                          | 1.06 |
| CP       | 21,537  | 118,993.5          | 763            | 0.64                     | 0.60                          | 0.69 |
| MM       | 5,389   | 62,074.4           | 938            | 1.51                     | 1.42                          | 1.61 |
| MP       | 15,840  | 106,188.0          | 808            | 0.76                     | 0.71                          | 0.81 |

### Revision vs Bearing Surfaces of Fully Cemented Prostheses

| Surfaces | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |       |
|----------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| CP       | 751     | 6,039.9            | 47             | 0.78                     | 0.57                          | 1.03  |
| MM       | 7       | 59.7               | 2              | 3.35                     | 0.41                          | 12.11 |
| MP       | 24,616  | 208,386.6          | 1,355          | 0.65                     | 0.62                          | 0.69  |

### Revision vs Bearing Surfaces of Hybrid Prostheses

| Surfaces | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |       |
|----------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| CC       | 2,994   | 19,272.5           | 91             | 0.47                     | 0.38                          | 0.58  |
| CM       | 20      | 41.4               | 1              | 2.41                     | 0.06                          | 13.44 |
| CP       | 10,263  | 62,679.5           | 383            | 0.61                     | 0.55                          | 0.67  |
| MM       | 739     | 7,531.6            | 89             | 1.18                     | 0.95                          | 1.45  |
| MP       | 39,122  | 275,052.6          | 1,856          | 0.67                     | 0.64                          | 0.71  |



### Summary for Revision vs Bearing Surfaces

| Surfaces | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| CC       | 13,783  | 93,689.5           | 536            | 0.57                     | 0.52                          | 0.62 |
| CM       | 500     | 3,919.1            | 30             | 0.77                     | 0.52                          | 1.09 |
| CP       | 32,551  | 187,712.9          | 1,193          | 0.64                     | 0.60                          | 0.67 |
| MM       | 6,135   | 69,665.7           | 1,029          | 1.48                     | 1.39                          | 1.57 |
| MP       | 79,578  | 589,627.2          | 4,019          | 0.68                     | 0.66                          | 0.70 |


### Revision vs Bearing Surface Options for 6 Acetabulae in common use

|                 |    | No. Ops | Observed comp. Yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |       |
|-----------------|----|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| RM Pressfit cup | M  | 333     | 3,252.3            | 24             | 0.74                     | 0.47                          | 1.10  |
|                 | PS | 6,131   | 41,856.2           | 240            | 0.57                     | 0.50                          | 0.65  |
|                 | PX | 5,521   | 19,049.0           | 100            | 0.52                     | 0.43                          | 0.64  |
|                 | P  | 11,652  | 60,905.2           | 340            | 0.56                     | 0.50                          | 0.62  |
| Pinnacle        | C  | 3,247   | 19,344.0           | 104            | 0.54                     | 0.44                          | 0.65  |
|                 | M  | 1,525   | 15,150.8           | 174            | 1.15                     | 0.98                          | 1.33  |
|                 | PS | 24      | 149.8              | 2              | 1.34                     | 0.16                          | 4.82  |
|                 | PX | 13,796  | 56,707.2           | 342            | 0.60                     | 0.54                          | 0.67  |
|                 | P  | 13,820  | 56,857.0           | 344            | 0.61                     | 0.54                          | 0.67  |
| R3 porous       | C  | 991     | 5,087.2            | 17             | 0.33                     | 0.19                          | 0.54  |
|                 | M  | 110     | 817.0              | 49             | 6.00                     | 4.44                          | 7.93  |
|                 | P  | 3,700   | 13,084.5           | 80             | 0.61                     | 0.48                          | 0.76  |
| Trident         | C  | 2,482   | 24,687.3           | 112            | 0.45                     | 0.37                          | 0.54  |
|                 | M  | 100     | 188.6              | 3              | 1.59                     | 0.22                          | 4.24  |
|                 | PS | 1       | 12.9               | 0              | 0.00                     | 0.00                          | 28.56 |
|                 | PX | 12,134  | 75,687.3           | 385            | 0.51                     | 0.46                          | 0.56  |
|                 | P  | 12,135  | 75,700.2           | 385            | 0.51                     | 0.46                          | 0.56  |
| Tritanium       | C  | 108     | 541.0              | 1              | 0.18                     | 0.00                          | 1.03  |
|                 | M  | 135     | 343.8              | 5              | 1.45                     | 0.47                          | 3.39  |
|                 | P  | 4,603   | 17,068.2           | 110            | 0.64                     | 0.53                          | 0.77  |
| Trilogy         | C  | 69      | 884.3              | 5              | 0.57                     | 0.18                          | 1.32  |
|                 | M  | 5       | 57.8               | 0              | 0.00                     | 0.00                          | 6.38  |
|                 | PS | 158     | 2,269.3            | 14             | 0.62                     | 0.34                          | 1.04  |
|                 | PX | 6,084   | 46,248.0           | 241            | 0.52                     | 0.46                          | 0.59  |
|                 | P  | 6,242   | 48,517.3           | 255            | 0.53                     | 0.46                          | 0.59  |

### Revision vs Monoblock Femoral Stems

| No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| 1,297   | 15,165             | 79             | 0.52                     | 0.41                          | 0.65 |





### Revision vs Acetabulum type

| Acetabulum type     | No. Ops. | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|---------------------|----------|--------------------|----------------|--------------------------|-------------------------------|------|
| Cemented            | 25,877   | 218,465.4          | 1,444          | 0.66                     | 0.63                          | 0.70 |
| Uncemented Liner    | 85,841   | 567,381.6          | 4,025          | 0.71                     | 0.69                          | 0.73 |
| Uncemented No Liner | 20,829   | 158,767.4          | 1,338          | 0.84                     | 0.80                          | 0.89 |

### Revision vs Age Bands

| Age Bands | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <55       | 19,673  | 161,726.1          | 1,688          | 1.04                     | 0.99                          | 1.09 |
| 55-64     | 33,927  | 264,056.1          | 2,212          | 0.84                     | 0.80                          | 0.87 |
| 65-74     | 45,626  | 330,429.4          | 2,051          | 0.62                     | 0.59                          | 0.65 |
| >=75      | 36,235  | 2159,26.1          | 1,014          | 0.47                     | 0.44                          | 0.50 |

### Revision vs Age Bands vs Bearing Surfaces

| Bearing Surface | Age Bands | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------------|-----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| CC              | <55       | 5,348   | 36,576.0           | 238            | 0.65                     | 0.57                          | 0.74 |
|                 | 55-64     | 5,579   | 38,901.2           | 194            | 0.50                     | 0.43                          | 0.57 |
|                 | 65-74     | 2,598   | 16,809.8           | 97             | 0.58                     | 0.47                          | 0.70 |
|                 | >=75      | 258     | 1,402.6            | 7              | 0.50                     | 0.20                          | 1.03 |
| CM              | <55       | 192     | 1,513.6            | 11             | 0.73                     | 0.36                          | 1.30 |
|                 | 55-64     | 212     | 1,720.0            | 15             | 0.87                     | 0.49                          | 1.44 |
|                 | 65-74     | 79      | 599.0              | 4              | 0.67                     | 0.18                          | 1.71 |
|                 | >=75      | 17      | 86.5               | 0              | 0.00                     | 0.00                          | 4.26 |
| CP              | <55       | 6,146   | 39,884.0           | 343            | 0.86                     | 0.77                          | 0.95 |
|                 | 55-64     | 11,380  | 68,720.8           | 433            | 0.63                     | 0.57                          | 0.69 |
|                 | 65-74     | 10,832  | 59,888.4           | 305            | 0.51                     | 0.45                          | 0.57 |
|                 | >=75      | 4,193   | 19,219.7           | 112            | 0.58                     | 0.48                          | 0.70 |
| MM              | <55       | 2,890   | 35,582.5           | 497            | 1.40                     | 1.28                          | 1.53 |
|                 | 55-64     | 2,394   | 26,778.9           | 437            | 1.63                     | 1.48                          | 1.79 |
|                 | 65-74     | 705     | 6,725.6            | 85             | 1.26                     | 1.01                          | 1.56 |
|                 | >=75      | 146     | 578.6              | 10             | 1.73                     | 0.83                          | 3.18 |
| MP              | <55       | 4,828   | 4,4793.6           | 563            | 1.26                     | 1.16                          | 1.37 |
|                 | 55-64     | 13,839  | 122,050.0          | 1,096          | 0.90                     | 0.85                          | 0.95 |
|                 | 65-74     | 30,352  | 235,623.5          | 1,503          | 0.64                     | 0.61                          | 0.67 |
|                 | >=75      | 30,559  | 187,160.2          | 857            | 0.46                     | 0.43                          | 0.49 |



### Revision vs Gender

| Sex | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| F   | 72,430  | 518,181.5          | 3,350          | 0.65                     | 0.62                          | 0.67 |
| M   | 63,031  | 453,956.2          | 3,615          | 0.80                     | 0.77                          | 0.82 |

### Revision vs Surgeon Annual Workload


| Operations per Year | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|---------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <10                 | 1,758   | 14,119.5           | 141            | 1.00                     | 0.84                          | 1.18 |
| 10-24               | 14,131  | 105,567.4          | 792            | 0.75                     | 0.70                          | 0.80 |
| 25-49               | 56,040  | 406,873.9          | 3,126          | 0.77                     | 0.74                          | 0.80 |
| 50-74               | 32,520  | 220,913.2          | 1,395          | 0.63                     | 0.60                          | 0.67 |
| 75-99               | 18,403  | 114,801.5          | 711            | 0.62                     | 0.57                          | 0.67 |
| >=100               | 12,609  | 109,862.3          | 800            | 0.73                     | 0.68                          | 0.78 |

### Revision vs Approach

| Approach  | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Anterior  | 4,461   | 38,954.9           | 303            | 0.78                     | 0.69                          | 0.87 |
| Posterior | 89,189  | 619,297.7          | 4,481          | 0.72                     | 0.70                          | 0.74 |
| Lateral   | 32,813  | 255,564.6          | 1,692          | 0.66                     | 0.63                          | 0.69 |
| Troch     | 215     | 1,692.6            | 23             | 1.36                     | 0.86                          | 2.04 |

### Revision vs Arthroplasty Fixation

| Fixation   | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Cemented   | 27,185  | 231,428.4          | 1,500          | 0.65                     | 0.62                          | 0.68 |
| Uncemented | 54,566  | 370,016.9          | 3,009          | 0.81                     | 0.78                          | 0.84 |
| Hybrid     | 53,710  | 370,692.4          | 2,456          | 0.66                     | 0.64                          | 0.69 |



### Revision by Arthroplasty Fixation vs Age Bands

| Age Bands  | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Cemented   |         |                    |                |                          |                               |      |
| <55        | 752     | 7,713.8            | 150            | 1.94                     | 1.64                          | 2.27 |
| 55-64      | 2,611   | 28,271.1           | 325            | 1.15                     | 1.03                          | 1.28 |
| 65-74      | 9,321   | 92,394.6           | 642            | 0.69                     | 0.64                          | 0.75 |
| >=75       | 14,501  | 103,048.8          | 383            | 0.37                     | 0.34                          | 0.41 |
| Uncemented |         |                    |                |                          |                               |      |
| <55        | 14,663  | 115,428.0          | 1,082          | 0.94                     | 0.88                          | 0.99 |
| 55-64      | 19,793  | 139,086.2          | 1,162          | 0.84                     | 0.79                          | 0.88 |
| 65-74      | 14,647  | 88,291.5           | 564            | 0.64                     | 0.59                          | 0.69 |
| >=75       | 5,463   | 27,211.2           | 201            | 0.74                     | 0.64                          | 0.85 |
| Hybrid     |         |                    |                |                          |                               |      |
| <55        | 4,258   | 38,584.2           | 456            | 1.18                     | 1.08                          | 1.30 |
| 55-64      | 11,523  | 96,698.7           | 725            | 0.75                     | 0.70                          | 0.81 |
| 65-74      | 21,658  | 149,743.4          | 845            | 0.56                     | 0.53                          | 0.60 |
| >=75       | 16,271  | 85,666.1           | 430            | 0.50                     | 0.46                          | 0.55 |

### Revision vs ASA Status

| ASA Class | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| 1         | 16,223  | 101,430.6          | 729            | 0.72                     | 0.67                          | 0.77 |
| 2         | 60,350  | 340,925.2          | 2,154          | 0.63                     | 0.61                          | 0.66 |
| 3         | 24,203  | 117,082.3          | 801            | 0.68                     | 0.64                          | 0.73 |
| 4         | 879     | 3,027.7            | 32             | 1.06                     | 0.71                          | 1.47 |


### Revision vs BMI Status

| BMI     | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|---------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| < 19    | 514     | 1,678.1            | 15             | 0.89                     | 0.48                          | 1.44 |
| 19 - 24 | 10,281  | 37,506.3           | 210            | 0.56                     | 0.49                          | 0.64 |
| 25 - 29 | 18,752  | 68,315.6           | 381            | 0.56                     | 0.50                          | 0.62 |
| 30 - 39 | 17,752  | 62,730.8           | 415            | 0.66                     | 0.60                          | 0.73 |
| 40+     | 1,928   | 6,342.0            | 71             | 1.12                     | 0.87                          | 1.41 |

### Revision for Deep Infection within six months vs Theatre Environment


| Theatre      | Total Number | Number revised | %     | Standard error |
|--------------|--------------|----------------|-------|----------------|
| Conventional | 79,409       | 135            | 0.170 | 0.0146         |
| Laminar flow | 48,214       | 130            | 0.270 | 0.0236         |


### % Revision for Deep infection within 6 months



|              |         | Total Number | Number revised | %     | Standard error |
|--------------|---------|--------------|----------------|-------|----------------|
| Conventional | Suit    | 10,971       | 22             | 0.201 | 0.043          |
|              | no suit | 68,438       | 113            | 0.165 | 0.016          |
| Laminar flow | Suit    | 24,975       | 64             | 0.256 | 0.032          |
|              | no suit | 23,239       | 66             | 0.284 | 0.035          |


### % Revision for Deep infection within 6 months






|         | Total Number | Number revised | %    | Standard error |
|---------|--------------|----------------|------|----------------|
| Suit    | 35,946       | 86             | 0.24 | 0.026          |
| No suit | 91,677       | 179            | 0.20 | 0.015          |

### % Revision for Deep infection within 6 months



### Comparison of Major vs Minor Revisions by Year




A major revision is defined as revision of acetabulum and/or femur including any of minor components and minor revision as change of head and/or liner only.

### Re-revisions for Major vs Minor Revisions

|       | No. Ops | Observed comp. yrs | Number Re-revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-------|---------|--------------------|-------------------|--------------------------|-------------------------------|------|
| Minor | 1,579   | 6,395.0            | 276               | 4.32                     | 3.82                          | 4.86 |
| Major | 5,349   | 22,431.4           | 730               | 3.25                     | 3.02                          | 3.50 |

### Percentage of hips revised in the first year



### Resurfacing Arthroplasty


| No. Ops | Observed comp. yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      |
|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| 1,877   | 14,163.3           | 150            | 1.06                     | 0.90                          | 1.24 |

### Resurfacing Prosthesis vs Revision Rate

| Prosthesis                 | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |       |
|----------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| Adept                      | 4       | 43.1               | 0              | 0.00                     | 0.00                          | 8.56  |
| ASR                        | 132     | 1,346.7            | 41             | 3.04                     | 2.18                          | 4.13  |
| BHR                        | 1,694   | 12,391.0           | 102            | 0.82                     | 0.67                          | 1.00  |
| BMHR                       | 28      | 216.8              | 1              | 0.46                     | 0.01                          | 2.57  |
| Conserve Superfinish       | 3       | 28.6               | 0              | 0.00                     | 0.00                          | 12.90 |
| Durom                      | 4       | 56.5               | 0              | 0.00                     | 0.00                          | 6.52  |
| Mitch TRH Resurfacing Head | 12      | 80.6               | 6              | 7.45                     | 2.73                          | 16.20 |


### Head size vs Revision Rate

| Head size | No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <=44      | 99      | 834.7              | 33             | 3.95                     | 2.72                          | 5.55 |
| 45-49     | 358     | 3,000.1            | 49             | 1.63                     | 1.21                          | 2.16 |
| 50-54     | 1,327   | 9,455.9            | 59             | 0.62                     | 0.47                          | 0.80 |
| >=55      | 93      | 872.6              | 9              | 1.03                     | 0.47                          | 1.96 |




## KAPLAN MEIER CURVES

The following Kaplan Meier survival analyses are for the 20 years 1999 – 2018 with deceased patients censored at time of death.



| Years | % Revision-free | No. in each year |
|-------|-----------------|------------------|
| 1     | 98.77           | 122,849          |
| 2     | 98.29           | 111,678          |
| 3     | 97.83           | 100,972          |
| 4     | 97.40           | 90,655           |
| 5     | 96.96           | 80,576           |
| 6     | 96.46           | 71,222           |
| 7     | 95.89           | 62,423           |
| 8     | 95.32           | 54,129           |
| 9     | 94.62           | 46,145           |
| 10    | 93.85           | 38,815           |
| 11    | 93.02           | 32,295           |
| 12    | 92.16           | 26,298           |
| 13    | 91.12           | 21,011           |
| 14    | 90.05           | 16,215           |
| 15    | 88.97           | 11,969           |
| 16    | 87.67           | 8,560            |
| 17    | 86.29           | 5,727            |
| 18    | 85.18           | 3,377            |
| 19    | 83.85           | 1,405            |



### Cemented

| Years | % Revision-free | No in each year |
|-------|-----------------|-----------------|
| 1     | 99.19           | 25,384          |
| 2     | 98.74           | 23,987          |
| 3     | 98.39           | 22,459          |
| 4     | 98.04           | 20,703          |
| 5     | 97.71           | 18,894          |
| 6     | 97.36           | 17,136          |
| 7     | 96.85           | 15,333          |
| 8     | 96.26           | 13,595          |
| 9     | 95.66           | 12,054          |
| 10    | 94.75           | 10,559          |
| 11    | 93.80           | 9,088           |
| 12    | 92.77           | 7,636           |
| 13    | 91.50           | 6,235           |
| 14    | 90.28           | 4,927           |
| 15    | 89.42           | 3,745           |
| 16    | 88.40           | 2,766           |
| 17    | 86.87           | 1,907           |
| 18    | 85.89           | 1,160           |
| 19    | 84.37           | 531             |


### Uncemented

| Years | % Revision-free | No in each year |
|-------|-----------------|-----------------|
| 1     | 98.54           | 49,189          |
| 2     | 97.98           | 44,314          |
| 3     | 97.43           | 39,856          |
| 4     | 96.84           | 35,603          |
| 5     | 96.28           | 31,503          |
| 6     | 95.61           | 27,649          |
| 7     | 94.95           | 24,117          |
| 8     | 94.28           | 20,653          |
| 9     | 93.44           | 16,985          |
| 10    | 92.69           | 13,625          |
| 11    | 91.88           | 10,875          |
| 12    | 91.11           | 8,547           |
| 13    | 90.19           | 6,599           |
| 14    | 89.41           | 4,988           |
| 15    | 88.40           | 3,638           |
| 16    | 86.94           | 2,539           |
| 17    | 85.71           | 1,674           |
| 18    | 84.75           | 981             |
| 19    | 83.62           | 404             |


### Hybrid

| Years | % Revision-free | No in each year |
|-------|-----------------|-----------------|
| 1     | 98.80           | 48,276          |
| 2     | 98.37           | 43,377          |
| 3     | 97.96           | 38,657          |
| 4     | 97.64           | 34,349          |
| 5     | 97.26           | 30,179          |
| 6     | 96.84           | 26,437          |
| 7     | 96.34           | 22,973          |
| 8     | 95.87           | 19,881          |
| 9     | 95.25           | 17,106          |
| 10    | 94.55           | 14,631          |
| 11    | 93.78           | 12,332          |
| 12    | 92.97           | 10,115          |
| 13    | 92.01           | 8,177           |
| 14    | 90.82           | 6,300           |
| 15    | 89.49           | 4,586           |
| 16    | 88.10           | 3,255           |
| 17    | 86.70           | 2,146           |
| 18    | 85.35           | 1,236           |
| 19    | 84.03           | 470             |







### Survival of Cemented vs Uncemented no Liner vs Uncemented with Liner




### Survival of Cemented vs Uncemented Acetabulae




### Survival of Cemented vs Uncemented Femoral components




### Survival of Head Sizes






### Survival of Bearing Surfaces





### Survival of Crosslinked vs Standard polyethylene






### Survival of combinations with > 2500 procedures






The following K M graphs are for the six main individual reasons for revision:

### Deep infection




### Dislocation







### Loosening acetabular component




### Loosening femoral component






### Fracture femur




### Pain







Survival for surgical approach




Survival for age bands








### Survival for surgeon annual output




### Survival male vs female





Survival vs BMI



Re-revisions of conventional hips

Analyses were undertaken of hip re-revisions.

There were 1,010 registered conventional hip replacements that had been revised twice, 238 that had been revised three times, 65 that had been revised four times, 20 that had been revised 5 times and 7 that had been revised 6 times. There was 1 patient who has now had 12 revisions.

Second revision


Time between the first and second revisions averaged 864 days, with a range of 0 – 6,257 and a standard deviation of 1,139. This compares to an average of 2,137 days between the primary and first revision.

Reason for revision

|                                |     |
|--------------------------------|-----|
| Deep infection                 | 305 |
| Dislocation                    | 285 |
| Loosening femoral component    | 123 |
| Loosening acetabulum component | 123 |
| Pain                           | 97  |
| Fracture femur                 | 75  |

Revision


|                      |     |
|----------------------|-----|
| Change of head       | 677 |
| Change of liner      | 462 |
| Change of acetabulum | 278 |
| Change of femoral    | 285 |
| Change of all        | 268 |



## Re-revisions

| No. Ops | Observed comp. yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| 6,965   | 29,075             | 1,010          | 3.47                     | 3.26                          | 3.69 |

| Years | Percentage re-revision free | No. in each year |
|-------|-----------------------------|------------------|
| 1     | 91.60                       | 5,162            |
| 2     | 88.80                       | 4,377            |
| 3     | 87.00                       | 3,684            |
| 4     | 85.70                       | 3,090            |
| 5     | 84.10                       | 2,515            |
| 6     | 82.50                       | 1,992            |
| 7     | 81.20                       | 1,520            |
| 8     | 80.10                       | 1,179            |
| 9     | 78.70                       | 899              |
| 10    | 77.40                       | 658              |
| 11    | 75.20                       | 470              |
| 12    | 73.70                       | 338              |
| 13    | 71.30                       | 226              |
| 14    | 70.50                       | 153              |
| 15    | 67.80                       | 91               |





### **Third revision**

The average time between second and third revisions for the 238 arthroplasties was 613 days with a range of 1 – 4,451 and a standard deviation of 829.

### **Fourth revision**

There were n = 65 registered with 4 revisions.

### **Fifth revision**

There were 20 registered 5 revisions.

### **Sixth revision**

There were 7 registered with 6 revisions.


One patient has had n = 12 revisions.

Overall it can be noted that the time between successive revisions steadily decreases.

### **Re- revisions of resurfacing hip replacements**

There have been 35 re-revisions.

The average time between the first and second revisions was 806 days, with a range of 11 – 3,036 and a standard deviation of 962. This compares with an average of 2,016 days between the primary resurfacing and the first revision.



## PATIENT BASED QUESTIONNAIRE OUTCOMES AT SIX MONTHS, FIVE YEARS, TEN YEARS, FIFTEEN YEARS AND TWENTY YEARS POST-SURGERY

### Questionnaires at six months post-surgery

At six months post-surgery a random selection of patients are sent the Oxford-12 questionnaire in order to achieve a response rate of 20% of the total which is deemed to be ample to provide powerful statistical analysis.

There are 12 questions with the scores now ranging from 4 to 0. A score of 48 is the best, indicating normal function. A score of 0 is the worst, indicating the most severe disability.

In addition we have grouped the questionnaire responses according to the classification system published by Kalairajah et al, 2005 (see appendix 1).

This groups each score into four categories:

|            |         |           |
|------------|---------|-----------|
| Category 1 | >41     | excellent |
| Category 2 | 34 – 41 | good      |
| Category 3 | 27 – 33 | fair      |
| Category 4 | < 27    | poor      |

For the twenty- year period, and as at July 2019, there were 32,754 primary hip questionnaire responses registered six months post-surgery. The average hip score was 40.39 (standard deviation 7.60, range 48 – 0).

|         |        |        |
|---------|--------|--------|
| Scoring | > 41   | 18,450 |
| Scoring | 34 -41 | 9,029  |
| Scoring | 27 -33 | 3,144  |
| Scoring | < 27   | 2,131  |

At six months post-surgery, 84% had an excellent or good score.

### Questionnaires at five years post-surgery

All patients who had a six month registered questionnaire, and who had not had revision surgery were sent a further questionnaire at five years post-surgery.

This dataset represents sequential Oxford hip scores for 12,115 individual patients.

At five years post-surgery, 89% of these patients achieved an excellent or good score and had an average of 42.40.

### Questionnaires at ten years post-surgery

All patients who had a six month registered questionnaire, and who had not had revision surgery were sent a further questionnaire at ten years post-surgery.

This dataset represents sequential Oxford hip scores for 8,164 individual patients.

At ten years post-surgery, 87% of these patients achieved an excellent or good score and had an average of 41.91.

### Questionnaires at fifteen years post-surgery

All patients who had a six month registered questionnaire, and who had not had revision surgery were sent a further questionnaire at 15 years post-surgery.

This dataset represents sequential Oxford hip scores for 3,170 individual patients.

At fifteen years post-surgery, 86% of these patients achieved an excellent or good score and had an average of 41.50.

### Questionnaires at twenty years post-surgery

All patients who had a six month registered questionnaire, and who had not had revision surgery were sent a further questionnaire at 20 years post-surgery.

This dataset represents sequential Oxford hip scores for 173 individual patients.

At twenty years post-surgery, 88% of these patients achieved an excellent or good score and had an average of 41.32

### Oxford Scores (at 6 m) vs BMI Status

| BMI          | Mean         | Standard Error of Mean | Number/ Group |
|--------------|--------------|------------------------|---------------|
| < 19         | 38.86        | 0.925                  | 76            |
| 19 - 24      | 40.93        | 0.165                  | 1,809         |
| 25 - 29      | 40.70        | 0.128                  | 3,102         |
| 30 - 39      | 39.31        | 0.153                  | 2,539         |
| 40+          | 37.15        | 0.567                  | 241           |
| <b>Total</b> | <b>40.17</b> | <b>0.084</b>           | <b>7,767</b>  |

### Revision hip questionnaire responses

There were 10,357 revision hip responses with 62% achieving an excellent or good score. This group includes all revision hip procedures including revisions of primary arthroplasties performed prior to 1999. The average revision hip score was 35.02 (standard deviation 9.87, range 48 – 2).




### OXFORD 12 SCORE AS A PREDICTOR OF HIP ARTHROPLASTY REVISION

A statistically significant relationship has been confirmed between the Oxford scores at six months, five and ten years' post-surgery and arthroplasty revision within two years of the Oxford 12 questionnaire date.

### Six month score and revision arthroplasty


By plotting the patients' six month scores in the Kalairajah groupings against the proportion of hips revised for that same group it demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score. A patient with a score below 27 has 13 times the risk of a revision within two years compared to a person with a score >41.




Revision risk versus Kalairajah groupings of Oxford scores within two years of the six month score date.

| Kalairajah Group | Number in Group | Number revised | %    | Standard error |
|------------------|-----------------|----------------|------|----------------|
| < 27             | 1,848           | 96             | 5.19 | 0.52           |
| 27_33            | 2,720           | 44             | 1.62 | 0.24           |
| 34_41            | 7,859           | 75             | 0.95 | 0.11           |
| 42+              | 16,186          | 67             | 0.41 | 0.05           |

In view of the large number of six month Oxford scores it is possible with statistical significance to further break down the score groupings to demonstrate an even more convincing relationship between score and risk of revision within two years.




Revision risk versus groupings of Oxford scores within two years of the six month score date



### Five year score and revision arthroplasty

As with the six month scores, plotting the patients' five year scores in the Kalairajah groupings against the proportion of hips revised for that same group demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score. A patient with a score below 27 has 10 times the risk of a revision within two years compared to a person with a score >41.



Revision risk versus Kalairajah groupings of Oxford scores within two years of the five year score date.

| Kalairajah Group | Number in Group | Number revised | %    | Standard error |
|------------------|-----------------|----------------|------|----------------|
| < 27             | 436             | 20             | 4.59 | 1.00           |
| 27_33            | 630             | 15             | 2.38 | 0.61           |
| 34_41            | 1,892           | 18             | 0.95 | 0.22           |
| 42+              | 6,949           | 33             | 0.47 | 0.08           |

### Ten year score and revision arthroplasty

As with the six month and five year scores, plotting the patients' ten year scores in the Kalairajah groupings against the proportion of hips revised for that same group demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score. A patient with a score below 27 has 7 times the risk of a revision within two years compared to a person with a score >41.



Revision risk versus Kalairajah groupings of Oxford scores within two years of the ten year score date



| Kalairajah Group | Number in Group | Number revised | %    | Standard error |
|------------------|-----------------|----------------|------|----------------|
| < 27             | 341             | 30             | 8.80 | 1.53           |
| 27_33            | 439             | 21             | 4.78 | 1.02           |
| 34_41            | 1,245           | 28             | 2.25 | 0.42           |
| 42+              | 4,190           | 51             | 1.22 | 0.17           |

**Fifteen year score and revision arthroplasty**

As with the six month, five year and 10 year scores, plotting the patients' fifteen year scores in the Kalairajah groupings against the proportion of hips revised for that same group demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score. A patient with a score below 27 has 7 times the risk of a revision within two years compared to a person with a score >41




| Kalairajah Group | Number in Group | Number revised | %    | Standard error |
|------------------|-----------------|----------------|------|----------------|
| < 27             | 136             | 13             | 9.56 | 2.52           |
| 27_33            | 141             | 6              | 4.26 | 1.70           |
| 34_41            | 392             | 12             | 3.06 | 0.87           |
| 42+              | 1,252           | 18             | 1.44 | 0.34           |

**Prediction of second revision from six month score following first revision**

Plotting the patients' six month scores, following their first revision in the Kalairajah groupings, against the proportion of hips revised for that same group, again demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score. A patient with a score below 27 has 6 times the risk of a revision within two years compared to a person with a score >41.






Second revision risk versus Kalairajah groupings of Oxford scores within two years of the six month post- first revision score date

| Kalairajah Group | Revision to 2 years | Number revised | %     | Standard error |
|------------------|---------------------|----------------|-------|----------------|
| < 27             | 1,452               | 148            | 10.19 | 0.79           |
| 27_33            | 1,404               | 70             | 4.99  | 0.58           |
| 34_41            | 2,561               | 68             | 2.66  | 0.32           |
| 42+              | 2,739               | 47             | 1.72  | 0.25           |

Mean Oxford scores at 6 months and 5 years for 8 hip combinations with > 2000 registrations.




# KNEE ARTHROPLASTY

## PRIMARY KNEE ARTHROPLASTY

The **twenty year** report analyses data for the period January 1999 – December 2018. There were 110,678 primary knee procedures registered, an additional 8,392 compared to last year's report.

The 110,678 includes 602 patello-femoral prostheses with 71 registered in 2018.



## Data Analysis

### Age and sex distribution

The average age for a knee replacement was 68 years, with a range of 8 – 100 years.

#### All knee arthroplasty

|               | Female | Male   |
|---------------|--------|--------|
| Number        | 57,152 | 53,529 |
| Percentage    | 51.64  | 48.36  |
| Mean age      | 68.52  | 67.86  |
| Maximum age   | 100.49 | 98.68  |
| Minimum age   | 10.17  | 8.19   |
| Standard dev. | 9.71   | 9.26   |

#### Conventional knee arthroplasty

|               | Female | Male   |
|---------------|--------|--------|
| Number        | 56,708 | 53,371 |
| Percentage    | 51.51  | 48.49  |
| Mean age      | 68.59  | 67.89  |
| Maximum age   | 100.49 | 98.68  |
| Minimum age   | 10.17  | 8.19   |
| Standard dev. | 9.67   | 9.24   |

### Patello-femoral arthroplasty


|               | Female | Male  |
|---------------|--------|-------|
| Number        | 444    | 158   |
| Percentage    | 73.75  | 26.25 |
| Mean age      | 60.03  | 60.13 |
| Maximum age   | 89.39  | 88.84 |
| Minimum age   | 31.15  | 31.25 |
| Standard dev. | 11.42  | 10.89 |

### Body Mass Index

For the nine-year period 2010 - 2018, there were 43,273 BMI registrations for primary knee replacements. The average was 31.30 (obese) with a range of 15 – 68.7 and a standard deviation of 6.00.

### Previous operation

|                         |        |
|-------------------------|--------|
| None                    | 92,959 |
| Meniscectomy            | 11,183 |
| Osteotomy               | 1,606  |
| Ligament reconstruction | 1,550  |
| Internal fixation       | 859    |
| Synovectomy             | 178    |



### Diagnosis

|                          |         |
|--------------------------|---------|
| Osteoarthritis           | 105,061 |
| Rheumatoid arthritis     | 2,362   |
| Post fracture            | 1,159   |
| Post ligament disruption | 931     |
| Other inflammatory       | 831     |
| Avascular necrosis       | 374     |
| Tumour                   | 96      |

### Approach


|                            |        |
|----------------------------|--------|
| Medial parapatellar        | 92,825 |
| Image guided               | 14,599 |
| Other                      | 2,653  |
| Lateral parapatellar       | 1,412  |
| Minimally invasive surgery | 227    |
| Robot navigation           | 44     |

### Bone graft


|                   |     |
|-------------------|-----|
| Femoral autograft | 314 |
| Femoral allograft | 17  |
| Femoral synthetic | 11  |
| Tibial autograft  | 105 |
| Tibial allograft  | 22  |
| Tibial synthetic  | 4   |

## Comparison of proportions of cemented vs. uncemented vs. hybrid by year


Hybrid knees have a cemented tibia and uncemented femur. Uncemented TKA fixation remains common in NZ, but in the last 3 years the previous downward trend has reversed.



## Proportion of fixed vs mobile knees by year



## Proportion of posterior stabilized vs cruciate retaining vs minimally stabilized knees by year



### Cement

|                      |         |     |
|----------------------|---------|-----|
| Femur cemented       | 101,886 | 92% |
| Antibiotic in cement | 68,784  | 68% |
| Tibia cemented       | 105,309 | 95% |
| Antibiotic in cement | 70,490  | 67% |

### Systemic antibiotic prophylaxis

|   |         |     |
|---|---------|-----|
| Patient number receiving at least one systemic antibiotic | 105,025 | 95% |
|---|---------|-----|

### Operating theatre

|              |        |
|--------------|--------|
| Conventional | 60,878 |
| Laminar flow | 48,884 |
| Space suits  | 37,360 |

### ASA Class

This was introduced with the updated forms at the beginning of 2005. For the fourteen year period 2005 – 2018, there were 88,309 (95%) primary knee procedures with the ASA class recorded.

### Definitions

**ASA class 1:** A healthy patient

**ASA class 2:** A patient with mild systemic disease

**ASA class 3:** A patient with severe systemic disease that limits activity but is not incapacitating

**ASA class 4:** A patient with an incapacitating disease that is a constant threat to life

| ASA | Number | Percentage |
|-----|--------|------------|
| 1   | 9,815  | 11         |
| 2   | 56,139 | 64         |
| 3   | 21,997 | 24         |
| 4   | 356    | 1          |

### Operative time (skin to skin in minutes)

|         |         |
|---------|---------|
| Average | 83 mins |
|---------|---------|

### Surgeon grade


The updated forms introduced in 2005 have separated advanced trainee into supervised and unsupervised. The following figures are for the fourteen-year period 2005 – 2018.

|                               |        |
|-------------------------------|--------|
| Consultant                    | 81,678 |
| Advanced trainee supervised   | 7,162  |
| Basic trainee                 | 1,798  |
| Advanced trainee unsupervised | 1,599  |


### Prosthesis usage

#### Patello-femoral prostheses used in 2018

|               |    |
|---------------|----|
| Gender        | 61 |
| Restoris Mako | 8  |
| Custom device | 2  |



### Patello- femoral prostheses used for five years 2014 - 2018



In 2018 there were 71 patello-femoral procedures registered to 27 surgeons.


### Conventional primary knees

#### Top ten knee prostheses used in 2018

|            |       |
|------------|-------|
| Triathlon  | 2,958 |
| Attune     | 2,015 |
| Persona    | 1,263 |
| Genesis II | 450   |
| Nexgen     | 362   |
| LCS        | 302   |
| PFC Sigma  | 293   |
| Balansys   | 148   |
| Sigma      | 141   |
| Vanguard   | 134   |



**Most Used Knee Prostheses per year for five years 2014 – 2018**



## Surgeon and hospital workload


### Surgeons

In 2018, 234 surgeons performed 8,371 total knee replacements, an average of 36 procedures per surgeon.

43 surgeons performed less than ten procedures and 73 performed more than 40.

### Hospitals

In 2018 primary knee replacement was performed in 51 hospitals. 27 were public hospitals and 24 were private.



## REVISION KNEE ARTHROPLASTY

Revision is defined by the Registry as a new operation in a previously replaced knee joint, during which one or more of the components is exchanged, removed, manipulated or added. It includes arthrodesis or amputation, but not soft tissue procedures. A two or more staged procedure is registered as one revision.

### Data analysis

For the twenty year period January 1999 – December 2018, there were 8,647 revision knee procedures registered. This is an additional 635 compared to last year's report.

The average age for a revision knee replacement was 70 years, with a range of 11 – 98 years.

### Revision knees

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 4,083  | 4,564 |
| Percentage    | 47.22  | 52.78 |
| Mean age      | 69.92  | 69.27 |
| Maximum age   | 95.80  | 98.39 |
| Minimum age   | 10.57  | 15.00 |
| Standard dev. | 10.24  | 10.10 |

### Body Mass Index

For the nine-year period 2010 - 2018, there were 1,843 BMI registrations for revision knee replacements. The average BMI was 31.40 (obese) with a range of 15 – 65 and a standard deviation of 6.10.

## REVISION OF REGISTERED PRIMARY KNEE ARTHROPLASTIES

This section analyses data for **revisions of the primary registered knee arthroplasties** for the twenty-year period.

There were 3,652 revisions of the 110,076 primary conventional total knee replacements (3.3%) and 58 revisions of the 602 patello-femoral prostheses (9.6%), a total of 3,710 revisions.

### Conventional knee replacement analysis

#### Time to revision

|                    |            |
|--------------------|------------|
| Average            | 1,514 days |
| Maximum            | 6,922 days |
| Minimum            | 1 day      |
| Standard deviation | 1,437 days |

#### Reason for revision

|                              |       |
|------------------------------|-------|
| Pain                         | 1,077 |
| Deep infection               | 974   |
| Primary patellar component   | 918   |
| Loosening tibial             | 849   |
| Loosening femoral component  | 395   |
| Loosening patellar component | 69    |
| Fracture femur               | 70    |
| Fracture tibia               | 42    |

*There is often more than one listed reason for revision and all are entered.*

### Analysis of the five main reasons for revision by year after primary procedure


*NB each year column does not add up to exactly 100% as often more than one cause for revision is listed and there are other reasons for revision other than the five above listed in the registry.*

| Years        | Loosening tibial component |      | Primary patellar component |      | Deep infection |      | Pain         |      | Loosening femoral component |      |
|--------------|----------------------------|------|----------------------------|------|----------------|------|--------------|------|-----------------------------|------|
|              | Count                      | %    | Count                      | %    | Count          | %    | Count        | %    | Count                       | %    |
| 0            | 43                         | 5.1  | 110                        | 12.0 | 384            | 39.4 | 142          | 13.2 | 16                          | 4.1  |
| 1            | 80                         | 9.4  | 244                        | 26.6 | 182            | 18.7 | 272          | 25.3 | 37                          | 9.4  |
| 2            | 109                        | 12.8 | 154                        | 16.8 | 99             | 10.2 | 170          | 15.8 | 34                          | 8.6  |
| 3            | 95                         | 11.2 | 91                         | 9.9  | 85             | 8.7  | 109          | 10.1 | 33                          | 8.4  |
| 4            | 78                         | 9.2  | 62                         | 6.8  | 51             | 5.2  | 73           | 6.8  | 43                          | 10.9 |
| 5            | 71                         | 8.4  | 39                         | 4.2  | 33             | 3.4  | 55           | 5.1  | 32                          | 8.1  |
| 6            | 77                         | 9.1  | 39                         | 4.2  | 33             | 3.4  | 42           | 3.9  | 30                          | 7.6  |
| 7            | 55                         | 6.5  | 36                         | 3.9  | 25             | 2.6  | 43           | 4.0  | 27                          | 6.8  |
| 8            | 41                         | 4.8  | 29                         | 3.2  | 18             | 1.8  | 36           | 3.3  | 22                          | 5.6  |
| 9            | 53                         | 6.2  | 26                         | 2.8  | 18             | 1.8  | 23           | 2.1  | 25                          | 6.3  |
| 10           | 32                         | 3.8  | 22                         | 2.4  | 12             | 1.2  | 32           | 3.0  | 18                          | 4.6  |
| >10          | 115                        | 13.5 | 66                         | 7.2  | 34             | 3.5  | 80           | 7.4  | 78                          | 19.7 |
| <b>Total</b> | <b>849</b>                 |      | <b>918</b>                 |      | <b>974</b>     |      | <b>1,077</b> |      | <b>395</b>                  |      |

## Analyses by numbers of the five main reasons for revision by year

|           | Loosening tibial component | Primary patellar component | Deep infection | Pain | Loosening femoral component |
|-----------|----------------------------|----------------------------|----------------|------|-----------------------------|
| Years     | No.                        | No.                        | No.            | No.  | No.                         |
| 1999-2007 | 140                        | 158                        | 169            | 213  | 76                          |
| 2008      | 42                         | 37                         | 47             | 55   | 25                          |
| 2009      | 52                         | 39                         | 54             | 51   | 24                          |
| 2010      | 53                         | 54                         | 40             | 61   | 20                          |
| 2011      | 52                         | 53                         | 44             | 70   | 24                          |
| 2012      | 54                         | 52                         | 68             | 63   | 21                          |
| 2013      | 62                         | 74                         | 73             | 78   | 30                          |
| 2014      | 63                         | 64                         | 85             | 81   | 39                          |
| 2015      | 59                         | 84                         | 91             | 97   | 24                          |
| 2016      | 90                         | 101                        | 115            | 105  | 41                          |
| 2017      | 86                         | 80                         | 112            | 103  | 37                          |
| 2018      | 96                         | 122                        | 76             | 100  | 34                          |

## Percentage of the 5 main reasons for revision by year



## Patello-Femoral Arthroplasty

### Revision of patello-femoral knees

Of the 602 registered, n = 58 have been revised.

### Time to revision

|                    |            |
|--------------------|------------|
| Average            | 1,779 days |
| Maximum            | 5,718 days |
| Minimum            | 108 days   |
| Standard deviation | 1,479 days |


### Reason for revision

|                    |    |
|--------------------|----|
| Pain               | 22 |
| Loosening patellar | 4  |
| Deep infection     | 4  |

## Patellar resurfacing

63 % of the 110,079 registered conventional primary knees did not have the patella resurfaced and 37% did have the patella resurfaced. Of the group that was not resurfaced, 914 subsequently had the patella resurfaced.





## Statistical note

In the table below there are two statistical terms readers may not be familiar with:

### i) Observed component years

This is the number of registered primary procedures multiplied by the number of years each component has been in situ.

### ii) Rate/100 component years

This is equivalent to the yearly revision rate expressed as a percent and is derived by dividing the number of prostheses revised by the observed component years multiplied by 100. It therefore allows for the number of years of post-operative follow up in calculating the revision rate. These rates are usually very low, hence it is expressed per 100 component years rather than per

component year. Statisticians consider that this is a more accurate way of deriving a revision rate for comparison when analysing data with widely varying follow up times. It is also important to note the confidence intervals. The closer they are to the estimated revision rate/100 component years, the more precise the estimate is.


## Statistical Significance

Where it is stated that a difference among results is significant the p value is 0.05 or less. In most of these situations this is because there is no overlap of the confidence intervals (CIs) but sometimes significance can apply in the presence of CI overlap.

### All Primary Conventional Knee Arthroplasties


| No. Ops | Observed comp. yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |      |
|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| 110,076 | 753,723            | 3,652          | 0.48                     | 0.47                          | 0.50 |


The figure below summarises the 16 Knee prostheses with >500 procedures. Showing the number of procedures for the history of the Registry and for the previous 2 years.





The figure below summarises the 16 Knee prostheses with >500 procedures. Showing the number of procedures for the previous 2 years and the historical revision rate.





## Revision Rate of Individual Knee Prostheses Sorted by Number of Arthroplasties

(Minimum of 50 arthroplasties)

| Prosthesis      | No. Ops | Observed comp. yrs | Number revised | Rate/100 component-years | Exact 95% confidence interval |       |
|-----------------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| Triathlon       | 23,650  | 116,467.5          | 501            | 0.430                    | 0.393                         | 0.469 |
| Nexgen          | 19,728  | 155,715.1          | 813            | 0.522                    | 0.487                         | 0.559 |
| LCS             | 14,769  | 139,891.7          | 670            | 0.479                    | 0.443                         | 0.517 |
| Genesis II      | 13,823  | 100,420.7          | 468            | 0.466                    | 0.425                         | 0.510 |
| PFC Sigma       | 10,789  | 91,493.9           | 356            | 0.389                    | 0.349                         | 0.431 |
| Attune          | 7,789   | 16,207.7           | 96             | 0.592                    | 0.480                         | 0.723 |
| Duracon         | 4,213   | 49,295.6           | 158            | 0.321                    | 0.272                         | 0.373 |
| Persona         | 3,854   | 6,842.5            | 49             | 0.716                    | 0.530                         | 0.947 |
| Vanguard        | 2,073   | 10,961.4           | 71             | 0.648                    | 0.502                         | 0.812 |
| Sigma           | 1,600   | 7,104.8            | 37             | 0.521                    | 0.367                         | 0.718 |
| Balansys        | 1,555   | 6,472.2            | 45             | 0.695                    | 0.507                         | 0.930 |
| Sigma CR150     | 997     | 6,197.2            | 27             | 0.000                    | 0.287                         | 0.634 |
| Scorpio         | 852     | 9,635.8            | 66             | 0.685                    | 0.530                         | 0.871 |
| Maxim           | 822     | 9,759.9            | 55             | 0.564                    | 0.425                         | 0.734 |
| Trekking        | 792     | 2,781.0            | 27             | 0.971                    | 0.625                         | 1.391 |
| Optetrak        | 661     | 5,798.3            | 57             | 0.983                    | 0.745                         | 1.274 |
| AGC             | 376     | 4,493.9            | 18             | 0.401                    | 0.237                         | 0.633 |
| Journey         | 330     | 1,526.3            | 17             | 1.114                    | 0.649                         | 1.783 |
| MBK             | 256     | 3,407.6            | 18             | 0.528                    | 0.313                         | 0.835 |
| Insall/Burstein | 249     | 2,988.9            | 47             | 1.572                    | 1.141                         | 2.072 |
| Legion          | 241     | 911.5              | 7              | 0.768                    | 0.309                         | 1.582 |
| Advance         | 157     | 1,804.4            | 6              | 0.333                    | 0.122                         | 0.724 |
| AMK             | 95      | 1,277.6            | 2              | 0.157                    | 0.019                         | 0.565 |
| Saiph           | 93      | 153.7              | 2              | 1.301                    | 0.158                         | 4.701 |
| ROCC            | 66      | 639.5              | 6              | 0.938                    | 0.344                         | 2.042 |

## Revision Rate of Individual Knee Prostheses Sorted by Revision Rate


(Minimum of 50 arthroplasties)

| Prosthesis       | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |       |
|------------------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| *Insall/Burstein | 249     | 2,988.9            | 47             | 1.572                    | 1.141                         | 2.072 |
| Saiph            | 93      | 153.7              | 2              | 1.301                    | 0.158                         | 4.701 |
| *#Journey        | 330     | 1,526.3            | 17             | 1.114                    | 0.649                         | 1.783 |
| *Optetrak        | 661     | 5,798.3            | 57             | 0.983                    | 0.745                         | 1.274 |
| *#Trekking       | 792     | 2,781.0            | 27             | 0.971                    | 0.625                         | 1.391 |
| ROCC             | 66      | 639.5              | 6              | 0.938                    | 0.344                         | 2.042 |
| Legion           | 241     | 911.5              | 7              | 0.768                    | 0.309                         | 1.582 |
| *#Persona        | 3,854   | 6,842.5            | 49             | 0.716                    | 0.530                         | 0.947 |
| *#Balansys       | 1,555   | 6,472.2            | 45             | 0.695                    | 0.507                         | 0.930 |
| *Scorpio         | 852     | 9,635.8            | 66             | 0.685                    | 0.530                         | 0.871 |
| Vanguard         | 2,073   | 10,961.4           | 71             | 0.648                    | 0.502                         | 0.812 |
| Attune           | 7,789   | 16,207.7           | 96             | 0.592                    | 0.480                         | 0.723 |
| Maxim            | 822     | 9,759.9            | 55             | 0.564                    | 0.425                         | 0.734 |
| MBK              | 256     | 3,407.6            | 18             | 0.528                    | 0.313                         | 0.835 |
| Nexgen           | 19,728  | 155,715.1          | 813            | 0.522                    | 0.487                         | 0.559 |
| Sigma            | 1,600   | 7,104.8            | 37             | 0.521                    | 0.367                         | 0.718 |
| LCS              | 14,769  | 139,891.7          | 670            | 0.479                    | 0.443                         | 0.517 |
| Genesis II       | 13,823  | 100,420.7          | 468            | 0.466                    | 0.425                         | 0.510 |
| Triathlon        | 23,650  | 116,467.5          | 501            | 0.430                    | 0.393                         | 0.469 |
| AGC              | 376     | 4,493.9            | 18             | 0.401                    | 0.237                         | 0.633 |
| PFC Sigma        | 10,789  | 91,493.9           | 356            | 0.389                    | 0.349                         | 0.431 |
| Advance          | 157     | 1,804.4            | 6              | 0.333                    | 0.122                         | 0.724 |
| Duracon          | 4,213   | 49,295.6           | 158            | 0.321                    | 0.272                         | 0.373 |
| AMK              | 95      | 1,277.6            | 2              | 0.157                    | 0.019                         | 0.565 |
| Sigma CR150      | 997     | 6,197.2            | 27             | 0.000                    | 0.287                         | 0.634 |

Those marked with an \* in the above table have revision rates significantly higher than the overall rate of 0.48/100 component years @ the 95% confidence interval. There are several other combinations with high revision rates, but without statistical significance because of the wide CIs.

Those marked with a # as well as an \* indicate those combinations used during 2018. The Persona and Balansys were both on the top 10 list for 2018.

It is to be noted that several variants of basically the same knee prosthesis type, e.g. Nexgen, LCS which are registered separately have been merged into the one group to enable comparable statistical analyses with other prostheses which may also have more than one variant, but are registered as one or two prostheses.



## Revision vs Arthroplasty Fixation for Fully Cemented Prostheses Sorted by Revision Rate

| Prosthesis      | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |       |
|-----------------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| Insall/Burstein | 249     | 2,988.9            | 47             | 1.572                    | 1.141                         | 2.072 |
| Saiph           | 93      | 153.7              | 2              | 1.301                    | 0.158                         | 4.701 |
| Optetrak        | 281     | 2,511.5            | 31             | 1.234                    | 0.822                         | 1.729 |
| Journey         | 330     | 1,526.3            | 17             | 1.114                    | 0.649                         | 1.783 |
| Trekking        | 792     | 2,781.0            | 27             | 0.971                    | 0.625                         | 1.391 |
| Legion          | 239     | 905.2              | 7              | 0.773                    | 0.311                         | 1.593 |
| Persona         | 3,854   | 6,842.5            | 49             | 0.716                    | 0.530                         | 0.947 |
| Balansys        | 1,555   | 6,472.2            | 45             | 0.695                    | 0.507                         | 0.930 |
| Scorpio         | 852     | 9,635.8            | 66             | 0.685                    | 0.530                         | 0.871 |
| Vanguard        | 2,054   | 10,857.3           | 69             | 0.636                    | 0.490                         | 0.799 |
| Attune          | 7,582   | 16,004.6           | 93             | 0.581                    | 0.466                         | 0.708 |
| Maxim           | 822     | 9,759.9            | 55             | 0.564                    | 0.425                         | 0.734 |
| MBK             | 247     | 3,299.1            | 18             | 0.546                    | 0.323                         | 0.862 |
| Nexgen          | 18,845  | 148,555.4          | 785            | 0.528                    | 0.492                         | 0.567 |
| Genesis II      | 13,770  | 99,816.7           | 463            | 0.464                    | 0.423                         | 0.508 |
| Sigma           | 1,365   | 6,286.5            | 29             | 0.461                    | 0.309                         | 0.663 |
| Sigma CR150     | 997     | 6,197.2            | 27             | 0.436                    | 0.287                         | 0.634 |
| Triathlon       | 23,011  | 114,501.4          | 487            | 0.425                    | 0.388                         | 0.464 |
| AGC             | 376     | 4,493.9            | 18             | 0.401                    | 0.237                         | 0.633 |
| LCS             | 9,566   | 95,641.4           | 378            | 0.395                    | 0.356                         | 0.437 |
| PFC Sigma       | 9,935   | 85,915.6           | 321            | 0.374                    | 0.333                         | 0.416 |
| Duracon         | 3,432   | 39,635.9           | 132            | 0.333                    | 0.279                         | 0.395 |
| Advance         | 157     | 1,804.4            | 6              | 0.333                    | 0.122                         | 0.724 |
| AMK             | 95      | 1,277.6            | 2              | 0.157                    | 0.019                         | 0.565 |

The Insall/Burstein, Trekking, Journey, Scorpio, Persona, Optetrak and Balansys have significantly higher revision rates than the overall rate of 0.48/100 component years at the 95% confidence interval. Balansys, Trekking and Persona prostheses were implanted in 2018.

## Revision vs Arthroplasty for Hybrid Fixation of Prostheses Sorted by Revision Rate

(Minimum of 50 primary registered arthroplasties)

| Femur Prosthesis | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |       |
|------------------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| Sigma            | 235     | 818.4              | 8              | 0.978                    | 0.422                         | 1.926 |
| Optetrak         | 380     | 3,286.8            | 26             | 0.791                    | 0.504                         | 1.141 |
| Triathlon        | 200     | 1,489.0            | 11             | 0.739                    | 0.369                         | 1.322 |
| Genesis II       | 51      | 594.7              | 4              | 0.673                    | 0.142                         | 1.599 |
| PFC Sigma        | 847     | 5,509.3            | 35             | 0.635                    | 0.443                         | 0.884 |
| LCS              | 2,235   | 20,265.3           | 95             | 0.469                    | 0.379                         | 0.573 |
| Nexgen           | 760     | 6,225.3            | 24             | 0.386                    | 0.247                         | 0.574 |
| Duracon          | 321     | 4,372.3            | 14             | 0.320                    | 0.167                         | 0.523 |

The Optetrak is the only hybrid fixation prosthesis with significantly higher revision rates than the overall rate of 0.48/100 component years at the 95% confidence interval.

## Revision vs Arthroplasty Fixation for Fully Uncemented Prostheses Sorted by Revision Rate

(Minimum of 50 primary registered arthroplasties)


| Femur Prosthesis | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |       |
|------------------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| Attune           | 199     | 193.8              | 3              | 1.548                    | 0.319                         | 4.524 |
| LCS              | 2,968   | 23,985.1           | 197            | 0.821                    | 0.711                         | 0.944 |
| Triathlon        | 439     | 477.2              | 3              | 0.629                    | 0.087                         | 1.678 |
| Nexgen           | 123     | 934.5              | 4              | 0.428                    | 0.117                         | 1.096 |
| Duracon          | 460     | 5,287.4            | 12             | 0.227                    | 0.111                         | 0.384 |

The uncemented LCS still implanted in 2018 and has a significantly higher revision rate than the overall rate of 0.48/100 component years at the 95% confidence interval.

## Revision Rates for Fixed vs Mobile Bearing Knees

| Femur Prosthesis | Mobile/Fixed | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |       |
|------------------|--------------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| AGC              | Fixed        | 376     | 4,493.9            | 18             | 0.401                    | 0.237                         | 0.633 |
| AMK              | Fixed        | 95      | 1,277.6            | 2              | 0.157                    | 0.019                         | 0.565 |
| Balansys         | Fixed        | 1,545   | 6,463.8            | 45             | 0.696                    | 0.508                         | 0.932 |
| Duracon          | Fixed        | 4,207   | 49,213.2           | 157            | 0.319                    | 0.270                         | 0.372 |
| Genesis II       | Fixed        | 13,811  | 100,399.5          | 468            | 0.466                    | 0.425                         | 0.510 |
| Insall/Burstein  | Fixed        | 249     | 2,988.9            | 47             | 1.572                    | 1.141                         | 2.072 |
| Journey          | Fixed        | 276     | 1,479.7            | 16             | 1.081                    | 0.618                         | 1.756 |
| LCS              | Mobile       | 14,767  | 139,888.4          | 670            | 0.479                    | 0.443                         | 0.517 |
| Maxim            | Fixed        | 822     | 9,759.9            | 55             | 0.564                    | 0.425                         | 0.734 |
| MBK              | Mobile       | 256     | 3,407.6            | 18             | 0.528                    | 0.313                         | 0.835 |
| Trekking         | Mobile       | 782     | 2,764.7            | 26             | 0.940                    | 0.614                         | 1.378 |
| Persona          | Fixed        | 3,844   | 6,832.8            | 49             | 0.717                    | 0.531                         | 0.948 |
| Nexgen           | Fixed        | 16,753  | 131,390.4          | 702            | 0.534                    | 0.495                         | 0.575 |
|                  | Mobile       | 2,715   | 22,702.6           | 94             | 0.414                    | 0.335                         | 0.507 |
| PFC Sigma        | Fixed        | 7,315   | 5,8632.2           | 233            | 0.397                    | 0.348                         | 0.452 |
|                  | Mobile       | 3,442   | 3,2603.6           | 122            | 0.374                    | 0.311                         | 0.447 |
| Scorpio          | Fixed        | 737     | 8346.3             | 57             | 0.683                    | 0.512                         | 0.878 |
|                  | Mobile       | 104     | 1216.4             | 6              | 0.493                    | 0.181                         | 1.074 |
| Sigma            | Fixed        | 575     | 2,531.7            | 10             | 0.395                    | 0.176                         | 0.701 |
|                  | Mobile       | 1,009   | 4,478.1            | 27             | 0.603                    | 0.397                         | 0.877 |
| Sigma CR150      | Fixed        | 188     | 1,218.6            | 10             | 0.821                    | 0.365                         | 1.456 |
|                  | Mobile       | 808     | 4,974.4            | 17             | 0.342                    | 0.199                         | 0.547 |
| Triathlon        | Fixed        | 18,862  | 109,867.2          | 468            | 0.426                    | 0.388                         | 0.466 |
|                  | Mobile       | 659     | 3,200.4            | 13             | 0.406                    | 0.216                         | 0.695 |
| Attune           | Fixed        | 3,300   | 7,644.8            | 43             | 0.562                    | 0.402                         | 0.750 |
|                  | Mobile       | 3,992   | 8,385.7            | 51             | 0.608                    | 0.453                         | 0.800 |

In prostheses with both fixed and mobile variants there are no differences in revision rates between the two designs at the 95% confidence interval.



## Overall Revision Rates for Fixed vs Mobile Bearing Knees

| Fixed/Mobile | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|--------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Fixed        | 73,050  | 502,907            | 2,385          | 0.47                     | 0.46                          | 0.49 |
| Mobile       | 28,563  | 223,752            | 1,044          | 0.47                     | 0.44                          | 0.50 |

There is no significant difference between the two groups. It was not possible to determine fixed or mobile categories for all registered knees, which accounts for the 5,813 versus the total number of TKA's.

## Revision Rates for Cruciate Retaining (CR) vs. Posterior Stabilised (PS)

| Femur Prosthesis | CR/PS | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |       |
|------------------|-------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| AGC              | PS    | 28      | 372.8              | 4              | 1.073                    | 0.292                         | 2.747 |
| Insall/Burstein  | PS    | 249     | 2,988.9            | 47             | 1.572                    | 1.141                         | 2.072 |
| LCS              | PS    | 70      | 468.9              | 2              | 0.426                    | 0.052                         | 1.541 |
| Legion           | PS    | 193     | 747.9              | 5              | 0.669                    | 0.217                         | 1.560 |
| Sigma CR150      | CR    | 997     | 6,197.2            | 27             | 0.436                    | 0.287                         | 0.634 |
| Attune           | CR    | 4,872   | 10,684.6           | 66             | 0.618                    | 0.478                         | 0.786 |
|                  | PS    | 2,903   | 5,513.5            | 30             | 0.544                    | 0.360                         | 0.766 |
| Balansys         | CR    | 1,431   | 6,036.6            | 38             | 0.629                    | 0.439                         | 0.854 |
|                  | PS    | 113     | 422.2              | 7              | 1.658                    | 0.593                         | 3.255 |
| Genesis II       | CR    | 7,357   | 58,585.7           | 197            | 0.336                    | 0.291                         | 0.387 |
|                  | PS    | 6,459   | 41,789.6           | 271            | 0.648                    | 0.572                         | 0.729 |
| Maxim            | CR    | 657     | 7,747.0            | 38             | 0.491                    | 0.347                         | 0.673 |
|                  | PS    | 165     | 2,012.9            | 17             | 0.845                    | 0.473                         | 1.322 |
| Nexgen           | CR    | 9,377   | 72,869.9           | 298            | 0.409                    | 0.363                         | 0.457 |
|                  | PS    | 10,009  | 81,279.2           | 490            | 0.603                    | 0.550                         | 0.658 |
| Optetrak         | CR    | 437     | 3,854.8            | 30             | 0.778                    | 0.525                         | 1.111 |
|                  | PS    | 224     | 1,943.5            | 27             | 1.389                    | 0.895                         | 1.991 |
| Persona          | CR    | 2,970   | 4,790.1            | 30             | 0.626                    | 0.414                         | 0.882 |
|                  | PS    | 874     | 2,042.7            | 19             | 0.930                    | 0.560                         | 1.453 |
| PFC Sigma        | CR    | 8,840   | 72,519.6           | 263            | 0.363                    | 0.320                         | 0.409 |
|                  | PS    | 1,888   | 18,490.5           | 93             | 0.503                    | 0.404                         | 0.613 |
| Scorpio          | CR    | 739     | 8,530.1            | 56             | 0.656                    | 0.496                         | 0.853 |
|                  | PS    | 111     | 1,089.6            | 10             | 0.918                    | 0.440                         | 1.688 |
| Sigma            | CR    | 278     | 1,053.5            | 0              | 0.000                    | 0.000                         | 0.350 |
|                  | PS    | 1,321   | 6,046.1            | 37             | 0.612                    | 0.431                         | 0.844 |
| Trekking         | CR    | 295     | 1,076.8            | 10             | 0.929                    | 0.445                         | 1.708 |
|                  | PS    | 487     | 1,687.9            | 16             | 0.948                    | 0.542                         | 1.539 |
| Triathlon        | CR    | 20,574  | 96,901.8           | 414            | 0.427                    | 0.387                         | 0.470 |
|                  | PS    | 3071    | 19,556.2           | 87             | 0.445                    | 0.356                         | 0.549 |
| Vanguard         | CR    | 1,486   | 7,924.2            | 46             | 0.581                    | 0.425                         | 0.774 |
|                  | PS    | 570     | 2,996.1            | 25             | 0.834                    | 0.540                         | 1.232 |

## Overall Revision Rates for Cruciate Retaining vs. Posterior Stabilised vs Minimally Stabilised Knees

| Prosthesis | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| CR         | 60,310  | 358,772.0          | 1,513          | 0.42                     | 0.40                          | 0.44 |
| Other      | 15,010  | 143,448.8          | 692            | 0.48                     | 0.45                          | 0.52 |
| PS         | 28,740  | 189,483.7          | 1,188          | 0.63                     | 0.59                          | 0.66 |

The LCS prostheses account for the majority of the minimally stabilised (MS). There is a significantly higher revision rate for the posterior and minimally stabilised compared to cruciate retaining knee prostheses.

## Revision vs. Arthroplasty Fixation

| Fixation   | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Cemented   | 100,752 | 679,552.5          | 3,204          | 0.47                     | 0.46                          | 0.49 |
| Uncemented | 4,202   | 30,981.7           | 220            | 0.71                     | 0.62                          | 0.81 |
| Hybrid     | 5,122   | 43,188.8           | 228            | 0.53                     | 0.46                          | 0.60 |

Uncemented knees have a significantly higher revision rate than either cemented or hybrid knees. Further analyses have shown that it is loosening of the uncemented tibial component that is responsible for the higher revision rate.

## Revision vs Age Bands

| Age Bands | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|-----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <55       | 9,108   | 68,055.9           | 613            | 0.90                     | 0.83                          | 0.97 |
| 55-64     | 30,725  | 219,557.2          | 1,346          | 0.61                     | 0.58                          | 0.65 |
| 65-74     | 42,298  | 291,931.2          | 1,246          | 0.43                     | 0.40                          | 0.45 |
| >=75      | 27,945  | 174,178.7          | 447            | 0.26                     | 0.23                          | 0.28 |

Each successive age band in ascending order has a significantly lower revision rate.

## Revision vs Gender


| Gender | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|--------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| F      | 56,707  | 395,714.5          | 1,756          | 0.44                     | 0.42                          | 0.47 |
| M      | 53,369  | 358,008.6          | 1,896          | 0.53                     | 0.51                          | 0.55 |

The revision rate for males is significantly higher than for females.

## Revision by Age Bands vs Arthroplasty Fixation

| Cemented | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <55      | 7,813   | 56,520.4           | 482            | 0.85                     | 0.78                          | 0.93 |
| 55-64    | 27,614  | 193,494.7          | 1,170          | 0.60                     | 0.57                          | 0.64 |
| 65-74    | 39,128  | 267,329.9          | 1,146          | 0.43                     | 0.40                          | 0.45 |
| >=75     | 26,197  | 162,207.5          | 406            | 0.25                     | 0.23                          | 0.28 |





### Revision by Age Bands vs Arthroplasty Fixation

| Uncemented | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <55        | 676     | 6,172.7            | 80             | 1.30                     | 1.03                          | 1.61 |
| 55-64      | 1,513   | 11,632.9           | 86             | 0.74                     | 0.59                          | 0.91 |
| 65-74      | 1,352   | 9,226.4            | 44             | 0.48                     | 0.35                          | 0.64 |
| >=75       | 661     | 3,949.6            | 10             | 0.25                     | 0.11                          | 0.45 |

| Hybrid | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|--------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <55    | 619     | 5,362.7            | 51             | 0.95                     | 0.71                          | 1.25 |
| 55-64  | 1,598   | 14,429.7           | 90             | 0.62                     | 0.50                          | 0.77 |
| 65-74  | 1,818   | 15,374.9           | 56             | 0.36                     | 0.27                          | 0.47 |
| >=75   | 1,087   | 8,021.6            | 31             | 0.39                     | 0.26                          | 0.55 |

### Revision vs Approach

| Approach | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Medial   | 99,008  | 677,499.6          | 3,243          | 0.48                     | 0.46                          | 0.50 |
| Lateral  | 1,379   | 11,055.4           | 77             | 0.70                     | 0.55                          | 0.87 |
| Other    | 2,387   | 18,199.5           | 72             | 0.40                     | 0.31                          | 0.50 |

The lateral approach has a significantly higher revision rate than the other two approaches.

### Revision vs. Image Guidance

| Image Guided | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|--------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| No           | 95,508  | 682,687.4          | 3,311          | 0.48                     | 0.47                          | 0.50 |
| Yes          | 14,568  | 71,035.7           | 341            | 0.48                     | 0.43                          | 0.53 |

There is no significant difference between the two groups.

### Revision vs Surgeon Annual Output

| Operations per year | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|---------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <10                 | 2,127   | 16,561.4           | 78             | 0.47                     | 0.37                          | 0.59 |
| 10-24               | 21,807  | 162,061.0          | 849            | 0.52                     | 0.49                          | 0.56 |
| 25-49               | 50,024  | 345,309.7          | 1,642          | 0.48                     | 0.45                          | 0.50 |
| 50-74               | 22,677  | 149,920.5          | 703            | 0.47                     | 0.43                          | 0.50 |
| 75-99               | 5,894   | 32,314.5           | 133            | 0.41                     | 0.34                          | 0.49 |
| >=100               | 7,547   | 47,556.0           | 247            | 0.52                     | 0.46                          | 0.59 |

There is no significant difference between the groups.



### Revision vs ASA Status

| ASA Class | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|-----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| 1         | 9,653   | 57,770.9           | 307            | 0.53                     | 0.47                          | 0.59 |
| 2         | 55,796  | 314,976.0          | 1,520          | 0.48                     | 0.46                          | 0.51 |
| 3         | 21,940  | 111,870.1          | 626            | 0.56                     | 0.52                          | 0.60 |
| 4         | 357     | 1,580.1            | 9              | 0.57                     | 0.26                          | 1.08 |

### Revision vs. BMI


(BMI has been collected by the NZJR since 2010)

| BMI     | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|---------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| < 19    | 89      | 331.6              | 0              | 0.00                     | 0.00                          | 1.11 |
| 19 - 24 | 4,740   | 16,766.4           | 111            | 0.66                     | 0.54                          | 0.80 |
| 25 - 29 | 14,021  | 50,420.8           | 286            | 0.57                     | 0.50                          | 0.64 |
| 30 - 39 | 20,027  | 70,423.0           | 413            | 0.59                     | 0.53                          | 0.65 |
| 40+     | 4,040   | 14,133.5           | 117            | 0.83                     | 0.68                          | 0.99 |


40+ group has a significantly higher revision rate than the two groups before it.

### Revision for Deep Infection within six months versus Theatre Environment


| Theatre Environment | Total number | Number revised | %       | Standard error |
|---------------------|--------------|----------------|---------|----------------|
| Conventional        | 57,988       | 101            | 0.17417 | 0.01732        |
| Laminar flow        | 46,453       | 154            | 0.33152 | 0.02667        |



As with hip arthroplasty, there is a significant difference in knee revision rates (2x) for deep infection within six months of surgery between conventional and laminar flow theatres.




| Theatre Environment  | Suit/No Suit | Total number | Number revised | %       | Standard error |
|----------------------|--------------|--------------|----------------|---------|----------------|
| Conventional-Suit    | Suit         | 9,665        | 27             | 0.27936 | 0.05369        |
| Conventional-No suit | no suit      | 48,323       | 74             | 0.15314 | 0.01779        |
| Laminar flow-Suit    | Suit         | 25,212       | 73             | 0.28954 | 0.03384        |
| Laminar flow-No suit | no suit      | 21,242       | 81             | 0.38132 | 0.04229        |




There is a significant difference in the revision rates between conventional/no suit and the conventional/suit and laminar/suit environments. See Tayton et al BJJ. 2016 98-B (3), 334-340 for a more detailed analysis of infection data.

|         | Total number | Number revised | %       | Standard error |
|---------|--------------|----------------|---------|----------------|
| Suit    | 34,877       | 100            | 0.28672 | 0.02863        |
| No suit | 69,565       | 155            | 0.22281 | 0.01788        |



## Comparison of Major vs Minor Revisions by Year


A major revision is defined as revision of tibial and/or femoral components, including any of minor components and minor revision as change of bearing and/or patellar components only.




## Re-revisions for major vs minor knee revisions

| Major/Minor | No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|-------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Minor       | 1,106   | 4,649              | 195            | 4.19                     | 3.63                          | 4.83 |
| Major       | 1,943   | 9,886              | 283            | 2.86                     | 2.54                          | 3.22 |

There is a significantly higher re-revision rate for minor compared to major revisions.



## Percentage of knees revised in the first year



## Patello-Femoral Arthroplasty

| No. Ops | Observed comp. yrs | Number revised | Rate/100 component years | Exact 95% confidence interval |      |
|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| 602     | 3,045              | 58             | 1.91                     | 1.45                          | 2.46 |

*The revision rate is nearly four times that for total knee arthroplasty.*

### Revised to:


|                  |    |
|------------------|----|
| Total            | 52 |
| Patello- Femoral | 3  |
| Uniknee          | 3  |



## KAPLAN MEIER CURVES


The following Kaplan Meier survival analyses are for the 20 years 1999 – 2018 with deceased patients censored at time of death.


All Knees




| Years | % Revision-free | No. in each year |
|-------|-----------------|------------------|
| 1     | 99.3            | 100,124          |
| 2     | 98.5            | 90,213           |
| 3     | 98.0            | 80,947           |
| 4     | 97.6            | 72,308           |
| 5     | 97.2            | 63,722           |
| 6     | 96.9            | 55,842           |
| 7     | 96.6            | 48,481           |
| 8     | 96.3            | 41,474           |
| 9     | 96.0            | 34,867           |
| 10    | 95.7            | 28,606           |
| 11    | 95.3            | 23,097           |
| 12    | 94.9            | 17,978           |
| 13    | 94.4            | 13,681           |
| 14    | 94.1            | 9,979            |
| 15    | 93.7            | 7,114            |
| 16    | 93.2            | 5,044            |
| 17    | 92.9            | 3,435            |
| 18    | 92.5            | 2,008            |
| 19    | 92.1            | 795              |

Cemented vs Uncemented vs Hybrid






### Fixed vs. Mobile knees




### Posterior Stabilised vs. Cruciate Retaining







Survival for age bands




Survival for male vs. female








### Survival for surgeon annual output




### Survival for BMI groups







### Tibial loosening




### Femoral loosening






## Deep infection






## Pain





## Patella





## KNEE RE-REVISIONS

Analysis was undertaken of re-revisions. There were 560 registered primary knee revisions that had been revised twice, 121 that had been revised three times, 31 that had been revised four times, 9 that had been revised five times and 3 that had been revised six times.

### Second revision

Time between the first and second revision for the 560 knee arthroplasties averaged 876 days, with a range of 1 – 5,398 and a standard deviation of 1,015 days. This compares to an average of 1,514 days between primary and first revision knee arthroplasty.

### Reason for revision

|                              |     |
|------------------------------|-----|
| Deep infection               | 279 |
| Pain                         | 114 |
| Loosening tibial component   | 79  |
| Loosening femoral component  | 66  |
| Loosening patellar component | 9   |
| Fracture femur               | 4   |
| Fracture tibia               | 1   |

### Second Revisions

| Number of primary revisions | Observed comp. yrs | Number of second re-revisions | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------------------------|--------------------|-------------------------------|--------------------------|-------------------------------|------|
| 3,307                       | 15,433             | 497                           | 3.22                     | 2.94                          | 3.52 |

### Third revision

The average time between second and third revisions for the 121 knee arthroplasties was 622 days, with a range of 5 – 5,185 and a standard deviation of 683 days.

### Fourth revision

The average time between third and fourth revisions for the 31 knee arthroplasties was 598 days, with a range of 10 – 3,136 and a standard deviation of 795 days.


### Fifth revision

The average time between fourth and fifth revisions for the 9 knee arthroplasties was 938 days.

### Sixth revision

The average time between the fifth and sixth revisions for the 3 knee arthroplasties was 544 days.

## KAPLAN MEIER SURVIVAL CURVE FOR FIRST REVISION KNEE ARTHROPLASTIES



| Years | Percentage re-revision free | No. in year |
|-------|-----------------------------|-------------|
| 1     | 93.08                       | 3,014       |
| 2     | 89.69                       | 2,552       |
| 3     | 87.47                       | 2,143       |
| 4     | 85.56                       | 1,776       |
| 5     | 84.00                       | 1,472       |
| 6     | 82.67                       | 1,203       |
| 7     | 81.94                       | 1,002       |
| 8     | 79.86                       | 810         |
| 9     | 79.20                       | 640         |
| 10    | 78.05                       | 487         |
| 11    | 77.32                       | 352         |
| 12    | 75.75                       | 247         |
| 13    | 74.36                       | 182         |
| 14    | 73.87                       | 130         |

## PATIENT BASED QUESTIONNAIRE OUTCOMES AT SIX MONTHS, FIVE YEARS, TEN YEARS, FIFTEEN YEARS AND TWENTY YEARS POST-SURGERY

### Questionnaires at six months post-surgery

At six months post-surgery a random selection of patients are sent the Oxford-12 questionnaire in order to achieve a response rate of 20% of the total which is deemed to be ample to provide powerful statistical analysis.

The scores now range from 4 to 0. A score of 48 is the best, indicating normal function. A score of 0 is the worst, indicating the most severe disability.

In addition we have grouped the questionnaire responses according to the classification system published by Kalairajah et al in 2005. (See appendix 1).

This groups each score into four categories:

|            |         |           |
|------------|---------|-----------|
| Category 1 | >41     | excellent |
| Category 2 | 34 – 41 | good      |
| Category 3 | 27 – 33 | fair      |
| Category 4 | < 27    | poor      |

For the twenty year period and as at July 2019, there were 29,816 primary knee questionnaire responses registered at six months post-surgery.

The average knee score was 37.65 (standard deviation 8.03, range 48 – 0).

|         |         |        |
|---------|---------|--------|
| Scoring | > 41    | 11,671 |
| Scoring | 34 – 41 | 10,532 |
| Scoring | 27 – 33 | 4,468  |
| Scoring | < 27    | 3,663  |

At six months post-surgery, 75% had an excellent or good score.

### Questionnaires at five years post-surgery

All patients who had a six month registered questionnaire, and who had not had revision surgery were sent a further questionnaire at five years post-surgery.

This dataset represents sequential Oxford knee scores for 11,768 individual patients.

At five years post-surgery, 84% of patients achieved an excellent or good score and had an average of 40.50.

### Questionnaires at ten years post-surgery

All patients who had a six month registered questionnaire, and who had not had revision surgery were sent a further questionnaire at ten years post-surgery.

This dataset represents sequential Oxford knee scores for 6,572 individual patients.

At ten years post-surgery, 82% of patients achieved an excellent or good score and had an average of 39.96.

### Questionnaires at fifteen years post-surgery

All patients who had a six month registered questionnaire, and who had not had revision surgery were sent a further questionnaire at fifteen years post-surgery.

This dataset represents sequential Oxford knee scores for 2,364 individual patients.

At fifteen years post-surgery, 79% of patients achieved an excellent or good score and had an average of 39.39.

### Questionnaires at twenty years post-surgery

All patients who had a six month registered questionnaire, and who had not had revision surgery were sent a further questionnaire at fifteen years post-surgery.

This dataset represents sequential Oxford knee scores for 128 individual patients.


At twenty years post-surgery, 75% of patients achieved an excellent or good score and had an average of 39.04.

### BMI vs Oxford score at six months

| BMI          | Mean         | Standard Error of Mean | Number       |
|--------------|--------------|------------------------|--------------|
| < 19         | 39.67        | 2.081                  | 15           |
| 19 - 24      | 39.81        | 0.212                  | 1,087        |
| 25 - 29      | 39.25        | 0.128                  | 3,109        |
| 30 - 39      | 37.85        | 0.127                  | 3,821        |
| 40+          | 36.11        | 0.312                  | 631          |
| <b>Total</b> | <b>38.47</b> | <b>0.081</b>           | <b>8,663</b> |

### Revision knee questionnaire responses

There were 4,714 revision knee responses with 54% achieving an excellent or good score. This group includes all revision knee procedures. The average revision knee score was 32.92 (standard deviation 10.18, range 2 – 48).




**OXFORD 12 SCORE AS A PREDICTOR OF KNEE ARTHROPLASTY REVISION**

A statistically significant relationship has been confirmed between the Oxford scores at six months, five and ten years' post-surgery and arthroplasty revision within two years of the Oxford 12 questionnaire date.

**Six month score and revision arthroplasty**

Plotting the patients' six month scores in the Kalairajah groupings against the proportion of knees revised for that same group demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score. A patient with a score below 27 has 12 times the risk of a revision within two years compared to a person with a score >41.




Revision risk versus Kalairajah groupings of Oxford scores within two years of the six month score date

| Score group | Revision to 2 years | Number revised | %    | Standard error |
|-------------|---------------------|----------------|------|----------------|
| < 27        | 2,804               | 154            | 5.49 | 0.43           |
| 27_33       | 3,949               | 59             | 1.49 | 0.19           |
| 34_41       | 9,283               | 68             | 0.73 | 0.09           |
| 42+         | 10,163              | 45             | 0.44 | 0.07           |

A person with an Oxford score > 42 has a 0.44 risk of revision within two years compared to a 5.49% risk with a score of 27 or less.

In view of the large number of six month Oxford scores it is possible with statistical significance to further break down the score groupings to demonstrate an even more convincing relationship between score and risk of revision within two years.




Revision risk versus groupings of Oxford scores within two years of the 6 month score date



Five year score and revision arthroplasty

As with the six month scores, plotting the patients' five year scores in the Kalairajah groupings against the proportion of knees revised for that same group demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score. A patient with a score below 27 has 13 times the risk of a revision within two years compared to a person with a score > 41.



Revision risk versus Kalairajah groupings of Oxford scores within two years of the five year score date.

| Score group | Revision to 2 years | Number revised | %    | Standard error |
|-------------|---------------------|----------------|------|----------------|
| < 27        | 672                 | 24             | 3.57 | 0.72           |
| 27_33       | 904                 | 12             | 1.33 | 0.38           |
| 34_41       | 2,438               | 12             | 0.49 | 0.14           |
| 42+         | 5,806               | 16             | 0.28 | 0.07           |


Ten year score and revision arthroplasty

As with the six month and five year scores, plotting the patients' ten year scores in the Kalairajah groupings against the proportion of knees revised for that same group demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score. A patient with a score below 27 has 10 times the risk of a revision within two years compared to a person with a score >41.



Revision risk versus Kalairajah groupings of Oxford scores within two years of the 10 year score date.





| Score group | Revision to 2 years | Number revised | %    | Standard error |
|-------------|---------------------|----------------|------|----------------|
| < 27        | 368                 | 18             | 4.89 | 1.12           |
| 27_33       | 473                 | 11             | 2.33 | 0.69           |
| 34_41       | 1,208               | 6              | 0.50 | 0.20           |
| 42+         | 2,648               | 13             | 0.49 | 0.14           |


A person with an Oxford score of > 41 has a 0.49% risk of revision within two years compared to a 4.89% risk with a score of 27 or less.

### Fifteen year score and revision arthroplasty

As with the six month, five year and ten year scores, plotting the patients' fifteen year scores in the Kalairajah groupings against the proportion of knees revised for that same group demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score. A patient with a score below 27 has 10 times the risk of a revision within two years compared to a person with a score >41.

| Score group | Revision to 2 years | Number revised | %    | Standard error |
|-------------|---------------------|----------------|------|----------------|
| < 27        | 139                 | 12             | 8.63 | 2.38           |
| 27_33       | 135                 | 0              | 0.00 | 0.00           |
| 34_41       | 326                 | 1              | 0.31 | 0.31           |
| 42+         | 715                 | 6              | 0.84 | 0.34           |


Revision (%) to 2 years - by Oxford score at 15 Years



### Prediction of second revision from six month score following first revision

Plotting the patients' six month scores following their first revision in the Kalairajah groupings against the proportion of knees revised for that same group again demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score. A patient with a score below 27 has 5 times the risk of a revision within two years compared to a person with a score >41.


Revision (%) to 2 years - by Oxford score at Revision



Second revision risk versus Kalairajah groupings of Oxford scores within two years of the six month post- first revision score date.


| Score group | Revision to 2 years | Number revised | %    | Standard error |
|-------------|---------------------|----------------|------|----------------|
| < 27        | 929                 | 83             | 8.93 | 0.94           |
| 27_33       | 698                 | 23             | 3.30 | 0.68           |
| 34_41       | 1,080               | 31             | 2.87 | 0.51           |
| 42+         | 921                 | 17             | 1.85 | 0.44           |

**Mean Oxford scores at six months and five years for six knee prostheses with minimum of 1,800 registrations**



**Oxford scores for 6 most common knee prostheses with 6m and 5 years Oxford scores**

| Oxford Score |                    | Duracon | Genesis II | LCS   | Nexgen | PFC Sigma | Triathlon |
|--------------|--------------------|---------|------------|-------|--------|-----------|-----------|
| 6 Month      | Mean               | 36.9    | 37.4       | 36.4  | 37.9   | 38.1      | 38.5      |
|              | Std. Error of Mean | 0.2     | 0.1        | 0.1   | 0.1    | 0.1       | 0.1       |
|              | Number             | 1,799   | 3,436      | 5,677 | 5,027  | 2,937     | 4,334     |
| 5 Year       | Mean               | 40.0    | 40.5       | 39.5  | 40.6   | 41.0      | 41.8      |
|              | Std. Error of Mean | 0.3     | 0.2        | 0.2   | 0.2    | 0.2       | 0.2       |
|              | Number             | 780     | 1,642      | 2,516 | 2,369  | 1,570     | 1,689     |



# UNICOMPARTMENTAL KNEE ARTHROPLASTY

## PRIMARY UNICOMPARTMENTAL KNEE ARTHROPLASTY

The **nineteen year** report analyses data for the period January 2000 – December 2018. There were 12,627 unicompartmental knee procedures registered with an additional 1,096 for 2018.

For the 2018 year the Oxford uncemented medial UKR remains the most commonly used prosthesis with 753 (69%), followed by the Zimmer UK 150 (14%). Smaller numbers of Persona Partial – 62 (5%), the Oxford lateral dome UKR at 59 (5%) and Restoris 45 (4%) are also being implanted.

## Data Analysis

### Age and sex distribution

The average age for a unicompartmental knee replacement was 66 years, with a range of 18 – 95 years.

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 5,773  | 6,854 |
| Percentage    | 45.72  | 54.28 |
| Mean age      | 65.90  | 66.28 |
| Maximum age   | 94.71  | 94.55 |
| Minimum age   | 18.28  | 31.62 |
| Standard dev. | 10.16  | 9.21  |

### Body Mass Index

For the nine- year period 2010 - 2018, there were 5,719 BMI registrations for unicompartmental knee replacements. The average was 29.90 with a range of 16.60 – 59.50 and a standard deviation of 5.00.

### Previous operation

|                         |        |
|-------------------------|--------|
| None                    | 10,207 |
| Meniscectomy            | 1,871  |
| Ligament reconstruction | 76     |
| Osteotomy               | 48     |
| Internal fixation       | 34     |
| Synovectomy             | 5      |

### Diagnosis

|                          |        |
|--------------------------|--------|
| Osteoarthritis           | 12,368 |
| Avascular necrosis       | 99     |
| Post ligament disruption | 61     |
| Rheumatoid arthritis     | 26     |
| Post fracture            | 25     |
| Other inflammatory       | 22     |
| Tumour                   | 2      |

### Approach

|                            |       |
|----------------------------|-------|
| Medial                     | 9,445 |
| Minimally invasive surgery | 3,017 |
| Lateral                    | 270   |
| Other                      | 216   |
| Image guided surgery       | 168   |
| Robot assisted             | 41    |

Image guided surgery was added to the updated forms at the beginning of 2005, but unlike in total knee arthroplasty, it has never become popular. Robot assisted is reported for the first time in this report.

### Cement

|                      |           |
|----------------------|-----------|
| Femur cemented       | 7,960 63% |
| Antibiotic in cement | 5,208 65% |
| Tibia cemented       | 8,431 67% |
| Antibiotic in cement | 5,536 66% |

### Systemic antibiotic prophylaxis

|   |            |
|---|------------|
| Patient number receiving at least one systemic antibiotic | 12,190 97% |
|---|------------|

### Operating theatre

|              |       |
|--------------|-------|
| Conventional | 8,755 |
| Laminar flow | 3,736 |
| Space suits  | 2,932 |

### ASA Class

This was introduced with the updated forms at the beginning of 2005.

For the fourteen- year period 2005 – 2018, there were 9,706 (97%) unicompartmental knee procedures with the ASA class recorded.

### Definitions

|                     |   |
|---------------------|---|
| <b>ASA class 1:</b> | A healthy patient   |
| <b>ASA class 2:</b> | A patient with mild systemic disease  |
| <b>ASA class 3:</b> | A patient with severe systemic disease that limits activity but is not incapacitating |
| <b>ASA class 4:</b> | A patient with an incapacitating disease that is a constant threat to life            |

| ASA | Number | Percentage |
|-----|--------|------------|
| 1   | 1,843  | 19         |
| 2   | 6,239  | 64         |
| 3   | 1,603  | 16         |
| 4   | 21     | 1          |

### Operative time (skin to skin)

|      |            |
|------|------------|
| Mean | 73 minutes |
|------|------------|

### Surgeon grade

The updated forms introduced in 2005 have separated advanced trainee into supervised and unsupervised.

The following figures are for the fourteen year period 2005 – 2018.

|                               |       |
|-------------------------------|-------|
| Consultant                    | 9,535 |
| Advanced trainee supervised   | 440   |
| Advanced trainee unsupervised | 63    |
| Basic trainee                 | 16    |


### Prosthesis usage

#### Unicompartmental knee prostheses used in 2018

|                     |     |
|---------------------|-----|
| Oxford 3 uncemented | 753 |
| Zimmer Uni          | 150 |
| Persona Partial     | 62  |
| Oxford 3 cemented   | 59  |
| Restoris            | 45  |
| Sigma HP Uni        | 15  |
| Triathlon PKR       | 7   |
| Journey Uni         | 5   |



### Most used Unicompartmental prostheses for 5 years (2014 – 2018)



### Surgeon and hospital workload


#### Surgeons

In 2018, 82 surgeons performed 1,096 unicompartmental knee replacements, an average of 13 procedures per surgeon.

25 surgeons performed less than 5 procedures (41 surgeons in 2017), 39 surgeons did 5-15 procedures with 18 surgeons performing more than 15 cases (16-157).

#### Hospitals

In 2018, unicompartmental knee replacements were performed in 38 hospitals; 21 were public and 17 were private.



## REVISION OF REGISTERED PRIMARY UNICOMPARTMENTAL ARTHROPLASTIES

This section analyses the data for revision of unicompartmental knee replacement over the nineteen-year period.

Revision is defined by the Registry as a new operation in a previously partially replaced knee joint during which one or more of the components are exchanged, removed, manipulated or added. It includes arthrodesis or amputation, but not soft tissue procedures. A two or more staged procedure is registered as one revision.

There were 1038 revisions of the 12,627 registered unicompartmental knee replacements (8%). A further 108 had a second revision, 17 a third revision, 1 a fourth revision and 1 a fifth revision.

839 of the 1038 (81%) were revised to total knee replacements and 199 (19%) were revised to further unicompartmental replacements.

Of the implants that were in common use in 2018, 124 medial Oxford UKR were revised (0.70/100 ocys), 29 Zimmer UKR (0.47/100 ocys), 9 Triathlon PKR (0.88/100 ocys) and 23 lateral dome Oxford UKR (1.67/100 ocys).

The observed revision rate remains higher for the more implanted Oxford compared to the Zimmer UK.

### Time to revision

|                    |            |
|--------------------|------------|
| Mean               | 2,080 days |
| Maximum            | 6,703 days |
| Minimum            | 4 days     |
| Standard deviation | 1,699 days |

### Reason for revision

|                            |     |
|----------------------------|-----|
| Pain                       | 317 |
| Loosening tibial component | 181 |
| Loosening femoral          | 129 |
| Deep infection             | 38  |
| Fracture tibia             | 26  |
| Fracture femur             | 4   |

There is sometimes more than one reason listed for revision and all are registered.

### Analysis of the three main reasons for revision by year after the primary procedure

| Years | Loosening femoral component |      | Loosening tibial component |      | Pain  |      |
|-------|-----------------------------|------|----------------------------|------|-------|------|
|       | Count                       | %    | Count                      | %    | Count | %    |
| 0     | 12                          | 9.3  | 32                         | 17.7 | 44    | 13.9 |
| 1     | 22                          | 17.1 | 36                         | 19.9 | 66    | 20.8 |
| 2     | 9                           | 7.0  | 14                         | 7.7  | 35    | 11.0 |
| 3     | 16                          | 12.4 | 14                         | 7.7  | 17    | 5.4  |
| 4     | 5                           | 3.9  | 10                         | 5.5  | 30    | 9.5  |
| 5     | 10                          | 7.8  | 8                          | 4.4  | 16    | 5.0  |
| 6     | 5                           | 3.9  | 12                         | 6.6  | 19    | 6.0  |
| 7     | 11                          | 8.5  | 9                          | 5.0  | 16    | 5.0  |
| 8     | 7                           | 5.4  | 6                          | 3.3  | 13    | 4.1  |
| 9     | 4                           | 3.1  | 11                         | 6.1  | 13    | 4.1  |
| 10    | 8                           | 6.2  | 6                          | 3.3  | 13    | 4.1  |
| 11+   | 20                          | 15.5 | 23                         | 12.7 | 35    | 11.0 |
| Total | 129                         |      | 181                        |      | 317   |      |

### Statistical note

In the tables below there are two statistical terms readers may not be familiar with:

#### i) Observed component years

This is the number of registered primary procedures multiplied by the number of years each component has been in place.

#### ii) Rate/100 component years

This is equivalent to the yearly revision rate expressed as a percent and is derived by dividing the number of prostheses revised by the observed component years multiplied by 100. It therefore allows for the number of years of post-operative follow-up in calculating the revision rate. These rates are usually very low, hence are expressed per

100 component years rather than per component year. Statisticians consider that this is a more accurate way of deriving a revision rate for comparison when analysing data with widely varying follow-up times. It is also important to note the confidence intervals. The closer they are to the estimated revision rate/100 component years, the more precise the estimate is.

### Statistical significance

Where it is stated that a difference among results is significant the p value is 0.05 or less. In most of these situations this is because there is no overlap of the confidence intervals (CIs) but sometimes significance can apply in the presence of CI overlap.




### All Primary Unicompartmental Knee Arthroplasties

| No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| 12,627  | 86,980             | 1,038          | 1.20                     | 1.12                          | 1.27 |

### Revision Rate of Individual Unicompartmental Knee Prostheses Sorted Alphabetically

|                               | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |          |
|-------------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|----------|
| EIUS Uni Knee                 | 22      | 234.8              | 2              | 0.85                     | 0.10                          | 3.08     |
| Freedom Active Uni            | 36      | 191.3              | 7              | 3.66                     | 1.47                          | 7.54     |
| Genesis Uni                   | 359     | 3,768.7            | 52             | 1.38                     | 1.02                          | 1.79     |
| HLS Uni Evolution             | 1       | 0.5                | 1              | 193.25                   | 4.89                          | 1,076.74 |
| Journey Uni                   | 12      | 26.5               | 1              | 3.77                     | 0.10                          | 20.99    |
| LCS Uni                       | 6       | 63.6               | 2              | 3.15                     | 0.38                          | 11.36    |
| Miller/Galante                | 710     | 8,072.3            | 82             | 1.02                     | 0.80                          | 1.25     |
| Optetrak Unicondylar Cemented | 101     | 838.9              | 11             | 1.31                     | 0.65                          | 2.35     |
| Oxford 3 cemented             | 4,139   | 40,162.4           | 564            | 1.40                     | 1.29                          | 1.53     |
| Oxford 3 uncemented           | 4,616   | 19,018.9           | 147            | 0.77                     | 0.65                          | 0.91     |
| Oxford TiNbN coated           | 1       | 7.5                | 0              | 0.00                     | 0.00                          | 49.50    |
| Oxinium Uni                   | 33      | 280.5              | 12             | 4.28                     | 0.00                          | 7.47     |
| Persona Partial cemented      | 62      | 18.5               | 1              | 5.40                     | 0.14                          | 30.11    |
| Preservation                  | 484     | 5,119.5            | 88             | 1.72                     | 1.38                          | 2.12     |
| Repicci II                    | 98      | 1,212.3            | 23             | 1.90                     | 1.20                          | 2.85     |
| Restoris MCK                  | 63      | 40.6               | 0              | 0.00                     | 0.00                          | 9.08     |
| Sigma HP Uni                  | 160     | 649.4              | 4              | 0.62                     | 0.17                          | 1.58     |
| Triathlon PKR                 | 232     | 1,026.7            | 9              | 0.88                     | 0.40                          | 1.66     |
| Unix Uni                      | 14      | 91.4               | 3              | 3.28                     | 0.68                          | 9.60     |
| Zimmer Unicompartmental Knee  | 1,478   | 6,155.6            | 29             | 0.47                     | 0.31                          | 0.67     |

| Oxford 3 uncemented | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|---------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Medial Oxford       | 4,292   | 17,645.29          | 124            | 0.70                     | 0.58                          | 0.84 |
| Lateral Dome Oxford | 324     | 1,375.87           | 23             | 1.67                     | 1.06                          | 2.51 |



### Revision vs Arthroplasty Fixation

| Fixation   | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Cemented   | 7,930   | 67,396.0           | 882            | 1.31                     | 1.22                          | 1.40 |
| Uncemented | 4,166   | 17,009.9           | 121            | 0.71                     | 0.59                          | 0.85 |
| Hybrid     | 531     | 2,573.8            | 35             | 1.36                     | 0.93                          | 1.87 |

### Revision vs Age Bands

| Age Bands | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <55       | 1,614   | 11,169.9           | 213            | 1.91                     | 1.66                          | 2.18 |
| 55-64     | 4,307   | 31,065.2           | 464            | 1.49                     | 1.36                          | 1.64 |
| 65-74     | 4,250   | 29,783.7           | 260            | 0.87                     | 0.77                          | 0.99 |
| >=75      | 2,456   | 14,961.0           | 101            | 0.68                     | 0.55                          | 0.82 |

### Revision vs Gender

| Gender | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|--------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| F      | 5,773   | 40,982             | 535            | 1.31                     | 1.20                          | 1.42 |
| M      | 6,854   | 45,998             | 503            | 1.09                     | 1.00                          | 1.19 |

### Revision vs Surgeon Annual Workload

| Consultant Number of ops/yr | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <10                         | 5,449   | 41,328             | 574            | 1.39                     | 1.28                          | 1.51 |
| >=10                        | 7,176   | 45,641             | 463            | 1.01                     | 0.92                          | 1.11 |

### Revision vs Surgical Approach


| Approach               | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Medial parapatellar    | 9,445   | 66,137             | 837            | 1.27                     | 1.18                          | 1.35 |
| Lateral parapatellar   | 270     | 1,928              | 31             | 1.61                     | 1.09                          | 2.28 |
| Not Minimally Invasive | 9,610   | 66,883             | 853            | 1.28                     | 1.19                          | 1.36 |
| Minimally Invasive     | 3,017   | 20,097             | 185            | 0.92                     | 0.79                          | 1.06 |
| Not Image guided       | 12,459  | 86,431             | 1,034          | 1.20                     | 1.12                          | 1.27 |
| Image guided           | 168     | 549                | 4              | 0.73                     | 0.00                          | 1.87 |



## KAPLAN MEIER CURVES


The following Kaplan Meier survival analyses are for the 19 years from 2000 to 2018, with deceased patients censored at time of death.

Unicompartmental Knees




| Years | % Revision-free | Number |
|-------|-----------------|--------|
| 1     | 98.6            | 11,317 |
| 2     | 97.1            | 10,047 |
| 3     | 96.2            | 9,019  |
| 4     | 95.6            | 8,054  |
| 5     | 94.7            | 7,206  |
| 6     | 93.9            | 6,361  |
| 7     | 93.0            | 5,558  |
| 8     | 92.0            | 4,878  |
| 9     | 91.0            | 4,193  |
| 10    | 89.8            | 3,523  |
| 11    | 88.6            | 2,956  |
| 12    | 87.2            | 2,407  |
| 13    | 85.6            | 1,873  |
| 14    | 83.9            | 1,424  |
| 15    | 82.1            | 972    |
| 16    | 80.3            | 611    |
| 17    | 78.9            | 329    |






### Survival curves for the 3 unicompartmental knees with the biggest number of implantations



#### Revision Rate for Re-revisions

| Re Revisions    | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Revised to full | 839     | 4,943.3            | 66             | 1.34                     | 1.03                          | 1.70 |
| Revised to Uni  | 199     | 859.6              | 42             | 4.89                     | 3.52                          | 6.60 |
| All             | 1,038   | 5,803.0            | 108            | 1.86                     | 1.53                          | 2.25 |



|             | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Total Knees | 110,076 | 753,723.0          | 3,652          | 0.48                     | 0.47                          | 0.50 |
| Uni Knees   | 12,627  | 86,979.7           | 1,038          | 1.19                     | 1.12                          | 1.27 |



**PATIENT BASED QUESTIONNAIRE OUTCOMES  
AT SIX MONTHS, FIVE YEARS, TEN YEARS  
AND FIFTEEN YEARS POST-SURGERY**

At six months post-surgery all patients are sent the Oxford-12 questionnaire.

There are 12 questions, with the scores now ranging from 4 to 0. A score of 48 is the best, indicating normal function. A score of 0 is the worst, indicating the most severe disability.

In addition we have grouped the questionnaire responses according to the classification system published by Kalairajah et al, 2005 (See appendix 1). This groups each score into four categories:

|            |         |           |
|------------|---------|-----------|
| Category 1 | >41     | excellent |
| Category 2 | 34 – 41 | good      |
| Category 3 | 27 – 33 | fair      |
| Category 4 | < 27    | poor      |

For the nineteen year period and as at July 2019, there were 8,166 unicompartmental knee questionnaire responses registered at six months post-surgery. The average unicompartmental knee score was 39.82 (standard deviation 7.15, range 3 – 48).

|                 |       |
|-----------------|-------|
| Scoring > 41    | 4,221 |
| Scoring 34 - 41 | 2,595 |
| Scoring 27- 33  | 857   |
| Scoring < 27    | 493   |

At six months post-surgery, 83% had an excellent or good score.

**Questionnaires at five years post surgery**

Patients who had a registered six month questionnaire and who had not had revision surgery were sent a further questionnaire at five years post-surgery.

This dataset represents sequential Oxford knee scores for 3,329 individual patients.

At five years post-surgery, 88% of patients had achieved an excellent or good score and had an average of 41.66.

**Questionnaires at ten years post-surgery**

All patients who had a six-month registered questionnaire, and who had not had revision surgery were sent a further questionnaire at ten years post-surgery.

This dataset represents sequential Oxford knee scores for 1,730 individual patients.


At ten years post-surgery, 84% of patients achieved an excellent or good score and had an average of 40.75.

**Questionnaires at fifteen years post-surgery**

All patients who had a six-month registered questionnaire, and who had not had revision surgery were sent a further questionnaire at fifteen years post-surgery.

This dataset represents sequential Oxford knee scores for 469 individual patients.

At fifteen years post-surgery, 84% of patients achieved an excellent or good score and had an average of 40.48.




## OXFORD 12 SCORE AS A PREDICTOR OF KNEE ARTHROPLASTY REVISION

A statistically significant relationship has been confirmed between the Oxford scores at six months, five years and ten years and arthroplasty revision within two years of the Oxford 12 questionnaire date.

### Six month score and revision arthroplasty

Plotting the patients' six month scores in the Kalairajah groupings against the proportion of knees revised for that same group demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score. A patient with a score below 27 has 18 times the risk of a revision within two years compared to a person with a score of >41.

Revision (%) to 2 years - by Oxford score at 6 months



Revision risk versus Kalairajah groupings of Oxford scores within two years of the six month score date

| Kalairajah group | Revision to 2 yrs | No. revised | %     | Std error |
|------------------|-------------------|-------------|-------|-----------|
| 0_26             | 406               | 75          | 18.47 | 1.93      |
| 27-33            | 720               | 33          | 4.58  | 0.78      |
| 34-41            | 2,126             | 30          | 1.41  | 0.26      |
| > 41             | 3,323             | 34          | 1.02  | 0.17      |

A person with an Oxford score >41 has a 1.17% risk of revision within two years compared to an 18.38% risk with a score of < 27.

### Five year score and revision arthroplasty

Plotting the patients' five year scores in the Kalairajah groupings against the proportion of knees revised for that same group demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score. A patient with a score below 27 has 17 times the risk of a revision within two years compared to a person with a score of >41.

Revision (%) to 2 years - by Oxford score at 5 Years






Revision risk versus Kalairajah groupings of Oxford scores within two years of the five year score date

| Kalairajah group | Revision to 2 yrs | No. revised | %     | Std error |
|------------------|-------------------|-------------|-------|-----------|
| 0_26             | 119               | 14          | 11.76 | 2.95      |
| 27-33            | 191               | 6           | 3.14  | 1.26      |
| 34-41            | 623               | 11          | 1.77  | 0.53      |
| > 41             | 1,718             | 12          | 0.70  | 0.20      |

Ten year score and revision arthroplasty


Plotting the patients' ten scores in the Kalairajah groupings against the proportion of knees revised for that same group demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score. A patient with a score below 27 has 11 times the risk of a revision within two years compared to a person with a score of >41.

Revision (%) to 2 years - by Oxford score at ten years



Revision risk versus Kalairajah groupings of Oxford scores within two years of the 10 year score date

| Kalairajah group | Revision to 2 yrs | No. revised | %     | Std error |
|------------------|-------------------|-------------|-------|-----------|
| 0_26             | 93                | 13          | 13.98 | 3.60      |
| 27-33            | 117               | 5           | 4.27  | 1.87      |
| 34-41            | 266               | 6           | 2.26  | 0.91      |
| > 41             | 718               | 9           | 1.25  | 0.42      |



# ANKLE ARTHROPLASTY

## PRIMARY ANKLE ARTHROPLASTY

The **nineteen year** report analyses data for the period January 2000 – December 2018. There were 1,619 primary ankle procedures registered, an additional 117 compared to last year's report.

## Data Analysis

### Age and sex distribution

The average age for an ankle replacement was 66 years, with a range of 32 – 96 years.

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 637    | 982   |
| Percentage    | 39.35  | 60.65 |
| Mean age      | 64.16  | 67.78 |
| Maximum age   | 95.52  | 90.78 |
| Minimum age   | 32.32  | 33.42 |
| Standard dev. | 9.79   | 8.52  |

### Body Mass Index

For the nine year period 2010 - 2018, there were 625 BMI registrations for primary ankle replacements. The average was 28.42 with a range of 17 – 54 and a standard deviation of 4.60.

### Previous operation

|  |       |
|--|-------|
| None   | 1,291 |
| Internal fixation for juxta articular fracture | 153   |
| Arthrodesis                                    | 46    |
| Osteotomy                                      | 23    |

### Diagnosis

|                      |       |
|----------------------|-------|
| Osteoarthritis       | 1,237 |
| Post trauma          | 252   |
| Rheumatoid arthritis | 131   |
| Other inflammatory   | 20    |
| Avascular necrosis   | 7     |

### Approach

|               |       |
|---------------|-------|
| Anterior      | 1,386 |
| Anterolateral | 48    |
| Other         | 25    |

### Bone graft

|                 |    |
|-----------------|----|
| Tibia autograft | 43 |
| Tibia allograft | 3  |
| Tibia synthetic | 1  |
| Talus autograft | 10 |
| Talus allograft | 3  |

### Cement

|                      |    |
|----------------------|----|
| Tibia cemented       | 27 |
| Antibiotic in cement | 17 |
| Talus cemented       | 19 |
| Antibiotic in cement | 12 |

### Systemic antibiotic prophylaxis

|   |             |
|---|-------------|
| Patient number receiving at least one systemic antibiotic | 1,558 (96%) |
|---|-------------|

### Operating theatre

|              |     |
|--------------|-----|
| Conventional | 803 |
| Laminar flow | 800 |
| Space suits  | 310 |

### ASA Class

This was introduced with the updated forms at the beginning of 2005.

For the fourteen year period 2005 -2018, there were 1,347 (92%) primary ankle procedures with the ASA class recorded.

### Definitions

|                     |   |
|---------------------|---|
| <b>ASA class 1:</b> | A healthy patient   |
| <b>ASA class 2:</b> | A patient with mild systemic disease  |
| <b>ASA class 3:</b> | A patient with severe systemic disease that limits activity but is not incapacitating |
| <b>ASA class 4:</b> | A patient with an incapacitating disease that is a constant threat to life            |

| ASA | Number |
|-----|--------|
| 1   | 257    |
| 2   | 824    |
| 3   | 261    |
| 4   | 5      |

### Operative time (skin to skin)

|      |             |
|------|-------------|
| Mean | 121 minutes |
|------|-------------|

### Surgeon grade

The updated forms introduced in 2005 have separated advanced trainee into supervised and unsupervised. The following figures are for the fourteen-year period 2005 -2018.


|                             |       |
|-----------------------------|-------|
| Consultant                  | 1,464 |
| Advanced trainee supervised | 11    |

### Prosthesis usage

#### Ankle prostheses used in 2018

|               |    |
|---------------|----|
| Infinity      | 45 |
| Salto Talaris | 32 |
| Salto         | 30 |
| Zimmer TM     | 19 |

## Most Used Ankle Prostheses 2014 – 2018



### Surgeon and hospital workload

#### Surgeons

In 2018, 18 surgeons performed 117 primary ankle procedures. 3 surgeons performed more than 15 procedures and 9 performed <5 procedures.

#### Hospitals

In 2018, primary ankle replacement was performed in 23 hospitals. 15 were public and 8 were private.

### REVISION ANKLE ARTHROPLASTY


Revision is defined by the Registry as a new operation in a previously replaced ankle joint, during which one or more of the components are exchanged, removed, manipulated or added. It includes arthrodesis or amputation, but not soft tissue procedures. A two or more staged procedure is registered as one revision.

### Data Analysis

For the nineteen year period January 2000–December 2018, there were 235 revision ankle procedures registered.

The average age for an ankle revision was 66 years, with a range of 35 – 85.

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 90     | 145   |
| Percentage    | 38.30  | 61.70 |
| Mean          | 64.00  | 66.71 |
| Maximum age   | 81.68  | 85.43 |
| Minimum age   | 42.13  | 34.55 |
| Standard dev. | 9.33   | 8.29  |



## REVISION OF REGISTERED PRIMARY ANKLE ARTHROPLASTIES

This section analyses data for revisions of primary ankle procedures for the nineteen-year period 2000 – 2018.

There were 181 revisions of the primary total ankle procedures of 1,619 (11%).

### Time to revision

|                    |            |
|--------------------|------------|
| Average            | 1,660 days |
| Maximum            | 5,173 days |
| Minimum            | 21 days    |
| Standard deviation | 1,226 days |

### Reason for revision

|                            |    |
|----------------------------|----|
| Pain                       | 79 |
| Loosening talar component  | 54 |
| Loosening tibial component | 40 |
| Deep infection             | 17 |
| Dislocation                | 4  |
| Fracture talus             | 3  |

### Ankle re-revisions

There were 19 registered primary ankle procedures that were revised twice and 2 procedures that were revised three times.

### Analysis of the four main reasons for revision by year after primary procedure

| Years        | Loosening talar component |      | Loosening tibial component |      | Pain      |      | Deep Infection |      |
|--------------|---------------------------|------|----------------------------|------|-----------|------|----------------|------|
|              | Count                     | %    | Count                      | %    | Count     | %    | Count          | %    |
| 0            | 3                         | 5.6  | 2                          | 5.0  | 4         | 5.1  | 8              | 47.1 |
| 1            | 7                         | 13.0 | 12                         | 30.0 | 16        | 20.3 | 2              | 11.8 |
| 2            | 8                         | 14.8 | 3                          | 7.5  | 10        | 12.7 | 2              | 11.8 |
| 3            | 8                         | 14.8 | 3                          | 7.5  | 10        | 12.7 | 2              | 11.8 |
| 4            | 8                         | 14.8 | 5                          | 12.5 | 12        | 15.2 | 1              | 5.9  |
| 5            | 4                         | 7.4  | 1                          | 2.5  | 5         | 6.3  | 0              | 0.0  |
| 6            | 3                         | 5.6  | 3                          | 7.5  | 5         | 6.3  | 0              | 0.0  |
| 7            | 2                         | 3.7  | 1                          | 2.5  | 4         | 5.1  | 1              | 5.9  |
| 8            | 2                         | 3.7  | 4                          | 10.0 | 4         | 5.1  | 0              | 0.0  |
| 9            | 3                         | 5.6  | 2                          | 5.0  | 3         | 3.8  | 0              | 0.0  |
| 10           | 2                         | 3.7  | 1                          | 2.5  | 3         | 3.8  | 0              | 0.0  |
| 11           | 2                         | 3.7  | 2                          | 5.0  | 3         | 3.8  | 1              | 5.9  |
| 12           | 0                         | 0.0  | 1                          | 2.5  | 0         | 0.0  | 0              | 0.0  |
| 13           | 1                         | 1.9  | 0                          | 0.0  | 0         | 0.0  | 0              | 0.0  |
| 14           | 1                         | 1.9  | 0                          | 0.0  | 0         | 0.0  | 0              | 0.0  |
| <b>Total</b> | <b>54</b>                 |      | <b>40</b>                  |      | <b>79</b> |      | <b>17</b>      |      |

### Statistical note

In the table below there are two statistical terms readers may not be familiar with:

#### i) Observed component years

This is the number of registered primary procedures multiplied by the number of years each component has been in place.

#### ii) Rate/100 component years

This is equivalent to the yearly revision rate expressed as a percent and is derived by dividing the number of prostheses revised by the observed component years multiplied by 100. It therefore allows for the number of years of post-operative follow up in calculating the revision rate.

These rates are usually very low, hence it is expressed per 100 component years rather than per component year. Statisticians consider that this is a more accurate way of deriving a revision rate for comparison when analysing data with widely varying follow-up times. It is also important to note the confidence intervals. The closer they are to the estimated revision rate/100 component years, the more precise the estimate is.

### Statistical significance

Where it is stated that a difference among results is significant the p value is 0.05 or less. In most of these situations this is because there is no overlap of the confidence intervals (CIs) but sometimes significance can apply in the presence of CI overlap.

### All Primary Ankle Arthroplasties

| No. Ops. | Observed comp. Yrs | Number Revised | Rate/100-component-years | Exact 95% confidence interval |      |
|----------|--------------------|----------------|--------------------------|-------------------------------|------|
| 1,619    | 10,037.6           | 181            | 1.80                     | 1.55                          | 2.09 |



### Revision vs Prosthesis Type Sorted in Alphabetical Order

| Prosthesis    | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |       |
|---------------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| Agility       | 119     | 1,320.5            | 35             | 2.65                     | 1.85                          | 3.69  |
| Box           | 6       | 36.5               | 2              | 5.48                     | 0.66                          | 19.80 |
| Hintegra      | 22      | 94.7               | 3              | 3.17                     | 0.65                          | 9.26  |
| Infinity      | 111     | 141.2              | 2              | 1.42                     | 0.00                          | 5.12  |
| Mobility      | 450     | 3,685.9            | 68             | 1.84                     | 1.43                          | 2.34  |
| Ramses        | 11      | 97.7               | 5              | 5.12                     | 1.66                          | 11.94 |
| Salto         | 721     | 3,953.9            | 53             | 1.34                     | 1.00                          | 1.75  |
| Salto Talaris | 98      | 192.1              | 0              | 0.00                     | 0.00                          | 1.92  |
| STAR          | 47      | 465.4              | 12             | 2.58                     | 1.33                          | 4.50  |
| Zimmer TM     | 34      | 49.7               | 1              | 2.01                     | 0.00                          | 11.21 |


### Revision vs Gender

| Gender  | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|---------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Females | 637     | 3,983.4            | 71             | 1.78                     | 1.39                          | 2.25 |
| Males   | 982     | 6,054.1            | 110            | 1.82                     | 1.49                          | 2.18 |

### Revision vs Age Bands


| Age Bands | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <55       | 170     | 1,138.5            | 35             | 3.07                     | 2.14                          | 4.28 |
| 55-64     | 502     | 3,522.5            | 81             | 2.30                     | 1.83                          | 2.86 |
| 65-74     | 661     | 3,958.6            | 57             | 1.44                     | 1.09                          | 1.87 |
| >=75      | 286     | 1,417.9            | 8              | 0.56                     | 0.24                          | 1.11 |





## KAPLAN MEIER CURVES

The following Kaplan Meier survival analyses are for the 19 years from 2000 to 2018, with deceased patients censored at time of death.



| Years | % Revision-free | No in each year |
|-------|-----------------|-----------------|
| 1     | 98.7            | 1,469           |
| 2     | 96.7            | 1,313           |
| 3     | 94.9            | 1,156           |
| 4     | 92.8            | 1,020           |
| 5     | 90.8            | 894             |
| 6     | 89.8            | 773             |
| 7     | 88.3            | 655             |
| 8     | 86.9            | 552             |
| 9     | 85.5            | 436             |
| 10    | 83.5            | 320             |
| 11    | 81.7            | 230             |
| 12    | 79.3            | 170             |
| 13    | 78.0            | 114             |
| 14    | 77.3            | 66              |

## PATIENT BASED QUESTIONNAIRE OUTCOMES AT SIX MONTHS POST-SURGERY

At six months post-surgery patients are sent an outcome questionnaire.

The non -validated ankle questionnaire used previously by the Registry was replaced by the validated Manchester-Oxford Foot Questionnaire towards the end of 2015.

This has 16 questions answered on a 5 point Likert scale, with each item scoring from 0 – 4, with 4 denoting “most severe”. Total score range from 0-64

For the 3 year period 2016 – 2018 there were 208 responses.

Average = 16.67, Maximum = 59, Minimum = 0 and Standard deviation = 13.82.

# SHOULDER ARTHROPLASTY

## PRIMARY SHOULDER ARTHROPLASTY

The **nineteen-year** report analyses data for the period January 2000 – December 2018. There were 10,324 primary shoulder procedures registered with an additional 1,066 registered in 2018.

Of the 10,324 shoulder registrations, 1,791 are hemi shoulder replacements, 3,449 are conventional total shoulder replacements, 4,681 are reverse shoulder replacements, 224 are partial resurfacing shoulder replacements, 178 are total resurfacing replacements and 1 is a humeral sphere.

## Data Analysis

### Age and sex distribution

The average age for all patients with a shoulder arthroplasty was 71 years, with a range of 15 – 99 years.

#### All shoulder arthroplasty

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 6,485  | 3,839 |
| Percentage    | 62.81  | 37.19 |
| Mean age      | 72.64  | 68.53 |
| Maximum age   | 97.71  | 99.36 |
| Minimum age   | 15.02  | 20.13 |
| Standard dev. | 9.40   | 10.15 |

#### Hemiarthroplasty

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 1,1165 | 625   |
| Percentage    | 65.08  | 34.92 |
| Mean age      | 71.17  | 64.63 |
| Maximum age   | 97.71  | 99.36 |
| Minimum age   | 15.02  | 20.13 |
| Standard dev. | 11.52  | 12.68 |

#### Conventional total shoulder arthroplasty

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 2,151  | 1,297 |
| Percentage    | 62.38  | 37.62 |
| Mean age      | 70.35  | 66.41 |
| Maximum age   | 95.43  | 89.11 |
| Minimum age   | 26.64  | 29.38 |
| Standard dev. | 8.74   | 8.79  |

#### Reverse shoulder arthroplasty

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 2,979  | 1,702 |
| Percentage    | 63.64  | 36.36 |
| Mean age      | 75.31  | 72.75 |
| Maximum age   | 96.82  | 92.65 |
| Minimum age   | 35.61  | 34.62 |
| Standard dev. | 7.74   | 7.73  |

## Partial resurfacing arthroplasty

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 80     | 144   |
| Percentage    | 35.71  | 64.29 |
| Mean age      | 58.46  | 55.86 |
| Maximum age   | 87.06  | 86.12 |
| Minimum age   | 20.70  | 21.83 |
| Standard dev. | 14.39  | 11.12 |

## Total resurfacing arthroplasty

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 109    | 71    |
| Percentage    | 60.55  | 39.45 |
| Mean age      | 71.17  | 66.36 |
| Maximum age   | 86.79  | 81.51 |
| Minimum age   | 47.24  | 23.67 |
| Standard dev. | 8.00   | 9.74  |

## Humeral sphere

One female patient aged 50.11 years.

## Previous operation

|                          |       |
|--------------------------|-------|
| None                     | 8,656 |
| Rotator cuff repair      | 661   |
| Internal fixation for    |       |
| Juxta articular fracture | 247   |
| Previous stabilisation   | 210   |
| Arthroscopic debridement | 55    |
| Osteotomy                | 6     |
| Arthrodesis              | 2     |

## Diagnosis

|                              |       |
|------------------------------|-------|
| Osteoarthritis               | 5,508 |
| Cuff tear arthropathy        | 2,393 |
| Acute fracture prox. humerus | 956   |
| Rheumatoid arthritis         | 657   |
| Post old trauma              | 566   |
| Avascular necrosis           | 204   |
| Post recurrent dislocation   | 153   |
| Other inflammatory           | 87    |

## Approach


|                               |       |
|-------------------------------|-------|
| Deltopectoral                 | 9,907 |
| Other including deltoid split | 280   |

## Bone graft

|                   |     |
|-------------------|-----|
| Humeral autograft | 121 |
| Humeral allograft | 25  |
| Humeral synthetic | 4   |
| Glenoid autograft | 170 |
| Glenoid allograft | 20  |

## Cement

|                      |       |
|----------------------|-------|
| Humerus cemented     | 1,778 |
| Antibiotic in cement | 1,114 |
| Glenoid cemented     | 2,578 |
| Antibiotic in cement | 1,807 |



### Systemic antibiotic prophylaxis

Patient number receiving at least one systemic antibiotic 9,720 (94%)

### Operating theatre

|              |       |
|--------------|-------|
| Conventional | 6,230 |
| Laminar flow | 3,959 |
| Space suits  | 1,779 |

### ASA Class

This was introduced with the updated forms at the beginning of 2005.

For the fourteen- year period 2005 – 2018 there were 9,342 (97%) shoulder procedures with the ASA class recorded.

### Definitions

**ASA class 1:** A healthy patient  
**ASA class 2:** A patient with mild systemic disease  
**ASA class 3:** A patient with severe systemic disease that limits activity but is not incapacitating  
**ASA class 4:** A patient with an incapacitating disease that is a constant threat to life

| ASA | Number | Percentage |
|-----|--------|------------|
| 1   | 764    | 8          |
| 2   | 5,090  | 56         |
| 3   | 3,078  | 34         |
| 4   | 104    | 2          |

### Operative time (skin to skin in minutes)

|                      | Mean |
|----------------------|------|
| Hemi Arthroplasty    | 110  |
| Conventional Total   | 126  |
| Partial Resurfacing  | 94   |
| Total Resurfacing    | 123  |
| Reverse Arthroplasty | 112  |

### Surgeon grade

The updated forms introduced in 2005 have separated advanced trainee into supervised and unsupervised.

The following figures are for the fourteen-year period 2005 – 2018.


|                               |       |
|-------------------------------|-------|
| Consultant                    | 8,903 |
| Advanced trainee supervised   | 53    |
| Advanced trainee unsupervised | 22    |
| Basic trainee                 | 5     |

### Top 10 shoulder prostheses 2018


|                       |     |
|-----------------------|-----|
| SMR Reverse           | 317 |
| Delta Xtend Reverse   | 233 |
| Aequalis reversed     | 109 |
| SMR conventional      | 54  |
| Aequalis conventional | 48  |
| Global Unite          | 38  |
| Affinis               | 38  |
| Comprehensive Reverse | 36  |
| Aequalis hemi         | 29  |
| Global AP             | 23  |

*Affinis is a new addition to the top 10 with some reshuffling of the order outside the top 3.*

## Most used shoulder prostheses for five years 2014 - 2018



## Percentages of the different types of shoulder prostheses used by year



### Surgeon and hospital workload


#### Surgeons

In 2018, 75 surgeons performed 1,066 shoulder procedures; an average of 14 procedures per surgeon. 19 surgeons performed more than 20 procedures and 8 surgeons each performed 1 procedure.

#### Hospitals

In 2018, shoulder replacement was performed in 48 hospitals. 27 were public and 21 were private.

For 2018, the average number of shoulder replacements per hospital was 22.



## REVISION SHOULDER ARTHROPLASTY

Revision is defined by the Registry as a new operation in a previously replaced shoulder joint during which one or more of the components are exchanged, removed, manipulated or added. It includes excision, arthrodesis or amputation, but not soft tissue procedures. A two or more staged procedure is registered as one revision.

### Data Analysis

For the nineteen- year period January 2000 – December 2018 there were 817 revision shoulder procedures registered.

The average age for a shoulder revision was 69 years with a range of 24 – 90 years.

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 476    | 341   |
| Percentage    | 58.26  | 41.74 |
| Mean          | 70.40  | 66.61 |
| Maximum age   | 89.95  | 88.46 |
| Minimum age   | 33.20  | 24.05 |
| Standard dev. | 9.96   | 10.18 |

## REVISION OF REGISTERED PRIMARY SHOULDER ARTHROPLASTIES

This section analyses data for revisions of primary shoulder procedures for the nineteen- year period January 2000 – December 2018.

There were 523 revisions of the primary group of 10,324 (5 %). There were 66 procedures that had been revised twice, 17 that had been revised three times and 4 revised 4 times.

### Time to revision

|                    |            |
|--------------------|------------|
| Average            | 1,142 days |
| Maximum            | 5,901 days |
| Minimum            | 0 days     |
| Standard deviation | 1,165 days |

### Reason for revision

|                                      |     |
|--------------------------------------|-----|
| Pain                                 | 108 |
| Sub acromial cuff impingement        | 80  |
| Dislocation/instability anterior     | 78  |
| Loosening glenoid                    | 78  |
| Deep infection                       | 41  |
| Loosening humeral                    | 24  |
| Instability posterior                | 17  |
| Fracture humerus                     | 11  |
| Sub acromial tuberosity impingement. | 7   |
| Loosening both                       | 7   |

### Analysis of the six main reasons for revision by year after primary procedure

|              | Loosening glenoid |      | Dislocation |      | Deep infection |      | Pain       |      | Sub acromial Cuff |      | Loosening Humeral Component |      |
|--------------|-------------------|------|-------------|------|----------------|------|------------|------|-------------------|------|-----------------------------|------|
| 0            | 19                | 24.4 | 47          | 60.3 | 14             | 34.1 | 22         | 20.4 | 19                | 23.8 | 6                           | 25.0 |
| 1            | 13                | 16.7 | 13          | 16.7 | 12             | 29.3 | 25         | 23.1 | 19                | 23.8 | 3                           | 12.5 |
| 2            | 8                 | 10.3 | 3           | 3.8  | 5              | 12.2 | 17         | 15.7 | 13                | 16.3 | 2                           | 8.3  |
| 3            | 5                 | 6.4  | 2           | 2.6  | 3              | 7.3  | 9          | 8.3  | 4                 | 5.0  | 3                           | 12.5 |
| 4            | 3                 | 3.8  | 4           | 5.1  | 4              | 9.8  | 10         | 9.3  | 5                 | 6.3  | 2                           | 8.3  |
| 5            | 5                 | 6.4  | 4           | 5.1  | 1              | 2.4  | 4          | 3.7  | 7                 | 8.8  | 3                           | 12.5 |
| 6            | 3                 | 3.8  | 1           | 1.3  | 0              | 0.0  | 4          | 3.7  | 2                 | 2.5  | 0                           | 0.0  |
| 7            | 1                 | 1.3  | 1           | 1.3  | 1              | 2.4  | 5          | 4.6  | 4                 | 5.0  | 0                           | 0.0  |
| 8            | 2                 | 2.6  | 2           | 2.6  | 0              | 0.0  | 3          | 2.8  | 1                 | 1.3  | 1                           | 4.2  |
| 9            | 7                 | 9.0  | 0           | 0.0  | 0              | 0.0  | 3          | 2.8  | 2                 | 2.5  | 2                           | 8.3  |
| 10           | 5                 | 6.4  | 0           | 0.0  | 0              | 0.0  | 1          | 0.9  | 3                 | 3.8  | 1                           | 4.2  |
| 11+          | 7                 | 9.0  | 1           | 1.3  | 1              | 2.4  | 5          | 4.6  | 1                 | 1.3  | 1                           | 4.2  |
| <b>Total</b> | <b>78</b>         |      | <b>78</b>   |      | <b>41</b>      |      | <b>108</b> |      | <b>80</b>         |      | <b>24</b>                   |      |

### Statistical note

In the table below there are two statistical terms readers may not be familiar with:

#### i) Observed component years

This is the number of registered primary procedures multiplied by the number of years each component has been in place.

#### ii) Rate/100 component years

This is equivalent to the yearly revision rate expressed as a percent and is derived by dividing the number of prostheses revised by the observed component years multiplied by 100. It therefore allows for the number of years of post-operative follow up in calculating the revision rate. These rates are usually very low, hence are expressed per 100 component years rather than per component year.

Statisticians consider that this is a more accurate way of deriving a revision rate for comparison when analysing data with widely varying follow up times. It is also important to note the confidence intervals. The closer they are to the estimated revision rate/100 component years, the more precise the estimate is.

### Statistical significance

Where it is stated that a difference among results is significant the p value is 0.05 or less. In most of these situations this is because there is no overlap of the confidence intervals (CIs) but sometimes significance can apply in the presence of CI overlap.

### All Total Shoulder Arthroplasties

| No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| 10,324  | 54,863.6           | 523            | 0.95                     | 0.87                          | 1.04 |


### Revision rate of Shoulder Prostheses vs Arthroplasty Type

| Operation Type      | No. Ops. | Observed | Number Revised | Rate/100 component-years | Exact 95% confidence interval |       |
|---------------------|----------|----------|----------------|--------------------------|-------------------------------|-------|
| Total               | 3,449    | 21,190.9 | 199            | 0.94                     | 0.81                          | 1.08  |
| Reverse             | 4,681    | 17,168.7 | 131            | 0.76                     | 0.64                          | 0.91  |
| Hemi                | 1,791    | 13,949.3 | 154            | 1.10                     | 0.94                          | 1.29  |
| Resurfacing         | 178      | 917.7    | 5              | 0.54                     | 0.18                          | 1.27  |
| Partial resurfacing | 224      | 1,631.9  | 34             | 2.08                     | 1.44                          | 2.91  |
| Humeral Sphere      | 1        | 5.1      | 0              | 0.00                     | 0.00                          | 72.75 |

There is a significantly higher revision rate for Partial Resurfacing compared to all the other types.

### Revision Rate of Individual Shoulder Prostheses Sorted on Alphabetical Order

| Operation Type | Prosthesis                  | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |        |
|----------------|-----------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|--------|
| Total          | Aequalis                    | 576     | 3141.1             | 16             | 0.51                     | 0.28                          | 0.81   |
|                | Affinis                     | 76      | 111.1              | 0              | 0.00                     | 0.00                          | 3.32   |
|                | Anatomical                  | 35      | 462.4              | 1              | 0.22                     | 0.01                          | 1.20   |
|                | Arthrex Eclipse             | 11      | 6.1                | 0              | 0.00                     | 0.00                          | 60.47  |
|                | Ascend TM                   | 2       | 10.4               | 0              | 0.00                     | 0.00                          | 35.49  |
|                | Bi-Angular                  | 8       | 82.5               | 0              | 0.00                     | 0.00                          | 4.47   |
|                | Bigliani/Flatow             | 301     | 2709.7             | 9              | 0.33                     | 0.14                          | 0.61   |
|                | Cofield 2                   | 21      | 241.3              | 0              | 0.00                     | 0.00                          | 1.53   |
|                | Comprehensive               | 43      | 89.6               | 0              | 0.00                     | 0.00                          | 4.12   |
|                | Epoca Humeral stem          | 4       | 32.0               | 0              | 0.00                     | 0.00                          | 11.51  |
|                | Equinox Preserve            | 4       | 1.4                | 0              | 0.00                     | 0.00                          | 258.12 |
|                | Global                      | 519     | 4629.8             | 24             | 0.52                     | 0.33                          | 0.77   |
|                | Global AP                   | 498     | 2551.4             | 8              | 0.31                     | 0.14                          | 0.62   |
|                | Global Icon                 | 3       | 1.5                | 0              | 0.00                     | 0.00                          | 245.42 |
|                | Global Unite                | 195     | 403.8              | 3              | 0.74                     | 0.15                          | 2.17   |
|                | Humeral stem                | 1       | 6.3                | 0              | 0.00                     | 0.00                          | 58.15  |
|                | Neer 3                      | 2       | 28.4               | 0              | 0.00                     | 0.00                          | 12.99  |
|                | Neer II                     | 12      | 156.7              | 1              | 0.64                     | 0.02                          | 3.56   |
|                | Osteonics humeral component | 49      | 512.0              | 6              | 1.17                     | 0.37                          | 2.42   |



| Operation Type | Prosthesis                 | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |        |
|----------------|----------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|--------|
|                | Sidus                      | 1       | 4.3                | 0              | 0.00                     | 0.00                          | 85.28  |
|                | Simpliciti TM              | 41      | 91.0               | 1              | 1.10                     | 0.00                          | 6.12   |
|                | SMR                        | 1041    | 5875.9             | 130            | 2.21                     | 1.84                          | 2.62   |
|                | Univers 3D                 | 5       | 40.5               | 0              | 0.00                     | 0.00                          | 9.11   |
|                | Univers Apex               | 1       | 1.6                | 0              | 0.00                     | 0.00                          | 229.53 |
| Reverse        | Aequalis                   | 346     | 758.3              | 7              | 0.92                     | 0.37                          | 1.90   |
|                | Aequalis Reversed          | 162     | 550.7              | 4              | 0.73                     | 0.20                          | 1.86   |
|                | Aequalis Reversed Fracture | 43      | 128.4              | 0              | 0.00                     | 0.00                          | 2.87   |
|                | Affinis                    | 22      | 41.2               | 1              | 2.43                     | 0.06                          | 13.52  |
|                | Arthrex Univers Revers     | 15      | 2.6                | 0              | 0.00                     | 0.00                          | 140.20 |
|                | Comprehensive              | 150     | 286.6              | 1              | 0.35                     | 0.00                          | 1.94   |
|                | Delta                      | 55      | 500.2              | 2              | 0.40                     | 0.05                          | 1.44   |
|                | Delta Xtend Reverse        | 1636    | 5954.3             | 55             | 0.92                     | 0.70                          | 1.20   |
|                | Equinox Preserve           | 14      | 3.9                | 0              | 0.00                     | 0.00                          | 95.35  |
|                | Global Unite               | 6       | 5.0                | 0              | 0.00                     | 0.00                          | 73.39  |
|                | Mutars                     | 1       | 0.6                | 0              | 0.00                     | 0.00                          | 663.73 |
|                | RSP                        | 2       | 3.8                | 0              | 0.00                     | 0.00                          | 95.97  |
|                | SMR                        | 2190    | 8783.3             | 59             | 0.67                     | 0.51                          | 0.86   |
|                | Trabecular Metal Reverse   | 38      | 142.1              | 2              | 1.41                     | 0.17                          | 5.08   |
|                | Vaios                      | 1       | 7.7                | 0              | 0.00                     | 0.00                          | 47.91  |
| Hemi           | Aequalis                   | 239     | 1394.6             | 15             | 1.08                     | 0.58                          | 1.73   |
|                | Aequalis Reversed          | 1       | 2.4                | 0              | 0.00                     | 0.00                          | 153.46 |
|                | Affinis                    | 9       | 25.8               | 1              | 3.87                     | 0.10                          | 21.58  |
|                | Anatomical                 | 19      | 246.0              | 0              | 0.00                     | 0.00                          | 1.50   |
|                | Arthrex Eclipse            | 3       | 20.3               | 0              | 0.00                     | 0.00                          | 18.18  |
|                | Ascend TM                  | 1       | 6.6                | 0              | 0.00                     | 0.00                          | 56.14  |
|                | Bi-Angular                 | 19      | 218.7              | 2              | 0.91                     | 0.11                          | 3.30   |
|                | Bigliani/Flatow            | 137     | 1352.3             | 15             | 1.11                     | 0.59                          | 1.78   |
|                | Bio-modular                | 1       | 7.1                | 1              | 14.00                    | 0.35                          | 78.03  |
|                | Cofield 2                  | 50      | 583.1              | 1              | 0.17                     | 0.00                          | 0.96   |
|                | Comprehensive              | 3       | 5.3                | 0              | 0.00                     | 0.00                          | 70.25  |
|                | Delta                      | 1       | 8.8                | 0              | 0.00                     | 0.00                          | 42.08  |
|                | Delta Xtend Reverse        | 27      | 110.0              | 4              | 3.64                     | 0.99                          | 9.31   |
|                | Global                     | 723     | 6492.7             | 57             | 0.88                     | 0.66                          | 1.14   |
|                | Global AP                  | 90      | 466.6              | 3              | 0.64                     | 0.09                          | 1.72   |
|                | Global Icon                | 1       | 0.8                | 0              | 0.00                     | 0.00                          | 449.12 |
|                | Global Unite               | 60      | 173.5              | 10             | 5.76                     | 2.76                          | 10.60  |
|                | MRS Humeral                | 4       | 18.9               | 0              | 0.00                     | 0.00                          | 19.47  |




| Operation Type      | Prosthesis                  | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |         |
|---------------------|-----------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|---------|
|                     | Neer II                     | 24      | 248.7              | 0              | 0.00                     | 0.00                          | 1.48    |
|                     | Osteonics humeral component | 43      | 403.8              | 2              | 0.50                     | 0.06                          | 1.79    |
|                     | Randelli                    | 1       | 8.2                | 0              | 0.00                     | 0.00                          | 44.82   |
|                     | Simpliciti TM               | 1       | 3.4                | 0              | 0.00                     | 0.00                          | 107.19  |
|                     | SMR                         | 332     | 2138.5             | 43             | 2.01                     | 1.46                          | 2.71    |
|                     | Trabecular Metal Reverse    | 1       | 9.2                | 0              | 0.00                     | 0.00                          | 39.96   |
|                     | Univers 3D                  | 1       | 3.8                | 0              | 0.00                     | 0.00                          | 96.59   |
| Total Resurfacing   | Aequalis Resurfacing Head   | 10      | 71.8               | 0              | 0.00                     | 0.00                          | 5.14    |
|                     | Epoca Head                  | 103     | 481.3              | 4              | 0.83                     | 0.18                          | 1.98    |
|                     | Global CAP Resurfacing      | 61      | 345.7              | 1              | 0.29                     | 0.01                          | 1.61    |
|                     | Hemicap Resurfacing         | 1       | 2.7                | 0              | 0.00                     | 0.00                          | 135.01  |
|                     | SMR Resurfacing             | 3       | 16.1               | 0              | 0.00                     | 0.00                          | 22.93   |
| Partial resurfacing | Aequalis Resurfacing Head   | 1       | 3.0                | 0              | 0.00                     | 0.00                          | 121.06  |
|                     | Arthrex Eclipse             | 3       | 11.9               | 2              | 16.76                    | 2.03                          | 60.55   |
|                     | Ascension                   | 20      | 120.7              | 2              | 1.66                     | 0.09                          | 5.99    |
|                     | Copeland Resurfacing        | 19      | 166.0              | 4              | 2.41                     | 0.66                          | 6.17    |
|                     | Custom Global Cap           | 1       | 7.4                | 0              | 0.00                     | 0.00                          | 49.77   |
|                     | Epoca Head                  | 21      | 109.8              | 2              | 1.82                     | 0.22                          | 6.58    |
|                     | Global AP CTA Humeral Head  | 1       | 0.5                | 1              | 188.27                   | 4.77                          | 1048.99 |
|                     | Global Cap CTA              | 1       | 0.5                | 0              | 0.00                     | 0.00                          | 677.07  |
|                     | Global CAP Resurfacing      | 96      | 807.1              | 13             | 1.61                     | 0.86                          | 2.75    |
|                     | Global Humeral Head         | 1       | 6.2                | 0              | 0.00                     | 0.00                          | 59.15   |
|                     | Hemicap Resurfacing         | 8       | 58.0               | 1              | 1.72                     | 0.04                          | 9.60    |
|                     | SMR Resurfacing             | 45      | 297.6              | 7              | 2.35                     | 0.95                          | 4.85    |
|                     | SMR Resurfacing CTA         | 7       | 43.0               | 2              | 4.65                     | 0.56                          | 16.79   |

**Revision vs Glenoid Fixation**  
(Conventional Total arthroplasties only)

|            | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Uncemented | 995     | 6,041.9            | 124            | 2.05                     | 1.71                          | 2.45 |
| Cemented   | 2,453   | 15,150.4           | 75             | 0.50                     | 0.39                          | 0.62 |





### Revision vs Prosthesis Group vs Age Bands

| Prosthesis          | Age Bands | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |       |
|---------------------|-----------|---------|--------------------|----------------|--------------------------|-------------------------------|-------|
| Total               | <55       | 219     | 1,180.1            | 25             | 2.12                     | 1.37                          | 3.13  |
|                     | 55-64     | 841     | 5,077.6            | 74             | 1.46                     | 1.14                          | 1.83  |
|                     | 65-74     | 1,517   | 9,501.5            | 74             | 0.78                     | 0.61                          | 0.98  |
|                     | >=75      | 871     | 5,433.1            | 26             | 0.48                     | 0.31                          | 0.70  |
| Reverse             | <55       | 55      | 135.4              | 3              | 2.22                     | 0.31                          | 5.91  |
|                     | 55-64     | 500     | 1,801.5            | 26             | 1.44                     | 0.92                          | 2.08  |
|                     | 65-74     | 1,802   | 6,577.3            | 55             | 0.84                     | 0.63                          | 1.09  |
|                     | >=75      | 2,324   | 8,654.5            | 47             | 0.54                     | 0.40                          | 0.72  |
| Hemi                | <55       | 237     | 1,809.8            | 28             | 1.55                     | 1.01                          | 2.20  |
|                     | 55-64     | 365     | 2,986.0            | 56             | 1.88                     | 1.40                          | 2.42  |
|                     | 65-74     | 538     | 4,567.0            | 44             | 0.96                     | 0.70                          | 1.29  |
|                     | >=75      | 650     | 4,581.9            | 26             | 0.57                     | 0.37                          | 0.83  |
| Resurfacing         | <55       | 8       | 36.0               | 1              | 2.78                     | 0.07                          | 15.50 |
|                     | 55-64     | 44      | 246.3              | 1              | 0.41                     | 0.01                          | 2.26  |
|                     | 65-74     | 82      | 410.8              | 3              | 0.73                     | 0.15                          | 2.13  |
|                     | >=75      | 46      | 227.9              | 0              | 0.00                     | 0.00                          | 1.62  |
| Partial resurfacing | <55       | 92      | 680.1              | 16             | 2.35                     | 1.34                          | 3.82  |
|                     | 55-64     | 72      | 558.1              | 9              | 1.61                     | 0.74                          | 3.06  |
|                     | 65-74     | 47      | 313.8              | 8              | 2.55                     | 1.10                          | 5.02  |
|                     | >=75      | 13      | 80.0               | 1              | 1.25                     | 0.03                          | 6.96  |

### Revision vs Age Bands

| Age Bands | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <55       | 612     | 3,846.4            | 73             | 1.90                     | 1.49                          | 2.39 |
| 55-64     | 1,822   | 10,669.5           | 166            | 1.56                     | 1.32                          | 1.81 |
| 65-74     | 3,986   | 21,370.4           | 184            | 0.86                     | 0.74                          | 0.99 |
| >=75      | 3,904   | 18,977.4           | 100            | 0.53                     | 0.43                          | 0.64 |

### Revision vs Gender

| Gender  | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|---------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Females | 6,485   | 35,033.5           | 308            | 0.88                     | 0.78                          | 0.98 |
| Males   | 3,839   | 19,830.1           | 215            | 1.08                     | 0.94                          | 1.24 |

### Revision vs Surgeon Annual Workload


| Consultant Number of ops/yr | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <10                         | 3,804   | 21,332.1           | 220            | 1.03                     | 0.90                          | 1.17 |
| >=10                        | 6,520   | 33,531.5           | 303            | 0.90                     | 0.80                          | 1.01 |




## KAPLAN MEIER CURVES

The following Kaplan Meier survival analyses are for the 19 years from 2000 to 2018, with deceased patients censored at time of death.


All Shoulders



| Years | % Revision-free | Number |
|-------|-----------------|--------|
| 1     | 98.4            | 8,990  |
| 2     | 97.2            | 7,741  |
| 3     | 96.3            | 6,606  |
| 4     | 95.6            | 5,468  |
| 5     | 94.8            | 4,550  |
| 6     | 94.0            | 3,745  |
| 7     | 93.7            | 3,031  |
| 8     | 93.2            | 2,477  |
| 9     | 92.8            | 2,002  |
| 10    | 91.9            | 1,566  |
| 11    | 91.4            | 1,189  |
| 12    | 91.2            | 870    |
| 13    | 90.3            | 615    |
| 14    | 90.3            | 446    |
| 15    | 89.4            | 292    |
| 16    | 88.7            | 175    |



Survival curves for different shoulder categories



## PATIENT BASED QUESTIONNAIRE OUTCOMES AT SIX MONTH, FIVE YEARS, TEN YEARS AND FIFTEEN YEARS POST-SURGERY

### Questionnaires at six months post-surgery

At six months post-surgery patients are sent the Oxford-12 questionnaire.

The scores now range from 4 to 0. A score of 48 is the best, indicating normal function. A score of 0 is the worst, indicating the most severe disability.

We have grouped the questionnaire responses based on the scoring system as published by Kalairajah et al, in 2005 (See appendix 1). This groups each score into four categories:

|            |         |           |
|------------|---------|-----------|
| Category 1 | >41     | excellent |
| Category 2 | 34 – 41 | good      |
| Category 3 | 27 – 33 | fair      |
| Category 4 | < 27    | poor      |

For the nineteen year period and as at July 2019, there were 6,564 shoulder questionnaire responses registered at six months post-surgery.

The average shoulder score was 36.48 (standard deviation 9.40, range 2 – 48)

|         |         |       |
|---------|---------|-------|
| Scoring | > 41    | 2,438 |
| Scoring | 34 - 41 | 2,122 |
| Scoring | 27 - 33 | 971   |
| Scoring | <27     | 1,033 |

At six months post-surgery, 69% had an excellent or good score.

### Questionnaires at five years post-surgery

All patients who had a six month registered questionnaire, and who had not had revision surgery, were sent a further questionnaire at five years post-surgery.

This dataset represents sequential Oxford shoulder scores for 2,131 individual patients.

At five years post-surgery, 80% of these patients achieved an excellent or good score and had an average of 39.90.

### Questionnaires at ten years post-surgery

All patients who had a six month registered questionnaire, and who had not had revision surgery, were sent a further questionnaire at ten years post-surgery.

This dataset represents sequential Oxford shoulder scores for 674 individual patients.

At ten years post-surgery, 77% of these patients achieved an excellent or good score and had an average of 39.11.

### Questionnaires at fifteen years post-surgery

All patients who had a six month registered questionnaire, and who had not had revision surgery, were sent a further questionnaire at ten years post-surgery.

This dataset represents sequential Oxford shoulder scores for 131 individual patients.

At fifteen years post-surgery, 77% of these patients achieved an excellent or good score and had an average of 39.14.

### Revision shoulder questionnaire responses


There were 428 revision shoulder responses with 46% achieving an excellent or good score. This group includes all revision shoulder responses. The average revision shoulder score was 30.93 (standard deviation 10.57 range 3 – 48).





Six Month and Five Year Oxford Scores for the different arthroplasty types

| Prosthesis type     | Time Post-Surgery | Mean Score | Std. Error | 95% Confidence Interval |             |
|---------------------|-------------------|------------|------------|-------------------------|-------------|
|                     |                   |            |            | Lower Bound             | Upper Bound |
| Total               | 6 Months          | 39.51      | 0.17       | 39.18                   | 39.84       |
|                     | 5 Years           | 42.07      | 0.25       | 41.59                   | 42.55       |
| Reverse             | 6 Months          | 35.51      | 0.18       | 35.17                   | 35.86       |
|                     | 5 Years           | 39.75      | 0.33       | 39.10                   | 40.41       |
| Hemi                | 6 Months          | 31.86      | 0.30       | 31.27                   | 32.46       |
|                     | 5 Years           | 35.56      | 0.45       | 34.67                   | 36.46       |
| Resurfacing         | 6 Months          | 42.22      | 0.45       | 41.32                   | 43.11       |
|                     | 5 Years           | 42.36      | 1.14       | 40.07                   | 44.65       |
| Partial Resurfacing | 6 Months          | 35.65      | 0.84       | 33.99                   | 37.31       |
|                     | 5 Years           | 39.16      | 1.21       | 36.73                   | 41.58       |

Comparison of six month and five year scores for different arthroplasty types







## OXFORD 12 SCORE AS A PREDICTOR OF SHOULDER ARTHROPLASTY REVISION

A statistically significant relationship has been confirmed between the Oxford scores at six months and five years and arthroplasty revision within two years of the Oxford 12 questionnaire date.

### Six month score and revision arthroplasty

Plotting the patients' six month scores in the Kalairajah groupings against the proportion of shoulders revised for that same group demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score. A patient with a score below 27 has 7 times the risk of a revision within two years compared to a person with a score of >41




Revision risk versus Kalairajah groupings of Oxford scores within two years of the six month score date

| Kalairajah group | No in group | No. revised | %    | Std error |
|------------------|-------------|-------------|------|-----------|
| 0_26             | 773         | 54          | 6.99 | 0.92      |
| 27-33            | 757         | 27          | 3.57 | 0.67      |
| 34-41            | 1,581       | 17          | 1.08 | 0.26      |
| > 41             | 1,874       | 18          | 0.96 | 0.23      |

### Five year score and revision arthroplasty

Plotting the patients' five year scores in the Kalairajah groupings against the proportion of shoulders revised for that same group demonstrates that there is an incremental increase in risk during the next two years related to the Oxford score, although it is not as clear cut as for the hips and knees. A patient with a score below 27 has 3 times the risk of a revision within two years compared to a person with a score of >41.






### Revision risk versus Kalairajah groupings of Oxford scores within two years of the 5 year score date

| Kalairajah group | No in group | No. revised | %    | Std error |
|------------------|-------------|-------------|------|-----------|
| 0-26             | 126         | 2           | 1.59 | 1.11      |
| 27-33            | 178         | 5           | 2.81 | 1.24      |
| 34-41            | 327         | 2           | 0.61 | 0.43      |
| > 41             | 828         | 4           | 0.48 | 0.24      |

A person with an Oxford score >41 has a 0.17% risk of revision within two years compared to a 2.24% risk with a score 27-33.

### Oxford score at 6 months by shoulder operation



| Operation types          | No. of operations | Mean        | Std. Error | 95% confidence interval |             |
|--------------------------|-------------------|-------------|------------|-------------------------|-------------|
| Total                    | 2,340             | 39.5        | 0.2        | 39.1                    | 39.9        |
| Reverse                  | 2,903             | 35.5        | 0.2        | 35.2                    | 35.8        |
| Hemi                     | 1,076             | 31.9        | 0.3        | 31.3                    | 32.4        |
| Resurfacing head         | 130               | 42.2        | 0.8        | 40.7                    | 43.8        |
| Partial resurfacing head | 114               | 35.6        | 0.8        | 34.0                    | 37.3        |
| <b>Total</b>             | <b>6,564</b>      | <b>36.5</b> | <b>0.1</b> | <b>36.3</b>             | <b>36.7</b> |




Oxford score at 5 Years by shoulder operation




| Operation types          | No. of operations | Mean | Std. Error | 95% confidence interval |      |
|--------------------------|-------------------|------|------------|-------------------------|------|
| Total                    | 936               | 42.1 | 0.3        | 41.5                    | 42.6 |
| Reverse                  | 630               | 39.8 | 0.3        | 39.1                    | 40.4 |
| Hemi                     | 467               | 35.6 | 0.4        | 34.8                    | 36.3 |
| Resurfacing head         | 47                | 42.4 | 1.2        | 40.0                    | 44.8 |
| Partial resurfacing head | 51                | 39.2 | 1.2        | 36.9                    | 41.5 |
| Total                    | 2,131             | 39.9 | 0.2        | 39.5                    | 40.3 |





### Oxford score at 10 Years by shoulder operation



| Operation types          | No. of operations | Mean | Std. Error | 95% confidence interval |      |
|--------------------------|-------------------|------|------------|-------------------------|------|
| Total                    | 335               | 41.0 | 0.5        | 40.0                    | 42.0 |
| Reverse                  | 104               | 39.4 | 0.9        | 37.7                    | 41.1 |
| Hemi                     | 217               | 36.4 | 0.6        | 35.2                    | 37.6 |
| Partial resurfacing head | 17                | 35.4 | 2.2        | 31.1                    | 39.6 |
| Total                    | 674               | 39.1 | 0.4        | 38.4                    | 39.8 |
| Total                    | 2,131             | 39.9 | 0.2        | 39.5                    | 40.3 |

# ELBOW ARTHROPLASTY

## PRIMARY ELBOW ARTHROPLASTY

The **nineteen-year** report analyses data for the period January 2000 – December 2018. There were 587 primary elbow procedures registered with an additional 27 registered in 2018.

### Data Analysis

#### Age and sex distribution

The average age for an elbow replacement was 67 years, with a range of 15 – 92 years.

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 450    | 137   |
| Percentage    | 76.66  | 23.34 |
| Mean age      | 67.86  | 65.42 |
| Maximum age   | 92.41  | 91.73 |
| Minimum age   | 36.38  | 15.16 |
| Standard dev. | 11.50  | 14.56 |

#### Previous operation

|  |     |
|--|-----|
| None   | 487 |
| Internal fixation for juxta articular fracture | 30  |
| Synovectomy+-removal radial head               | 22  |
| Debridement                                    | 15  |
| Osteotomy                                      | 3   |
| Ligament reconstruction                        | 3   |
| Interposition arthroplasty                     | 2   |

#### Diagnosis

|                          |     |
|--------------------------|-----|
| Rheumatoid arthritis     | 290 |
| Post fracture            | 179 |
| Osteoarthritis           | 90  |
| Other inflammatory       | 13  |
| Post dislocation         | 10  |
| Post ligament disruption | 6   |

#### Approach

|           |     |
|-----------|-----|
| Posterior | 370 |
| Medial    | 106 |
| Lateral   | 39  |

#### Bone graft

|                   |    |
|-------------------|----|
| Humeral autograft | 37 |
| Humeral allograft | 3  |
| Humeral synthetic | 1  |
| Ulnar autograft   | 3  |

#### Cement

|                      |           |
|----------------------|-----------|
| Humerus cemented     | 532       |
| Antibiotic in cement | 412 (77%) |
| Ulna cemented        | 503       |
| Antibiotic in cement | 386 (77%) |
| Radius cemented      | 27        |
| Antibiotic in cement | 26 (96%)  |

#### Systemic antibiotic prophylaxis

|   |           |
|---|-----------|
| Patient number receiving at least one systemic antibiotic | 587 (94%) |
|---|-----------|

#### Operating theatre

|              |     |
|--------------|-----|
| Conventional | 389 |
| Laminar flow | 193 |
| Space suits  | 81  |

#### ASA Class

This was introduced with the updated forms at the beginning of 2005.

For the fourteen- year period 2005 – 2018, there were 433 (95%) primary elbow procedures with the ASA class recorded.

#### Definitions

**ASA class 1:** A healthy patient

**ASA class 2:** A patient with mild systemic disease

**ASA class 3:** A patient with severe systemic disease that limits activity but is not incapacitating

**ASA class 4:** A patient with an incapacitating disease that is a constant threat to life

| ASA | Number |
|-----|--------|
| 1   | 21     |
| 2   | 191    |
| 3   | 213    |
| 4   | 8      |

#### Operative time (skin to skin)

|      |             |
|------|-------------|
| Mean | 146 minutes |
|------|-------------|

#### Surgeon grade

The updated forms introduced in 2005 have separated advanced trainee into supervised and unsupervised.

The following figures are for the fourteen- year period 2005 – 2018.

|                               |     |
|-------------------------------|-----|
| Consultant                    | 450 |
| Advanced trainee supervised   | 8   |
| Advanced trainee unsupervised | 4   |

#### Surgeon and hospital workload

In 2018, 15 surgeons performed 27 primary elbow procedures. These ranged from 1 to 4 per surgeon, with 8 performing 1 elbow procedure.


#### Hospitals

In 2018, primary elbow replacement was performed in 13 hospitals, of which 11 were public and 2 were private.


#### Prosthesis usage

##### Elbow prostheses used in 2018

|                |    |
|----------------|----|
| Zimmer Nexel   | 21 |
| Latitude       | 3  |
| Coonrad/Morrey | 2  |
| Evolve         | 1  |



## MOST USED ELBOW PROSTHESES FOR FIVE YEARS 2014 – 2018



## REVISION ELBOW ARTHROPLASTY

Revision is defined by the Registry as a new operation in a previously replaced elbow joint during which one or more of the components are exchanged, removed, manipulated or added. It includes arthrodesis or amputation, but not soft tissue procedures. A two or more staged procedure is registered as one revision.

### Data Analysis

For the nineteen-year period January 2000 – December 2018, there were 102 revision elbow procedures registered.

The average age for a revision elbow replacement was 66 years, with a range of 30 – 91 years.

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 72     | 30    |
| Percentage    | 70.59  | 29.41 |
| Mean          | 66.47  | 64.53 |
| Maximum age   | 89.08  | 90.50 |
| Minimum age   | 42.23  | 30.34 |
| Standard dev. | 9.91   | 14.98 |

## REVISION OF REGISTERED PRIMARY ELBOW ARTHROPLASTIES

This section analyses data for revisions of primary elbow procedures for the nineteen- year period January 2000 – December 2018.

There were 42 revisions of the primary group of 587 (7.2%).

There were 6 that had been revised twice and 1 that had been revised 3 times.

### Time to revision


|                    |            |
|--------------------|------------|
| Average            | 1,614 days |
| Maximum            | 5,174 days |
| Minimum            | 62 days    |
| Standard deviation | 1,377 days |

### Reason for revision

|                                 |    |
|---------------------------------|----|
| Loosening humeral component     | 5  |
| Deep infection                  | 13 |
| Loosening ulnar component       | 12 |
| Pain                            | 5  |
| Fracture humerus                | 4  |
| Loosening radial head component | 4  |
| Dislocation                     | 2  |
| Fracture ulna                   | 2  |

### Analysis of the three main reasons for revision by year after primary procedure

| Years        | Loosening humeral component |      | Loosening Ulnar component |      | Deep infection |      |
|--------------|-----------------------------|------|---------------------------|------|----------------|------|
|              | Count                       | %    | Count                     | %    | Count          | %    |
| 0            | 1                           | 6.7  | 1                         | 7.7  | 2              | 16.7 |
| 1            | 2                           | 13.3 | 0                         | 0.0  | 4              | 33.3 |
| 2            | 3                           | 20.0 | 4                         | 30.8 | 3              | 25.0 |
| 3            | 3                           | 20.0 | 2                         | 15.4 | 0              | 0.0  |
| 4            | 1                           | 6.7  | 0                         | 0.0  | 0              | 0.0  |
| 5            | 0                           | 0.0  | 0                         | 0.0  | 0              | 0.0  |
| 6            | 0                           | 0.0  | 0                         | 0.0  | 1              | 8.3  |
| 7            | 1                           | 6.7  | 1                         | 7.7  | 0              | 0.0  |
| 8            | 1                           | 6.7  | 1                         | 7.7  | 1              | 8.3  |
| 9            | 1                           | 6.7  | 1                         | 7.7  | 0              | 0.0  |
| 10           | 1                           | 6.7  | 2                         | 15.4 | 0              | 0.0  |
| 11           | 1                           | 6.7  | 1                         | 7.7  | 0              | 0.0  |
| 12           | 0                           | 0.0  | 0                         | 0.0  | 0              | 0.0  |
| 13           | 0                           | 0.0  | 0                         | 0.0  | 0              | 0.0  |
| 14           | 0                           | 0.0  | 0                         | 0.0  | 1              | 8.3  |
| <b>Total</b> | <b>15</b>                   |      | <b>13</b>                 |      | <b>12</b>      |      |



## Statistical note

In the table below there are two statistical terms readers may not be familiar with:

### i) Observed component years

This is the number of registered primary procedures multiplied by the number of years each component has been in place.

### ii) Rate/100 component years

This is equivalent to the yearly revision rate expressed as a percent and is derived by dividing the number of prostheses revised by the observed component years multiplied by 100. It therefore allows for the number of years of post-operative follow up in calculating the revision rate. These rates are usually very low, hence it is expressed per

100 component years rather than per component year. Statisticians consider that this is a more accurate way of deriving a revision rate for comparison when analysing data with widely varying follow-up times. It is also important to note the confidence intervals. The closer they are to the estimated revision rate/100 component years, the more precise the estimate is.

### Statistical Significance

Where it is stated that a difference among results is significant the p value is 0.05 or less. In most of these situations this is because there is no overlap of the confidence intervals (CIs) but sometimes significance can apply in the presence of CI overlap.

## All Primary Total Elbow Replacements

| No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| 587     | 3,819.1            | 42             | 1.10                     | 0.79                          | 1.49 |

## Revision Rate of Individual Prostheses Sorted in Alphabetic Order

| Prosthesis              | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |        |
|-------------------------|---------|--------------------|----------------|--------------------------|-------------------------------|--------|
| Acclaim                 | 16      | 154.1              | 6              | 3.89                     | 1.43                          | 8.47   |
| Coonrad/Morrey          | 347     | 2,651.2            | 17             | 0.64                     | 0.36                          | 1.00   |
| Evolve Stem             | 20      | 102.3              | 2              | 1.96                     | 0.00                          | 7.07   |
| Kudo                    | 18      | 166.7              | 4              | 2.40                     | 0.65                          | 6.14   |
| Latitude                | 96      | 577.8              | 11             | 1.90                     | 0.95                          | 3.41   |
| Mutars                  | 1       | 2.9                | 0              | 0.00                     | 0.00                          | 129.31 |
| Sorbie Questor          | 1       | 6.8                | 0              | 0.00                     | 0.00                          | 54.09  |
| Stanmore custom implant | 1       | 8.4                | 0              | 0.00                     | 0.00                          | 43.75  |
| Zimmer Nexel            | 87      | 148.87             | 2              | 1.34                     | 0.16                          | 4.85   |

## Revision vs Gender

| Gender  | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|---------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| Females | 450     | 3081.0             | 27             | 0.88                     | 0.58                          | 1.28 |
| Males   | 137     | 738.1              | 15             | 2.03                     | 1.14                          | 3.35 |


## Revision vs Age Bands

| Age Bands | No. Ops | Observed comp. Yrs | Number Revised | Rate/100 component-years | Exact 95% confidence interval |      |
|-----------|---------|--------------------|----------------|--------------------------|-------------------------------|------|
| <55       | 99      | 807.6              | 13             | 1.61                     | 0.81                          | 2.67 |
| 55-64     | 140     | 1,085.5            | 11             | 1.01                     | 0.47                          | 1.75 |
| 65-74     | 175     | 1,038.5            | 12             | 1.16                     | 0.60                          | 2.02 |
| >=75      | 173     | 887.4              | 6              | 0.68                     | 0.21                          | 1.39 |



KAPLAN MEIER CURVES

The following Kaplan Meier survival analyses are for the 19 years from 2000 to 2018, with deceased patients censored at time of death.



| Years | % Revision-free | Number |
|-------|-----------------|--------|
| 1     | 98.9            | 534    |
| 2     | 97.5            | 467    |
| 3     | 95.5            | 402    |
| 4     | 94.3            | 346    |
| 5     | 93.7            | 311    |
| 6     | 93.7            | 288    |
| 7     | 93.0            | 255    |
| 8     | 92.3            | 211    |
| 9     | 91.3            | 184    |
| 10    | 90.8            | 150    |
| 11    | 88.8            | 116    |
| 12    | 87.9            | 86     |
| 13    | 86.8            | 69     |

PATIENT BASED QUESTIONNAIRE  
OUTCOMES AT SIX-MONTHS POST SURGERY

Questionnaires at six months post-surgery

At six months post-surgery patients are sent an outcome questionnaire.

This was replaced by the validated Oxford Elbow score at the end of 2015.


There are 12 questions and each response is scores from 4-0 with 0 representing the greatest severity.

Total score range 0-48

For the 3 year period 2016 – 2018 there were n = 54 responses.

For the 2 year period 2016 – 2017  
there were n = 41 responses.

|         |       |
|---------|-------|
| Average | 32.26 |
| Maximum | 48    |
| Minimum | 8     |



# LUMBAR DISC REPLACEMENT

## PRIMARY LUMBAR DISC REPLACEMENT

This report analyses data for the **seventeen-year** period January 2002 – December 2018. There were 163 lumbar disc replacements registered, an additional 3 compared to last year.

### Data Analysis

The average age for a lumbar disc replacement was 40 years, with a range of 22 – 62 years.

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 74     | 89    |
| Percentage    | 45.40  | 54.60 |
| Mean age      | 40.57  | 39.75 |
| Maximum age   | 62.19  | 60.71 |
| Minimum age   | 24.07  | 22.25 |
| Standard dev. | 8.66   | 7.67  |

### Disc replacement levels

|       |     |
|-------|-----|
| L3/4  | 21  |
| L4/5  | 111 |
| L5/S1 | 33  |

### Fusion levels

|       |    |
|-------|----|
| L3/4  | 5  |
| L4/5  | 21 |
| L5/S1 | 37 |

### Previous operation

|            |    |
|------------|----|
| Discectomy | 29 |
| L3/4       | 0  |
| L4/5       | 11 |
| L5/S1      | 17 |

### Diagnosis

#### Degenerative disc disease

|       |    |
|-------|----|
| L3/4  | 12 |
| L4/5  | 61 |
| L5/S1 | 85 |

#### Annular tear MRI scan

|       |    |
|-------|----|
| L3/4  | 13 |
| L4/5  | 70 |
| L5/S1 | 26 |

#### Discogenic pain on discography

|       |    |
|-------|----|
| L3/4  | 20 |
| L4/5  | 86 |
| L5/S1 | 63 |

### Approach

|                         |     |
|-------------------------|-----|
| Retroperitoneal midline | 145 |
| Retroperitoneal lateral | 3   |
| Transperitoneal         | 2   |

### Intraoperative complications

|                       |    |
|-----------------------|----|
| Damage to major veins | 13 |
| Subsidence            | 1  |

### Systemic antibiotic prophylaxis

|  |     |
|--|-----|
| Patient number receiving systemic antibiotic prophylaxis | 135 |
|--|-----|

### Operating theatre

|              |    |
|--------------|----|
| Conventional | 90 |
| Laminar flow | 72 |
| Spacesuits   | 2  |

### Operative time (skin to skin)

|      |             |
|------|-------------|
| Mean | 139 minutes |
|------|-------------|

### Surgeon grade

|            |     |
|------------|-----|
| Consultant | 163 |
|------------|-----|



REVISION OF REGISTERED PRIMARY LUMBAR DISC REPLACEMENTS

This section analyses data for revisions of primary lumbar disc replacements for the seventeen-year period.

There were 3 revisions of the primary group of 163 lumbar disc replacements and 1 re-revision.

Time to revision

|         |            |
|---------|------------|
| Mean    | 1,841 days |
| Maximum | 4,528 days |
| Minimum | 242 days   |

Reason for revision

|                          |   |
|--------------------------|---|
| Pain                     | 2 |
| Loss of spinal alignment | 1 |

Oswestry Disability Index

There are 10 sections. For each section, the total score is 5: if the first statement is marked the score = 0; if the last statement is marked, the score = 5. Intervening statements are scored according to rank.

If more than one box is marked in each section, take the highest score.


If all 10 sections are completed, the score is calculated as follows:

Example: 16 (total scored)/50(total possible score) x 100 = 32%

Pre operative scores

|                           |       |
|---------------------------|-------|
| Oswestry Disability Index | n =21 |
| Average                   | 25    |





# CERVICAL DISC REPLACEMENT

This report analyses data for the **fifteen-year** period January 2004 – December 2018. There were 453 primary cervical disc replacements, an additional 60 from the previous year.

## Data Analysis

The average age for a cervical disc replacement was 45 years, with a range of 23 – 66 years.

|               | Female | Male  |
|---------------|--------|-------|
| Number        | 189    | 264   |
| Percentage    | 41.72  | 58.28 |
| Mean age      | 46.37  | 44.59 |
| Maximum age   | 65.79  | 68.29 |
| Minimum age   | 23.26  | 23.54 |
| Standard dev. | 8.08   | 8.86  |

## Disc replacement levels

|      |     |
|------|-----|
| C3/4 | 12  |
| C4/5 | 46  |
| C5/6 | 243 |
| C6/7 | 215 |
| C7T1 | 8   |

## Previous operation

|                                  |    |
|----------------------------------|----|
| Foraminotomy                     | 16 |
| Adjacent level fusion            | 22 |
| Adjacent level disc arthroplasty | 2  |

## Diagnosis

|                     |     |
|---------------------|-----|
| Acute disc prolapse | 308 |
| Chronic spondylosis | 43  |
| Neck pain           | 27  |

## Approach

|                |     |
|----------------|-----|
| Anterior right | 269 |
| Anterior left  | 83  |

## Intra operative complications

|                    |   |
|--------------------|---|
| Equipment failure  | 1 |
| Removal of implant | 1 |
| Tear jugular vein  | 1 |

## Systemic antibiotic prophylaxis

|  |     |
|--|-----|
| Patient number receiving systemic antibiotic prophylaxis | 393 |
|--|-----|

## Operating theatre

|              |     |
|--------------|-----|
| Conventional | 230 |
| Laminar flow | 217 |
| Spacesuits   | 1   |

## Operative time (skin to skin)

|      |             |
|------|-------------|
| Mean | 112 minutes |
|------|-------------|

## Surgeon grade

|                             |     |
|-----------------------------|-----|
| Consultant                  | 451 |
| Advanced trainee supervised | 2   |

## Revision Cervical disc replacement

There were 3 revisions registered.

## Neck Disability Index Scoring

There are 10 sections. For each section, the total score is 5: if the first statement is marked the score = 0; if the last statement is marked, the score = 5. Intervening statements are scored according to rank.

If more than one box is marked in each section, take the highest score.

If all 10 sections are completed, the score is calculated as follows:

### Example:

$$16 \text{ (total scored)} / 50 \text{ (total possible score)} \times 100 = 32\%$$

If one section is missed (or not applicable) the score is calculated:

### Example:

$$16 \text{ (total scored)} / 45 \text{ (total possible score)} \times 100 = 35.5\%$$

0 is the best score and 100 is the worst score.

## Post-operative score

|                       |       |
|-----------------------|-------|
| Neck Disability Index | 182   |
| Mean                  | 19.89 |

# RE-OPERATION WITHOUT REPLACEMENT OR REMOVAL OF ANY PROSTHETIC COMPONENTS

The re-operation form was introduced in December 2015.

For the period 2015 – 2018 there were 280 re-operations registered, 70 more than last year.

## Reason for Re-operation

|                        |    |
|------------------------|----|
| Deep infection         | 88 |
| Dislocation of joint   | 31 |
| Dislocation of bearing | 6  |
| Fracture               | 43 |
| Instability            | 5  |
| Malalignment           | 0  |
| Impingement            | 9  |
| Stiffness              | 53 |
| Haematoma evacuation   | 18 |
| Arthrofibrosis         | 3  |

## Procedure

|                                 |     |
|---------------------------------|-----|
| Open lavage                     | 105 |
| Arthroscopic lavage             | 6   |
| Closed reduction of dislocation | 20  |
| Open reduction of dislocation   | 9   |
| Fracture fixation               | 35  |
| Soft tissue procedure           | 19  |
| Ligament reconstruction         | 4   |
| Osteotomy                       | 2   |
| Bone debridement                | 14  |
| Arthrolysis                     | 3   |
| MUA                             | 52  |

| ASA | Number |
|-----|--------|
| 1   | 21     |
| 2   | 127    |
| 3   | 100    |
| 4   | 17     |

## Surgeon grade

|                               |     |
|-------------------------------|-----|
| Consultant                    | 214 |
| Advanced trainee supervised   | 18  |
| Advanced trainee unsupervised | 41  |
| Basic trainee                 | 5   |

## APPENDIX 1 - OXFORD 12 QUESTIONNAIRE REFERENCES

Murray, D.W et al, *The use of the Oxford hip and knee scores.* J Bone Joint Surg (Br) 2007; 89-B: 1010-14

*Questionnaire on the perceptions of patients about shoulder surgery* Jill Dawson, Ray Fitzpatrick, Andrew Carr. J Bone Joint Surg B. 1996 July; 78(4) 593-600

Kalairajah, Y et al, *Health outcome measures in the evaluation of total hip arthroplasties: a comparison between the Harris hip score and the Oxford hip score.* J Arthroplasty 2005; 20: 1037-41

### **Publications in Peer Reviewed Journals**

Murray, D.W et al, *The use of the Oxford hip and knee scores.* J Bone Joint Surg (Br) 2007; 89-B: 1010-14

*Questionnaire on the perceptions of patients about shoulder surgery.* Jill Dawson, Ray Fitzpatrick, Andrew Carr. J Bone Joint Surg B. 1996 July; 78(4) 593-600

Kalairajah, Y et al, *Health outcome measures in the evaluation of total hip arthroplasties: a comparison between the Harris hip score and the Oxford hip score.* J Arthroplasty 2005; 20: 1037-41

## APPENDIX 3 - DATA FORMS

## Data Forms

**DO NOT PLACE IN PATIENT NOTES**

**TO BE RETAINED IN THEATRE SUITE**

|  |  |  |  |
|--|--|--|--|
| <p align="center"><b>NEW ZEALAND JOINT REGISTRY</b></p> <p align="center"><b>Primary Replacement Hip</b></p> <p>Free Phone 0800-274-989      Total Hip Arthroplasty <input type="checkbox"/>      Resurfacing Arthroplasty <input type="checkbox"/></p> <p>31.05.2010</p>  |  |  |  |
| <p>Date: .....</p> <p>BMI:.....</p> <p>Side:..... **</p>   | <p>Patient Name: .....</p> <p>Address: .....</p> <p>d.o.b. .... NHI: .....</p> <p align="center"><b>Attach Patient Label</b></p> |  | <p>Consultant: .....</p> <p>[If different from patient label]</p> <p>hospital: .....</p> <p align="center">Town/City</p> |
| <p><i>Tick Appropriate Boxes</i></p>   |  |  |  |
| <p><b>PREVIOUS OPERATION ON INDEX JOINT</b></p> <p><input type="checkbox"/> None      <input type="checkbox"/> Arthrodesis</p> <p><input type="checkbox"/> Internal fixation for juxtarticular fractures      <input type="checkbox"/> Other: .....</p> <p><input type="checkbox"/> Osteotomy      .....</p>   |  |  |  |
| <p><b>DIAGNOSIS</b></p> <p><input type="checkbox"/> Osteoarthritis      <input type="checkbox"/> Old fracture NOF</p> <p><input type="checkbox"/> Rheumatoid arthritis      <input type="checkbox"/> Post-acute dislocation</p> <p><input type="checkbox"/> Other inflammatory      <input type="checkbox"/> Avascular necrosis</p> <p><input type="checkbox"/> Acute fracture NOF      <input type="checkbox"/> Tumour</p> <p><input type="checkbox"/> Developmental dysplasia/dislocation      <input type="checkbox"/> Other: Name: .....</p> |  |  |  |
| <p><b>APPROACH</b>      <input type="checkbox"/> Image guided surgery      <input type="checkbox"/> Minimally invasive surgery</p> <p><input type="checkbox"/> Anterior      <input type="checkbox"/> Posterior      <input type="checkbox"/> Lateral      <input type="checkbox"/> Trochanteric</p> <p>osteotomy</p>  |  |  |  |
| <p><b>FEMUR</b></p> <p align="center">Please do not fold<br/>bar-coded label</p>   |  | <p><b>ACETABULUM</b></p> <p align="center">Please do not fold<br/>bar-coded label</p>  |  |
| <p align="center"><b>STICK EXTRA LABELS ON REVERSE SIDE</b></p>  |  |  |  |
| <p><b>BONE GRAFT - FEMUR</b></p> <p><input type="checkbox"/> Allograft      <input type="checkbox"/> Synthetic</p> <p><input type="checkbox"/> Autograft</p>   |  | <p><b>BONE GRAFT - ACETABULUM</b></p> <p><input type="checkbox"/> Allograft      <input type="checkbox"/></p> <p><input type="checkbox"/> Autograft      <input type="checkbox"/></p> <p>Synthetic</p> |  |
| <p><b>FEMORAL HEAD</b></p> <p align="center">Please do not fold<br/>bar-coded label</p>  |  | <p><b>AUGMENTS</b></p> <p align="center">Please do not fold<br/>bar-coded label</p>  |  |
| <p align="center"><b>STICK EXTRA LABELS ON REVERSE SIDE</b></p>  |  |  |  |
| <p><b>CEMENT</b></p> <p><input type="checkbox"/> Femur      <input type="checkbox"/> Acetabulum      <input type="checkbox"/> Antibiotic brand: .....</p>  |  |  |  |
| <p><input type="checkbox"/> <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b></p> <p>Name: ..... ASA Class:    1    2    3    4    (please circle one)</p>  |  |  |  |
| <p><b>OPERATING THEATRE</b></p> <p><input type="checkbox"/> Conventional      <input type="checkbox"/> Laminar flow or similar      <input type="checkbox"/> Space suits</p>   |  |  |  |
| <p><b>SKIN TO SKIN TIME mins</b>      Start skin.....      Finish skin.....</p>  |  |  |  |
| <p><b>PRIMARY OPERATING SURGEON</b></p> <p><input type="checkbox"/> Consultant      <input type="checkbox"/> Adv Trainee Unsupervised      Year.....      <input type="checkbox"/> Basic Trainee</p> <p><input type="checkbox"/> Adv Trainee Supervised</p>  |  |  |  |

**\*\*NB**

*If bilateral procedure two completed forms are required*

DO NOT PLACE IN PATIENT NOTES TO BE RETAINED IN THEATRE SUITE

| <b>NEW ZEALAND JOINT REGISTRY</b><br><b>Revision Hip Joint</b>  |  |   |  |
|---|--|---|--|
| Free Phone 0800-274-989<br>07.04.2005   |  |   |  |
| Date: .....<br>Side: ..... **   | <div style="border: 1px solid black; padding: 5px;">             Patient Name: .....<br/>             Address: .....<br/> <br/>             d.o.b. .... NHI: .....<br/> <b>Attach Patient Label</b> </div> | Consultant: .....<br>[on patient label]<br>Hospital: .....<br>Town/City: .....  |  |
| <b>Tick Appropriate Boxes</b>   |  |   |  |
| <b>REASON FOR REVISION</b><br><input type="checkbox"/> Loosening acetabular component<br><input type="checkbox"/> Loosening femoral component<br><input type="checkbox"/> Dislocation<br><input type="checkbox"/> Pain  |  | <input type="checkbox"/> Previous hemiarthroplasty<br><input type="checkbox"/> Deep infection<br><input type="checkbox"/> Fracture femur<br><input type="checkbox"/> Removal of components<br><input type="checkbox"/> Other: Name: ..... |  |
| Date Index Operation: .....<br><b>REVISION</b><br><input type="checkbox"/> Change of femoral component<br><input type="checkbox"/> Change of acetabular component<br><input type="checkbox"/> Change of head  |  | If re-revision - Date previous revision: .....<br><input type="checkbox"/> Change of liner<br><input type="checkbox"/> Change of all components   |  |
| <b>APPROACH</b> <input type="checkbox"/> Image guided surgery <input type="checkbox"/> Minimally invasive surgery<br><input type="checkbox"/> Anterior <input type="checkbox"/> Posterior <input type="checkbox"/> Lateral <input type="checkbox"/> Trochanteric<br>osteotomy |  |   |  |
| <b>FEMUR</b><br><div style="border: 1px solid black; height: 60px; text-align: center; padding: 10px;"> <b>Please do not fold<br/>bar-coded label</b> </div>  |  | <b>ACETABULUM</b><br><div style="border: 1px solid black; height: 60px; text-align: center; padding: 10px;"> <b>Please do not fold<br/>bar-coded label</b> </div>   |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |  |   |  |
| <b>BONE GRAFT - FEMUR</b><br><input type="checkbox"/> Allograft <input type="checkbox"/> Synthetic<br><input type="checkbox"/> Autograft  |  | <b>BONE GRAFT - ACETABULUM</b><br><input type="checkbox"/> Allograft <input type="checkbox"/> Synthetic<br><input type="checkbox"/> Autograft   |  |
| <b>FEMORAL HEAD</b><br><div style="border: 1px solid black; height: 60px; text-align: center; padding: 10px;"> <b>Please do not fold<br/>bar-coded label</b> </div>   |  | <b>AUGMENTS</b><br><div style="border: 1px solid black; height: 60px; text-align: center; padding: 10px;"> <b>Please do not fold<br/>bar-coded label</b> </div>   |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |  |   |  |
| <b>CEMENT</b><br><input type="checkbox"/> Femur <input type="checkbox"/> Acetabulum <input type="checkbox"/> Antibiotic brand: .....  |  |   |  |
| <input type="checkbox"/> <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b><br>Name ..... ASA Class:    1    2    3    4    (please circle one)   |  |   |  |
| <b>OPERATING THEATRE</b><br><input type="checkbox"/> Conventional <input type="checkbox"/> Laminar flow or similar <input type="checkbox"/> Space suits   |  |   |  |
| <b>SKIN TO SKIN TIME mins</b> Start skin.....    Finish skin.....   |  |   |  |
| <b>PRIMARY OPERATING SURGEON</b><br><input type="checkbox"/> Consultant <input type="checkbox"/> Adv Trainee Supervised    Year..... <input type="checkbox"/> Basic Trainee   |  |   |  |

\*\*NB

If bilateral procedure two completed forms are required

DO NOT PLACE IN PATIENT NOTES TO BE RETAINED IN THEATRE SUITE

| <b>NEW ZEALAND JOINT REGISTRY</b><br><b>Primary Replacement Knee</b>   |  |   |  |
|--|--|---|--|
| Free Phone 0800-274-989 <input type="checkbox"/> Total Knee Arthroplasty <input type="checkbox"/> Unicompartmental <input type="checkbox"/> Patellofemoral<br>31.05.2010   |  |   |  |
| Date: .....<br>BMI:.....<br>Side:..... **  | Patient Name:<br>Address:<br><br>d.o.b.                      NHI:<br><div style="text-align: center; border: 1px solid black; padding: 2px;"><b>Attach Patient Label</b></div> | Consultant: .....<br>[If different from patient label]<br>Hospital: .....<br>Town/City:.....  |  |
| <b>Tick Appropriate Boxes</b>  |  |   |  |
| <b>PREVIOUS OPERATION ON INDEX JOINT</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> <input type="checkbox"/> None<br/> <input type="checkbox"/> Internal fixation for juxtarticular fracture<br/> <input type="checkbox"/> Ligament reconstruction<br/> <input type="checkbox"/> Meniscectomy             </div> <div style="width: 48%;"> <input type="checkbox"/> Synovectomy<br/> <input type="checkbox"/> Osteotomy<br/> <input type="checkbox"/> Other: Name: .....             </div> </div>                              |  |   |  |
| <b>DIAGNOSIS</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> <input type="checkbox"/> Osteoarthritis<br/> <input type="checkbox"/> Rheumatoid arthritis<br/> <input type="checkbox"/> Other inflammatory<br/> <input type="checkbox"/> Tumour             </div> <div style="width: 48%;"> <input type="checkbox"/> Post fracture<br/> <input type="checkbox"/> Post ligament disrupt/reconstruction<br/> <input type="checkbox"/> Avascular necrosis<br/> <input type="checkbox"/> Other: Name: .....             </div> </div> |  |   |  |
| <b>APPROACH</b> <input type="checkbox"/> Image guided surgery <input type="checkbox"/> Minimally invasive surgery<br><input type="checkbox"/> Medial parapatellar <input type="checkbox"/> Lateral parapatellar <input type="checkbox"/> Other   |  |   |  |
| <b>FEMUR</b><br><div style="border: 1px solid black; height: 40px; display: flex; align-items: center; justify-content: center; margin-top: 10px;"> <b>Please do not fold<br/>bar-coded label</b> </div>   |  | <b>TIBIA</b><br><div style="border: 1px solid black; height: 40px; display: flex; align-items: center; justify-content: center; margin-top: 10px;"> <b>Please do not fold<br/>bar-coded label</b> </div>    |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>  |  |   |  |
| <b>BONE GRAFT - FEMUR</b><br><input type="checkbox"/> Allograft <input type="checkbox"/> Synthetic<br><input type="checkbox"/> Autograft   |  | <b>BONE GRAFT - TIBIA</b><br><input type="checkbox"/> Allograft <input type="checkbox"/><br><input type="checkbox"/> Autograft <input type="checkbox"/> Synthetic   |  |
| <b>PATELLA</b><br><div style="border: 1px solid black; height: 40px; display: flex; align-items: center; justify-content: center; margin-top: 10px;"> <b>Please do not fold<br/>bar-coded label</b> </div>   |  | <b>AUGMENTS</b><br><div style="border: 1px solid black; height: 40px; display: flex; align-items: center; justify-content: center; margin-top: 10px;"> <b>Please do not fold<br/>bar-coded label</b> </div> |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>  |  |   |  |
| <b>CEMENT</b><br><input type="checkbox"/> Femur <input type="checkbox"/> Tibia <input type="checkbox"/> Patella <input type="checkbox"/> Antibiotic brand: .....   |  |   |  |
| <input type="checkbox"/> <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b><br>Name .....                      ASA Class:    1    2    3    4    (please circle one)   |  |   |  |
| <b>OPERATING THEATRE</b><br><input type="checkbox"/> Conventional <input type="checkbox"/> Laminar flow or similar <input type="checkbox"/> Space suits  |  |   |  |
| <b>SKIN TO SKIN TIME mins</b> Start skin.....                      Finish skin.....  |  |   |  |
| <b>PRIMARY OPERATING SURGEON</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 45%;"> <input type="checkbox"/> Consultant                      <input type="checkbox"/> Adv Trainee Unsupervised<br/> <input type="checkbox"/> Trainee                      <input type="checkbox"/> Adv Trainee Supervised             </div> <div style="width: 45%;">               Year.....                      <input type="checkbox"/> Basic             </div> </div>  |  |   |  |

\*\*NB    If bilateral procedure two completed forms are required

DO NOT PLACE IN PATIENT NOTES

TO BE RETAINED IN THEATRE SUITE

| <b>NEW ZEALAND JOINT REGISTRY</b><br><b>Revision Knee Joint</b>   |   |  |   |
|---|---|--|---|
| Free Phone 0800-274-989<br>07.04.2005   |   |  |   |
| Date: .....<br><b>Side:..... **</b>   | <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="padding: 2px;">               Patient Name:<br/>               Address:<br/>               d.o.b.                      NHI:<br/> <b>Attach Patient Label</b> </td> <td style="padding: 2px;"> <b>Consultant: .....</b><br/>               [If different from patient label]<br/> <b>Hospital: .....</b><br/> <b>Town/City:.....</b> </td> </tr> </table> | Patient Name:<br>Address:<br>d.o.b.                      NHI:<br><b>Attach Patient Label</b> | <b>Consultant: .....</b><br>[If different from patient label]<br><b>Hospital: .....</b><br><b>Town/City:.....</b> |
| Patient Name:<br>Address:<br>d.o.b.                      NHI:<br><b>Attach Patient Label</b>  | <b>Consultant: .....</b><br>[If different from patient label]<br><b>Hospital: .....</b><br><b>Town/City:.....</b>   |  |   |
| <i>Tick Appropriate Boxes</i>   |   |  |   |
| <b>REASON FOR REVISION</b><br><input type="checkbox"/> Loosening femoral component<br><input type="checkbox"/> Loosening tibial component<br><input type="checkbox"/> Loosening patellar component<br><input type="checkbox"/> Pain   | <input type="checkbox"/> Previous Unicompartmental<br><input type="checkbox"/> Deep infection<br><input type="checkbox"/> Fracture femur<br><input type="checkbox"/> Fracture tibia<br><input type="checkbox"/> Other details: .....  |  |   |
| Date Index Operation: .....<br><b>REVISION</b><br><input type="checkbox"/> Change of femoral component<br><input type="checkbox"/> Change of tibial component<br><input type="checkbox"/> Change of patellar component<br><input type="checkbox"/> Addition of patellar component | If re-revision - Date previous revision: .....<br><input type="checkbox"/> Change of tibial polyethylene only<br><input type="checkbox"/> Change of all components<br><input type="checkbox"/> Removal of components<br><input type="checkbox"/> Other  |  |   |
| <b>APPROACH</b> <input type="checkbox"/> Image guided surgery <input type="checkbox"/> Minimally invasive surgery<br><input type="checkbox"/> Medial parapatellar <input type="checkbox"/> Lateral parapatellar <input type="checkbox"/> Other                                    |   |  |   |
| <b>FEMUR</b><br><div style="border: 1px solid black; height: 40px; margin-top: 10px; text-align: center; padding: 5px;"> <b>Please do not fold<br/>bar-coded label</b> </div>   | <b>TIBIA</b><br><div style="border: 1px solid black; height: 40px; margin-top: 10px; text-align: center; padding: 5px;"> <b>Please do not fold<br/>bar-coded label</b> </div>   |  |   |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |   |  |   |
| <b>BONE GRAFT - FEMUR</b><br><input type="checkbox"/> Allograft <input type="checkbox"/> Synthetic<br><input type="checkbox"/> Autograft  | <b>BONE GRAFT - TIBIA</b><br><input type="checkbox"/> Allograft <input type="checkbox"/> Synthetic<br><input type="checkbox"/> Autograft  |  |   |
| <b>PATELLA</b><br><div style="border: 1px solid black; height: 40px; margin-top: 10px; text-align: center; padding: 5px;"> <b>Please do not fold<br/>bar-coded label</b> </div>   | <b>AUGMENTS</b><br><div style="border: 1px solid black; height: 40px; margin-top: 10px; text-align: center; padding: 5px;"> <b>Please do not fold<br/>bar-coded label</b> </div>  |  |   |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |   |  |   |
| <b>CEMENT</b><br><input type="checkbox"/> Femur <input type="checkbox"/> Tibia <input type="checkbox"/> Patella <input type="checkbox"/> Antibiotic brand: .....  |   |  |   |
| <input type="checkbox"/> <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b><br>Name .....                      ASA Class:    1    2    3    4    (please circle one)  |   |  |   |
| <b>OPERATING THEATRE</b><br><input type="checkbox"/> Conventional <input type="checkbox"/> Laminar flow or similar <input type="checkbox"/> Space suits   |   |  |   |
| <b>SKIN TO SKIN TIME mins</b> Start skin.....                      Finish skin.....   |   |  |   |
| <b>PRIMARY OPERATING SURGEON</b><br><input type="checkbox"/> Consultant <input type="checkbox"/> Adv Trainee Unsupervised <input type="checkbox"/> Adv Trainee Supervised    Year..... <input type="checkbox"/> Basic Trainee   |   |  |   |

**\*\*NB**    *If bilateral procedure two completed forms are required*



| <b>NEW ZEALAND JOINT REGISTRY</b><br><b>Primary Replacement Shoulder</b><br><input type="checkbox"/> Total shoulder Arthroplasty <input type="checkbox"/> Hemiarthroplasty <input type="checkbox"/> Reverse Shoulder |  |   |  |
|--|--|---|--|
| 24.03.2016   |  |   |  |
| <b>Date:</b> .....   |  | <b>Consultant:</b> .....<br>[If different from patient label]   |  |
| <b>BMI:</b> .....  |  | <div style="border: 1px solid black; padding: 10px; margin: 0 auto; width: 80%;">           Patient Name:<br/>           Address:<br/> <br/>           d.o.b.                      NHI:<br/> <b>Attach Patient Label</b> </div>                 |  |
| <b>Hospital:</b> .....   |  |   |  |
| <b>Side:</b> ..... **  |  |   |  |
| <b>Town/City:</b>  |  |   |  |
| <b>Tick Appropriate Boxes</b>  |  |   |  |
| <b>PREVIOUS OPERATION ON INDEX JOINT</b>   |  |   |  |
| <input type="checkbox"/> None<br><input type="checkbox"/> Internal fixation for juxtarticular fracture<br><input type="checkbox"/> Previous stabilisation<br><input type="checkbox"/> Rotator Cuff Repair            |  | <input type="checkbox"/> Osteotomy<br><input type="checkbox"/> Arthrodesis<br><input type="checkbox"/> Arthroscopic debridement/compression<br><input type="checkbox"/> Other: Name: .....  |  |
| <b>DIAGNOSIS</b>   |  |   |  |
| <input type="checkbox"/> Rheumatoid arthritis<br><input type="checkbox"/> Osteoarthritis<br><input type="checkbox"/> Other inflammatory<br><input type="checkbox"/> Acute fracture proximal humerus<br>.....         |  | <input type="checkbox"/> Post recurrent dislocation<br><input type="checkbox"/> Avascular necrosis<br><input type="checkbox"/> Cuff tear arthropathy<br><input type="checkbox"/> Post old trauma<br><input type="checkbox"/> Other: Name: ..... |  |
| <b>APPROACH</b>  |  |   |  |
| <input type="checkbox"/> Deltopectoral   |  | <input type="checkbox"/> Other : specify  |  |
| <b>HUMERUS</b>   |  | <b>GLENOID</b>  |  |
| <div style="border: 1px solid black; padding: 20px; min-height: 100px;"> <b>Please do not fold<br/>bar-coded label</b> </div>  |  | <div style="border: 1px solid black; padding: 20px; min-height: 100px;"> <b>Please do not fold<br/>bar-coded label</b> </div>   |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>  |  |   |  |
| <b>BONE GRAFT - HUMERUS</b>  |  | <b>BONE GRAFT - GLENOID</b>   |  |
| <input type="checkbox"/> Allograft <input type="checkbox"/> Synthetic<br><input type="checkbox"/> Autograft  |  | <input type="checkbox"/> Allograft <input type="checkbox"/> Synthetic<br><input type="checkbox"/> Autograft   |  |
| <b>HUMERAL HEAD</b>  |  | <b>AUGMENTS</b>   |  |
| <div style="border: 1px solid black; padding: 20px; min-height: 100px;"> <b>Please do not fold<br/>bar-coded label</b> </div>  |  | <div style="border: 1px solid black; padding: 20px; min-height: 100px;"> <b>Please do not fold<br/>bar-coded label</b> </div>   |  |
| <b>STICK ALL LABELS ON REVERSE SIDE</b>  |  |   |  |
| <b>CEMENT</b>  |  |   |  |
| <input type="checkbox"/> Humerus <input type="checkbox"/> Glenoid  |  | <input type="checkbox"/> Antibiotic brand: .....  |  |
| <input type="checkbox"/> <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b>  |  |   |  |
| <b>Name:</b> .....   |  | <b>ASA Class:</b> 1    2    3    4    (please circle one)   |  |
| <b>OPERATING THEATRE</b>   |  |   |  |
| <input type="checkbox"/> Conventional  |  | <input type="checkbox"/> Laminar flow or similar <input type="checkbox"/> Space suits   |  |
| <b>SKIN TO SKIN TIME mins</b> <b>Start skin</b> ..... <b>Finish skin</b> .....   |  |   |  |
| <b>**NB</b> <i>If bilateral procedure two completed forms are required</i>   |  |   |  |

# APPENDIX 3 - DATA FORMS

DO NOT PLACE IN PATIENT NOTES TO BE RETAINED IN THEATRE SUITE

| <b>NEW ZEALAND JOINT REGISTRY</b><br><b>Revision Shoulder</b>   |  |  |  |
|---|--|--|--|
| <b>Free Phone 0800-274-989</b><br><b>07.04.2005</b>   |  |  |  |
| <b>Date:</b> .....<br><b>Side:</b> ..... **   | <div style="border: 1px solid black; padding: 5px;"> <b>Patient Name:</b><br/> <b>Address:</b><br/><br/>                 d.o.b.                      NHI:             </div> | <b>Consultant:</b> .....<br>[If different from patient label]<br><b>Hospital:</b> .....<br><b>Town/City:</b> .....   |  |
| <b>Tick Appropriate Boxes</b> <b>Attach Patient Label</b>   |  |  |  |
| <b>REASON FOR REVISION</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> <input type="checkbox"/> Loosening glenoid component<br/> <input type="checkbox"/> Loosening humeral component<br/> <input type="checkbox"/> Loosening both components<br/> <input type="checkbox"/> Dislocation/instability anterior<br/> <input type="checkbox"/> Instability posterior                 </div> <div style="width: 48%;"> <input type="checkbox"/> Subacromial tuberosity impingement<br/> <input type="checkbox"/> Subacromial cuff impingement/tear<br/> <input type="checkbox"/> Fracture humerus<br/> <input type="checkbox"/> Deep infection<br/> <input type="checkbox"/> Pain<br/> <input type="checkbox"/> Other: Name: .....                 </div> </div> |  |  |  |
| <b>Date Index Operation:</b> ..... <b>If re-revision - Date previous revision:</b> .....  |  |  |  |
| <b>REVISION</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> <input type="checkbox"/> Change of head only<br/> <input type="checkbox"/> Change of humeral component<br/> <input type="checkbox"/> Change of glenoid component<br/> <input type="checkbox"/> Change of liner (glenoid non cemented)                 </div> <div style="width: 48%;"> <input type="checkbox"/> Change of all components<br/> <input type="checkbox"/> Remove glenoid<br/> <input type="checkbox"/> Remove humerus<br/> <input type="checkbox"/> Removal of components<br/> <input type="checkbox"/> Other Specify: .....                 </div> </div>   |  |  |  |
| <b>APPROACH</b><br><input type="checkbox"/> Deltopectoral <input type="checkbox"/> Other: specify   |  |  |  |
| <b>HUMERUS</b><br><div style="border: 1px solid black; height: 40px; text-align: center; margin-top: 10px;">                 Please do not fold<br/>bar-coded labels             </div>   |  | <b>GLENOID</b><br><div style="border: 1px solid black; height: 40px; text-align: center; margin-top: 10px;">                 Please do not fold<br/>bar-coded labels             </div>  |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |  |  |  |
| <b>BONE GRAFT - HUMERUS</b><br><input type="checkbox"/> Allograft <input type="checkbox"/> Synthetic<br><input type="checkbox"/> Autograft  |  | <b>BONE GRAFT - GLENOID</b><br><input type="checkbox"/> Allograft <input type="checkbox"/> Synthetic<br><input type="checkbox"/> Autograft   |  |
| <b>HUMERAL HEAD</b><br><div style="border: 1px solid black; height: 40px; text-align: center; margin-top: 10px;">                 Please do not fold<br/>bar-coded labels             </div>  |  | <b>AUGMENTS</b><br><div style="border: 1px solid black; height: 40px; text-align: center; margin-top: 10px;">                 Please do not fold<br/>bar-coded labels             </div> |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |  |  |  |
| <b>CEMENT</b><br><input type="checkbox"/> Humerus <input type="checkbox"/> Glenoid <input type="checkbox"/> Antibiotic brand: .....   |  |  |  |
| <input type="checkbox"/> <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b><br>Name .....                      ASA Class:    1    2    3    4    (please circle one)  |  |  |  |
| <b>OPERATING THEATRE</b><br><input type="checkbox"/> Conventional <input type="checkbox"/> Laminar flow or similar <input type="checkbox"/> Space suits   |  |  |  |
| <b>SKIN TO SKIN TIME mins</b> <b>Start skin:</b> ..... <b>Finish skin:</b> .....  |  |  |  |
| <b>PRIMARY OPERATING SURGEON</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> <input type="checkbox"/> Adv Trainee Unsupervised<br/> <input type="checkbox"/> Supervised Year:.....                 </div> <div style="width: 48%;"> <input type="checkbox"/> Consultant                      <input type="checkbox"/> Adv Trainee<br/> <input type="checkbox"/> Basic Trainee                 </div> </div>   |  |  |  |

**\*\*NB**                      *If bilateral procedure two completed forms are required*

| <b>NEW ZEALAND JOINT REGISTRY</b><br><b>Primary Replacement Ankle</b>   |   |  |  |
|---|---|--|--|
| <b>Free Phone 0800-274-989</b><br><b>31.05.2010</b>   |   |  |  |
| <b>Date:</b> .....  | <div style="border: 1px solid black; padding: 5px;"> <b>Patient Name:</b><br/> <b>Address:</b><br/><br/> <b>d.o.b.</b> ..... <b>NHI:</b> .....         </div> | <b>Consultant:</b> .....<br><b>[If different from patient label]</b><br><b>Hospital:</b> .....<br><br><b>Town/City:</b> .....  |  |
| <b>BMI:</b> .....<br><br><b>Side:</b> ..... **  | <b>Tick Appropriate Boxes</b>   |  |  |
| <b>PREVIOUS OPERATION ON INDEX JOINT</b>  |   |  |  |
| <input type="checkbox"/> <b>None</b><br><input type="checkbox"/> <b>Internal fixation for juxtaarticular fractures</b><br><input type="checkbox"/> <b>Osteotomy</b>               |   | <input type="checkbox"/> <b>Arthrodesis</b><br><input type="checkbox"/> <b>Other: Name:</b> .....  |  |
| <b>DIAGNOSIS</b>  |   |  |  |
| <input type="checkbox"/> <b>Osteoarthritis</b><br><input type="checkbox"/> <b>Rheumatoid arthritis</b><br><input type="checkbox"/> <b>Other inflammatory</b><br>.....             |   | <input type="checkbox"/> <b>Post trauma</b><br><input type="checkbox"/> <b>Avascular necrosis talus</b><br><input type="checkbox"/> <b>Other: Name:</b> .....                  |  |
| <b>APPROACH</b>   |   |  |  |
| <input type="checkbox"/> <b>Anterior</b>  |   | <input type="checkbox"/> <b>Anterio-lateral</b>  |  |
| <input type="checkbox"/> <b>Other</b>   |   | <input type="checkbox"/> <b>Other</b>  |  |
| <b>TIBIA</b><br><div style="border: 1px solid black; height: 60px; margin-top: 10px; text-align: center; padding: 10px;"> <b>Please do not fold<br/>bar-coded label</b> </div>    |   | <b>TALUS</b><br><div style="border: 1px solid black; height: 60px; margin-top: 10px; text-align: center; padding: 10px;"> <b>Please do not fold<br/>bar-coded label</b> </div> |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |   |  |  |
| <b>BONE GRAFT - TIBIA</b><br><input type="checkbox"/> <b>Allograft</b><br><input type="checkbox"/> <b>Autograft</b> <input type="checkbox"/> <b>Synthetic</b>                     |   | <b>BONE GRAFT - TALUS</b><br><input type="checkbox"/> <b>Allograft</b><br><input type="checkbox"/> <b>Autograft</b> <input type="checkbox"/> <b>Synthetic</b>                  |  |
| <b>AUGMENTS</b><br><div style="border: 1px solid black; height: 60px; margin-top: 10px; text-align: center; padding: 10px;"> <b>Please do not fold<br/>bar-coded label</b> </div> |   | <b>FUSION DISTAL TFJ</b>   |  |
| <b>STICK ALL LABELS ON REVERSE SIDE</b>   |   |  |  |
| <b>CEMENT</b>   |   |  |  |
| <input type="checkbox"/> <b>Tibia</b>   |   | <input type="checkbox"/> <b>Talus</b>  |  |
| <input type="checkbox"/> <b>Antibiotic Brand:</b> .....   |   | <input type="checkbox"/> <b>Antibiotic Brand:</b> .....  |  |
| <input type="checkbox"/> <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b>   |   |  |  |
| <b>Name:</b> .....  |   | <b>ASA Class:</b> 1   2   3   4 (please circle one)  |  |
| <b>OPERATING THEATRE</b>  |   |  |  |
| <input type="checkbox"/> <b>Conventional</b>  |   | <input type="checkbox"/> <b>Laminar flow or similar</b>  |  |
| <input type="checkbox"/> <b>Space suits</b>   |   | <input type="checkbox"/> <b>Space suits</b>  |  |
| <b>SKIN TO SKIN TIME mins</b> <b>Start skin:</b> ..... <b>Finish skin:</b> .....  |   | <b>SKIN TO SKIN TIME mins</b> <b>Start skin:</b> ..... <b>Finish skin:</b> .....   |  |
| <b>PRIMARY OPERATING SURGEON</b>  |   |  |  |
| <input type="checkbox"/> <b>Consultant</b>  |   | <input type="checkbox"/> <b>Adv Trainee Unsupervised</b>   |  |
| <input type="checkbox"/> <b>Trainee</b>   |   | <input type="checkbox"/> <b>Adv Trainee Supervised</b> <b>Year:</b> ..... <input type="checkbox"/> <b>Basic</b>  |  |

P.163

DO NOT PLACE IN PATIENT NOTES

TO BE RETAINED IN THEATRE SUITE

| <b>NEW ZEALAND JOINT REGISTRY</b><br><b>Revision Ankle Joint</b>   |   |  |  |
|--|---|--|--|
| Free Phone 0800-274-989  |   | 07.04.2005   |  |
| Date: .....<br><br>Side:..... **   | <div style="border: 1px solid black; padding: 5px;">             Patient Name:<br/>             Address:<br/><br/>             d.o.b.                      NHI:<br/> <b>Attach Patient Label</b> </div> |  | Consultant: .....<br>[If different from patient label]<br><br>Hospital:.....<br><br>Town/City: ..... |
| <i>Tick Appropriate Boxes</i>  |   |  |  |
| <b>REASON FOR REVISION</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> <input type="checkbox"/> Loosening talar component<br/> <input type="checkbox"/> Loosening tibial component<br/> <input type="checkbox"/> Dislocation<br/> <input type="checkbox"/> Pain             </div> <div style="width: 48%;"> <input type="checkbox"/> Deep infection<br/> <input type="checkbox"/> Fracture talus<br/> <input type="checkbox"/> Fracture tibia<br/> <input type="checkbox"/> Dislocations<br/> <input type="checkbox"/> Other details: .....             </div> </div> |   |  |  |
| Date Index Operation: .....  |   | If re-revision - Date previous revision: .....   |  |
| <b>REVISION</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> <input type="checkbox"/> Change of talar component<br/> <input type="checkbox"/> Change of tibial component<br/> <input type="checkbox"/> Change of polyethylene only             </div> <div style="width: 48%;"> <input type="checkbox"/> Change of all components<br/> <input type="checkbox"/> Removal of components<br/> <input type="checkbox"/> Other Name: .....             </div> </div>   |   |  |  |
| <b>APPROACH</b><br><input type="checkbox"/> Anterior <input type="checkbox"/> Anterio-lateral <input type="checkbox"/> Posterior   |   |  |  |
| <b>TIBIA</b><br><br><div style="border: 1px solid black; padding: 10px; text-align: center;"> <b>Please do not fold<br/>bar-coded label</b> </div>   |   | <b>TALUS</b><br><br><div style="border: 1px solid black; padding: 10px; text-align: center;"> <b>Please do not fold<br/>bar-coded label</b> </div> |  |
| <b>STICK ALL LABELS ON REVERSE SIDE</b>  |   |  |  |
| <b>BONE GRAFT - TIBIA</b><br><input type="checkbox"/> Allograft <input type="checkbox"/> Autograft <input type="checkbox"/> Synthetic  |   | <b>BONE GRAFT - TALUS</b><br><input type="checkbox"/> Allograft <input type="checkbox"/> Autograft <input type="checkbox"/> Synthetic              |  |
| <b>AUGUMENTS</b><br><br><div style="border: 1px solid black; padding: 10px; text-align: center;"> <b>Please do not fold<br/>bar-coded label</b> </div>   |   | <b>FUSION DISTAL TFJ</b><br><br>Yes <input type="checkbox"/> No <input type="checkbox"/>   |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>  |   |  |  |
| <b>CEMENT</b><br><br><input type="checkbox"/> Talus <input type="checkbox"/> Tibia <input type="checkbox"/> Antibiotic brand: .....  |   |  |  |
| <input type="checkbox"/> <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b><br><br>Name .....                      ASA Class:    1    2    3    4    (please circle one)   |   |  |  |
| <b>OPERATING THEATRE</b><br><input type="checkbox"/> Conventional <input type="checkbox"/> Laminar flow or similar <input type="checkbox"/> Space suits  |   |  |  |
| <b>SKIN TO SKIN TIME mins</b> Start skin.....                      Finish skin.....  |   |  |  |
| <b>PRIMARY OPERATING SURGEON</b><br><input type="checkbox"/> Consultant <input type="checkbox"/> Adv Trainee Supervised Year..... <input type="checkbox"/> Basic Trainee   |   |  |  |

**\*\*NB** If bilateral procedure two completed forms are required

DO NOT PLACE IN PATIENT NOTES TO BE RETAINED IN THEATRE SUITE

| <b>NEW ZEALAND JOINT REGISTRY</b><br><b>Primary Replacement Elbow</b>   |  |   |  |
|---|--|---|--|
|   |  |   | <b>Free Phone 0800-274-989</b><br>07.04.2005   |
| <b>Date:</b> .....  |  |   |  |
| <b>Side:</b> ..... **   |  | <b>Patient Name:</b><br><b>Address:</b><br><br><b>d.o.b.</b> <b>NHI:</b>  | <b>Consultant:</b> .....<br>[If different from patient label]<br><b>Hospital:</b> .....<br><b>Town/City:</b> ..... |
| <i>Tick Appropriate Boxes</i>   |  |   |  |
| <b>PREVIOUS OPERATION ON INDEX JOINT</b>  |  |   |  |
| <input type="checkbox"/> None<br><input type="checkbox"/> Internal fixation for juxtarticular fracture<br><input type="checkbox"/> Ligament reconstruction<br><input type="checkbox"/> Interposition arthroplasty |  | <input type="checkbox"/> Debridement<br><input type="checkbox"/> Synovectomy + removal radial head<br><input type="checkbox"/> Osteotomy<br><input type="checkbox"/> Other: Name: ..... |  |
| <b>DIAGNOSIS</b>  |  |   |  |
| <input type="checkbox"/> Rheumatoid arthritis<br><input type="checkbox"/> Osteoarthritis<br><input type="checkbox"/> Other inflammatory<br><input type="checkbox"/> Post dislocation                              |  | <input type="checkbox"/> Post fracture<br><input type="checkbox"/> Post ligament disruption<br><input type="checkbox"/> Other: Name: .....  |  |
| <b>APPROACH</b>   |  |   |  |
| <input type="checkbox"/> Medial   |  | <input type="checkbox"/> Lateral  |  |
|   |  | <input type="checkbox"/> Posterior  |  |
| <b>HUMERUS</b>  |  | <b>ULNA</b>   |  |
| Please do not fold<br>bar-coded label   |  | Please do not fold<br>bar-coded label   |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |  |   |  |
| <b>BONE GRAFT - HUMERUS</b>   |  | <b>BONE GRAFT - ULNA</b>  |  |
| <input type="checkbox"/> Allograft<br><input type="checkbox"/> Autograft <input type="checkbox"/><br>Synthetic  |  | <input type="checkbox"/> Allograft<br><input type="checkbox"/> Autograft <input type="checkbox"/> Synthetic   |  |
| <b>RADIAL HEAD</b>  |  | <b>AUGMENTS</b>   |  |
| Please do not fold<br>bar-coded label   |  | Please do not fold<br>bar-coded label   |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |  |   |  |
| <b>CEMENT</b>   |  |   |  |
| <input type="checkbox"/> Humerus <input type="checkbox"/> Ulna <input type="checkbox"/> Radius <input type="checkbox"/> Antibiotic brand: .....   |  |   |  |
| <input type="checkbox"/> <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b>   |  |   |  |
| Name .....                      ASA Class: 1   2   3   4   (please circle one)  |  |   |  |
| <b>OPERATING THEATRE</b>  |  |   |  |
| <input type="checkbox"/> Conventional   |  | <input type="checkbox"/> Laminar flow or similar  |  |
|   |  | <input type="checkbox"/> Space suits  |  |
| <b>SKIN TO SKIN TIME mins</b> Start skin.....                      Finish skin.....   |  |   |  |
| <b>PRIMARY OPERATING SURGEON</b>  |  |   |  |
| <input type="checkbox"/> Consultant   |  | <input type="checkbox"/> Adv Trainee Unsupervised   |  |
|   |  | <input type="checkbox"/> Adv Trainee Supervised      Year..... <input type="checkbox"/> Basic Trainee   |  |

**\*\*NB**      *If bilateral procedure two completed forms are required*

DO NOT PLACE IN PATIENT NOTES

TO BE RETAINED IN THEATRE SUITE

| <b>NEW ZEALAND JOINT REGISTRY</b><br><b>Revision Elbow Joint</b>  |   |   |  |
|---|---|---|--|
| Free Phone 0800-274-989   |   | 07.04.2005  |  |
| Date: .....<br><br>Side: ..... **   | <div style="border: 1px solid black; padding: 5px;">             Patient Name:<br/>             Address:<br/><br/>             d.o.b.                      NHI:<br/> <b>Attach Patient Label</b> </div> |   | Consultant: .....<br>[If different from patient label]<br>Hospital: .....<br>Town/City: .....                                  |
| <i>Tick Appropriate Boxes</i>   |   |   |  |
| <b>REASON FOR REVISION</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> <input type="checkbox"/> Loosening humeral component<br/> <input type="checkbox"/> Loosening ulnar component<br/> <input type="checkbox"/> Loosening radial head component<br/> <input type="checkbox"/> Pain               </div> <div style="width: 48%;"> <input type="checkbox"/> Deep infection<br/> <input type="checkbox"/> Fracture humerus<br/> <input type="checkbox"/> Fracture ulna<br/> <input type="checkbox"/> Dislocations<br/> <input type="checkbox"/> Other Name: .....               </div> </div> |   |   |  |
| Date Index Operation: .....   |   | If re-revision - Date previous revision: .....  |  |
| <b>REVISION</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> <input type="checkbox"/> Change of humeral component<br/> <input type="checkbox"/> Change of ulnar component<br/> <input type="checkbox"/> Change of radial head component               </div> <div style="width: 48%;"> <input type="checkbox"/> Change of all components<br/> <input type="checkbox"/> Removal of components<br/> <input type="checkbox"/> Other Name: .....               </div> </div>   |   |   |  |
| <b>APPROACH</b><br><input type="checkbox"/> Medial <input type="checkbox"/> Lateral <input type="checkbox"/> Posterior  |   |   |  |
| <div style="border: 1px solid black; padding: 10px; text-align: center;"> <b>Please do not fold<br/>bar-coded label</b> </div>  | <div style="border: 1px solid black; padding: 10px; text-align: center;"> <b>Please do not fold<br/>bar-coded label</b> </div>  |   | <div style="border: 1px solid black; padding: 10px; text-align: center;"> <b>Please do not fold<br/>bar-coded label</b> </div> |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |   |   |  |
| <b>BONE GRAFT - HUMERUS</b><br><input type="checkbox"/> Allograft <input type="checkbox"/> Synthetic<br><input type="checkbox"/> Autograft  |   | <b>BONE GRAFT - ULNA</b><br><input type="checkbox"/> Allograft <input type="checkbox"/> Synthetic<br><input type="checkbox"/> Autograft           |  |
| <b>RADIAL HEAD</b><br><div style="border: 1px solid black; padding: 10px; text-align: center;"> <b>Please do not fold<br/>bar-coded label</b> </div>  |   | <b>AUGMENTS</b><br><div style="border: 1px solid black; padding: 10px; text-align: center;"> <b>Please do not fold<br/>bar-coded label</b> </div> |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |   |   |  |
| <b>CEMENT</b><br><input type="checkbox"/> Humerus <input type="checkbox"/> Ulna <input type="checkbox"/> Radius <input type="checkbox"/> Antibiotic brand: .....  |   |   |  |
| <input type="checkbox"/> <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b><br>Name .....                      ASA Class:    1    2    3    4    (please circle one)  |   |   |  |
| <b>OPERATING THEATRE</b><br><input type="checkbox"/> Conventional <input type="checkbox"/> Laminar flow or similar <input type="checkbox"/> Space suits   |   |   |  |
| <b>SKIN TO SKIN TIME mins</b> Start skin.....    Finish skin.....   |   |   |  |
| <b>PRIMARY OPERATING SURGEON</b><br><div style="display: flex; justify-content: space-between;"> <div> <input type="checkbox"/> Consultant    <input type="checkbox"/> Adv Trainee Supervised    Year.....               </div> <div> <input type="checkbox"/> Adv Trainee Unsupervised    <input type="checkbox"/> Basic Trainee               </div> </div>   |   |   |  |

\*\*NB If bilateral procedure two completed forms are required

| <b>NEW ZEALAND JOINT REGISTRY</b><br><b>Primary Cervical Disc Replacement</b>   |  |   |  | 14.08.2008 |
|---|--|---|--|------------|
| Free Phone 0800-274-989   |  |   |  |            |
| Date: .....   | <div style="border: 1px solid black; padding: 5px;">           Patient Name:<br/>           Address:<br/><br/>           DOB: ..... NHI: .....<br/> <b>Attach Patient Label</b> </div> | Consultant: .....<br>[If different from patient label]<br>Hospital: .....<br>Town/City: .....<br>ACC <input type="checkbox"/> ACC Claim |  |            |
| Tick Appropriate Boxes<br>No: .....   |  |   |  |            |
| <b>LEVELS OF DISC REPLACEMENT</b><br><br>.....<br><input type="checkbox"/> C3/4 <input type="checkbox"/> C6/7<br><input type="checkbox"/> C4/5 <input type="checkbox"/> C7/T1<br><input type="checkbox"/> C5/6      Other .....     |  | <b>PRE OP PATIENT SCORE</b><br><b>(NECK DISABILITY INDEX)</b>   |  |            |
| <b>PREVIOUS OPERATION</b><br><input type="checkbox"/> Foreminotomy <input type="checkbox"/> Adjacent Level Disc Arthroplasty<br><input type="checkbox"/> Adjacent Level Fusion <input type="checkbox"/> Other.....                  |  |   |  |            |
| <b>DIAGNOSIS</b><br><input type="checkbox"/> Acute Disc Prolapse<br><input type="checkbox"/> Chronic Spondylosis<br><input type="checkbox"/> Neck Pain<br><input type="checkbox"/> Other .....                                      |  |   |  |            |
| <b>APPROACH</b><br><input type="checkbox"/> Anterior Right <input type="checkbox"/> Anterior Left <input type="checkbox"/> Other .....  |  |   |  |            |
| <b>IMPLANTS</b>   |  |   |  |            |
| <div style="border: 1px solid black; padding: 20px; min-height: 100px;"> <b>Affix Supplier Label</b> </div>   |  | <div style="border: 1px solid black; padding: 20px; min-height: 100px;"> <b>Affix Supplier Label</b> </div>                             |  |            |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |  |   |  |            |
| <div style="border: 1px solid black; padding: 20px; min-height: 100px;"> <b>Affix Supplier Label</b> </div>   |  | <div style="border: 1px solid black; padding: 20px; min-height: 100px;"> <b>Affix Supplier Label</b> </div>                             |  |            |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |  |   |  |            |
| <b>INTRAOPERATIVE COMPLICATIONS</b><br>.....<br>.....   |  |   |  |            |
| <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b><br><input type="checkbox"/> Yes <input type="checkbox"/> No  |  |   |  |            |
| <b>OPERATIVE THEATRE</b><br><input type="checkbox"/> Conventional <input type="checkbox"/> Laminar flow or similar <input type="checkbox"/> Space suits   |  |   |  |            |
| <b>SKIN TO SKIN TIME mins</b> Start skin.....      Finish skin.....   |  |   |  |            |
| <b>PRIMARY OPERATING SURGEON</b><br><input type="checkbox"/> Consultant <input type="checkbox"/> Adv Trainee Unsupervised      Year .....<br><input type="checkbox"/> Adv Trainee Supervised <input type="checkbox"/> Basic Trainee |  |   |  |            |

DO NOT PLACE IN PATIENT NOTES TO BE RETAINED IN THEATRE SUITE

| <b>NEW ZEALAND JOINT REGISTRY</b><br><b>Revision Cervical Disc Replacement</b>  |   |   |  |
|---|---|---|--|
| Free Phone 0800-274-989<br>14.08.2008   |   |   |  |
| Date: .....<br><br>LEVEL OF REVISION<br>.....<br><input type="checkbox"/> C3/4 <input type="checkbox"/> C6/7<br><input type="checkbox"/> C4/5 <input type="checkbox"/> C7/T1<br><br><input type="checkbox"/> C5/6 <input type="checkbox"/> Other:   | <div style="border: 1px solid black; padding: 5px;">           Patient Name:<br/>           Address:<br/><br/>           DOB:                      NHI:<br/> <b>Attach Patient Label</b> </div> | Consultant: .....<br>[If different from patient label]<br>Hospital:<br><br>Town/City: ..... | ACC <input type="checkbox"/> ACC Claim No: ..... |
| <b>REASON FOR REVISION</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> <input type="checkbox"/> Dislocation of component<br/> <input type="checkbox"/> Failure of component<br/> <input type="checkbox"/> Infection<br/> <input type="checkbox"/> Pain (Neck)             </div> <div style="width: 48%;"> <input type="checkbox"/> Adjacent level surgery<br/> <input type="checkbox"/> Additional decompression required<br/> <input type="checkbox"/> Heterotopic calcification<br/> <input type="checkbox"/> Other: Name: .....             </div> </div> |   |   |  |
| <div style="display: flex; justify-content: space-between;"> <div style="width: 48%;">           Date Index Operation: .....<br/> <b>REVISION</b><br/> <input type="checkbox"/> Replace disc prosthesis (same)<br/> <input type="checkbox"/> Replace disc prosthesis (different)<br/> <input type="checkbox"/> Fuse         </div> <div style="width: 48%;">           If re-revision - Date previous revision: ...<br/> <input type="checkbox"/> Removal only<br/> <input type="checkbox"/> Other: .....         </div> </div>   |   |   |  |
| <b>APPROACH</b> <input type="checkbox"/> Image guided surgery <input type="checkbox"/> Minimally invasive surgery<br><input type="checkbox"/> Anterior <input type="checkbox"/> Posterior <input type="checkbox"/> Lateral <input type="checkbox"/> Trochanteric<br><b>Osteotomy</b>  |   |   |  |
| <b>IMPLANTS</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 48%; text-align: center; border: 1px solid black; padding: 20px;">           Please do not fold<br/>bar-coded label         </div> <div style="width: 48%; text-align: center; border: 1px solid black; padding: 20px;">           Please do not fold<br/>bar-coded label         </div> </div>  |   |   |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |   |   |  |
| <div style="display: flex; justify-content: space-between;"> <div style="width: 48%; text-align: center; border: 1px solid black; padding: 20px;">           Please do not fold<br/>bar-coded label         </div> <div style="width: 48%; text-align: center; border: 1px solid black; padding: 20px;">           Please do not fold<br/>bar-coded label         </div> </div>   |   |   |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>   |   |   |  |
| <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b><br>Name .....  |   |   |  |
| <b>OPERATING THEATRE</b><br><input type="checkbox"/> Conventional <input type="checkbox"/> Laminar flow or similar <input type="checkbox"/> Space suits   |   |   |  |
| <b>SKIN TO SKIN TIME mins</b> Start skin.....    Finish skin.....   |   |   |  |
| <b>PRIMARY OPERATING SURGEON</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 40%;"> <input type="checkbox"/> Consultant    <input type="checkbox"/> Adv Trainee Supervised             </div> <div style="width: 20%;">           Year.....         </div> <div style="width: 40%;"> <input type="checkbox"/> Adv Trainee Unsupervised    <input type="checkbox"/> Basic Trainee             </div> </div>   |   |   |  |



DO NOT PLACE IN PATIENT NOTES TO BE RETAINED IN THEATRE SUITE

| <b>NEW ZEALAND JOINT REGISTRY</b><br><b>Primary Lumbar Disc Replacement</b>  |   |  |
|--|---|--|
| Free Phone 0800-274-989<br>14.08.2008  |   |  |
| Date: .....  | <div style="border: 1px solid black; padding: 5px;">             Patient Name:<br/>             Address:<br/><br/>             d.o.b.                      NHI:<br/> <b>Attach Patient Label</b> </div> | Consultant: .....<br>[If different from patient label]<br>Hospital: .....<br><br>Town/City.....  |
| Tick Appropriate Boxes                      ACC <input type="checkbox"/> ACC Claim No. ....  |   |  |
| <b>DISC REPLACEMENT Levels</b><br><input type="checkbox"/> L3/4<br>responses.....<br><input type="checkbox"/> L4/5<br><input type="checkbox"/> L5/S1   | <b>FUSION Levels</b><br><input type="checkbox"/> L3/4<br><input type="checkbox"/> L4/5<br>Percentage score  | <b>PRE OP PATIENT SCORE</b><br><i>Modified Roland and Morris</i><br>Total number of "Yes"<br><br><i>Oswestry Score</i> <input type="checkbox"/> L5/S1<br>Other ..... |
| <b>PREVIOUS OPERATION</b><br><input type="checkbox"/> Discectomy <input type="checkbox"/> L3/4 <input type="checkbox"/> L4/5 <input type="checkbox"/> L5/S1 <input type="checkbox"/> Other .....<br><input type="checkbox"/> Other ..... <input type="checkbox"/> L3/4 <input type="checkbox"/> L4/5 <input type="checkbox"/> L5/S1  |   |  |
| <b>DIAGNOSIS</b><br>1. Degenerative Disc disease <input type="checkbox"/> L3/4 <input type="checkbox"/> L4/5 <input type="checkbox"/> L5/S1 <input type="checkbox"/> Other .....<br>(plain x-ray changes present)<br>2. Annular tear MRI scan <input type="checkbox"/> L3/4 <input type="checkbox"/> L4/5 <input type="checkbox"/> L5/S1 <input type="checkbox"/> Other .....<br>(normal plain x-ray)<br>3. Discogenic pain on discography <input type="checkbox"/> L3/4 <input type="checkbox"/> L4/5 <input type="checkbox"/> L5/S1 <input type="checkbox"/> Other ..... |   |  |
| <b>APPROACH</b><br><input type="checkbox"/> Retroperitoneal midline abdominal wall incision <input type="checkbox"/> Transperitoneal<br><input type="checkbox"/> Retroperitoneal lateral abdominal wall incision <input type="checkbox"/> Other .....  |   |  |
| <b>IMPLANTS</b>  |   |  |
| Affix Supplier Label   | Affix Supplier Label  |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>  |   |  |
| Affix Supplier Label   | Affix Supplier Label  |  |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>  |   |  |
| <b>INTRAOPERATIVE COMPLICATIONS</b><br>.....   |   |  |
| <input type="checkbox"/> <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b><br>Yes <input type="checkbox"/> No <input type="checkbox"/>  |   |  |
| <b>OPERATIVE THEATRE</b><br><input type="checkbox"/> Conventional <input type="checkbox"/> Laminar flow or similar <input type="checkbox"/> Space suits  |   |  |
| <b>SKIN TO SKIN TIME mins</b> Start skin .....                      Finish skin .....  |   |  |
| <b>PRIMARY OPERATING SURGEON</b><br><input type="checkbox"/> Consultant <input type="checkbox"/> Adv Trainee                      Year..... <input type="checkbox"/> Basic Trainee   |   |  |

## APPENDIX 3 - DATA FORMS

DO NOT PLACE IN PATIENT NOTES TO BE RETAINED IN THEATRE SUITE

| <b>NEW ZEALAND JOINT REGISTRY</b><br><b>Revision Lumbar Disc Replacement</b>   |   |   |  |   |   |
|--|---|---|--|---|---|
| Free Phone 0800-274-989<br>14.08.2008  |   |   |  |   |   |
| Date: .....  | <table border="1" style="width: 100%; border-collapse: collapse;"> <tr> <td style="padding: 2px;">                             Patient Name:<br/>                             Address:<br/><br/>                             d.o.b.                      NHI:<br/> <b>Attach Patient Label</b> </td> </tr> </table> | Patient Name:<br>Address:<br><br>d.o.b.                      NHI:<br><b>Attach Patient Label</b>  | Consultant: .....<br>[If different from patient label]<br>Hospital: .....<br>Town/City: .....  |   |   |
| Patient Name:<br>Address:<br><br>d.o.b.                      NHI:<br><b>Attach Patient Label</b>   |   |   |  |   |   |
| Tick Appropriate Boxes                      ACC <input type="checkbox"/> ACC Claim No: .....   |   |   |  |   |   |
| <b>REASON FOR REVISION</b><br><table style="width: 100%;"> <tr> <td style="width: 50%; vertical-align: top;"> <input type="checkbox"/> Loosening of components<br/> <input type="checkbox"/> Dislocation of articulating core<br/> <input type="checkbox"/> Loss of spinal alignment<br/> <input type="checkbox"/> Pain                         </td> <td style="width: 50%; vertical-align: top;"> <input type="checkbox"/> Deep infection<br/> <input type="checkbox"/> Fracture of vertebra<br/> <input type="checkbox"/> Removal of components<br/> <input type="checkbox"/> Other: Name: .....                         </td> </tr> </table>   |   |   | <input type="checkbox"/> Loosening of components<br><input type="checkbox"/> Dislocation of articulating core<br><input type="checkbox"/> Loss of spinal alignment<br><input type="checkbox"/> Pain                    | <input type="checkbox"/> Deep infection<br><input type="checkbox"/> Fracture of vertebra<br><input type="checkbox"/> Removal of components<br><input type="checkbox"/> Other: Name: ..... |   |
| <input type="checkbox"/> Loosening of components<br><input type="checkbox"/> Dislocation of articulating core<br><input type="checkbox"/> Loss of spinal alignment<br><input type="checkbox"/> Pain  | <input type="checkbox"/> Deep infection<br><input type="checkbox"/> Fracture of vertebra<br><input type="checkbox"/> Removal of components<br><input type="checkbox"/> Other: Name: .....   |   |  |   |   |
| <table style="width: 100%;"> <tr> <td style="width: 50%;">Date Index Operation: .....</td> <td style="width: 50%;">If re-revision - Date previous revision: .....</td> </tr> </table>  |   |   | Date Index Operation: .....  | If re-revision - Date previous revision: .....  |   |
| Date Index Operation: .....  | If re-revision - Date previous revision: .....  |   |  |   |   |
| <b>REVISION</b><br><table style="width: 100%;"> <tr> <td style="width: 50%; vertical-align: top;"> <input type="checkbox"/> Change of TDR components<br/> <input type="checkbox"/> Change to Anterior Fusion                         </td> <td style="width: 50%; vertical-align: top;"> <input type="checkbox"/> Change of articulating core<br/> <input type="checkbox"/> In-situ posterior instrumented fusion                         </td> </tr> </table>   |   |   | <input type="checkbox"/> Change of TDR components<br><input type="checkbox"/> Change to Anterior Fusion  | <input type="checkbox"/> Change of articulating core<br><input type="checkbox"/> In-situ posterior instrumented fusion  |   |
| <input type="checkbox"/> Change of TDR components<br><input type="checkbox"/> Change to Anterior Fusion  | <input type="checkbox"/> Change of articulating core<br><input type="checkbox"/> In-situ posterior instrumented fusion  |   |  |   |   |
| <b>APPROACH</b><br><table style="width: 100%;"> <tr> <td style="width: 50%; vertical-align: top;"> <input type="checkbox"/> Retroperitoneal midline abdominal wall incision<br/> <input type="checkbox"/> Retroperitoneal lateral abdominal wall incision<br/> <input type="checkbox"/> Posterior Approach for in-situ fusion                         </td> <td style="width: 50%; vertical-align: top;"> <input type="checkbox"/> Transperitoneal<br/> <input type="checkbox"/> Other .....                         </td> </tr> </table>  |   |   | <input type="checkbox"/> Retroperitoneal midline abdominal wall incision<br><input type="checkbox"/> Retroperitoneal lateral abdominal wall incision<br><input type="checkbox"/> Posterior Approach for in-situ fusion | <input type="checkbox"/> Transperitoneal<br><input type="checkbox"/> Other .....  |   |
| <input type="checkbox"/> Retroperitoneal midline abdominal wall incision<br><input type="checkbox"/> Retroperitoneal lateral abdominal wall incision<br><input type="checkbox"/> Posterior Approach for in-situ fusion   | <input type="checkbox"/> Transperitoneal<br><input type="checkbox"/> Other .....  |   |  |   |   |
| <table style="width: 100%;"> <tr> <td style="width: 33%; vertical-align: top;"> <b>NEW DISC REPLACEMENT Levels</b><br/> <input type="checkbox"/> L3/4<br/> <input type="checkbox"/> L4/5<br/> <input type="checkbox"/> L5/S1                         </td> <td style="width: 33%; vertical-align: top;"> <b>NEW FUSION Levels</b><br/> <input type="checkbox"/> L3/4<br/> <input type="checkbox"/> L4/5<br/> <input type="checkbox"/> L5/S1                         </td> <td style="width: 34%; vertical-align: top;"> <b>PRE OP PATIENT SCORE</b><br/> <i>Modified Roland and Morris</i><br/>                     Total number of "Yes" responses.....<br/> <i>Oswestry Score</i><br/>                     Percentage score                 </td> </tr> </table> |   |   | <b>NEW DISC REPLACEMENT Levels</b><br><input type="checkbox"/> L3/4<br><input type="checkbox"/> L4/5<br><input type="checkbox"/> L5/S1   | <b>NEW FUSION Levels</b><br><input type="checkbox"/> L3/4<br><input type="checkbox"/> L4/5<br><input type="checkbox"/> L5/S1  | <b>PRE OP PATIENT SCORE</b><br><i>Modified Roland and Morris</i><br>Total number of "Yes" responses.....<br><i>Oswestry Score</i><br>Percentage score |
| <b>NEW DISC REPLACEMENT Levels</b><br><input type="checkbox"/> L3/4<br><input type="checkbox"/> L4/5<br><input type="checkbox"/> L5/S1   | <b>NEW FUSION Levels</b><br><input type="checkbox"/> L3/4<br><input type="checkbox"/> L4/5<br><input type="checkbox"/> L5/S1  | <b>PRE OP PATIENT SCORE</b><br><i>Modified Roland and Morris</i><br>Total number of "Yes" responses.....<br><i>Oswestry Score</i><br>Percentage score |  |   |   |
| Other .....  |   |   |  |   |   |
| <b>IMPLANTS</b><br><table style="width: 100%;"> <tr> <td style="width: 50%; text-align: center; padding: 20px;"> <div style="border: 1px solid black; width: 100%; height: 40px; margin: 0 auto;"></div> <b>Affix Supplier Label</b> </td> <td style="width: 50%; text-align: center; padding: 20px;"> <div style="border: 1px solid black; width: 100%; height: 40px; margin: 0 auto;"></div> <b>Affix Supplier Label</b> </td> </tr> </table>  |   |   | <div style="border: 1px solid black; width: 100%; height: 40px; margin: 0 auto;"></div> <b>Affix Supplier Label</b>  | <div style="border: 1px solid black; width: 100%; height: 40px; margin: 0 auto;"></div> <b>Affix Supplier Label</b>   |   |
| <div style="border: 1px solid black; width: 100%; height: 40px; margin: 0 auto;"></div> <b>Affix Supplier Label</b>  | <div style="border: 1px solid black; width: 100%; height: 40px; margin: 0 auto;"></div> <b>Affix Supplier Label</b>   |   |  |   |   |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>  |   |   |  |   |   |
| <table style="width: 100%;"> <tr> <td style="width: 50%; text-align: center; padding: 20px;"> <div style="border: 1px solid black; width: 100%; height: 40px; margin: 0 auto;"></div> <b>Affix Supplier Label</b> </td> <td style="width: 50%; text-align: center; padding: 20px;"> <div style="border: 1px solid black; width: 100%; height: 40px; margin: 0 auto;"></div> <b>Affix Supplier Label</b> </td> </tr> </table>   |   |   | <div style="border: 1px solid black; width: 100%; height: 40px; margin: 0 auto;"></div> <b>Affix Supplier Label</b>  | <div style="border: 1px solid black; width: 100%; height: 40px; margin: 0 auto;"></div> <b>Affix Supplier Label</b>   |   |
| <div style="border: 1px solid black; width: 100%; height: 40px; margin: 0 auto;"></div> <b>Affix Supplier Label</b>  | <div style="border: 1px solid black; width: 100%; height: 40px; margin: 0 auto;"></div> <b>Affix Supplier Label</b>   |   |  |   |   |
| <b>STICK EXTRA LABELS ON REVERSE SIDE</b>  |   |   |  |   |   |
| <b>INTRAOPERATIVE COMPLICATIONS</b><br>.....<br>.....  |   |   |  |   |   |
| <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b><br>Yes <input type="checkbox"/> No <input type="checkbox"/>   |   |   |  |   |   |
| <b>OPERATIVE THEATRE</b><br><input type="checkbox"/> Conventional <input type="checkbox"/> Laminar flow or similar <input type="checkbox"/> Space suits  |   |   |  |   |   |
| <b>SKIN TO SKIN TIME mins</b> Start skin .....      Finish skin .....  |   |   |  |   |   |
| <b>PRIMARY OPERATING SURGEON</b><br><input type="checkbox"/> Consultant <input type="checkbox"/> Adv Trainee      Year..... <input type="checkbox"/> Basic Trainee   |   |   |  |   |   |

| <b><u>NEW ZEALAND JOINT REGISTRY</u></b><br><br><b>REOPERATION WITHOUT REPLACEMENT or<br/>REMOVAL OF ANY PROSTHETIC COMPONENTS</b>  |  |
|---|--|
| <b>Patient label</b><br><br><div style="border: 1px solid black; padding: 10px; margin: 0 auto; width: 80%;">           Patient Name:<br/>           Address:<br/><br/>           D.O.B. _____ NHI: _____<br/> <b><i>Attach Patient Label</i></b> </div>  |  |
| Date: .....   | Consultant: .....<br>(if different from label) |
| Side: .....   | Hospital: .....                                |
| Town/City: .....  |  |
| <b>Tick Appropriate Boxes</b><br><input type="checkbox"/> Hip <input type="checkbox"/> Knee <input type="checkbox"/> Ankle <input type="checkbox"/> Shoulder <input type="checkbox"/> Elbow   |  |
| <b>REASON FOR REOPERATION</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> <input type="checkbox"/> Deep Infection<br/> <input type="checkbox"/> Dislocation of joint<br/> <input type="checkbox"/> Dislocation of bearing<br/> <input type="checkbox"/> Fracture<br/> <input type="checkbox"/> Instability             </div> <div style="width: 48%;"> <input type="checkbox"/> Malalignment<br/> <input type="checkbox"/> Impingement<br/> <input type="checkbox"/> Stiffness<br/> <input type="checkbox"/> Haematoma evacuation<br/> <input type="checkbox"/> Arthrofibrosis<br/> <input type="checkbox"/> Other             </div> </div>   |  |
| Date Index Operation: .....   |  |
| <b>PROCEDURE</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 48%;"> <input type="checkbox"/> Open lavage<br/> <input type="checkbox"/> Arthroscopic lavage<br/> <input type="checkbox"/> Closed reduction of dislocation<br/> <input type="checkbox"/> Open reduction of dislocation<br/> <input type="checkbox"/> Fracture fixation<br/> <input type="checkbox"/> Soft tissue procedure             </div> <div style="width: 48%;"> <input type="checkbox"/> Ligament reconstruction<br/> <input type="checkbox"/> Osteotomy<br/> <input type="checkbox"/> Bone debridement<br/> <input type="checkbox"/> Arthrolysis<br/> <input type="checkbox"/> M. U. A.<br/> <input type="checkbox"/> Other             </div> </div> |  |
| <input type="checkbox"/> <b>SYSTEMIC ANTIBIOTIC PROPHYLAXIS</b><br>Name ..... ASA Class:    1    2    3    4    (please circle one)   |  |
| <b>PRIMARY OPERATING SURGEON</b><br><div style="display: flex; justify-content: space-between;"> <div style="width: 40%;"> <input type="checkbox"/> Consultant    <input type="checkbox"/> Adv Trainee Unsupervised<br/> <input type="checkbox"/> Adv Trainee Supervised             </div> <div style="width: 20%;">Year.....</div> <div style="width: 40%; text-align: right;"> <input type="checkbox"/> Basic Trainee             </div> </div>  |  |

**TOTAL HIP REPLACEMENT - QUESTIONNAIRE****Patient Name:** .....**Date of Birth:** .....**Patient Address:** .....**Operating Surgeon:** .....**Date of Surgery:** .....

We would like you to score yourself on the following 12 questions. Each question is scored from 4 to 0, from least to most difficulty or severity: 4 being the least difficult/severe and 0 being the most difficult/severe.

Please circle the number which best describes yourself **OVER THE LAST 4 WEEKS**

| Please circle the SIDE on which you had your surgery performed |  | Left | Right  |
|--|--|------|--|
| 1  | How would you describe the pain you usually had from your operated on hip?<br>4 None<br>3 Very mild<br>2 Mild<br>1 Moderate<br>0 Severe  | 8    | After a meal (sat at a table), how painful has it been for you to stand up from a chair because of your operated on hip?<br>4 Not at all painful<br>3 Slightly painful<br>2 Moderately painful<br>1 Very painful<br>0 Unbearable |
| 2  | For how long have you been able to walk before the pain from your operated on hip becomes severe? (with or without a stick)<br>4 No pain/more than 30 minutes<br>3 16 to 30 minutes<br>2 5 to 15 minutes<br>1 Around the house only<br>0 Unable to walk because of severe pain | 9    | Have you had any sudden, severe pain - 'shooting', 'stabbing' or 'spasms' - from the affected operated on hip?<br>4 No days<br>3 Only 1 or 2 days<br>2 Some days<br>1 Most days<br>0 Every day                                   |
| 3  | Have you had any trouble getting in and out of a car or using public transport because of your operated on hip?<br>4 No trouble at all<br>3 Very little trouble<br>2 Moderate trouble<br>1 Extreme difficulty<br>0 Impossible to do  | 10   | Have you been limping when walking, because of your operated on hip?<br>4 Rarely/never<br>3 Sometimes or just at first<br>2 Often, not just at first<br>1 Most of the time<br>0 All of the time                                  |
|  | 4 Have you been able to put on a pair of socks, stockings or tights?<br>4 Yes, easily<br>3 With little difficulty<br>2 With moderate difficulty<br>1 With extreme difficulty<br>0 No, impossible   | 11   | Have you been able to climb a flight of stairs?<br>4 Yes, easily<br>3 With little difficulty<br>2 With moderate difficulty<br>1 With extreme difficulty<br>0 No, impossible  |
| 5  | Could you do the household shopping on your own?<br>4 Yes, easily<br>3 With little difficulty<br>2 With moderate difficulty<br>1 With extreme difficulty<br>0 No, impossible   | 12   | Have you been troubled by pain from your operated on hip in bed at night?<br>4 No nights<br>3 Only 1 or 2 nights<br>2 Some nights<br>1 Most nights<br>0 Every night  |
| 6  | Have you had any trouble with washing and drying yourself (all over) because of your operated on hip?<br>4 No trouble at all<br>3 Very little trouble<br>2 Moderate trouble<br>1 Extreme difficulty<br>0 Impossible to do  |      |  |
| 7  | How much has pain from your operated on hip interfered with your usual work (including housework)?<br>4 Not at all<br>3 A little bit<br>2 Moderately<br>1 Greatly<br>0 Totally   |      |  |

- ☐ I wish to receive a progress report on the study. **NB:** If there are reasons other than the operation which would stop you doing one of the tasks listed; try to answer the question from the joint replacement aspect alone.

## REVISION HIP REPLACEMENT - QUESTIONNAIRE

Patient Name: .....

Date of Birth: .....

Patient Address: .....

Operating Surgeon: .....

Date of Surgery: .....

We would like you to score yourself on the following 12 questions. Each question is scored from 4 to 0, from least to most difficulty or severity: 4 being the least difficult/severe and 0 being the most difficult/severe.

Please circle the number which best describes yourself **OVER THE LAST 4 WEEKS**

**Please circle the SIDE on which you had your surgery performed**      **Left**      **Right**

|  |  |
|--|--|
| <p>1 How would you describe the pain you usually had from your operated on hip?</p> <p>4 None</p> <p>3 Very mild</p> <p>2 Mild</p> <p>1 Moderate</p> <p>0 Severe</p> <p>2 For how long have you been able to walk before the pain from your operated on hip becomes severe? (with or without a stick)</p> <p>4 No pain/more than 30 minutes</p> <p>3 16 to 30 minutes</p> <p>2 5 to 15 minutes</p> <p>1 Around the house only</p> <p>0 Unable to walk because of severe pain</p> <p>3 Have you had any trouble getting in and out of a car or using public transport because of your operated on hip?</p> <p>4 No trouble at all</p> <p>3 Very little trouble</p> <p>2 Moderate trouble</p> <p>1 Extreme difficulty</p> <p>0 Impossible to do</p> <p>4 Have you been able to put on a pair of socks, stockings or tights?</p> <p>4 Yes, easily</p> <p>3 With little difficulty</p> <p>2 With moderate difficulty</p> <p>1 With extreme difficulty</p> <p>0 No, impossible</p> <p>5 Could you do the household shopping on your own?</p> <p>4 Yes, easily</p> <p>3 With little difficulty</p> <p>2 With moderate difficulty</p> <p>1 With extreme difficulty</p> <p>0 No, impossible</p> <p>6 Have you had any trouble with washing and drying yourself (all over) because of your operated on hip?</p> <p>4 No trouble at all</p> <p>3 Very little trouble</p> <p>2 Moderate trouble</p> <p>1 Extreme difficulty</p> <p>0 Impossible to do</p> <p>7 How much has pain from your operated on hip interfered with your usual work (including housework)?</p> <p>4 Not at all</p> <p>3 A little bit</p> <p>2 Moderately</p> <p>1 Greatly</p> <p>0 Totally</p> | <p>8 After a meal (sat at a table), how painful has it been for you to stand up from a chair because of your operated on hip?</p> <p>4 Not at all painful</p> <p>3 Slightly painful</p> <p>2 Moderately painful</p> <p>1 Very painful</p> <p>0 Unbearable</p> <p>9 Have you had any sudden, severe pain - 'shooting', 'stabbing' or 'spasms' - from the affected operated on hip?</p> <p>4 No days</p> <p>3 Only 1 or 2 days</p> <p>2 Some days</p> <p>1 Most days</p> <p>0 Every day</p> <p>10 Have you been limping when walking, because of your operated on hip?</p> <p>4 Rarely/never</p> <p>3 Sometimes, or just at first</p> <p>2 Often, not just at first</p> <p>1 Most of the time</p> <p>0 All of the time</p> <p>11 Have you been able to climb a flight of stairs?</p> <p>4 Yes, easily</p> <p>3 With little difficulty</p> <p>2 With moderate difficulty</p> <p>1 With extreme difficulty</p> <p>0 No, impossible</p> <p>12 Have you been troubled by pain from your operated on hip in bed at night?</p> <p>4 No nights</p> <p>3 Only 1 or 2 nights</p> <p>2 Some nights</p> <p>1 Most nights</p> <p>0 Every night</p> |
|--|--|

- ☐ I wish to receive a progress report on the study. **NB:** If there are reasons other than the operation which would stop you doing one of the tasks listed; try to answer the question from the joint replacement aspect alone.

## APPENDIX 4 - OXFORD QUESTIONNAIRE FORMS

### TOTAL KNEE REPLACEMENT - QUESTIONNAIRE

**Patient Name:** ..... **Date of Birth:** .....  
**Patient Address:** ..... **Operating Surgeon:** .....  
**Date of Surgery:** .....

We would like you to score yourself on the following 12 questions. Each question is scored from 4 to 0, from least to most difficulty or severity: 4 being the least difficult/severe and 0 being the most difficult/severe.

Please circle the number which best describes yourself **OVER THE LAST 4 WEEKS**

**Please circle the SIDE on which you had your surgery performed**      **Left**      **Right**

|   |   |
|---|---|
| <p>1 How would you describe the pain you usually have from your operated on knee?</p> <p>4 None<br/>3 Very mild<br/>2 Mild<br/>1 Moderate<br/>0 Severe</p> <p>2 For how long have you been able to walk before the pain from your operated on knee becomes severe? (with or without a stick)</p> <p>4 No pain/more than 30 minutes<br/>3 16 to 30 minutes<br/>2 5 to 15 minutes<br/>1 Around the house only<br/>0 Unable to walk because of severe pain</p> <p>3 Have you had any trouble getting in and out of a car or using public transport because of your operated on knee?</p> <p>4 No trouble at all<br/>3 Very little trouble<br/>2 Moderate trouble<br/>1 Extreme difficulty<br/>0 Impossible to do</p> <p>4 Could you kneel down and get up again afterwards on your operated knee?</p> <p>4 Yes, easily<br/>3 With little difficulty<br/>2 With moderate difficulty<br/>1 With extreme difficulty<br/>0 No, impossible</p> <p>5 Could you do the household shopping on your own?</p> <p>4 Yes, easily<br/>3 With little difficulty<br/>2 With moderate difficulty<br/>1 With extreme difficulty<br/>0 No, impossible</p> <p>6 Have you had any trouble with washing and drying yourself (all over) because of your operated on knee?</p> <p>4 No trouble at all<br/>3 Very little trouble<br/>2 Moderate trouble<br/>1 Extreme difficulty<br/>0 Impossible to do</p> <p>7 How much has pain from your operated on knee interfered with your usual work (including housework)?</p> <p>4 Not at all<br/>3 A little bit<br/>2 Moderately<br/>1 Greatly<br/>0 Totally</p> | <p>8 After a meal (sat at a table), how painful has it been for you to stand up from a chair because of your operated on knee?</p> <p>4 Not at all painful<br/>3 Slightly painful<br/>2 Moderately painful<br/>1 Very painful<br/>0 Unbearable</p> <p>9 Have you felt that your operated on knee might suddenly "give way" or let you down?</p> <p>4 Rarely/never<br/>3 Sometimes, or just at first<br/>2 Often, not just at first<br/>1 Most of the time<br/>0 All of the time</p> <p>10 Have you been limping when walking, because of your operated on knee?</p> <p>4 Rarely/never<br/>3 Sometimes, or just at first<br/>2 Often, not just at first<br/>1 Most of the time<br/>0 All of the time</p> <p>11 Could you walk down one flight of stairs?</p> <p>4 Yes, easily<br/>3 With little difficulty<br/>2 With moderate difficulty<br/>1 With extreme difficulty<br/>0 No, impossible</p> <p>12 Have you been troubled by pain from your operated on knee in bed at night?</p> <p>4 No nights<br/>3 Only 1 or 2 nights<br/>2 Some nights<br/>1 Most nights<br/>0 Every night</p> <p>.....</p> |
|---|---|

- ☐ I wish to receive a progress report on the study. **NB:** If there are reasons other than the operation which would stop you doing one of the tasks listed; try to answer the question from the joint replacement aspect alone.

## REVISION KNEE REPLACEMENT - QUESTIONNAIRE

**Patient Name:** ..... **Date of Birth:** .....  
**Patient Address:** ..... **Operating Surgeon:** .....  
..... **Date of Surgery:** .....

We would like you to score yourself on the following 12 questions. Each question is scored from 4 to 0, from least to most difficulty or severity: 4 being the least difficult/severe and 0 being the most difficult/severe.

Please circle the number which best describes yourself **OVER THE LAST 4 WEEKS**

**Please circle the SIDE on which you had your surgery performed Left Right**

|   |   |
|---|---|
| <p>1 How would you describe the pain you usually have from your operated on knee?</p> <p>4 None</p> <p>3 Very mild</p> <p>2 Mild</p> <p>1 Moderate</p> <p>0 Severe</p> <p>2 For how long have you been able to walk before the pain from your operated on knee becomes severe? (with or without a stick)</p> <p>4 No pain/more than 30 minutes</p> <p>3 16 to 30 minutes</p> <p>2 5 to 15 minutes</p> <p>1 Around the house only</p> <p>0 Unable to walk because of severe pain</p> <p>3 Have you had any trouble getting in and out of a car or using public transport because of your operated on knee?</p> <p>4 No trouble at all</p> <p>3 Very little trouble</p> <p>2 Moderate trouble</p> <p>1 Extreme difficulty</p> <p>0 Impossible to do</p> <p>4 Could you kneel down and get up again afterwards?</p> <p>4 Yes, easily</p> <p>3 With little difficulty</p> <p>2 With moderate difficulty</p> <p>1 With extreme difficulty</p> <p>0 No, impossible</p> <p>5 Could you do the household shopping on your own?</p> <p>4 Yes, easily</p> <p>3 With little difficulty</p> <p>2 With moderate difficulty</p> <p>1 With extreme difficulty</p> <p>0 No, impossible</p> <p>6 Have you had any trouble with washing and drying yourself (all over) because of your operated on knee?</p> <p>4 No trouble at all</p> <p>3 Very little trouble</p> <p>2 Moderate trouble</p> <p>1 Extreme difficulty</p> <p>0 Impossible to do</p> <p>7 How much has pain from your operated on knee interfered with your usual work (including housework)?</p> <p>4 Not at all</p> <p>3 A little bit</p> <p>2 Moderately</p> <p>1 Greatly</p> <p>0 Totally</p> | <p>8 After a meal (sat at a table), how painful has it been for you to stand up from a chair because of your operated on knee?</p> <p>4 Not at all painful</p> <p>3 Slightly painful</p> <p>2 Moderately painful</p> <p>1 Very painful</p> <p>0 Unbearable</p> <p>9 Have you felt that your operated on knee might suddenly "give way" or let you down?</p> <p>4 Rarely/never</p> <p>3 Sometimes, or just at first</p> <p>2 Often, not just at first</p> <p>1 Most of the time</p> <p>0 All of the time</p> <p>10 Have you been limping when walking, because of your operated on knee?</p> <p>4 Rarely/never</p> <p>3 Sometimes, or just at first</p> <p>2 Often, not just at first</p> <p>1 Most of the time</p> <p>0 All of the time</p> <p>11 Could you walk down one flight of stairs?</p> <p>4 Yes, easily</p> <p>3 With little difficulty</p> <p>2 With moderate difficulty</p> <p>1 With extreme difficulty</p> <p>0 No, impossible</p> <p>12 Have you been troubled by pain from your operated on knee in bed at night?</p> <p>4 No nights</p> <p>3 Only 1 or 2 nights</p> <p>2 Some nights</p> <p>1 Most nights</p> <p>0 Every night</p> <p><b>Additional Information</b></p> |
|---|---|

- ☐ I wish to receive a progress report on the study. **NB:** If there are reasons other than the operation which would stop you doing one of the tasks listed; try to answer the question from the joint replacement aspect alone.

## Manchester-Oxford Foot Questionnaire (MOxFQ)

Circle as appropriate Right / Left

Full Name \_\_\_\_\_

Please tick (✓) one for each statement

1. I have pain in my foot/ankle  

|                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None of the<br>Time      | Rarely                   | Some of the<br>time      | Most of the<br>time      | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
2. During the past 4 weeks this has applied to me:  
I avoid walking long distances because of pain in my foot/ankle  

|                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None of the<br>Time      | Rarely                   | Some of the<br>time      | Most of the<br>time      | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
3. During the past 4 weeks this has applied to me:  
I change the way I walk due to pain in my foot/ankle  

|                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None of the<br>Time      | Rarely                   | Some of the<br>time      | Most of the<br>time      | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
4. During the past 4 weeks this has applied to me:  
I walk slowly because of pain in my foot/ankle  

|                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None of the<br>Time      | Rarely                   | Some of the<br>time      | Most of the<br>time      | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
5. During the past 4 weeks this has applied to me:  
I have to stop and rest my foot/ankle because of pain  

|                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None of the<br>Time      | Rarely                   | Some of the<br>time      | Most of the<br>time      | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
6. During the past 4 weeks this has applied to me:  
I avoid some hard or rough surfaces because of pain in my foot/ankle  

|                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None of the<br>Time      | Rarely                   | Some of the<br>time      | Most of the<br>time      | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
7. During the past 4 weeks this has applied to me:  
I avoid standing for a long time because of pain in my foot/ankle  

|                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None of the<br>Time      | Rarely                   | Some of the<br>time      | Most of the<br>time      | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
8. During the past 4 weeks this has applied to me:  
I catch the bus or use the car instead of walking, because of pain in my foot/ankle  

|                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None of the<br>Time      | Rarely                   | Some of the<br>time      | Most of the<br>time      | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
9. During the past 4 weeks this has applied to me:  
I feel self-conscious about my foot/ankle  

|                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None of the<br>Time      | Rarely                   | Some of the<br>time      | Most of the<br>time      | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
10. During the past 4 weeks this has applied to me:  
I feel self-conscious about the shoes I have to wear  

|                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None of the<br>Time      | Rarely                   | Some of the<br>time      | Most of the<br>time      | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |



11. During the past 4 weeks this has applied to me:  
The pain in my foot/ankle is more painful in the evening
- |                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None of the<br>Time      | Rarely                   | Some of the<br>time      | Most of the<br>time      | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
12. During the past 4 weeks this has applied to me:  
I get shooting pains in my foot/ankle
- |                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None of the<br>Time      | Rarely                   | Some of the<br>time      | Most of the<br>time      | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
13. During the past 4 weeks this has applied to me:  
The pain in my foot/ankle prevents me from carrying out my work/everyday activities
- |                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None of the<br>Time      | Rarely                   | Some of the<br>time      | Most of the<br>time      | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
14. During the past 4 weeks this has applied to me:  
I am unable to do all my social or recreational activities because of pain in my foot/ankle
- |                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None of the<br>Time      | Rarely                   | Some of the<br>time      | Most of the<br>time      | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
15. During the past 4 weeks....  
How would you describe the pain you usually have in your foot/ankle?
- |                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| None                     | Very mild                | Mild                     | Moderate                 | Severe                   |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |
16. During the past 4 weeks....  
Have you been troubled by pain from your foot/ankle in bed at night?
- |                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| No nights                | Only 1 or 2<br>nights    | Some nights              | Most nights              | Every night              |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

## APPENDIX 4 - OXFORD QUESTIONNAIRE FORMS

### TOTAL SHOULDER REPLACEMENT - QUESTIONNAIRE

**Patient Name:** ..... **Date of Birth:** .....  
**Patient Address:** ..... **Operating Surgeon:** .....  
..... **Date of Surgery:** .....

We would like you to score yourself on the following 12 questions. Each question is scored from 4 to 0, from least to most difficulty or severity: 4 being the least difficult/severe and 0 being the most difficult/severe.

Please **circle the number** which best describes yourself **OVER THE LAST 4 WEEKS** **Which is your dominant arm?**

| Please circle the SIDE on which you had your surgery performed |  | Left   | Right   |
|--|--|--|---|
| 1  | How would you describe the <b>worst</b> pain you have had from your operated on shoulder?                            | 4 None<br>3 Mild<br>2 Moderate<br>1 Severe<br>0 Unbearable   | 8 Have you had any trouble dressing yourself because of your operated on shoulder?<br>4 No trouble at all<br>3 A little bit of trouble<br>2 Moderate trouble<br>1 Extreme difficulty<br>0 Impossible to do                |
| 2  | How would you describe the pain you <b>usually</b> have from your operated on shoulder?                              | 4 None<br>3 Very mild<br>2 Mild<br>1 Moderate<br>0 Severe  | 9 Could you hang your clothes up in a wardrobe – using the operated on arm?<br>4 Yes, easily<br>3 With little difficulty<br>2 With moderate difficulty<br>1 With extreme difficulty<br>0 No, impossible                   |
| 3  | Have you had any trouble getting in and out of a car or using public transport because of your operated on shoulder? | 4 No trouble at all<br>3 A little bit of trouble<br>2 Moderate trouble<br>1 Extreme difficulty<br>0 Impossible to do     | 10 Have you been able to wash and dry yourself under both arms?<br>4 Yes, easily<br>3 With little difficulty<br>2 With moderate difficulty<br>1 With extreme difficulty<br>0 No, impossible                               |
| 4  | Have you been able to use a knife and fork at the same time?   | 4 Yes, easily<br>3 With little difficulty<br>2 With moderate difficulty<br>1 With extreme difficulty<br>0 No, impossible | 11 How much has pain from your operated on shoulder interfered with your usual work hobbies or recreational activities (including housework)?<br>4 Not at all<br>3 A little bit<br>2 Moderately<br>1 Greatly<br>0 Totally |
| 5  | Could you do the household shopping on your own?   | 4 Yes, easily<br>3 With little difficulty<br>2 With moderate difficulty<br>1 With extreme difficulty<br>0 No, impossible | 12 Have you been troubled by pain from your operated on shoulder in bed at night?<br>4 No nights<br>3 Only 1 or 2 nights<br>2 Some nights<br>1 Most nights<br>0 Every night<br>.....                                      |
| 6  | Could you carry a tray containing a plate of food across a room?   | 4 Yes, easily<br>3 With little difficulty<br>2 With moderate difficulty<br>1 With extreme difficulty<br>0 No, impossible |   |
| 7  | Could you brush/comb your hair with the operated on arm?   | 4 Yes, easily<br>3 With little difficulty<br>2 With moderate difficulty<br>1 With extreme difficulty<br>0 No, Impossible |   |

☐ I wish to receive a progress report on the study. **NB:** If there are reasons other than the operation which would stop you doing one of the tasks listed; try to answer the question from the joint replacement aspect alone.

## REVISION SHOULDER REPLACEMENT - QUESTIONNAIRE

Patient Name: .....

Date of Birth: .....

Patient Address: .....

Operating surgeon: .....

.....

Date of Surgery: .....

We would like you to score yourself on the following 12 questions. Each question is scored from 4 to 0, from least to most difficulty or severity: 4 being the least difficult/severe and 0 being the most difficult/severe.

Please circle the number which best describes yourself **OVER THE LAST 4 WEEKS** Which is your dominant arm?      **Left**      **Right**

Please circle the SIDE on which you had your surgery performed      **Left**      **Right**

|  |  |
|--|--|
| <p>1 How would you describe the <b>worst</b> pain you have had from your operated on shoulder?</p> <p>4 None</p> <p>3 Mild</p> <p>2 Moderate</p> <p>1 Severe</p> <p>0 Unbearable</p> <p>2 How would you describe the pain you <b>usually</b> have from your operated on shoulder?</p> <p>4 None</p> <p>3 Very mild</p> <p>2 Mild</p> <p>1 Moderate</p> <p>0 Severe</p> <p>3 Have you had any trouble getting in and out of a car or using public transport because of your operated on shoulder?</p> <p>4 No trouble at all</p> <p>3 A little bit of trouble</p> <p>2 Moderate trouble</p> <p>1 Extreme difficulty</p> <p>0 Impossible to do</p> <p>4 Have you been able to use a knife and fork at the same time?</p> <p>4 Yes, easily</p> <p>3 With little difficulty</p> <p>2 With moderate difficulty</p> <p>1 With extreme difficulty</p> <p>0 No, impossible</p> <p>5 Could you do the household shopping on your own?</p> <p>4 Yes, easily</p> <p>3 With little difficulty</p> <p>2 With moderate difficulty</p> <p>1 With extreme difficulty</p> <p>0 No, impossible</p> <p>6 Could you carry a tray containing a plate of food across a room?</p> <p>4 Yes, easily</p> <p>3 With little difficulty</p> <p>2 With moderate difficulty</p> <p>1 With extreme difficulty</p> <p>0 No, impossible</p> <p>7 Could you brush/comb your hair with the operated on arm?</p> <p>4 Yes, easily</p> <p>3 With little difficulty</p> <p>2 With moderate difficulty</p> <p>1 With extreme difficulty</p> <p>0 No, Impossible</p> | <p>8 Have you had any trouble dressing yourself because of your operated on shoulder?</p> <p>4 No trouble at all</p> <p>3 A little bit of trouble</p> <p>2 Moderate trouble</p> <p>1 Extreme difficulty</p> <p>0 Impossible to do</p> <p>9 Could you hang your clothes up in a wardrobe – using the operated on arm?</p> <p>4 Yes, easily</p> <p>3 With little difficulty</p> <p>2 With moderate difficulty</p> <p>1 With extreme difficulty</p> <p>0 No, impossible</p> <p>10 Have you been able to wash and dry yourself under both arms?</p> <p>4 Yes, easily</p> <p>3 With little difficulty</p> <p>2 With moderate difficulty</p> <p>1 With extreme difficulty</p> <p>0 No, impossible</p> <p>11 How much has pain from your operated on shoulder interfered with your usual work hobbies or recreational activities (including housework)?</p> <p>4 Not at all</p> <p>3 A little bit</p> <p>2 Moderately</p> <p>1 Greatly</p> <p>0 Totally</p> <p>12 Have you been troubled by pain from your operated on shoulder in bed at night?</p> <p>4 No nights</p> <p>3 Only 1 or 2 nights</p> <p>2 Some nights</p> <p>1 Most nights</p> <p>0 Every night</p> <p>.....</p> |
|--|--|

☐ I wish to receive a progress report on the study. **NB:** If there are reasons other than the operation which would stop you doing one of the tasks listed; try to answer the question from the joint replacement aspect alone.

**Oxford Elbow Score (OES)****Problems with your elbow****Full Name** \_\_\_\_\_**Circle as appropriate****Right / Left****Please tick (✓) one box for every question****1. During the past 4 weeks:**Have you had difficulty lifting things in your home, such as putting out the rubbish, because of your elbow problem?No  
difficulty  
☐A little bit of  
difficulty  
☐Moderate  
difficulty  
☐Extreme  
difficulty  
☐Impossible  
to do  
☐**2. During the past 4 weeks:**Have you had difficulty carrying bags of shopping, because of your elbow problem?No  
difficulty  
☐A little bit of  
difficulty  
☐Moderate  
difficulty  
☐Extreme  
difficulty  
☐Impossible  
to do  
☐**3. During the past 4 weeks:**Have you had any difficulty washing yourself all over, because of your elbow problem?No  
difficulty  
☐A little bit of  
difficulty  
☐Moderate  
difficulty  
☐Extreme  
difficulty  
☐Impossible  
to do  
☐**4. During the past 4 weeks:**Have you had any difficulty dressing yourself, because of your elbow problem?No  
difficulty  
☐A little bit of  
difficulty  
☐Moderate  
difficulty  
☐Extreme  
difficulty  
☐Impossible  
to do  
☐**5. During the past 4 weeks:**

Have you felt that your elbow problem is "controlling your life"?

No, not at all  
☐Occasionally  
☐Some days  
☐Most days  
☐Every day  
☐**6. During the past 4 weeks:**

How much has your elbow problem "been on your mind"?

Not at all  
☐A little  
of the time  
☐Some  
of the time  
☐Most  
of the time  
☐All  
of the time  
☐**7. During the past 4 weeks:**

Have you been troubled by pain from your elbow in bed at night?

Not at all  
☐1 or 2  
nights  
☐Some  
nights  
☐Most  
nights  
☐Every  
night  
☐**8. During the past 4 weeks:**

How often has your elbow pain interfered with your sleeping?

Not at all  
☐Occasionally  
☐Some  
of the time  
☐Most  
of the time  
☐All  
of the time  
☐**9. During the past 4 weeks:**

How much has your elbow problem interfered with your usual work or everyday activities?

Not at all  
☐A little bit  
☐Moderately  
☐Greatly  
☐Totally  
☐

**10** During the past 4 weeks:

Has your elbow problem limited your ability to take part in leisure activities that you enjoy doing?

|                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| No, not at all           | Occasionally             | Some of the time         | Most of the time         | All of the time          |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

**11** During the past 4 weeks:

How would you describe the worst pain you have from your elbow?

|                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| No pain                  | Mild pain                | Moderate pain            | Severe pain              | Unbearable               |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

**12** During the past 4 weeks:

How would you describe the pain you usually have from your elbow?

|                          |                          |                          |                          |                          |
|--------------------------|--------------------------|--------------------------|--------------------------|--------------------------|
| No pain                  | Mild pain                | Moderate pain            | Severe pain              | Unbearable               |
| <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |





